


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071123\
Data File : BG058152.D
Acq On : 11 Jul 2023 8:39
Operator : CG/JU
Sample : DFTPP516
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP516

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Title : SVOA CALIBRATION
Last Update : Mon Jul 10 16:52:40 2023



AutoFind: Scans 2497, 2498, 2499; Background Corrected with Scan 2492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.3	37525	PASS
68	69	0.00	2	1.4	475	PASS
69	198	0.00	100	32.7	33789	PASS
70	69	0.00	2	0.5	157	PASS
127	198	10	80	39.0	40355	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	103400	PASS
199	198	5	9	6.7	6972	PASS
275	198	10	60	27.5	28416	PASS
365	198	1	100	3.4	3486	PASS
441	198	0.01	100	13.4	13868	PASS
442	198	50	100	82.4	85248	PASS
443	442	15	24	18.8	16044	PASS

m/z	Abundance
41.10	161.0
44.00	1254.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

m/z	Abundance
37.00	158.0
37.90	362.0
39.00	2575.0
39.80	303.0
44.00	1075.0
49.00	226.0
50.00	8655.0
51.00	35120.0
52.00	2099.0
54.90	154.0
55.90	976.0
56.90	2298.0
60.90	333.0
61.90	444.0
63.00	1301.0
63.90	184.0
65.00	900.0
68.00	626.0
68.90	32208.0
69.90	259.0
73.10	230.0
74.00	2791.0
75.00	4716.0
76.10	1808.0
77.00	39592.0
78.00	2550.0
79.00	2032.0
80.00	1500.0
80.90	2650.0
81.90	509.0
82.90	551.0
85.00	482.0
85.90	776.0
87.00	413.0
91.00	557.0
92.00	752.0
93.00	3285.0
94.00	286.0
95.90	157.0
96.10	157.0
97.90	2620.0
99.00	2654.0
100.00	406.0
101.00	1277.0
102.90	373.0
103.90	778.0
104.90	843.0
106.90	10850.0
108.00	1873.0
108.90	315.0
109.90	20768.0
110.90	2805.0
111.90	543.0
112.90	158.0
115.90	584.0
117.00	6652.0
118.00	417.0
121.90	544.0
122.90	1152.0
123.90	557.0
124.90	572.0
127.00	35184.0
128.00	2911.0
129.00	12130.0
129.90	985.0
131.00	209.0
132.00	153.0
133.90	284.0
135.00	950.0
135.90	433.0
136.90	613.0
140.90	1572.0
141.90	585.0
143.00	398.0
146.00	393.0
146.90	847.0
148.00	1801.0
149.00	378.0
151.90	180.0
153.00	559.0
153.90	463.0
155.00	813.0
156.00	1479.0
157.10	337.0
157.90	384.0
158.90	263.0
159.90	586.0
160.90	834.0
162.00	176.0
164.90	535.0
165.90	563.0
167.00	3702.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

168.00	1844.0
168.90	405.0
170.00	164.0
170.90	214.0
172.00	378.0
172.90	302.0
174.00	866.0
175.00	1353.0
175.90	431.0
176.90	594.0
177.90	407.0
178.90	2974.0
180.00	1811.0
181.00	939.0
184.00	385.0
184.90	1358.0
186.00	11315.0
187.00	3067.0
188.00	500.0
188.90	624.0
191.10	362.0
191.90	1047.0
193.00	1074.0
194.00	240.0
196.00	2362.0
196.60	571.0
197.90	83672.0
198.90	5834.0
200.00	411.0
201.40	314.0
203.00	540.0
204.00	3252.0
205.00	5097.0
206.00	22408.0
207.00	3400.0
207.90	656.0
208.90	340.0
209.90	289.0
211.00	850.0
214.80	231.0
216.00	408.0
216.90	5457.0
218.00	640.0
221.00	3327.0
221.70	879.0
223.00	1260.0
224.00	12179.0
225.00	3279.0
226.00	354.0
227.00	4815.0
228.00	787.0
228.90	1339.0
231.00	463.0
234.00	253.0
234.90	298.0
236.00	229.0
237.00	418.0
239.00	180.0
240.90	253.0
242.00	590.0
243.00	727.0
244.00	9582.0
245.00	1283.0
245.90	1721.0
247.00	285.0
248.90	378.0
250.90	174.0
253.00	167.0
255.00	48960.0
256.00	7407.0
257.00	665.0
258.00	2627.0
259.00	496.0
264.90	1039.0
273.00	1327.0
274.00	4324.0
275.00	20888.0
276.00	2882.0
276.90	1876.0
277.90	314.0
283.00	283.0
284.90	455.0
290.80	219.0
292.80	376.0
296.00	5980.0
297.00	785.0
303.00	910.0
304.10	170.0
310.10	153.0
314.00	497.0
314.90	727.0
316.10	487.0
320.80	178.0
321.00	170.0
323.00	2127.0
323.90	261.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

327.10	322.0
328.00	205.0
332.00	166.0
332.90	156.0
334.00	1250.0
334.90	334.0
341.00	220.0
346.00	490.0
352.00	887.0
352.90	634.0
353.90	621.0
355.00	170.0
364.90	2454.0
366.00	455.0
372.00	1131.0
373.00	302.0
383.00	232.0
402.00	526.0
403.00	650.0
403.90	294.0
420.90	522.0
422.10	453.0
423.00	3564.0
424.00	814.0
441.00	8330.0
442.00	55576.0
443.00	10965.0
444.00	874.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

m/z	Abundance
36.90	164.0
37.90	428.0
39.00	3039.0
40.00	271.0
43.90	1139.0
48.90	257.0
50.00	10557.0
51.00	44920.0
52.00	2476.0
54.90	190.0
55.90	1260.0
57.00	3152.0
57.90	220.0
60.90	686.0
62.00	569.0
63.00	1562.0
63.90	293.0
65.00	940.0
68.00	798.0
68.90	39880.0
69.90	212.0
73.00	385.0
74.00	4276.0
75.00	7056.0
76.00	1958.0
77.00	48360.0
78.00	3414.0
79.00	2905.0
79.90	2362.0
80.90	3270.0
82.00	683.0
83.00	971.0
85.00	844.0
86.00	912.0
86.90	292.0
90.90	725.0
91.90	731.0
93.00	4685.0
93.90	299.0
96.00	227.0
98.00	3484.0
99.00	3244.0
99.90	386.0
100.90	1795.0
102.90	528.0
103.90	1107.0
104.90	877.0
106.00	482.0
107.00	13728.0
108.00	2472.0
109.00	387.0
109.90	27072.0
111.00	3988.0
112.00	354.0
113.00	164.0
116.00	910.0
116.90	9443.0
118.00	724.0
120.00	196.0
122.00	758.0
123.00	1122.0
123.90	500.0
124.90	732.0
127.00	48168.0
127.90	3539.0
128.90	15930.0
129.90	1615.0
131.00	343.0
132.00	278.0
133.90	327.0
134.90	1619.0
136.10	616.0
137.00	777.0
138.10	194.0
140.00	193.0
141.00	1940.0
141.90	829.0
143.00	526.0
143.90	168.0
145.90	388.0
147.00	1139.0
148.00	2910.0
149.00	572.0
150.00	253.0
151.00	364.0
151.70	253.0
153.00	676.0
153.90	639.0
155.00	1539.0
156.00	2309.0
157.00	449.0
157.90	495.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

158.90	333.0
159.90	687.0
161.00	1093.0
161.80	248.0
164.90	1070.0
166.00	816.0
167.00	4965.0
168.00	2464.0
168.90	521.0
170.00	188.0
170.80	206.0
172.00	590.0
172.90	554.0
173.90	1194.0
175.00	1915.0
176.00	571.0
177.00	975.0
177.90	367.0
179.00	3720.0
180.00	2906.0
181.00	1242.0
182.00	229.0
183.90	359.0
185.00	1796.0
186.00	15467.0
187.00	4777.0
188.00	473.0
188.90	949.0
190.90	506.0
191.90	1175.0
193.00	1513.0
193.70	301.0
194.90	214.0
196.00	3479.0
197.90	120120.0
198.90	7776.0
200.00	538.0
201.20	498.0
203.00	958.0
204.00	4702.0
205.00	7144.0
206.00	31792.0
207.00	4123.0
208.00	832.0
209.00	353.0
211.00	1156.0
214.90	401.0
216.00	712.0
216.90	7655.0
218.00	1142.0
221.00	4122.0
222.90	1785.0
224.00	16872.0
225.00	3890.0
226.00	398.0
227.00	6986.0
228.00	1017.0
228.90	1595.0
230.00	280.0
231.00	731.0
233.90	589.0
234.90	578.0
236.00	280.0
236.90	672.0
238.90	227.0
239.90	262.0
240.90	414.0
242.00	1050.0
243.00	1069.0
244.00	14129.0
245.00	1910.0
246.00	2435.0
247.00	694.0
249.00	421.0
249.90	157.0
251.00	202.0
252.20	188.0
252.90	446.0
255.00	72456.0
256.00	11252.0
256.90	812.0
258.00	3771.0
258.80	554.0
261.00	172.0
264.90	1520.0
266.00	321.0
273.00	2492.0
274.00	5684.0
275.00	32120.0
276.00	4341.0
277.00	2159.0
277.90	461.0
283.00	346.0
283.80	228.0
285.10	441.0
290.00	151.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

292.90	631.0
293.90	182.0
296.00	8813.0
297.00	1204.0
301.00	204.0
302.00	226.0
303.00	1028.0
304.00	300.0
308.90	219.0
313.90	422.0
314.90	938.0
316.00	603.0
321.00	323.0
321.90	164.0
323.00	3380.0
324.00	647.0
326.90	653.0
328.00	355.0
331.90	365.0
332.90	429.0
334.00	2346.0
334.90	635.0
341.00	449.0
345.90	501.0
352.00	884.0
353.00	720.0
354.00	939.0
354.90	173.0
365.00	3899.0
365.80	543.0
366.00	533.0
370.90	337.0
372.00	1688.0
373.00	477.0
382.90	462.0
389.90	253.0
392.00	153.0
401.10	196.0
401.90	876.0
402.90	1328.0
403.90	264.0
421.00	1004.0
422.00	705.0
423.00	5553.0
424.00	1316.0
425.10	164.0
441.00	16298.0
442.00	96384.0
443.00	18432.0
444.00	1977.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

m/z	Abundance
37.00	178.0
38.00	340.0
39.00	2167.0
40.80	163.0
43.90	1148.0
49.00	348.0
50.00	8296.0
51.00	32536.0
52.00	2056.0
54.90	177.0
55.90	806.0
56.90	2214.0
60.90	345.0
62.00	621.0
63.00	1151.0
63.90	272.0
65.00	905.0
68.90	29280.0
72.80	164.0
74.00	2604.0
75.00	4947.0
76.00	1531.0
77.00	38416.0
78.00	2510.0
79.00	2124.0
79.90	1757.0
80.90	2368.0
82.00	707.0
83.00	577.0
85.00	471.0
86.00	771.0
86.90	338.0
91.00	482.0
92.00	694.0
93.00	3429.0
94.00	315.0
96.00	264.0
97.90	2384.0
98.90	2463.0
100.00	211.0
100.90	1794.0
102.90	577.0
104.00	831.0
105.00	792.0
105.90	388.0
107.00	10881.0
108.00	1759.0
108.90	335.0
109.90	21584.0
111.00	2890.0
112.00	334.0
116.00	522.0
117.00	7928.0
118.00	739.0
122.00	833.0
122.90	1386.0
123.90	324.0
125.00	499.0
127.00	37712.0
128.00	2895.0
129.00	14149.0
130.00	1016.0
130.80	167.0
133.90	381.0
135.00	1025.0
135.90	366.0
137.00	655.0
140.00	263.0
141.00	1944.0
142.00	639.0
143.00	401.0
145.90	279.0
147.00	993.0
147.90	2013.0
148.90	408.0
150.00	179.0
150.90	234.0
151.70	300.0
153.00	581.0
153.90	650.0
155.00	1278.0
156.00	1460.0
157.10	384.0
158.00	514.0
159.00	276.0
160.00	549.0
161.00	1036.0
162.10	205.0
165.00	660.0
165.90	607.0
167.00	3694.0
168.00	2319.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

169.00	298.0
170.90	166.0
171.90	315.0
172.90	602.0
173.90	993.0
175.00	1865.0
176.00	423.0
176.90	761.0
178.00	441.0
178.90	3266.0
180.00	2323.0
181.00	1362.0
184.00	174.0
185.00	1516.0
186.00	13028.0
187.00	3977.0
188.00	560.0
188.90	663.0
189.90	177.0
191.00	413.0
191.90	1309.0
193.00	1256.0
194.00	267.0
196.00	3007.0
197.90	106408.0
198.90	7306.0
200.00	559.0
201.40	445.0
203.00	599.0
204.00	3368.0
205.00	6289.0
206.00	28536.0
207.00	3451.0
208.00	664.0
209.00	188.0
210.00	356.0
210.90	1000.0
212.00	193.0
214.90	281.0
216.00	565.0
216.90	7116.0
217.90	748.0
220.10	171.0
221.00	4078.0
223.00	1712.0
224.00	16273.0
225.00	4667.0
226.00	498.0
227.00	6963.0
228.00	958.0
229.00	1233.0
229.90	248.0
231.00	631.0
233.90	393.0
235.00	583.0
235.90	314.0
236.90	535.0
239.00	214.0
240.90	367.0
241.90	968.0
243.00	801.0
244.00	12708.0
245.00	1885.0
246.00	2089.0
246.90	569.0
249.00	514.0
250.80	188.0
253.00	262.0
255.00	67656.0
256.00	9937.0
257.00	846.0
258.00	3410.0
259.00	501.0
261.00	151.0
265.00	1704.0
265.90	161.0
271.00	185.0
271.90	220.0
273.00	2180.0
274.00	4798.0
275.00	32240.0
276.00	4152.0
277.00	2505.0
278.00	440.0
283.00	333.0
284.00	219.0
285.00	453.0
291.90	159.0
292.90	604.0
293.90	167.0
294.90	153.0
296.00	9171.0
296.90	1126.0
302.10	152.0
303.10	957.0
304.00	374.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516

313.90	504.0
314.90	863.0
316.00	496.0
320.90	350.0
321.90	233.0
323.00	3072.0
324.00	516.0
326.90	720.0
328.00	436.0
332.00	267.0
332.90	272.0
334.00	2000.0
335.00	517.0
341.00	285.0
345.90	625.0
352.00	877.0
353.00	860.0
354.00	1128.0
355.00	249.0
365.00	4105.0
365.90	706.0
370.90	219.0
372.10	1675.0
373.00	436.0
382.90	543.0
383.90	254.0
390.00	243.0
392.10	164.0
402.00	651.0
402.90	1062.0
404.00	501.0
420.90	995.0
421.90	991.0
423.00	6089.0
424.00	1087.0
425.00	190.0
441.00	16976.0
442.00	103784.0
443.00	18736.0
444.00	1584.0

Instrument :
BNA_G
ClientSampleId :
DFTPP516