

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG071218\
 Data File : BG035596.D
 Acq On : 12 Jul 2018 15:59
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 13 09:17:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 12 14:43:32 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	37.356	6.6	101	0.00
3	Pyridine	40.000	40.355	-0.9	98	0.00
4	n-Nitrosodimethylamine	40.000	37.789	5.5	96	0.00
5 S	2-Fluorophenol	80.000	78.286	2.1	98	0.00
6	Aniline	40.000	39.353	1.6	99	0.00
7 S	Phenol-d6	80.000	77.540	3.1	97	0.00
8	2-Chlorophenol	40.000	39.265	1.8	98	0.00
9	Benzaldehyde	40.000	40.579	-1.4	100	0.00
10 C	Phenol	40.000	39.772	0.6	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.613	1.0	100	0.00
12	1,3-Dichlorobenzene	40.000	39.379	1.6	101	0.00
13 C	1,4-Dichlorobenzene	40.000	39.775	0.6	101	0.00
14	1,2-Dichlorobenzene	40.000	39.668	0.8	101	0.00
15	Benzyl Alcohol	40.000	40.761	-1.9	102	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	47.008	-17.5	119	0.00
17	2-Methylphenol	40.000	39.563	1.1	97	0.00
18	Hexachloroethane	40.000	37.656	5.9	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.465	1.3	101	0.00
20	3+4-Methylphenols	40.000	41.347	-3.4	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	37.799	5.5	98	0.00
23 S	Nitrobenzene-d5	80.000	78.139	2.3	99	0.00
24	Nitrobenzene	40.000	38.471	3.8	99	0.00
25	Isophorone	40.000	38.934	2.7	100	0.00
26 C	2-Nitrophenol	40.000	41.027	-2.6	102	0.00
27	2,4-Dimethylphenol	40.000	40.184	-0.5	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.314	1.7	100	0.00
29 C	2,4-Dichlorophenol	40.000	39.310	1.7	96	0.00
30	1,2,4-Trichlorobenzene	40.000	39.440	1.4	102	0.00
31	Naphthalene	40.000	38.618	3.5	102	0.00
32	Benzoic acid	40.000	34.112	14.7	86	0.00
33	4-Chloroaniline	40.000	39.168	2.1	102	0.00
34 C	Hexachlorobutadiene	40.000	39.418	1.5	102	0.00
35	Caprolactam	40.000	40.318	-0.8	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.505	3.7	96	0.00
37	2-Methylnaphthalene	40.000	39.943	0.1	102	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	37.287	6.8	103	0.00
40 P	Hexachlorocyclopentadiene	40.000	37.938	5.2	101	0.00
41 S	2,4,6-Tribromophenol	80.000	77.212	3.5	104	0.00
42 C	2,4,6-Trichlorophenol	40.000	39.016	2.5	101	0.00
43	2,4,5-Trichlorophenol	40.000	38.745	3.1	108	0.00
44 S	2-Fluorobiphenyl	80.000	73.715	7.9	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG071218\
 Data File : BG035596.D
 Acq On : 12 Jul 2018 15:59
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 13 09:17:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 12 14:43:32 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	37.171	7.1	101	0.00
46	2-Chloronaphthalene	40.000	36.572	8.6	101	0.00
47	2-Nitroaniline	40.000	39.123	2.2	105	0.00
48	Acenaphthylene	40.000	37.052	7.4	102	0.00
49	Dimethylphthalate	40.000	37.400	6.5	105	0.00
50	2,6-Dinitrotoluene	40.000	37.342	6.6	100	0.00
51 C	Acenaphthene	40.000	36.781	8.0	102	0.00
52	3-Nitroaniline	40.000	38.904	2.7	103	0.00
53 P	2,4-Dinitrophenol	40.000	35.676	10.8	100	0.00
54	Dibenzofuran	40.000	36.670	8.3	101	0.00
55 P	4-Nitrophenol	40.000	39.795	0.5	106	0.00
56	2,4-Dinitrotoluene	40.000	38.723	3.2	100	0.00
57	Fluorene	40.000	36.872	7.8	103	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	38.294	4.3	103	0.00
59	Diethylphthalate	40.000	36.807	8.0	101	0.00
60	4-Chlorophenyl-phenylether	40.000	37.001	7.5	103	0.00
61	4-Nitroaniline	40.000	37.955	5.1	100	0.00
62	Azobenzene	40.000	36.914	7.7	102	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
64	4,6-Dinitro-2-methylphenol	40.000	37.923	5.2	103	0.00
65 c	n-Nitrosodiphenylamine	40.000	38.384	4.0	103	0.00
66	4-Bromophenyl-phenylether	40.000	38.753	3.1	103	0.00
67	Hexachlorobenzene	40.000	38.254	4.4	104	0.00
68	Atrazine	40.000	38.193	4.5	100	0.00
69 C	Pentachlorophenol	40.000	38.275	4.3	100	0.00
70	Phenanthrene	40.000	37.091	7.3	102	0.00
71	Anthracene	40.000	37.257	6.9	101	0.00
72	Carbazole	40.000	37.495	6.3	101	0.00
73	Di-n-butylphthalate	40.000	37.160	7.1	100	0.00
74 C	Fluoranthene	40.000	36.667	8.3	99	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
76	Benzidine	40.000	35.123	12.2	95	0.00
77	Pyrene	40.000	38.188	4.5	101	0.00
78 S	Terphenyl-d14	80.000	76.580	4.3	100	0.00
79	Butylbenzylphthalate	40.000	39.034	2.4	99	0.00
80	Benzo(a)anthracene	40.000	38.264	4.3	101	0.00
81	3,3'-Dichlorobenzidine	40.000	38.931	2.7	99	0.00
82	Chrysene	40.000	38.522	3.7	101	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	39.067	2.3	99	0.00
84 c	Di-n-octyl phthalate	40.000	39.555	1.1	99	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	39.200	2.0	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Benzo(b)fluoranthene	40.000	37.766	5.6	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG071218\
Data File : BG035596.D
Acq On : 12 Jul 2018 15:59
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 13 09:17:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 12 14:43:32 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	38.637	3.4	101	0.00
89 C	Benzo(a)pyrene	40.000	38.349	4.1	100	0.00
90	Dibenzo(a,h)anthracene	40.000	38.231	4.4	99	0.00
91	Benzo(a,h,i)perylene	40.000	38.542	3.6	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0