



Benzoic acid



$\text{Response} = 1.502\text{e-}001 * \text{Amt} - 6.935\text{e-}002$
 Coef of Det (r^2) = 0.993289 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG071222.M
 Calibration Table Last Updated: Tue Jul 12 15:49:20 2022


Hexachlorocyclopentadiene

Response Ratio




$R = 4.298e-002 A^2 + 1.424e-001 A - 4.164e-002$
Coef of Det (r^2) = 0.999332 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG071222.M
Calibration Table Last Updated: Tue Jul 12 15:49:20 2022

2,4-Dinitrophenol



Response = $2.030 \times 10^{-1} \times \text{Amt} - 5.873 \times 10^{-2}$
 Coef of Det (r^2) = 0.996639 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG071222.M
 Calibration Table Last Updated: Tue Jul 12 15:49:20 2022

Pentachlorophenol



Response = $1.464 \times 10^{-1} \times \text{Amt} - 3.318 \times 10^{-2}$
 Coef of Det (r^2) = 0.999207 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG071222.M
 Calibration Table Last Updated: Tue Jul 12 15:49:20 2022