

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071222\
 Data File : BG054257.D
 Acq On : 12 Jul 2022 15:29
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG071222

Quant Time: Jul 12 16:55:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 12 15:49:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	36.737	8.2	102	0.00
3	Pyridine	40.000	43.256	-8.1	102	0.00
4	n-Nitrosodimethylamine	40.000	43.485	-8.7	106	0.00
5 S	2-Fluorophenol	80.000	83.580	-4.5	102	0.00
6	Aniline	40.000	41.235	-3.1	102	0.00
7 S	Phenol-d6	80.000	81.928	-2.4	100	0.00
8	2-Chlorophenol	40.000	42.012	-5.0	103	0.00
9	Benzaldehyde	40.000	35.693	10.8	104	0.00
10 C	Phenol	40.000	40.673	-1.7	99	0.00
11	bis(2-Chloroethyl)ether	40.000	40.496	-1.2	97	0.00
12	1,3-Dichlorobenzene	40.000	40.767	-1.9	100	0.00
13 C	1,4-Dichlorobenzene	40.000	41.873	-4.7	104	0.00
14	1,2-Dichlorobenzene	40.000	40.330	-0.8	101	0.00
15	Benzyl Alcohol	40.000	40.081	-0.2	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.221	4.4	97	0.00
17	2-Methylphenol	40.000	40.330	-0.8	99	0.00
18	Hexachloroethane	40.000	41.867	-4.7	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.020	-0.1	100	0.00
20	3+4-Methylphenols	40.000	40.076	-0.2	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	41.746	-4.4	98	0.00
23 S	Nitrobenzene-d5	80.000	84.841	-6.1	100	0.00
24	Nitrobenzene	40.000	42.155	-5.4	99	0.00
25	Isophorone	40.000	41.208	-3.0	97	0.00
26 C	2-Nitrophenol	40.000	42.797	-7.0	99	0.00
27	2,4-Dimethylphenol	40.000	42.625	-6.6	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.105	-2.8	98	0.00
29 C	2,4-Dichlorophenol	40.000	43.252	-8.1	101	0.00
30	1,2,4-Trichlorobenzene	40.000	42.142	-5.4	99	0.00
31	Naphthalene	40.000	41.541	-3.9	98	0.00
32	Benzoic acid	40.000	37.496	6.3	96	0.00
33	4-Chloroaniline	40.000	42.361	-5.9	101	0.00
34 C	Hexachlorobutadiene	40.000	43.130	-7.8	100	0.00
35	Caprolactam	40.000	41.275	-3.2	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.133	-5.3	100	0.00
37	2-Methylnaphthalene	40.000	41.539	-3.8	98	0.00
38	1-Methylnaphthalene	40.000	41.244	-3.1	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.527	-3.8	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.628	-1.6	101	0.00
42 S	2,4,6-Tribromophenol	80.000	84.870	-6.1	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.932	-4.8	100	0.00
44	2,4,5-Trichlorophenol	40.000	43.134	-7.8	103	0.00
45 S	2-Fluorobiphenyl	80.000	82.628	-3.3	99	0.00
46	1,1'-Biphenyl	40.000	40.827	-2.1	98	0.00
47	2-Chloronaphthalene	40.000	41.056	-2.6	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071222\
 Data File : BG054257.D
 Acq On : 12 Jul 2022 15:29
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG071222

Quant Time: Jul 12 16:55:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 12 15:49:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.869	-4.7	101	0.00
49	Acenaphthylene	40.000	40.861	-2.2	99	0.00
50	Dimethylphthalate	40.000	40.558	-1.4	100	0.00
51	2,6-Dinitrotoluene	40.000	41.238	-3.1	98	0.00
52 C	Acenaphthene	40.000	40.603	-1.5	99	0.00
53	3-Nitroaniline	40.000	41.135	-2.8	99	0.00
54 P	2,4-Dinitrophenol	40.000	38.593	3.5	102	0.00
55	Dibenzofuran	40.000	41.110	-2.8	99	0.00
56 P	4-Nitrophenol	40.000	44.687	-11.7	107	0.00
57	2,4-Dinitrotoluene	40.000	41.293	-3.2	100	0.00
58	Fluorene	40.000	40.507	-1.3	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.562	-3.9	99	0.00
60	Diethylphthalate	40.000	40.806	-2.0	101	0.00
61	4-Chlorophenyl-phenylether	40.000	41.303	-3.3	100	0.00
62	4-Nitroaniline	40.000	42.282	-5.7	104	0.00
63	Azobenzene	40.000	40.209	-0.5	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.983	-5.0	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.613	-1.5	101	0.00
67	4-Bromophenyl-phenylether	40.000	41.271	-3.2	102	0.00
68	Hexachlorobenzene	40.000	40.765	-1.9	101	0.00
69	Atrazine	40.000	42.372	-5.9	105	0.00
70 C	Pentachlorophenol	40.000	39.586	1.0	103	0.00
71	Phenanthrene	40.000	41.212	-3.0	102	0.00
72	Anthracene	40.000	41.041	-2.6	102	0.00
73	Carbazole	40.000	41.059	-2.6	102	0.00
74	Di-n-butylphthalate	40.000	40.959	-2.4	103	0.00
75 C	Fluoranthene	40.000	41.362	-3.4	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	38.329	4.2	95	0.00
78	Pyrene	40.000	41.235	-3.1	105	0.00
79 S	Terphenyl-d14	80.000	83.641	-4.6	107	0.00
80	Butylbenzylphthalate	40.000	41.846	-4.6	108	0.00
81	Benzo(a)anthracene	40.000	41.754	-4.4	108	0.00
82	3,3'-Dichlorobenzidine	40.000	41.799	-4.5	105	0.00
83	Chrysene	40.000	41.424	-3.6	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.212	-3.0	106	0.00
85 c	Di-n-octyl phthalate	40.000	41.352	-3.4	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.058	-5.1	109	-0.01
88	Benzo(b)fluoranthene	40.000	42.195	-5.5	110	0.00
89	Benzo(k)fluoranthene	40.000	41.092	-2.7	107	0.00
90 C	Benzo(a)pyrene	40.000	41.638	-4.1	109	0.00
91	Dibenzo(a,h)anthracene	40.000	42.371	-5.9	110	0.00
92	Benzo(g,h,i)perylene	40.000	41.995	-5.0	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071222\
Data File : BG054257.D
Acq On : 12 Jul 2022 15:29
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG071222

Quant Time: Jul 12 16:55:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 12 15:49:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0