

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071323\
 Data File : BG058253.D
 Acq On : 15 Jul 2023 7:44
 Operator : CG/JU
 Sample : PB154013BS
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS013

Quant Time: Jul 15 22:35:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 13 14:01:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.902	152	54771	20.000	ng/u1	0.00
20) Naphthalene-d8	10.704	136	260486	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.541	164	184065	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.285	188	428407	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	370192	20.000	ng/u1	0.00
88) Perylene-d12	24.682	264	469033	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	7964	5.092	ng/uL	0.00
4) Pyridine-d5	3.754	84	135827	28.604	ng/u1	0.00
7) Phenol-d5	7.068	99	186319	33.310	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.232	67	108013	32.044	ng/u1	0.00
11) 2-Chlorophenol-d4	7.438	132	128751	33.374	ng/u1	0.00
15) 4-Methylphenol-d8	8.613	113	138467	32.808	ng/u1	0.00
21) Nitrobenzene-d5	9.065	128	66298	31.227	ng/u1	0.00
24) 2-Nitrophenol-d4	9.788	143	75862	33.999	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	143964	33.633	ng/u1	0.00
31) 4-Chloroaniline-d4	10.840	131	188014	28.971	ng/u1	0.00
46) Dimethylphthalate-d6	13.948	166	441822	30.422	ng/u1	0.00
49) Acenaphthylene-d8	14.236	160	514726	32.479	ng/u1	0.00
54) 4-Nitrophenol-d4	14.753	143	77146	30.066	ng/u1	0.00
60) Fluorene-d10	15.534	176	397026	32.190	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.657	200	77184	33.179	ng/u1	0.00
73) Anthracene-d10	17.385	188	630602	31.722	ng/u1	0.00
81) Pyrene-d10	19.676	212	740398	31.756	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.471	264	806568	33.753	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.378	88	22965	11.885	ng/uL#	89
5) Pyridine	3.771	79	152831	31.412	ng/u1	98
6) Benzaldehyde	7.038	77	117415	62.856	ng/u1	97
8) Phenol	7.097	94	208494	36.732	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.320	93	150304	34.295	ng/u1	97
12) 2-Chlorophenol	7.467	128	146574	36.511	ng/u1	97
13) 2-Methylphenol	8.343	108	160436	35.990	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.431	45	247705	36.717	ng/u1	99
16) Acetophenone	8.724	105	246043	35.137	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.707	70	127203	33.981	ng/u1	97
18) 4-Methylphenol	8.677	108	176266	36.568	ng/u1	96
19) Hexachloroethane	8.983	117	55656	36.164	ng/u1	99
22) Nitrobenzene	9.106	77	190675	34.596	ng/u1	98
23) Isophorone	9.629	82	387692	32.966	ng/u1	99
25) 2-Nitrophenol	9.817	139	88316	36.886	ng/u1	95
26) 2,4-Dimethylphenol	9.876	107	167467	31.105	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.111	93	218298	32.940	ng/u1	98
29) 2,4-Dichlorophenol	10.352	162	148448	35.669	ng/u1	92
30) Naphthalene	10.757	128	500116	34.164	ng/u1	99
32) 4-Chloroaniline	10.869	127	207519	31.272	ng/u1	99
33) Hexachlorobutadiene	11.033	225	101690	34.513	ng/u1	96
34) Caprolactam	11.633	113	40898	24.443	ng/u1	92
35) 4-Chloro-3-methylphenol	11.997	107	190160	37.196	ng/u1	99
36) 2-Methylnaphthalene	12.367	142	339139	34.368	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071323\
 Data File : BG058253.D
 Acq On : 15 Jul 2023 7:44
 Operator : CG/JU
 Sample : PB154013BS
 Misc :
 ALS Vial : 62 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SLCS013


Quant Time: Jul 15 22:35:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 13 14:01:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.585	142	341767	34.537	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.732	216	200051	35.664	ng/ul	98
40) Hexachlorocyclopentadiene	12.708	237	78308	22.164	ng/ul	97
41) 2,4,6-Trichlorophenol	12.978	196	141147	37.127	ng/ul	96
42) 2,4,5-Trichlorophenol	13.049	196	151422	36.829	ng/ul	98
43) 1,1'-Biphenyl	13.372	154	456209	34.393	ng/ul	100
44) 2-Chloronaphthalene	13.419	162	370409	34.794	ng/ul	100
45) 2-Nitroaniline	13.625	65	138959	37.543	ng/ul	98
47) Dimethylphthalate	13.995	163	489223	33.529	ng/ul	100
48) 2,6-Dinitrotoluene	14.118	165	109816	36.667	ng/ul	94
50) Acenaphthylene	14.265	152	590890	34.527	ng/ul	98
51) 3-Nitroaniline	14.447	138	103211	40.779	ng/ul	99
52) Acenaphthene	14.606	153	408183	34.758	ng/ul	98
53) 2,4-Dinitrophenol	14.659	184	49142	31.473	ng/ul	92
55) 4-Nitrophenol	14.764	109	62490	32.876	ng/ul	87
56) Dibenzofuran	14.941	168	578421	34.649	ng/ul	99
57) 2,4-Dinitrotoluene	14.911	165	157919	37.216	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.170	232	136449	36.455	ng/ul	99
59) Diethylphthalate	15.358	149	507314	33.313	ng/ul	99
61) Fluorene	15.593	166	467866	34.158	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.581	204	251439	34.693	ng/ul	97
63) 4-Nitroaniline	15.616	138	97396	42.957	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.675	198	86576	36.393	ng/ul	98
67) N-Nitrosodiphenylamine	15.798	169	403975	35.432	ng/ul	98
68) 4-Bromophenyl-phenylether	16.474	248	177621	37.771	ng/ul	98
69) Hexachlorobenzene	16.592	284	196463	37.315	ng/ul	96
70) Atrazine	16.744	200	161402	34.618	ng/ul	95
71) Pentachlorophenol	16.938	266	105289	33.310	ng/ul	97
72) Phenanthrene	17.332	178	758319	34.798	ng/ul	99
74) Anthracene	17.420	178	753568	34.172	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.343	216	205281	36.130	ng/uL	97
76) Pentachlorobenzene	14.858	250	224381	36.978	ng/uL	98
77) Carbazole	17.690	167	714359	35.818	ng/ul	99
78) Di-n-butylphthalate	18.243	149	877351	34.534	ng/ul	99
80) Fluoranthene	19.341	202	924226	34.210	ng/ul	99
82) Pyrene	19.706	202	908368	33.403	ng/ul	98
83) Butylbenzylphthalate	20.587	149	401720	36.429	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.439	252	365012	40.729	ng/ul	97
85) Benzo(a)anthracene	21.521	228	943825	35.508	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.421	149	571128	36.141	ng/ul	97
87) Chrysene	21.586	228	904698	36.308	ng/ul	98
89) Di-n-octyl phthalate	22.602	149	1020102	35.398	ng/ul	100
90) Benzo(b)fluoranthene	23.689	252	1013441	34.896	ng/ul	99
91) Benzo(k)fluoranthene	23.754	252	999786	35.123	ng/ul	99
93) Benzo(a)pyrene	24.547	252	910299	35.414	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.272	276	1207585	35.378	ng/ul	99
95) Dibenzo(a,h)anthracene	28.331	278	1011378	35.932	ng/ul	99
96) Benzo(g,h,i)perylene	29.394	276	970862	34.823	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 15 22:35:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 13 14:01:20 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.885 ng/uL
RT: 3.378 min Scan# 49
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion: 88 Resp: 22965
Ion Ratio Lower Upper
88 100
43 35.1 20.6 30.8#
58 90.1 65.5 98.3



#5
Pyridine
Concen: 31.412 ng/uL
RT: 3.771 min Scan# 116
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion: 79 Resp: 152831
Ion Ratio Lower Upper
79 100
52 85.5 68.6 102.8
51 40.2 29.4 44.0









#6
Benzaldehyde
Concen: 62.856 ng/ul
RT: 7.038 min Scan# 61
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion: 77 Resp: 117415
Ion Ratio Lower Upper
77 100
105 95.5 74.7 112.1
106 91.3 75.9 113.9



#8
Phenol
Concen: 36.732 ng/ul
RT: 7.097 min Scan# 682
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion: 94 Resp: 208494
Ion Ratio Lower Upper
94 100
65 29.8 22.2 33.4
66 36.0 28.6 43.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 34.295 ng/ul
 RT: 7.320 min Scan# 720
 Delta R.T. -0.004 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44




Instrument :
 BNA_G
 ClientSampleId :
 SLCS013


Tgt Ion: 93 Resp: 150304
 Ion Ratio Lower Upper
 93 100
 63 89.7 69.4 104.0
 95 33.0 27.0 40.6



#12
 2-Chlorophenol
 Concen: 36.511 ng/ul
 RT: 7.467 min Scan# 745
 Delta R.T. 0.002 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Tgt Ion:128 Resp: 146574
 Ion Ratio Lower Upper
 128 100
 64 57.0 44.3 66.5
 130 31.5 27.3 40.9





#13

2-Methylphenol

Concen: 35.990 ng/ul

RT: 8.343 min Scan# 894

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :





SLCS013

Tgt Ion:108 Resp: 160436

Ion Ratio Lower Upper

108 100

107 88.5 75.0 112.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 36.717 ng/ul

RT: 8.431 min Scan# 909

Delta R.T. 0.008 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44




Tgt Ion: 45 Resp: 247705


Ion Ratio Lower Upper

45 100

77 11.7 9.4 14.2

79 9.6 8.3 12.5









#16
Acetophenone
Concen: 35.137 ng/ul
RT: 8.724 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:105 Resp: 246043
Ion Ratio Lower Upper
105 100
77 77.2 61.2 91.8
51 30.1 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 33.981 ng/ul
RT: 8.707 min Scan# 956
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion: 70 Resp: 127203
Ion Ratio Lower Upper
70 100
42 62.3 48.0 72.0
101 9.6 7.4 11.2
130 18.1 13.4 20.0









#18
4-Methylphenol
Concen: 36.568 ng/ul
RT: 8.677 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:108 Resp: 176266
Ion Ratio Lower Upper
108 100
107 109.5 91.1 136.7



#19
Hexachloroethane
Concen: 36.164 ng/ul
RT: 8.983 min Scan# 1003
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:117 Resp: 55656
Ion Ratio Lower Upper
117 100
201 117.5 94.0 141.0
199 73.3 59.8 89.6









#22
Nitrobenzene
Concen: 34.596 ng/ul
RT: 9.106 min Scan# 1102
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion: 77 Resp: 190675
Ion Ratio Lower Upper
77 100
123 42.6 33.3 49.9
65 13.6 10.4 15.6



#23
Isophorone
Concen: 32.966 ng/ul
RT: 9.629 min Scan# 1113
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion: 82 Resp: 387692
Ion Ratio Lower Upper
82 100
95 6.6 4.8 7.2
138 14.7 11.4 17.0









#25
2-Nitrophenol
Concen: 36.886 ng/ul
RT: 9.817 min Scan# 1145
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:139 Resp: 88316
Ion Ratio Lower Upper
139 100
65 56.4 42.6 64.0
109 29.0 20.7 31.1



#26
2,4-Dimethylphenol
Concen: 31.105 ng/ul
RT: 9.876 min Scan# 1155
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:107 Resp: 167467
Ion Ratio Lower Upper
107 100
121 47.8 40.6 60.8
122 82.1 67.1 100.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.940 ng/ul
 RT: 10.111 min Scan# 1195
 Delta R.T. 0.002 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Instrument :
 BNA_G
 ClientSampleId :
 SLCS013

Tgt Ion: 93 Resp: 218298




Ion	Ratio	Lower	Upper
93	100		
95	31.7	26.4	39.6
123	9.8	7.9	11.9




#29
 2,4-Dichlorophenol
 Concen: 35.669 ng/ul
 RT: 10.352 min Scan# 1236
 Delta R.T. -0.004 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Tgt Ion: 162 Resp: 148448

Ion	Ratio	Lower	Upper
162	100		
164	69.9	51.1	76.7
98	39.7	27.9	41.9









#30
Naphthalene
Concen: 34.164 ng/ul
RT: 10.757 min Scan# 11
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:128 Resp: 500116
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 13.0 10.2 15.2



#32
4-Chloroaniline
Concen: 31.272 ng/ul
RT: 10.869 min Scan# 1324
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:127 Resp: 207519
Ion Ratio Lower Upper
127 100
129 32.2 26.0 39.0





#33

Hexachlorobutadiene

Concen: 34.513 ng/ul

RT: 11.033 min Scan# 11

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:225 Resp: 101690

Ion Ratio Lower Upper

225 100

223 62.8 53.2 79.8

227 62.3 51.4 77.2



#34

Caprolactam

Concen: 24.443 ng/ul

RT: 11.633 min Scan# 1454

Delta R.T. 0.014 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44




Tgt Ion:113 Resp: 40898


Ion Ratio Lower Upper

113 100

55 240.6 183.4 275.2

56 165.9 123.0 184.6









#35
4-Chloro-3-methylphenol
Concen: 37.196 ng/ul
RT: 11.997 min Scan# 1516
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:107 Resp: 190160
Ion Ratio Lower Upper
107 100
144 23.9 19.0 28.4
142 73.7 58.4 87.6



#36
2-Methylnaphthalene
Concen: 34.368 ng/ul
RT: 12.367 min Scan# 1579
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:142 Resp: 339139
Ion Ratio Lower Upper
142 100
141 90.7 69.4 104.2





#37

1-Methylnaphthalene

Concen: 34.537 ng/ul

RT: 12.585 min Scan# 1

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:142 Resp: 341767

Ion Ratio Lower Upper

142 100

141 94.2 73.8 110.8

116 4.0 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 35.664 ng/ul

RT: 12.732 min Scan# 1641

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Tgt Ion:216 Resp: 200051




Ion Ratio Lower Upper


216 100

214 78.8 64.1 96.1

179 18.6 16.0 24.0

108 16.2 12.1 18.1









#40
Hexachlorocyclopentadiene
Concen: 22.164 ng/ul
RT: 12.708 min Scan# 1637
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:237	Resp:	78308
Ion Ratio	Lower	Upper
237	100	
235	64.3	49.2 73.8
272	13.7	10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 37.127 ng/ul
RT: 12.978 min Scan# 1683
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:196	Resp:	141147
Ion Ratio	Lower	Upper
196	100	
198	97.1	73.5 110.3
200	30.3	24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 36.829 ng/ul
RT: 13.049 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:196 Resp: 151422
Ion Ratio Lower Upper
196 100
198 94.4 77.2 115.8
200 28.6 23.8 35.6



#43
1,1'-Biphenyl
Concen: 34.393 ng/ul
RT: 13.372 min Scan# 1750
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:154 Resp: 456209
Ion Ratio Lower Upper
154 100
153 41.1 32.6 49.0
76 14.4 11.6 17.4









#44
2-Chloronaphthalene
Concen: 34.794 ng/ul
RT: 13.419 min Scan# 1758
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:162 Resp: 370409
Ion Ratio Lower Upper
162 100
164 34.6 27.7 41.5
127 38.7 30.9 46.3



#45
2-Nitroaniline
Concen: 37.543 ng/ul
RT: 13.625 min Scan# 1793
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion: 65 Resp: 138959
Ion Ratio Lower Upper
65 100
92 64.0 53.2 79.8
138 90.5 71.1 106.7









#47
Dimethylphthalate
Concen: 33.529 ng/ul
RT: 13.995 min Scan# 1856
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:163 Resp: 489223
Ion Ratio Lower Upper
163 100
194 3.7 3.0 4.4
164 10.7 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 36.667 ng/ul
RT: 14.118 min Scan# 1877
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:165 Resp: 109816
Ion Ratio Lower Upper
165 100
89 58.8 42.7 64.1
121 19.6 14.8 22.2









#50
Acenaphthylene
Concen: 34.527 ng/ul
RT: 14.265 min Scan# 1902
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:152 Resp: 590890
Ion Ratio Lower Upper
152 100
151 20.6 15.6 23.4
153 13.3 10.5 15.7



#51
3-Nitroaniline
Concen: 40.779 ng/ul
RT: 14.447 min Scan# 1933
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:138 Resp: 103211
Ion Ratio Lower Upper
138 100
108 11.5 8.2 12.2
92 125.0 100.2 150.4





#52

Acenaphthene

Concen: 34.758 ng/ul

RT: 14.606 min Scan# 1960

Delta R.T. 0.002 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:153 Resp: 408183

Ion Ratio Lower Upper

153 100

152 48.3 36.5 54.7

154 87.3 70.5 105.7



#53

2,4-Dinitrophenol

Concen: 31.473 ng/ul

RT: 14.659 min Scan# 1969

Delta R.T. 0.002 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44




Tgt Ion:184 Resp: 49142


Ion Ratio Lower Upper

184 100

63 68.7 51.1 76.7

154 60.9 43.3 64.9









#55
4-Nitrophenol
Concen: 32.876 ng/ul
RT: 14.764 min Scan# 1986
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:109 Resp: 62490
Ion Ratio Lower Upper
109 100
139 133.4 119.0 178.4
65 151.5 135.5 203.3



#56
Dibenzofuran
Concen: 34.649 ng/ul
RT: 14.941 min Scan# 2017
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:168 Resp: 578421
Ion Ratio Lower Upper
168 100
139 34.9 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 37.216 ng/ul

RT: 14.911 min Scan# 2011

Delta R.T. 0.008 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:165 Resp: 157919

Ion Ratio Lower Upper

165 100

63 44.5 35.8 53.8

182 2.1 1.6 2.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 36.455 ng/ul

RT: 15.170 min Scan# 2056

Delta R.T. 0.002 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Tgt Ion:232 Resp: 136449




Ion Ratio Lower Upper


232 100

131 38.9 30.5 45.7

130 2.0 1.4 2.0

166 28.4 22.6 34.0









#59
Diethylphthalate
Concen: 33.313 ng/ul
RT: 15.358 min Scan# 2088
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:149 Resp: 507314
Ion Ratio Lower Upper
149 100
177 22.9 17.8 26.8
150 12.9 10.1 15.1



#61
Fluorene
Concen: 34.158 ng/ul
RT: 15.593 min Scan# 2128
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:166 Resp: 467866
Ion Ratio Lower Upper
166 100
165 100.5 79.1 118.7
167 12.9 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 34.693 ng/ul
RT: 15.581 min Scan# 2126
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:204 Resp: 251439
Ion Ratio Lower Upper
204 100
206 33.7 25.8 38.6
141 50.4 41.8 62.6



#63
4-Nitroaniline
Concen: 42.957 ng/ul
RT: 15.616 min Scan# 2132
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:138 Resp: 97396
Ion Ratio Lower Upper
138 100
92 54.4 43.8 65.8
108 79.9 58.8 88.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 36.393 ng/ul
 RT: 15.675 min Scan# 2142
 Delta R.T. 0.008 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44




Instrument :
 BNA_G
 ClientSampleId :
 SLCS013


Tgt Ion:198 Resp: 86576
 Ion Ratio Lower Upper
 198 100
 182 3.5 3.1 4.7
 77 21.6 17.9 26.9



#67
 N-Nitrosodiphenylamine
 Concen: 35.432 ng/ul
 RT: 15.798 min Scan# 2163
 Delta R.T. 0.002 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Tgt Ion:169 Resp: 403975
 Ion Ratio Lower Upper
 169 100
 168 69.6 54.2 81.2
 167 38.4 30.8 46.2









#68
4-Bromophenyl-phenylether
Concen: 37.771 ng/ul
RT: 16.474 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:248 Resp: 177621
Ion Ratio Lower Upper
248 100
250 100.0 78.3 117.5
141 61.9 49.1 73.7



#69
Hexachlorobenzene
Concen: 37.315 ng/ul
RT: 16.592 min Scan# 2298
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:284 Resp: 196463
Ion Ratio Lower Upper
284 100
142 29.8 21.9 32.9
249 33.0 25.0 37.6





#70

Atrazine

Concen: 34.618 ng/ul

RT: 16.744 min Scan# 21

Delta R.T. 0.002 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:200 Resp: 161402

Ion Ratio Lower Upper

200 100

173 26.6 20.0 30.0

215 49.0 35.8 53.8



#71

Pentachlorophenol

Concen: 33.310 ng/ul

RT: 16.938 min Scan# 2357

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44




Tgt Ion:266 Resp: 105289


Ion Ratio Lower Upper

266 100

264 61.1 51.5 77.3

268 62.5 49.0 73.4









#72
Phenanthrene
Concen: 34.798 ng/ul
RT: 17.332 min Scan# 2424
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:178 Resp: 758319
Ion Ratio Lower Upper
178 100
179 16.3 12.8 19.2
176 20.3 15.7 23.5



#74
Anthracene
Concen: 34.172 ng/ul
RT: 17.420 min Scan# 2439
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:178 Resp: 753568
Ion Ratio Lower Upper
178 100
179 16.5 12.6 18.8
176 19.7 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.130 ng/uL
RT: 13.343 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:216 Resp: 205281
Ion Ratio Lower Upper
216 100
214 77.8 64.8 97.2
218 47.0 39.1 58.7



#76
Pentachlorobenzene
Concen: 36.978 ng/uL
RT: 14.858 min Scan# 2003
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:250 Resp: 224381
Ion Ratio Lower Upper
250 100
248 64.7 52.8 79.2
252 64.4 52.4 78.6









#77
 Carbazole
 Concen: 35.818 ng/ul
 RT: 17.690 min Scan# 2485
 Delta R.T. -0.004 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Instrument :
 BNA_G
 ClientSampleId :
 SLCS013

Tgt Ion:167 Resp: 714359




Ion	Ratio	Lower	Upper
167	100		
166	22.6	17.5	26.3
139	12.4	9.5	14.3




#78
 Di-n-butylphthalate
 Concen: 34.534 ng/ul
 RT: 18.243 min Scan# 2579
 Delta R.T. -0.004 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Tgt Ion:149 Resp: 877351

Ion	Ratio	Lower	Upper
149	100		
150	9.6	7.4	11.2
104	5.1	3.9	5.9









#80
Fluoranthene
Concen: 34.210 ng/ul
RT: 19.341 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:202 Resp: 924226
Ion Ratio Lower Upper
202 100
101 11.2 9.1 13.7
100 9.1 7.1 10.7



#82
Pyrene
Concen: 33.403 ng/ul
RT: 19.706 min Scan# 2828
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:202 Resp: 908368
Ion Ratio Lower Upper
202 100
101 13.7 10.2 15.2
100 11.3 8.9 13.3





#83

Butylbenzylphthalate

Concen: 36.429 ng/ul

RT: 20.587 min Scan# 2978

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:149 Resp: 401720

Ion Ratio Lower Upper

149 100

91 71.7 56.5 84.7

206 21.8 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 40.729 ng/ul

RT: 21.439 min Scan# 3123

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44




Tgt Ion:252 Resp: 365012


Ion Ratio Lower Upper

252 100

254 65.4 50.7 76.1

126 9.3 6.6 10.0









#85
Benzo(a)anthracene
Concen: 35.508 ng/ul
RT: 21.521 min Scan# 3137
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:228 Resp: 943825
Ion Ratio Lower Upper
228 100
229 20.6 16.0 24.0
226 28.3 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 36.141 ng/ul
RT: 21.421 min Scan# 3120
Delta R.T. -0.004 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:149 Resp: 571128
Ion Ratio Lower Upper
149 100
167 27.7 20.9 31.3
279 3.7 3.1 4.7





#87

Chrysene

Concen: 36.308 ng/ul

RT: 21.586 min Scan# 3148

Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:228 Resp: 904698

Ion Ratio Lower Upper

228 100

226 31.1 24.4 36.6

229 20.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 35.398 ng/ul




RT: 22.602 min Scan# 3321


Delta R.T. -0.004 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Tgt Ion:149 Resp: 1020102









#90
Benzo(b)fluoranthene
Concen: 34.896 ng/ul
RT: 23.689 min Scan# 3505
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44




Instrument :
BNA_G
ClientSampleId :
SLCS013


Tgt Ion:252 Resp: 1013441
Ion Ratio Lower Upper
252 100
253 22.7 18.1 27.1
125 8.7 7.5 11.3



#91
Benzo(k)fluoranthene
Concen: 35.123 ng/ul
RT: 23.754 min Scan# 3517
Delta R.T. 0.002 min
Lab File: BG058253.D
Acq: 15 Jul 2023 7:44

Tgt Ion:252 Resp: 999786
Ion Ratio Lower Upper
252 100
253 22.7 17.7 26.5
125 9.0 7.4 11.2





#93

Benzo(a)pyrene

Concen: 35.414 ng/ul

RT: 24.547 min Scan# 3

Delta R.T. 0.008 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44

Instrument :

BNA_G

ClientSampleId :

SLCS013





Tgt Ion:252 Resp: 910299

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.6

125 9.7 8.3 12.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 35.378 ng/ul

RT: 28.272 min Scan# 4286

Delta R.T. 0.002 min

Lab File: BG058253.D

Acq: 15 Jul 2023 7:44




Tgt Ion:276 Resp: 1207585


Ion Ratio Lower Upper

276 100

138 17.1 14.0 21.0

277 24.6 19.4 29.2









#95
 Dibenzo(a,h)anthracene
 Concen: 35.932 ng/ul
 RT: 28.331 min Scan# 41
 Delta R.T. 0.008 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Instrument :
 BNA_G
 ClientSampleId :
 SLCS013

Tgt Ion:278 Resp: 1011378
 Ion Ratio Lower Upper
 278 100
 139 13.2 11.3 16.9
 279 24.1 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 34.823 ng/ul
 RT: 29.394 min Scan# 4477
 Delta R.T. 0.002 min
 Lab File: BG058253.D
 Acq: 15 Jul 2023 7:44

Tgt Ion:276 Resp: 970862
 Ion Ratio Lower Upper
 276 100
 138 18.5 14.2 21.2
 277 23.6 18.6 27.8

