

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071323\
 Data File : BG058261.D
 Acq On : 15 Jul 2023 14:38
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020518

Quant Time: Jul 15 22:36:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 13 14:01:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.901	152	53808	20.000	ng/ul	0.00
20) Naphthalene-d8	10.704	136	263640	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.541	164	185037	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.285	188	427445	20.000	ng/ul	0.00
79) Chrysene-d12	21.538	240	378885	20.000	ng/ul	0.00
88) Perylene-d12	24.682	264	476091	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	13033	8.482	ng/uL	0.00
4) Pyridine-d5	3.753	84	99202	21.265	ng/ul	0.00
7) Phenol-d5	7.067	99	122055	22.211	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.232	67	71961	21.730	ng/ul	0.00
11) 2-Chlorophenol-d4	7.437	132	82852	21.861	ng/ul	0.00
15) 4-Methylphenol-d8	8.606	113	91366	22.035	ng/ul	0.00
21) Nitrobenzene-d5	9.065	128	45009	20.946	ng/ul	0.00
24) 2-Nitrophenol-d4	9.787	143	50883	22.531	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	97045	22.400	ng/ul	0.00
31) 4-Chloroaniline-d4	10.839	131	132703	20.203	ng/ul	0.00
46) Dimethylphthalate-d6	13.941	166	284794	19.507	ng/ul	0.00
49) Acenaphthylene-d8	14.235	160	336606	21.128	ng/ul	0.00
54) 4-Nitrophenol-d4	14.746	143	45122	17.493	ng/ul	0.00
60) Fluorene-d10	15.534	176	258896	20.880	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	44893	19.342	ng/ul	0.00
73) Anthracene-d10	17.384	188	417185	21.033	ng/ul	0.00
81) Pyrene-d10	19.670	212	480581	20.140	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.458	264	495658	20.435	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.383	88	14510	7.644	ng/uL	95
5) Pyridine	3.777	79	103900	21.737	ng/ul	97
6) Benzaldehyde	7.044	77	47851	26.075	ng/ul	96
8) Phenol	7.091	94	127999	22.954	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.320	93	92209	21.416	ng/ul	99
12) 2-Chlorophenol	7.467	128	88497	22.439	ng/ul	95
13) 2-Methylphenol	8.342	108	97327	22.224	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.430	45	152965	23.079	ng/ul	99
16) Acetophenone	8.724	105	150912	21.937	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.706	70	78009	21.212	ng/ul	97
18) 4-Methylphenol	8.671	108	104877	22.147	ng/ul	97
19) Hexachloroethane	8.982	117	34405	22.756	ng/ul	98
22) Nitrobenzene	9.106	77	116521	20.889	ng/ul	97
23) Isophorone	9.623	82	236734	19.889	ng/ul	98
25) 2-Nitrophenol	9.817	139	52625	21.716	ng/ul#	91
26) 2,4-Dimethylphenol	9.876	107	116255	21.334	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.111	93	134341	20.029	ng/ul	97
29) 2,4-Dichlorophenol	10.351	162	93770	22.261	ng/ul	94
30) Naphthalene	10.757	128	311334	21.013	ng/ul	97
32) 4-Chloroaniline	10.863	127	137382	20.455	ng/ul	95
33) Hexachlorobutadiene	11.033	225	65714	22.036	ng/ul	97
34) Caprolactam	11.621	113	31472	18.584	ng/ul	90
35) 4-Chloro-3-methylphenol	11.991	107	113151	21.868	ng/ul	97
36) 2-Methylnaphthalene	12.361	142	210108	21.037	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071323\
 Data File : BG058261.D
 Acq On : 15 Jul 2023 14:38
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020518

Quant Time: Jul 15 22:36:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 13 14:01:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.584	142	214763	21.443	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.731	216	127528	22.616	ng/ul	98
40) Hexachlorocyclopentadiene	12.708	237	65263	18.375	ng/ul	98
41) 2,4,6-Trichlorophenol	12.972	196	85375	22.339	ng/ul	97
42) 2,4,5-Trichlorophenol	13.048	196	90660	21.934	ng/ul	99
43) 1,1'-Biphenyl	13.371	154	281366	21.100	ng/ul	99
44) 2-Chloronaphthalene	13.418	162	225686	21.088	ng/ul	99
45) 2-Nitroaniline	13.624	65	80387	21.604	ng/ul	97
47) Dimethylphthalate	13.988	163	289854	19.761	ng/ul	99
48) 2,6-Dinitrotoluene	14.118	165	63059	20.944	ng/ul	96
50) Acenaphthylene	14.265	152	361666	21.022	ng/ul	99
51) 3-Nitroaniline	14.447	138	49621	19.502	ng/ul#	98
52) Acenaphthene	14.605	153	248425	21.043	ng/ul	98
53) 2,4-Dinitrophenol	14.658	184	31649	20.163	ng/ul	94
55) 4-Nitrophenol	14.758	109	33091	17.318	ng/ul	91
56) Dibenzofuran	14.940	168	352200	20.987	ng/ul	99
57) 2,4-Dinitrotoluene	14.905	165	89627	21.011	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.163	232	79499	21.128	ng/ul#	96
59) Diethylphthalate	15.351	149	298330	19.487	ng/ul	99
61) Fluorene	15.586	166	283922	20.620	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.575	204	152800	20.972	ng/ul	99
63) 4-Nitroaniline	15.610	138	33382	14.646	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.669	198	45501	19.170	ng/ul	98
67) N-Nitrosodiphenylamine	15.792	169	241972	21.271	ng/ul	97
68) 4-Bromophenyl-phenylether	16.474	248	103054	21.964	ng/ul	99
69) Hexachlorobenzene	16.591	284	115956	22.073	ng/ul	99
70) Atrazine	16.738	200	95425	20.513	ng/ul	97
71) Pentachlorophenol	16.938	266	58675	18.605	ng/ul	96
72) Phenanthrene	17.326	178	451900	20.784	ng/ul	99
74) Anthracene	17.420	178	460920	20.948	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.342	216	131567	23.209	ng/uL	95
76) Pentachlorobenzene	14.858	250	139044	22.966	ng/uL	98
77) Carbazole	17.690	167	416736	20.942	ng/ul	98
78) Di-n-butylphthalate	18.242	149	519552	20.496	ng/ul	100
80) Fluoranthene	19.341	202	552625	19.986	ng/ul	98
82) Pyrene	19.705	202	559339	20.096	ng/ul	99
83) Butylbenzylphthalate	20.587	149	236402	20.946	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.438	252	206817	22.548	ng/ul	99
85) Benzo(a)anthracene	21.521	228	566174	20.812	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.421	149	342129	21.153	ng/ul	98
87) Chrysene	21.579	228	535201	20.986	ng/ul	99
89) Di-n-octyl phthalate	22.596	149	604133	20.653	ng/ul	100
90) Benzo(b)fluoranthene	23.677	252	597818	20.280	ng/ul	100
91) Benzo(k)fluoranthene	23.742	252	586659	20.304	ng/ul	99
93) Benzo(a)pyrene	24.529	252	528485	20.255	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.254	276	694519	20.045	ng/ul	99
95) Dibenzo(a,h)anthracene	28.313	278	574252	20.100	ng/ul	96
96) Benzo(g,h,i)perylene	29.370	276	538901	19.043	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.644 ng/uL
RT: 3.383 min Scan# 50
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion: 88 Resp: 14510
Ion Ratio Lower Upper
88 100
43 29.9 20.6 30.8
58 85.6 65.5 98.3



#5
Pyridine
Concen: 21.737 ng/uL
RT: 3.777 min Scan# 117
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion: 79 Resp: 103900
Ion Ratio Lower Upper
79 100
52 83.9 68.6 102.8
51 40.1 29.4 44.0









#6
Benzaldehyde
Concen: 26.075 ng/ul
RT: 7.044 min Scan# 672
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion: 77 Resp: 47851
Ion Ratio Lower Upper
77 100
105 99.6 74.7 112.1
106 96.9 75.9 113.9



#8
Phenol
Concen: 22.954 ng/ul
RT: 7.091 min Scan# 681
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion: 94 Resp: 127999
Ion Ratio Lower Upper
94 100
65 30.1 22.2 33.4
66 35.9 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 21.416 ng/ul
RT: 7.320 min Scan# 720
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion: 93 Resp: 92209
Ion Ratio Lower Upper
93 100
63 86.4 69.4 104.0
95 32.6 27.0 40.6



#12
2-Chlorophenol
Concen: 22.439 ng/ul
RT: 7.467 min Scan# 745
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:128 Resp: 88497
Ion Ratio Lower Upper
128 100
64 59.4 44.3 66.5
130 32.3 27.3 40.9





#13

2-Methylphenol

Concen: 22.224 ng/ul

RT: 8.342 min Scan# 894

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :





SSTD020518

Tgt Ion:108 Resp: 97327

Ion Ratio Lower Upper

108 100

107 89.7 75.0 112.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 23.079 ng/ul

RT: 8.430 min Scan# 909

Delta R.T. 0.008 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38




Tgt Ion: 45 Resp: 152965


Ion Ratio Lower Upper

45 100

77 12.4 9.4 14.2

79 10.3 8.3 12.5









#16
Acetophenone
Concen: 21.937 ng/ul
RT: 8.724 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:105 Resp: 150912
Ion Ratio Lower Upper
105 100
77 77.8 61.2 91.8
51 27.6 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 21.212 ng/ul
RT: 8.706 min Scan# 956
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion: 70 Resp: 78009
Ion Ratio Lower Upper
70 100
42 58.0 48.0 72.0
101 9.3 7.4 11.2
130 18.5 13.4 20.0









#18
4-Methylphenol
Concen: 22.147 ng/ul
RT: 8.671 min Scan# 91
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:108 Resp: 104877
Ion Ratio Lower Upper
108 100
107 110.1 91.1 136.7



#19
Hexachloroethane
Concen: 22.756 ng/ul
RT: 8.982 min Scan# 1003
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:117 Resp: 34405
Ion Ratio Lower Upper
117 100
201 119.0 94.0 141.0
199 77.9 59.8 89.6





#22

Nitrobenzene

Concen: 20.889 ng/ul

RT: 9.106 min Scan# 11024

Delta R.T. 0.002 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion: 77 Resp: 116521

Ion Ratio Lower Upper

77 100

123 44.0 33.3 49.9

65 13.3 10.4 15.6



#23

Isophorone

Concen: 19.889 ng/ul

RT: 9.623 min Scan# 11112

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38




Tgt Ion: 82 Resp: 236734


Ion Ratio Lower Upper

82 100

95 6.3 4.8 7.2

138 14.9 11.4 17.0









#25
2-Nitrophenol
Concen: 21.716 ng/ul
RT: 9.817 min Scan# 1145
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:139 Resp: 52625
Ion Ratio Lower Upper
139 100
65 58.3 42.6 64.0
109 31.9 20.7 31.1#



#26
2,4-Dimethylphenol
Concen: 21.334 ng/ul
RT: 9.876 min Scan# 1155
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:107 Resp: 116255
Ion Ratio Lower Upper
107 100
121 48.1 40.6 60.8
122 81.9 67.1 100.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.029 ng/ul
 RT: 10.111 min Scan# 1195
 Delta R.T. 0.002 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020518

Tgt Ion: 93 Resp: 134341




Ion	Ratio	Lower	Upper
93	100		
95	30.7	26.4	39.6
123	10.3	7.9	11.9




#29
 2,4-Dichlorophenol
 Concen: 22.261 ng/ul
 RT: 10.351 min Scan# 1236
 Delta R.T. -0.004 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38

Tgt Ion:162 Resp: 93770

Ion	Ratio	Lower	Upper
162	100		
164	68.7	51.1	76.7
98	38.7	27.9	41.9









#30
Naphthalene
Concen: 21.013 ng/ul
RT: 10.757 min Scan# 11
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:128 Resp: 311334
Ion Ratio Lower Upper
128 100
129 11.9 8.6 12.8
127 14.1 10.2 15.2



#32
4-Chloroaniline
Concen: 20.455 ng/ul
RT: 10.863 min Scan# 1323
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:127 Resp: 137382
Ion Ratio Lower Upper
127 100
129 29.7 26.0 39.0









#33
Hexachlorobutadiene
Concen: 22.036 ng/ul
RT: 11.033 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:225 Resp: 65714
Ion Ratio Lower Upper
225 100
223 62.3 53.2 79.8
227 63.6 51.4 77.2



#34
Caprolactam
Concen: 18.584 ng/ul
RT: 11.621 min Scan# 1452
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:113 Resp: 31472
Ion Ratio Lower Upper
113 100
55 236.7 183.4 275.2
56 176.1 123.0 184.6





#35

4-Chloro-3-methylphenol

Concen: 21.868 ng/ul

RT: 11.991 min Scan# 11

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:107 Resp: 113151

Ion Ratio Lower Upper

107 100

144 25.1 19.0 28.4

142 75.7 58.4 87.6



#36

2-Methylnaphthalene

Concen: 21.037 ng/ul

RT: 12.361 min Scan# 1578

Delta R.T. -0.004 min

Lab File: BG058261.D




Acq: 15 Jul 2023 14:38


Tgt Ion:142 Resp: 210108

Ion Ratio Lower Upper

142 100

141 86.6 69.4 104.2





#37

1-Methylnaphthalene

Concen: 21.443 ng/ul

RT: 12.584 min Scan# 1

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:142 Resp: 214763

Ion Ratio Lower Upper

142 100

141 91.9 73.8 110.8

116 4.0 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.616 ng/ul

RT: 12.731 min Scan# 1641

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Tgt Ion:216 Resp: 127528




Ion Ratio Lower Upper


216 100

214 77.9 64.1 96.1

179 19.8 16.0 24.0

108 14.3 12.1 18.1









#40
Hexachlorocyclopentadiene
Concen: 18.375 ng/ul
RT: 12.708 min Scan# 1637
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:237 Resp: 65263
Ion Ratio Lower Upper
237 100
235 62.6 49.2 73.8
272 12.3 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 22.339 ng/ul
RT: 12.972 min Scan# 1682
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:196 Resp: 85375
Ion Ratio Lower Upper
196 100
198 94.4 73.5 110.3
200 29.7 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 21.934 ng/ul
RT: 13.048 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:196 Resp: 90660
Ion Ratio Lower Upper
196 100
198 97.1 77.2 115.8
200 31.9 23.8 35.6



#43
1,1'-Biphenyl
Concen: 21.100 ng/ul
RT: 13.371 min Scan# 1750
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:154 Resp: 281366
Ion Ratio Lower Upper
154 100
153 40.3 32.6 49.0
76 14.3 11.6 17.4









#44
2-Chloronaphthalene
Concen: 21.088 ng/ul
RT: 13.418 min Scan# 1758
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:162 Resp: 225686
Ion Ratio Lower Upper
162 100
164 35.0 27.7 41.5
127 39.1 30.9 46.3



#45
2-Nitroaniline
Concen: 21.604 ng/ul
RT: 13.624 min Scan# 1793
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion: 65 Resp: 80387
Ion Ratio Lower Upper
65 100
92 63.2 53.2 79.8
138 90.1 71.1 106.7









#47
Dimethylphthalate
Concen: 19.761 ng/ul
RT: 13.988 min Scan# 1855
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:163 Resp: 289854
Ion Ratio Lower Upper
163 100
194 3.2 3.0 4.4
164 11.1 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 20.944 ng/ul
RT: 14.118 min Scan# 1877
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:165 Resp: 63059
Ion Ratio Lower Upper
165 100
89 56.8 42.7 64.1
121 20.0 14.8 22.2









#50
Acenaphthylene
Concen: 21.022 ng/ul
RT: 14.265 min Scan# 1902
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:152 Resp: 361666
Ion Ratio Lower Upper
152 100
151 20.4 15.6 23.4
153 13.4 10.5 15.7



#51
3-Nitroaniline
Concen: 19.502 ng/ul
RT: 14.447 min Scan# 1933
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:138 Resp: 49621
Ion Ratio Lower Upper
138 100
108 12.4 8.2 12.2#
92 127.3 100.2 150.4





#52

Acenaphthene

Concen: 21.043 ng/ul

RT: 14.605 min Scan# 1960

Delta R.T. 0.002 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:153 Resp: 248425

Ion Ratio Lower Upper

153 100

152 47.3 36.5 54.7

154 86.0 70.5 105.7



#53

2,4-Dinitrophenol

Concen: 20.163 ng/ul

RT: 14.658 min Scan# 1969

Delta R.T. 0.002 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38




Tgt Ion:184 Resp: 31649


Ion Ratio Lower Upper

184 100

63 71.2 51.1 76.7

154 55.1 43.3 64.9









#55
4-Nitrophenol
Concen: 17.318 ng/ul
RT: 14.758 min Scan# 1986
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:109 Resp: 33091
Ion Ratio Lower Upper
109 100
139 140.3 119.0 178.4
65 153.9 135.5 203.3



#56
Dibenzofuran
Concen: 20.987 ng/ul
RT: 14.940 min Scan# 1717
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:168 Resp: 352200
Ion Ratio Lower Upper
168 100
139 35.3 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 21.011 ng/ul

RT: 14.905 min Scan# 2011

Delta R.T. 0.002 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:165 Resp: 89627

Ion Ratio Lower Upper

165 100

63 44.0 35.8 53.8

182 2.3 1.6 2.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.128 ng/ul

RT: 15.163 min Scan# 2055

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Tgt Ion:232 Resp: 79499




Ion Ratio Lower Upper


232 100

131 40.6 30.5 45.7

130 2.1 1.4 2.0#

166 30.8 22.6 34.0









#59
Diethylphthalate
Concen: 19.487 ng/ul
RT: 15.351 min Scan# 2087
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:149 Resp: 298330
Ion Ratio Lower Upper
149 100
177 22.8 17.8 26.8
150 12.0 10.1 15.1



#61
Fluorene
Concen: 20.620 ng/ul
RT: 15.586 min Scan# 2127
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:166 Resp: 283922
Ion Ratio Lower Upper
166 100
165 97.9 79.1 118.7
167 13.8 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 20.972 ng/ul
RT: 15.575 min Scan# 2125
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:204 Resp: 152800
Ion Ratio Lower Upper
204 100
206 31.9 25.8 38.6
141 51.1 41.8 62.6



#63
4-Nitroaniline
Concen: 14.646 ng/ul
RT: 15.610 min Scan# 2131
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:138 Resp: 33382
Ion Ratio Lower Upper
138 100
92 53.3 43.8 65.8
108 75.3 58.8 88.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.170 ng/ul
 RT: 15.669 min Scan# 2141
 Delta R.T. 0.002 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020518


Tgt Ion:198 Resp: 45501
 Ion Ratio Lower Upper
 198 100
 182 3.8 3.1 4.7
 77 23.5 17.9 26.9



#67
 N-Nitrosodiphenylamine
 Concen: 21.271 ng/ul
 RT: 15.792 min Scan# 2162
 Delta R.T. -0.004 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38

Tgt Ion:169 Resp: 241972
 Ion Ratio Lower Upper
 169 100
 168 71.0 54.2 81.2
 167 38.1 30.8 46.2









#68
4-Bromophenyl-phenylether
Concen: 21.964 ng/ul
RT: 16.474 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:248 Resp: 103054
Ion Ratio Lower Upper
248 100
250 99.1 78.3 117.5
141 60.8 49.1 73.7



#69
Hexachlorobenzene
Concen: 22.073 ng/ul
RT: 16.591 min Scan# 2298
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:284 Resp: 115956
Ion Ratio Lower Upper
284 100
142 26.9 21.9 32.9
249 31.9 25.0 37.6





#70

Atrazine

Concen: 20.513 ng/ul

RT: 16.738 min Scan# 21

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:200 Resp: 95425

Ion Ratio Lower Upper

200 100

173 26.1 20.0 30.0

215 47.1 35.8 53.8



#71

Pentachlorophenol

Concen: 18.605 ng/ul

RT: 16.938 min Scan# 2357

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38




Tgt Ion:266 Resp: 58675


Ion Ratio Lower Upper

266 100

264 59.5 51.5 77.3

268 61.8 49.0 73.4









#72
Phenanthrene
Concen: 20.784 ng/ul
RT: 17.326 min Scan# 2424
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:178 Resp: 451900
Ion Ratio Lower Upper
178 100
179 15.8 12.8 19.2
176 20.3 15.7 23.5



#74
Anthracene
Concen: 20.948 ng/ul
RT: 17.420 min Scan# 2439
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:178 Resp: 460920
Ion Ratio Lower Upper
178 100
179 16.0 12.6 18.8
176 19.7 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.209 ng/uL
RT: 13.342 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:216 Resp: 131567
Ion Ratio Lower Upper
216 100
214 75.7 64.8 97.2
218 47.1 39.1 58.7



#76
Pentachlorobenzene
Concen: 22.966 ng/uL
RT: 14.858 min Scan# 2003
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:250 Resp: 139044
Ion Ratio Lower Upper
250 100
248 63.3 52.8 79.2
252 65.0 52.4 78.6









#77
 Carbazole
 Concen: 20.942 ng/ul
 RT: 17.690 min Scan# 2485
 Delta R.T. -0.004 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38




Instrument :
 BNA_G
 ClientSampleId :
 SST020518


Tgt Ion:167 Resp: 416736
 Ion Ratio Lower Upper
 167 100
 166 22.9 17.5 26.3
 139 12.2 9.5 14.3



#78
 Di-n-butylphthalate
 Concen: 20.496 ng/ul
 RT: 18.242 min Scan# 2579
 Delta R.T. -0.004 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38

Tgt Ion:149 Resp: 519552
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.2
 104 4.8 3.9 5.9









#80
Fluoranthene
Concen: 19.986 ng/ul
RT: 19.341 min Scan# 2766
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:202 Resp: 552625
Ion Ratio Lower Upper
202 100
101 10.5 9.1 13.7
100 8.1 7.1 10.7



#82
Pyrene
Concen: 20.096 ng/ul
RT: 19.705 min Scan# 2828
Delta R.T. 0.002 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:202 Resp: 559339
Ion Ratio Lower Upper
202 100
101 13.0 10.2 15.2
100 10.8 8.9 13.3









#83
Butylbenzylphthalate
Concen: 20.946 ng/ul
RT: 20.587 min Scan# 2978
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:149 Resp: 236402
Ion Ratio Lower Upper
149 100
91 71.9 56.5 84.7
206 21.9 17.4 26.0



#84
3,3'-Dichlorobenzidine
Concen: 22.548 ng/ul
RT: 21.438 min Scan# 3123
Delta R.T. -0.004 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:252 Resp: 206817
Ion Ratio Lower Upper
252 100
254 64.4 50.7 76.1
126 8.7 6.6 10.0





#85

Benzo(a)anthracene

Concen: 20.812 ng/ul

RT: 21.521 min Scan# 3137

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:228 Resp: 566174

Ion Ratio Lower Upper

228 100

229 19.2 16.0 24.0

226 27.0 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.153 ng/ul

RT: 21.421 min Scan# 3120

Delta R.T. -0.004 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38




Tgt Ion:149 Resp: 342129


Ion Ratio Lower Upper

149 100

167 27.0 20.9 31.3

279 4.1 3.1 4.7





#87

Chrysene

Concen: 20.986 ng/ul

RT: 21.579 min Scan# 3148

Delta R.T. -0.010 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:228 Resp: 535201

Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 20.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 20.653 ng/ul




RT: 22.596 min Scan# 3320


Delta R.T. -0.010 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Tgt Ion:149 Resp: 604133





#90

Benzo(b)fluoranthene

Concen: 20.280 ng/ul

RT: 23.677 min Scan# 3504

Delta R.T. -0.010 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD020518





Tgt Ion:252 Resp: 597818

Ion Ratio Lower Upper

252 100

253 22.6 18.1 27.1

125 9.3 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 20.304 ng/ul

RT: 23.742 min Scan# 3515

Delta R.T. -0.010 min

Lab File: BG058261.D

Acq: 15 Jul 2023 14:38




Tgt Ion:252 Resp: 586659


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 9.0 7.4 11.2









#93
Benzo(a)pyrene
Concen: 20.255 ng/ul
RT: 24.529 min Scan# 3649
Delta R.T. -0.010 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD020518


Tgt Ion:252 Resp: 528485
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 9.9 8.3 12.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.045 ng/ul
RT: 28.254 min Scan# 4283
Delta R.T. -0.016 min
Lab File: BG058261.D
Acq: 15 Jul 2023 14:38

Tgt Ion:276 Resp: 694519
Ion Ratio Lower Upper
276 100
138 17.3 14.0 21.0
277 24.6 19.4 29.2









#95
 Dibenzo(a,h)anthracene
 Concen: 20.100 ng/ul
 RT: 28.313 min Scan# 41
 Delta R.T. -0.010 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38

Instrument :
 BNA_G
 ClientSampleId :
 SST020518

Tgt Ion:278 Resp: 574252
 Ion Ratio Lower Upper
 278 100
 139 12.1 11.3 16.9
 279 25.0 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 19.043 ng/ul
 RT: 29.370 min Scan# 4473
 Delta R.T. -0.022 min
 Lab File: BG058261.D
 Acq: 15 Jul 2023 14:38

Tgt Ion:276 Resp: 538901
 Ion Ratio Lower Upper
 276 100
 138 17.7 14.2 21.2
 277 23.4 18.6 27.8

