

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071723\
 Data File : BG058283.D
 Acq On : 17 Jul 2023 15:11
 Operator : CG/JU
 Sample : SST00592
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST005442

Quant Time: Jul 18 01:29:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 18 01:24:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	46260	20.000	ng/ul	0.00
20) Naphthalene-d8	10.702	136	200845	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.539	164	147075	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.282	188	372644	20.000	ng/ul	0.00
79) Chrysene-d12	21.530	240	327685	20.000	ng/ul	0.00
88) Perylene-d12	24.668	264	385392	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	77	0.058	ng/uL	0.04
4) Pyridine-d5	3.751	84	13808	3.443	ng/ul	0.00
7) Phenol-d5	7.065	99	18660	3.950	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.224	67	13136	4.614	ng/ul	0.00
11) 2-Chlorophenol-d4	7.429	132	14013	4.301	ng/ul	0.00
15) 4-Methylphenol-d8	8.604	113	13996	3.926	ng/ul	0.00
21) Nitrobenzene-d5	9.063	128	6705	4.096	ng/ul	0.00
24) 2-Nitrophenol-d4	9.785	143	6939	4.033	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.326	165	13901	4.212	ng/ul	0.00
31) 4-Chloroaniline-d4	10.843	131	21872	4.371	ng/ul	0.00
46) Dimethylphthalate-d6	13.939	166	58880	5.074	ng/ul	0.00
49) Acenaphthylene-d8	14.233	160	57638	4.552	ng/ul	0.00
54) 4-Nitrophenol-d4	14.738	143	5965	2.909	ng/ul	0.00
60) Fluorene-d10	15.532	176	46041	4.672	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.649	200	6667	3.295	ng/ul	0.00
73) Anthracene-d10	17.382	188	80469	4.654	ng/ul	0.00
81) Pyrene-d10	19.668	212	95533	4.629	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.450	264	89035	4.535	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.375	88	2564	1.571	ng/uL#	80
5) Pyridine	3.775	79	17880	4.351	ng/ul#	86
6) Benzaldehyde	7.041	77	4721	2.992	ng/ul#	86
8) Phenol	7.094	94	20183	4.210	ng/ul#	90
10) Bis(2-Chloroethyl)ether	7.318	93	15993	4.320	ng/ul	94
12) 2-Chlorophenol	7.465	128	15051	4.439	ng/ul	98
13) 2-Methylphenol	8.340	108	15268	4.055	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.428	45	27472	4.821	ng/ul	97
16) Acetophenone	8.722	105	27081	4.579	ng/ul#	92
17) N-Nitroso-di-n-propyla...	8.698	70	14542	4.599	ng/ul	98
18) 4-Methylphenol	8.669	108	16284	4.000	ng/ul	97
19) Hexachloroethane	8.980	117	6072	4.671	ng/ul#	80
22) Nitrobenzene	9.104	77	18473	4.347	ng/ul	96
23) Isophorone	9.621	82	44040	4.857	ng/ul	99
25) 2-Nitrophenol	9.815	139	7517	4.072	ng/ul#	91
26) 2,4-Dimethylphenol	9.873	107	18238	4.393	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.108	93	22847	4.471	ng/ul	97
29) 2,4-Dichlorophenol	10.355	162	13389	4.172	ng/ul	94
30) Naphthalene	10.749	128	53764	4.763	ng/ul	98
32) 4-Chloroaniline	10.866	127	22037	4.307	ng/ul	98
33) Hexachlorobutadiene	11.031	225	10202	4.491	ng/ul	94
34) Caprolactam	11.613	113	5282	4.094	ng/ul	93
35) 4-Chloro-3-methylphenol	11.983	107	16537	4.195	ng/ul	97
36) 2-Methylnaphthalene	12.365	142	33838	4.447	ng/ul	95


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071723\
 Data File : BG058283.D
 Acq On : 17 Jul 2023 15:11
 Operator : CG/JU
 Sample : SSTD00592
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005442

Quant Time: Jul 18 01:29:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 18 01:24:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.582	142	36095	4.731	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.729	216	20788	4.638	ng/ul#	94
40) Hexachlorocyclopentadiene	12.705	237	6632	2.349	ng/ul	95
41) 2,4,6-Trichlorophenol	12.970	196	12982	4.274	ng/ul	94
42) 2,4,5-Trichlorophenol	13.046	196	13728	4.179	ng/ul	98
43) 1,1'-Biphenyl	13.369	154	49940	4.712	ng/ul	97
44) 2-Chloronaphthalene	13.416	162	38873	4.570	ng/ul	95
45) 2-Nitroaniline	13.622	65	11652	3.940	ng/ul	99
47) Dimethylphthalate	13.986	163	59838	5.133	ng/ul	100
48) 2,6-Dinitrotoluene	14.110	165	9040	3.778	ng/ul#	93
50) Acenaphthylene	14.257	152	63980	4.679	ng/ul	98
51) 3-Nitroaniline	14.445	138	7052	3.487	ng/ul	98
52) Acenaphthene	14.603	153	44996	4.795	ng/ul	98
53) 2,4-Dinitrophenol	14.662	184	2926	2.345	ng/ul	98
55) 4-Nitrophenol	14.756	109	3738	2.461	ng/ul	92
56) Dibenzofuran	14.932	168	63880	4.789	ng/ul	99
57) 2,4-Dinitrotoluene	14.897	165	13671	4.032	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.161	232	12894	4.311	ng/ul#	90
59) Diethylphthalate	15.349	149	60760	4.993	ng/ul	98
61) Fluorene	15.584	166	54321	4.963	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.579	204	27810	4.802	ng/ul	94
63) 4-Nitroaniline	15.602	138	2594	1.432	ng/ul#	88
66) 4,6-Dinitro-2-methylph...	15.661	198	6688	3.232	ng/ul#	92
67) N-Nitrosodiphenylamine	15.790	169	44327	4.470	ng/ul	99
68) 4-Bromophenyl-phenylether	16.472	248	18497	4.522	ng/ul	97
69) Hexachlorobenzene	16.589	284	21557	4.707	ng/ul	96
70) Atrazine	16.736	200	17812	4.392	ng/ul	98
71) Pentachlorophenol	16.936	266	8965	3.261	ng/ul	94
72) Phenanthrene	17.324	178	89150	4.703	ng/ul	96
74) Anthracene	17.412	178	91460	4.768	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.340	216	21576	4.366	ng/uL	93
76) Pentachlorobenzene	14.856	250	24118	4.569	ng/uL	96
77) Carbazole	17.688	167	80027	4.613	ng/ul	100
78) Di-n-butylphthalate	18.240	149	95780	4.334	ng/ul	99
80) Fluoranthene	19.333	202	110184	4.608	ng/ul	99
82) Pyrene	19.697	202	115308	4.790	ng/ul	98
83) Butylbenzylphthalate	20.584	149	41234	4.224	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.430	252	35409	4.464	ng/ul	95
85) Benzo(a)anthracene	21.513	228	109145	4.639	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.413	149	60591	4.332	ng/ul	99
87) Chrysene	21.577	228	101674	4.610	ng/ul	98
89) Di-n-octyl phthalate	22.588	149	101219	4.275	ng/ul	100
90) Benzo(b)fluoranthene	23.663	252	105685	4.429	ng/ul	98
91) Benzo(k)fluoranthene	23.734	252	111108	4.750	ng/ul	99
93) Benzo(a)pyrene	24.521	252	95881	4.540	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.228	276	123260	4.395	ng/ul	98
95) Dibenzo(a,h)anthracene	28.293	278	97612	4.221	ng/ul	99
96) Benzo(g,h,i)perylene	29.339	276	98182	4.286	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane

Concen: 1.571 ng/uL

RT: 3.375 min Scan# 49

Delta R.T. 0.000 min

Lab File: BG058283.D


Acq: 17 Jul 2023 15:11

Instrument :




BNA_G

ClientSampleId :

SSTD005442



Tgt Ion:	88	Resp:	2564
Ion Ratio	Lower	Upper	
88	100		
43	30.4	19.6	29.4#
58	66.7	70.2	105.4#



#5

Pyridine

Concen: 4.351 ng/uL




RT: 3.775 min Scan# 117


Delta R.T. 0.006 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Tgt Ion:	79	Resp:	17880
Ion Ratio	Lower	Upper	
79	100		
52	82.4	72.3	108.5
51	57.5	33.4	50.0#









#6
Benzaldehyde
Concen: 2.992 ng/ul
RT: 7.041 min Scan# 61
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion: 77 Resp: 4721
Ion Ratio Lower Upper
77 100
105 99.3 78.2 117.4
106 115.4 71.6 107.4#



#8
Phenol
Concen: 4.210 ng/ul
RT: 7.094 min Scan# 682
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion: 94 Resp: 20183
Ion Ratio Lower Upper
94 100
65 28.1 23.8 35.8
66 29.8 31.2 46.8#









#10
Bis(2-Chloroethyl)ether
Concen: 4.320 ng/ul
RT: 7.318 min Scan# 720
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion: 93	Resp:	15993
Ion Ratio	Lower	Upper
93	100	
63	91.1	67.7 101.5
95	31.9	27.0 40.6



#12
2-Chlorophenol
Concen: 4.439 ng/ul
RT: 7.465 min Scan# 745
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion: 128	Resp:	15051
Ion Ratio	Lower	Upper
128	100	
64	55.4	45.6 68.4
130	30.1	25.2 37.8









#13
2-Methylphenol
Concen: 4.055 ng/ul
RT: 8.340 min Scan# 894
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:108 Resp: 15268
Ion Ratio Lower Upper
108 100
107 86.3 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.821 ng/ul
RT: 8.428 min Scan# 909
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion: 45 Resp: 27472
Ion Ratio Lower Upper
45 100
77 13.2 9.4 14.2
79 10.3 7.4 11.2









#16
Acetophenone
Concen: 4.579 ng/ul
RT: 8.722 min Scan# 911
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:105 Resp: 27081
Ion Ratio Lower Upper
105 100
77 80.8 60.7 91.1
51 36.5 23.5 35.3#



#17
N-Nitroso-di-n-propylamine
Concen: 4.599 ng/ul
RT: 8.698 min Scan# 955
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion: 70 Resp: 14542
Ion Ratio Lower Upper
70 100
42 56.7 46.6 69.8
101 8.4 7.3 10.9
130 17.9 14.2 21.2









#18
4-Methylphenol
Concen: 4.000 ng/ul
RT: 8.669 min Scan# 91
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:108 Resp: 16284
Ion Ratio Lower Upper
108 100
107 116.8 90.6 136.0



#19
Hexachloroethane
Concen: 4.671 ng/ul
RT: 8.980 min Scan# 1003
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:117 Resp: 6072
Ion Ratio Lower Upper
117 100
201 102.8 103.6 155.4#
199 62.2 60.1 90.1









#22
Nitrobenzene
Concen: 4.347 ng/ul
RT: 9.104 min Scan# 11024
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion: 77 Resp: 18473
Ion Ratio Lower Upper
77 100
123 44.8 34.0 51.0
65 16.3 11.0 16.6



#23
Isophorone
Concen: 4.857 ng/ul
RT: 9.621 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion: 82 Resp: 44040
Ion Ratio Lower Upper
82 100
95 7.4 5.1 7.7
138 14.8 11.9 17.9









#25
2-Nitrophenol
Concen: 4.072 ng/ul
RT: 9.815 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:139 Resp: 7517
Ion Ratio Lower Upper
139 100
65 67.1 49.2 73.8
109 33.1 21.6 32.4#



#26
2,4-Dimethylphenol
Concen: 4.393 ng/ul
RT: 9.873 min Scan# 1155
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:107 Resp: 18238
Ion Ratio Lower Upper
107 100
121 52.9 39.1 58.7
122 84.9 68.9 103.3





#27

Bis(2-Chloroethoxy)methane

Concen: 4.471 ng/ul

RT: 10.108 min Scan# 1195

Delta R.T. 0.001 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion: 93 Resp: 22847

Ion Ratio Lower Upper

93 100

95 34.0 25.8 38.8

123 10.3 9.3 13.9



#29

2,4-Dichlorophenol

Concen: 4.172 ng/ul

RT: 10.355 min Scan# 1237

Delta R.T. 0.006 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11




Tgt Ion:162 Resp: 13389


Ion Ratio Lower Upper

162 100

164 65.1 58.4 87.6

98 36.8 29.6 44.4









#30
Naphthalene
Concen: 4.763 ng/ul
RT: 10.749 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:128 Resp: 53764
Ion Ratio Lower Upper
128 100
129 10.1 8.8 13.2
127 13.2 10.8 16.2



#32
4-Chloroaniline
Concen: 4.307 ng/ul
RT: 10.866 min Scan# 1324
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:127 Resp: 22037
Ion Ratio Lower Upper
127 100
129 33.8 26.1 39.1









#33
Hexachlorobutadiene
Concen: 4.491 ng/ul
RT: 11.031 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:225 Resp: 10202
Ion Ratio Lower Upper
225 100
223 68.5 49.7 74.5
227 64.1 53.2 79.8



#34
Caprolactam
Concen: 4.094 ng/ul
RT: 11.613 min Scan# 1451
Delta R.T. -0.023 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:113 Resp: 5282
Ion Ratio Lower Upper
113 100
55 213.6 179.8 269.6
56 174.1 132.2 198.4









#35
4-Chloro-3-methylphenol
Concen: 4.195 ng/ul
RT: 11.983 min Scan# 1514
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:107 Resp: 16537
Ion Ratio Lower Upper
107 100
144 20.8 19.0 28.6
142 71.2 55.8 83.6



#36
2-Methylnaphthalene
Concen: 4.447 ng/ul
RT: 12.365 min Scan# 1579
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:142 Resp: 33838
Ion Ratio Lower Upper
142 100
141 95.2 72.6 108.8





#37

1-Methylnaphthalene

Concen: 4.731 ng/ul

RT: 12.582 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion:142 Resp: 36095

Ion Ratio Lower Upper

142 100

141 97.2 75.8 113.6

116 4.1 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.638 ng/ul

RT: 12.729 min Scan# 1641

Delta R.T. 0.001 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Tgt Ion:216 Resp: 20788




Ion Ratio Lower Upper


216 100

214 73.4 62.2 93.4

179 15.5 16.2 24.2#

108 17.4 12.2 18.4









#40
Hexachlorocyclopentadiene
Concen: 2.349 ng/ul
RT: 12.705 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:	237	Resp:	6632
Ion Ratio	Lower	Upper	
237	100		
235	59.5	50.9	76.3
272	12.8	10.8	16.2



#41
2,4,6-Trichlorophenol
Concen: 4.274 ng/ul
RT: 12.970 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:	196	Resp:	12982
Ion Ratio	Lower	Upper	
196	100		
198	90.7	77.7	116.5
200	32.9	25.0	37.6









#42
 2,4,5-Trichlorophenol
 Concen: 4.179 ng/ul
 RT: 13.046 min Scan# 1695
 Delta R.T. 0.000 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005442


Tgt Ion:196 Resp: 13728
 Ion Ratio Lower Upper
 196 100
 198 95.6 74.8 112.2
 200 30.3 24.4 36.6



#43
 1,1'-Biphenyl
 Concen: 4.712 ng/ul
 RT: 13.369 min Scan# 1750
 Delta R.T. 0.001 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11

Tgt Ion:154 Resp: 49940
 Ion Ratio Lower Upper
 154 100
 153 41.5 31.8 47.6
 76 13.5 11.7 17.5









#44
2-Chloronaphthalene
Concen: 4.570 ng/ul
RT: 13.416 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:162	Resp:	38873
Ion	Ratio	Lower Upper
162	100	
164	37.9	26.2 39.2
127	38.2	29.8 44.6



#45
2-Nitroaniline
Concen: 3.940 ng/ul
RT: 13.622 min Scan# 1793
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion: 65	Resp:	11652
Ion	Ratio	Lower Upper
65	100	
92	65.8	51.4 77.2
138	88.9	71.6 107.4









#47
Dimethylphthalate
Concen: 5.133 ng/ul
RT: 13.986 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:163 Resp: 59838
Ion Ratio Lower Upper
163 100
194 3.4 2.6 3.8
164 10.1 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 3.778 ng/ul
RT: 14.110 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:165 Resp: 9040
Ion Ratio Lower Upper
165 100
89 65.8 48.9 73.3
121 24.6 16.2 24.4#









#50
Acenaphthylene
Concen: 4.679 ng/ul
RT: 14.257 min Scan# 1901
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:152 Resp: 63980
Ion Ratio Lower Upper
152 100
151 19.1 16.4 24.6
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 3.487 ng/ul
RT: 14.445 min Scan# 1933
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:138 Resp: 7052
Ion Ratio Lower Upper
138 100
108 11.0 8.6 13.0
92 137.9 108.4 162.6









#52
Acenaphthene
Concen: 4.795 ng/ul
RT: 14.603 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:	153	Resp:	44996
Ion Ratio	Lower	Upper	
153	100		
152	48.1	39.1	58.7
154	83.9	68.7	103.1



#53
2,4-Dinitrophenol
Concen: 2.345 ng/ul
RT: 14.662 min Scan# 1970
Delta R.T. 0.012 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:	184	Resp:	2926
Ion Ratio	Lower	Upper	
184	100		
63	64.6	52.9	79.3
154	58.9	47.8	71.8









#55
4-Nitrophenol
Concen: 2.461 ng/ul
RT: 14.756 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:109 Resp: 3738
Ion Ratio Lower Upper
109 100
139 130.6 100.3 150.5
65 164.2 120.6 180.8



#56
Dibenzofuran
Concen: 4.789 ng/ul
RT: 14.932 min Scan# 2016
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:168 Resp: 63880
Ion Ratio Lower Upper
168 100
139 34.2 27.8 41.8





#57

2,4-Dinitrotoluene

Concen: 4.032 ng/ul

RT: 14.897 min Scan# 2011

Delta R.T. -0.005 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion:165 Resp: 13671

Ion Ratio Lower Upper

165 100

63 51.7 33.8 50.6#

182 2.1 1.2 1.8#



#58

2,3,4,6-Tetrachlorophenol

Concen: 4.311 ng/ul

RT: 15.161 min Scan# 2055

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Tgt Ion:232 Resp: 12894




Ion Ratio Lower Upper


232 100

131 43.1 29.0 43.6

130 0.0 1.1 1.7#

166 33.2 23.0 34.6









#59
Diethylphthalate
Concen: 4.993 ng/ul
RT: 15.349 min Scan# 2087
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:149 Resp: 60760
Ion Ratio Lower Upper
149 100
177 22.1 18.4 27.6
150 12.0 10.0 15.0



#61
Fluorene
Concen: 4.963 ng/ul
RT: 15.584 min Scan# 2127
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:166 Resp: 54321
Ion Ratio Lower Upper
166 100
165 102.4 80.2 120.2
167 12.2 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 4.802 ng/ul
RT: 15.579 min Scan# 2125
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:204 Resp: 27810
Ion Ratio Lower Upper
204 100
206 35.5 26.0 39.0
141 47.7 42.2 63.2



#63
4-Nitroaniline
Concen: 1.432 ng/ul
RT: 15.602 min Scan# 2130
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:138 Resp: 2594
Ion Ratio Lower Upper
138 100
92 55.9 45.0 67.4
108 93.6 60.8 91.2#









#66
4,6-Dinitro-2-methylphenol
Concen: 3.232 ng/ul
RT: 15.661 min Scan# 2141
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:198 Resp: 6688
Ion Ratio Lower Upper
198 100
182 7.2 3.4 5.2#
77 27.0 18.4 27.6



#67
N-Nitrosodiphenylamine
Concen: 4.470 ng/ul
RT: 15.790 min Scan# 2162
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:169 Resp: 44327
Ion Ratio Lower Upper
169 100
168 69.4 55.7 83.5
167 37.5 31.2 46.8









#68
4-Bromophenyl-phenylether
Concen: 4.522 ng/ul
RT: 16.472 min Scan# 21
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:248 Resp: 18497
Ion Ratio Lower Upper
248 100
250 93.9 73.8 110.6
141 56.0 48.2 72.2



#69
Hexachlorobenzene
Concen: 4.707 ng/ul
RT: 16.589 min Scan# 2298
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:284 Resp: 21557
Ion Ratio Lower Upper
284 100
142 28.4 20.5 30.7
249 30.1 25.2 37.8





#70

Atrazine

Concen: 4.392 ng/ul

RT: 16.736 min Scan# 21

Delta R.T. 0.001 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion:200 Resp: 17812

Ion Ratio Lower Upper

200 100

173 24.3 19.8 29.8

215 47.4 36.2 54.4



#71

Pentachlorophenol

Concen: 3.261 ng/ul

RT: 16.936 min Scan# 2357

Delta R.T. 0.006 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11




Tgt Ion:266 Resp: 8965


Ion Ratio Lower Upper

266 100

264 55.7 46.6 70.0

268 52.8 47.6 71.4





#72

Phenanthrene

Concen: 4.703 ng/ul

RT: 17.324 min Scan# 2423

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion:178 Resp: 89150

Ion Ratio Lower Upper

178 100

179 16.6 11.9 17.9

176 21.4 15.8 23.8



#74

Anthracene

Concen: 4.768 ng/ul

RT: 17.412 min Scan# 2438

Delta R.T. -0.005 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11




Tgt Ion:178 Resp: 91460


Ion Ratio Lower Upper

178 100

179 15.8 13.4 20.2

176 17.2 16.2 24.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.366 ng/uL
RT: 13.340 min Scan# 1
Delta R.T. 0.006 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:216 Resp: 21576
Ion Ratio Lower Upper
216 100
214 80.6 62.2 93.4
218 42.0 40.9 61.3



#76
Pentachlorobenzene
Concen: 4.569 ng/uL
RT: 14.856 min Scan# 2003
Delta R.T. 0.001 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:250 Resp: 24118
Ion Ratio Lower Upper
250 100
248 61.9 51.6 77.4
252 64.1 54.4 81.6





#77

Carbazole

Concen: 4.613 ng/ul

RT: 17.688 min Scan# 24

Delta R.T. 0.006 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion:167 Resp: 80027

Ion Ratio Lower Upper

167 100

166 22.1 17.8 26.8

139 11.8 9.4 14.0



#78

Di-n-butylphthalate

Concen: 4.334 ng/ul

RT: 18.240 min Scan# 2579

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11




Tgt Ion:149 Resp: 95780


Ion Ratio Lower Upper

149 100

150 8.9 7.2 10.8

104 4.5 3.9 5.9









#80
 Fluoranthene
 Concen: 4.608 ng/ul
 RT: 19.333 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005442


Tgt Ion:202 Resp: 110184
 Ion Ratio Lower Upper
 202 100
 101 10.3 8.6 12.8
 100 8.5 7.2 10.8



#82
 Pyrene
 Concen: 4.790 ng/ul
 RT: 19.697 min Scan# 2827
 Delta R.T. 0.000 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11

Tgt Ion:202 Resp: 115308
 Ion Ratio Lower Upper
 202 100
 101 12.7 10.2 15.2
 100 12.6 8.9 13.3









#83
 Butylbenzylphthalate
 Concen: 4.224 ng/ul
 RT: 20.584 min Scan# 2978
 Delta R.T. 0.006 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005442


Tgt Ion:	149	Resp:	41234
Ion Ratio	Lower	Upper	
149	100		
91	70.3	59.9	89.9
206	21.6	16.9	25.3



#84
 3,3'-Dichlorobenzidine
 Concen: 4.464 ng/ul
 RT: 21.430 min Scan# 3122
 Delta R.T. 0.000 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11

Tgt Ion:	252	Resp:	35409
Ion Ratio	Lower	Upper	
252	100		
254	66.9	50.5	75.7
126	9.5	6.5	9.7





#85

Benzo(a)anthracene

Concen: 4.639 ng/ul

RT: 21.513 min Scan# 3136

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion:228 Resp: 109145

Ion Ratio Lower Upper

228 100

229 19.8 15.7 23.5

226 27.2 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.332 ng/ul

RT: 21.413 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11




Tgt Ion:149 Resp: 60591


Ion Ratio Lower Upper

149 100

167 27.2 21.6 32.4

279 4.3 3.0 4.4









#87
Chrysene
Concen: 4.610 ng/ul
RT: 21.577 min Scan# 3147
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:228 Resp: 101674
Ion Ratio Lower Upper
228 100
226 31.4 24.0 36.0
229 20.5 15.8 23.8



#89
Di-n-octyl phthalate
Concen: 4.275 ng/ul
RT: 22.588 min Scan# 3319
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:149 Resp: 101219









#90
Benzo(b)fluoranthene
Concen: 4.429 ng/ul
RT: 23.663 min Scan# 3503
Delta R.T. -0.005 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11




Instrument :
BNA_G
ClientSampleId :
SSTD005442


Tgt Ion:252 Resp: 105685
Ion Ratio Lower Upper
252 100
253 23.1 17.9 26.9
125 9.0 6.7 10.1



#91
Benzo(k)fluoranthene
Concen: 4.750 ng/ul
RT: 23.734 min Scan# 3514
Delta R.T. 0.000 min
Lab File: BG058283.D
Acq: 17 Jul 2023 15:11

Tgt Ion:252 Resp: 111108
Ion Ratio Lower Upper
252 100
253 22.1 18.3 27.5
125 9.4 7.4 11.0





#93

Benzo(a)pyrene

Concen: 4.540 ng/ul

RT: 24.521 min Scan# 3

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11

Instrument :

BNA_G

ClientSampleId :

SSTD005442





Tgt Ion:252 Resp: 95881

Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.0

125 9.6 7.7 11.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.395 ng/ul

RT: 28.228 min Scan# 4279

Delta R.T. 0.000 min

Lab File: BG058283.D

Acq: 17 Jul 2023 15:11




Tgt Ion:276 Resp: 123260


Ion Ratio Lower Upper

276 100

138 18.7 13.8 20.6

277 24.3 19.2 28.8









#95
 Dibenzo(a,h)anthracene
 Concen: 4.221 ng/ul
 RT: 28.293 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005442

Tgt Ion:278 Resp: 97612
 Ion Ratio Lower Upper
 278 100
 139 12.4 10.4 15.6
 279 22.6 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 4.286 ng/ul
 RT: 29.339 min Scan# 4468
 Delta R.T. -0.011 min
 Lab File: BG058283.D
 Acq: 17 Jul 2023 15:11

Tgt Ion:276 Resp: 98182
 Ion Ratio Lower Upper
 276 100
 138 15.7 13.5 20.3
 277 25.7 19.4 29.2

