

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071723\
 Data File : BG058285.D
 Acq On : 17 Jul 2023 16:34
 Operator : CG/JU
 Sample : SST02094
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020444

Quant Time: Jul 18 01:30:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 18 01:24:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	46509	20.000	ng/ul	0.00
20) Naphthalene-d8	10.701	136	228362	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.538	164	172269	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.282	188	422846	20.000	ng/ul	0.00
79) Chrysene-d12	21.530	240	367370	20.000	ng/ul	0.00
88) Perylene-d12	24.667	264	448361	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.339	96	10387	7.820	ng/uL	0.00
4) Pyridine-d5	3.751	84	76984	19.092	ng/ul	0.00
7) Phenol-d5	7.064	99	86493	18.210	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.223	67	54668	19.099	ng/ul	0.00
11) 2-Chlorophenol-d4	7.429	132	61573	18.796	ng/ul	0.00
15) 4-Methylphenol-d8	8.604	113	67679	18.884	ng/ul	0.00
21) Nitrobenzene-d5	9.062	128	32873	17.662	ng/ul	0.00
24) 2-Nitrophenol-d4	9.779	143	36161	18.486	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.325	165	69088	18.411	ng/ul	0.00
31) 4-Chloroaniline-d4	10.837	131	106108	18.650	ng/ul	0.00
46) Dimethylphthalate-d6	13.939	166	249818	18.379	ng/ul	0.00
49) Acenaphthylene-d8	14.233	160	264734	17.848	ng/ul	0.00
54) 4-Nitrophenol-d4	14.738	143	37668	15.686	ng/ul	0.00
60) Fluorene-d10	15.525	176	209036	18.109	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.649	200	39257	17.098	ng/ul	0.00
73) Anthracene-d10	17.382	188	349259	17.800	ng/ul	0.00
81) Pyrene-d10	19.667	212	410275	17.732	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.450	264	394637	17.276	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.375	88	11694	7.127	ng/uL	100
5) Pyridine	3.768	79	80834	19.565	ng/ul	100
6) Benzaldehyde	7.035	77	26432	16.664	ng/ul	100
8) Phenol	7.088	94	91898	19.066	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.317	93	71949	19.333	ng/ul	100
12) 2-Chlorophenol	7.464	128	64835	19.019	ng/ul	100
13) 2-Methylphenol	8.339	108	71563	18.905	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.428	45	116874	20.401	ng/ul	100
16) Acetophenone	8.721	105	120170	20.210	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.704	70	64925	20.425	ng/ul	100
18) 4-Methylphenol	8.668	108	78236	19.114	ng/ul	100
19) Hexachloroethane	8.980	117	25031	19.154	ng/ul	100
22) Nitrobenzene	9.103	77	88848	18.388	ng/ul	100
23) Isophorone	9.620	82	197318	19.138	ng/ul	100
25) 2-Nitrophenol	9.814	139	38835	18.501	ng/ul	100
26) 2,4-Dimethylphenol	9.873	107	85997	18.220	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.108	93	102991	17.727	ng/ul	100
29) 2,4-Dichlorophenol	10.349	162	66987	18.360	ng/ul	100
30) Naphthalene	10.754	128	231942	18.073	ng/ul	100
32) 4-Chloroaniline	10.860	127	110866	19.057	ng/ul	100
33) Hexachlorobutadiene	11.030	225	45744	17.709	ng/ul	100
34) Caprolactam	11.636	113	28185	19.214	ng/ul	100
35) 4-Chloro-3-methylphenol	11.988	107	84183	18.783	ng/ul	100
36) 2-Methylnaphthalene	12.358	142	160124	18.509	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071723\
 Data File : BG058285.D
 Acq On : 17 Jul 2023 16:34
 Operator : CG/JU
 Sample : SSTD02094
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020444

Quant Time: Jul 18 01:30:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 18 01:24:07 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.582	142	161392	18.604	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.728	216	93052	17.725	ng/ul	100
40) Hexachlorocyclopentadiene	12.705	237	46049	13.926	ng/ul	100
41) 2,4,6-Trichlorophenol	12.969	196	64205	18.045	ng/ul	100
42) 2,4,5-Trichlorophenol	13.046	196	67271	17.482	ng/ul	100
43) 1,1'-Biphenyl	13.369	154	218577	17.607	ng/ul	100
44) 2-Chloronaphthalene	13.416	162	174464	17.510	ng/ul	100
45) 2-Nitroaniline	13.616	65	60877	17.574	ng/ul	100
47) Dimethylphthalate	13.986	163	251924	18.448	ng/ul	100
48) 2,6-Dinitrotoluene	14.109	165	49496	17.658	ng/ul	100
50) Acenaphthylene	14.262	152	286586	17.893	ng/ul	100
51) 3-Nitroaniline	14.444	138	35558	15.011	ng/ul	100
52) Acenaphthene	14.603	153	195636	17.800	ng/ul	100
53) 2,4-Dinitrophenol	14.650	184	23597	16.148	ng/ul	100
55) 4-Nitrophenol	14.755	109	29423	16.539	ng/ul	100
56) Dibenzofuran	14.938	168	281144	17.995	ng/ul	100
57) 2,4-Dinitrotoluene	14.902	165	72766	18.322	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.161	232	63166	18.032	ng/ul	100
59) Diethylphthalate	15.349	149	260387	18.269	ng/ul	100
61) Fluorene	15.584	166	231189	18.034	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.572	204	121977	17.983	ng/ul	100
63) 4-Nitroaniline	15.607	138	30292	14.275	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.666	198	41481	17.666	ng/ul	100
67) N-Nitrosodiphenylamine	15.790	169	200238	17.794	ng/ul	100
68) 4-Bromophenyl-phenylether	16.465	248	83292	17.945	ng/ul	100
69) Hexachlorobenzene	16.589	284	96095	18.492	ng/ul	100
70) Atrazine	16.735	200	81163	17.637	ng/ul	100
71) Pentachlorophenol	16.929	266	51193	16.409	ng/ul	100
72) Phenanthrene	17.323	178	383186	17.815	ng/ul	100
74) Anthracene	17.417	178	385317	17.703	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.334	216	98097	17.493	ng/ul	100
76) Pentachlorobenzene	14.855	250	109162	18.226	ng/ul	100
77) Carbazole	17.681	167	355923	18.081	ng/ul	100
78) Di-n-butylphthalate	18.240	149	433637	17.293	ng/ul	100
80) Fluoranthene	19.332	202	470855	17.563	ng/ul	100
82) Pyrene	19.697	202	475756	17.629	ng/ul	100
83) Butylbenzylphthalate	20.578	149	185636	16.963	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.430	252	159087	17.888	ng/ul	100
85) Benzo(a)anthracene	21.512	228	461981	17.514	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.412	149	263812	16.822	ng/ul	100
87) Chrysene	21.577	228	440134	17.800	ng/ul	100
89) Di-n-octyl phthalate	22.587	149	456967	16.588	ng/ul	100
90) Benzo(b)fluoranthene	23.669	252	463518	16.696	ng/ul	100
91) Benzo(k)fluoranthene	23.733	252	467313	17.174	ng/ul	100
93) Benzo(a)pyrene	24.520	252	421719	17.163	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.228	276	548309	16.804	ng/ul	100
95) Dibenzo(a,h)anthracene	28.293	278	456527	16.967	ng/ul	100
96) Benzo(g,h,i)perylene	29.350	276	443670	16.647	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071723\
 Data File : BG058285.D
 Acq On : 17 Jul 2023 16:34
 Operator : CG/JU
 Sample : SST02094
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020444

Quant Time: Jul 18 01:30:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 18 01:24:07 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.127 ng/uL
RT: 3.375 min Scan# 49
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:	88	Resp:	11694
Ion	Ratio	Lower	Upper
88	100		
43	24.5	19.6	29.4
58	87.8	70.2	105.4



#5
Pyridine
Concen: 19.565 ng/uL
RT: 3.768 min Scan# 116
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:	79	Resp:	80834
Ion	Ratio	Lower	Upper
79	100		
52	90.4	72.3	108.5
51	41.7	33.4	50.0









#6
Benzaldehyde
Concen: 16.664 ng/ul
RT: 7.035 min Scan# 672
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion: 77 Resp: 26432
Ion Ratio Lower Upper
77 100
105 97.8 78.2 117.4
106 89.5 71.6 107.4



#8
Phenol
Concen: 19.066 ng/ul
RT: 7.088 min Scan# 681
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion: 94 Resp: 91898
Ion Ratio Lower Upper
94 100
65 29.8 23.8 35.8
66 39.0 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 19.333 ng/ul
RT: 7.317 min Scan# 720
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion: 93 Resp: 71949
Ion Ratio Lower Upper
93 100
63 84.6 67.7 101.5
95 33.8 27.0 40.6



#12
2-Chlorophenol
Concen: 19.019 ng/ul
RT: 7.464 min Scan# 745
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion: 128 Resp: 64835
Ion Ratio Lower Upper
128 100
64 57.0 45.6 68.4
130 31.5 25.2 37.8





#13

2-Methylphenol

Concen: 18.905 ng/ul

RT: 8.339 min Scan# 894

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :





SSTD020444

Tgt Ion:108 Resp: 71563

Ion Ratio Lower Upper

108 100

107 89.4 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.401 ng/ul

RT: 8.428 min Scan# 909

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34




Tgt Ion: 45 Resp: 116874


Ion Ratio Lower Upper

45 100

77 11.8 9.4 14.2

79 9.3 7.4 11.2









#16
Acetophenone
Concen: 20.210 ng/ul
RT: 8.721 min Scan# 959
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:105 Resp: 120170
Ion Ratio Lower Upper
105 100
77 75.9 60.7 91.1
51 29.4 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.425 ng/ul
RT: 8.704 min Scan# 956
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion: 70 Resp: 64925
Ion Ratio Lower Upper
70 100
42 58.2 46.6 69.8
101 9.1 7.3 10.9
130 17.7 14.2 21.2









#18
4-Methylphenol
Concen: 19.114 ng/ul
RT: 8.668 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:108 Resp: 78236
Ion Ratio Lower Upper
108 100
107 113.3 90.6 136.0



#19
Hexachloroethane
Concen: 19.154 ng/ul
RT: 8.980 min Scan# 1003
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:117 Resp: 25031
Ion Ratio Lower Upper
117 100
201 129.5 103.6 155.4
199 75.1 60.1 90.1









#22
Nitrobenzene
Concen: 18.388 ng/ul
RT: 9.103 min Scan# 1102
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion: 77 Resp: 88848
Ion Ratio Lower Upper
77 100
123 42.5 34.0 51.0
65 13.8 11.0 16.6



#23
Isophorone
Concen: 19.138 ng/ul
RT: 9.620 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion: 82 Resp: 197318
Ion Ratio Lower Upper
82 100
95 6.4 5.1 7.7
138 14.9 11.9 17.9





#25

2-Nitrophenol

Concen: 18.501 ng/ul

RT: 9.814 min Scan# 1145

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:139 Resp: 38835

Ion Ratio Lower Upper

139 100

65 61.5 49.2 73.8

109 27.0 21.6 32.4



#26

2,4-Dimethylphenol

Concen: 18.220 ng/ul

RT: 9.873 min Scan# 1155

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34




Tgt Ion:107 Resp: 85997


Ion Ratio Lower Upper

107 100

121 48.9 39.1 58.7

122 86.1 68.9 103.3





#27

Bis(2-Chloroethoxy)methane

Concen: 17.727 ng/ul

RT: 10.108 min Scan# 1195

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion: 93 Resp: 102991

Ion Ratio Lower Upper

93 100

95 32.3 25.8 38.8

123 11.6 9.3 13.9



#29

2,4-Dichlorophenol

Concen: 18.360 ng/ul

RT: 10.349 min Scan# 1236

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34




Tgt Ion:162 Resp: 66987


Ion Ratio Lower Upper

162 100

164 73.0 58.4 87.6

98 37.0 29.6 44.4









#30
Naphthalene
Concen: 18.073 ng/ul
RT: 10.754 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:128 Resp: 231942
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.5 10.8 16.2



#32
4-Chloroaniline
Concen: 19.057 ng/ul
RT: 10.860 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:127 Resp: 110866
Ion Ratio Lower Upper
127 100
129 32.6 26.1 39.1









#33
Hexachlorobutadiene
Concen: 17.709 ng/ul
RT: 11.030 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:225 Resp: 45744
Ion Ratio Lower Upper
225 100
223 62.1 49.7 74.5
227 66.5 53.2 79.8



#34
Caprolactam
Concen: 19.214 ng/ul
RT: 11.636 min Scan# 1455
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:113 Resp: 28185
Ion Ratio Lower Upper
113 100
55 224.7 179.8 269.6
56 165.3 132.2 198.4









#35
4-Chloro-3-methylphenol
Concen: 18.783 ng/ul
RT: 11.988 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:107 Resp: 84183
Ion Ratio Lower Upper
107 100
144 23.8 19.0 28.6
142 69.7 55.8 83.6



#36
2-Methylnaphthalene
Concen: 18.509 ng/ul
RT: 12.358 min Scan# 1578
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:142 Resp: 160124
Ion Ratio Lower Upper
142 100
141 90.7 72.6 108.8









#37
1-Methylnaphthalene
Concen: 18.604 ng/ul
RT: 12.582 min Scan# 1616
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:142 Resp: 161392
Ion Ratio Lower Upper
142 100
141 94.7 75.8 113.6
116 4.0 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.725 ng/ul
RT: 12.728 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:216 Resp: 93052
Ion Ratio Lower Upper
216 100
214 77.8 62.2 93.4
179 20.2 16.2 24.2
108 15.3 12.2 18.4









#40
Hexachlorocyclopentadiene
Concen: 13.926 ng/ul
RT: 12.705 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:237	Resp:	46049
Ion Ratio	Lower	Upper
237	100	
235	63.6	50.9 76.3
272	13.5	10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 18.045 ng/ul
RT: 12.969 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:196	Resp:	64205
Ion Ratio	Lower	Upper
196	100	
198	97.1	77.7 116.5
200	31.3	25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 17.482 ng/ul
RT: 13.046 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:196 Resp: 67271
Ion Ratio Lower Upper
196 100
198 93.5 74.8 112.2
200 30.5 24.4 36.6



#43
1,1'-Biphenyl
Concen: 17.607 ng/ul
RT: 13.369 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:154 Resp: 218577
Ion Ratio Lower Upper
154 100
153 39.7 31.8 47.6
76 14.6 11.7 17.5









#44
2-Chloronaphthalene
Concen: 17.510 ng/ul
RT: 13.416 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:162 Resp: 174464
Ion Ratio Lower Upper
162 100
164 32.7 26.2 39.2
127 37.2 29.8 44.6



#45
2-Nitroaniline
Concen: 17.574 ng/ul
RT: 13.616 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion: 65 Resp: 60877
Ion Ratio Lower Upper
65 100
92 64.3 51.4 77.2
138 89.5 71.6 107.4









#47
Dimethylphthalate
Concen: 18.448 ng/ul
RT: 13.986 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:163 Resp: 251924
Ion Ratio Lower Upper
163 100
194 3.2 2.6 3.8
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 17.658 ng/ul
RT: 14.109 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:165 Resp: 49496
Ion Ratio Lower Upper
165 100
89 61.1 48.9 73.3
121 20.3 16.2 24.4









#50
Acenaphthylene
Concen: 17.893 ng/ul
RT: 14.262 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:152 Resp: 286586
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6
153 13.3 10.6 16.0



#51
3-Nitroaniline
Concen: 15.011 ng/ul
RT: 14.444 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:138 Resp: 35558
Ion Ratio Lower Upper
138 100
108 10.8 8.6 13.0
92 135.5 108.4 162.6









#52
Acenaphthene
Concen: 17.800 ng/ul
RT: 14.603 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:153 Resp: 195636
Ion Ratio Lower Upper
153 100
152 48.9 39.1 58.7
154 85.9 68.7 103.1



#53
2,4-Dinitrophenol
Concen: 16.148 ng/ul
RT: 14.650 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:184 Resp: 23597
Ion Ratio Lower Upper
184 100
63 66.1 52.9 79.3
154 59.8 47.8 71.8









#55
4-Nitrophenol
Concen: 16.539 ng/ul
RT: 14.755 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:109 Resp: 29423
Ion Ratio Lower Upper
109 100
139 125.4 100.3 150.5
65 150.7 120.6 180.8



#56
Dibenzofuran
Concen: 17.995 ng/ul
RT: 14.938 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:168 Resp: 281144
Ion Ratio Lower Upper
168 100
139 34.8 27.8 41.8





#57

2,4-Dinitrotoluene

Concen: 18.322 ng/ul

RT: 14.902 min Scan# 2011

Delta R.T. 0.000 min

Lab File: BG058285.D


Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444






Tgt Ion:165 Resp: 72766

Ion Ratio Lower Upper

165 100

63 42.2 33.8 50.6

182 1.5 1.2 1.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.032 ng/ul

RT: 15.161 min Scan# 2055

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Tgt Ion:232 Resp: 63166




Ion Ratio Lower Upper


232 100

131 36.3 29.0 43.6

130 1.4 1.1 1.7

166 28.8 23.0 34.6









#59
Diethylphthalate
Concen: 18.269 ng/ul
RT: 15.349 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:149 Resp: 260387
Ion Ratio Lower Upper
149 100
177 23.0 18.4 27.6
150 12.5 10.0 15.0



#61
Fluorene
Concen: 18.034 ng/ul
RT: 15.584 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:166 Resp: 231189
Ion Ratio Lower Upper
166 100
165 100.2 80.2 120.2
167 13.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 17.983 ng/ul
RT: 15.572 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:204 Resp: 121977
Ion Ratio Lower Upper
204 100
206 32.5 26.0 39.0
141 52.7 42.2 63.2



#63
4-Nitroaniline
Concen: 14.275 ng/ul
RT: 15.607 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:138 Resp: 30292
Ion Ratio Lower Upper
138 100
92 56.2 45.0 67.4
108 76.0 60.8 91.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.666 ng/ul
 RT: 15.666 min Scan# 2141
 Delta R.T. 0.000 min
 Lab File: BG058285.D
 Acq: 17 Jul 2023 16:34




Instrument :
 BNA_G
 ClientSampleId :
 SST020444


Tgt Ion:198 Resp: 41481
 Ion Ratio Lower Upper
 198 100
 182 4.3 3.4 5.2
 77 23.0 18.4 27.6



#67
 N-Nitrosodiphenylamine
 Concen: 17.794 ng/ul
 RT: 15.790 min Scan# 2162
 Delta R.T. 0.000 min
 Lab File: BG058285.D
 Acq: 17 Jul 2023 16:34

Tgt Ion:169 Resp: 200238
 Ion Ratio Lower Upper
 169 100
 168 69.6 55.7 83.5
 167 39.0 31.2 46.8









#68
4-Bromophenyl-phenylether
Concen: 17.945 ng/ul
RT: 16.465 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:248 Resp: 83292
Ion Ratio Lower Upper
248 100
250 92.2 73.8 110.6
141 60.2 48.2 72.2



#69
Hexachlorobenzene
Concen: 18.492 ng/ul
RT: 16.589 min Scan# 2298
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:284 Resp: 96095
Ion Ratio Lower Upper
284 100
142 25.6 20.5 30.7
249 31.5 25.2 37.8





#70

Atrazine

Concen: 17.637 ng/ul

RT: 16.735 min Scan# 21

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:200 Resp: 81163

Ion Ratio Lower Upper

200 100

173 24.8 19.8 29.8

215 45.3 36.2 54.4



#71

Pentachlorophenol

Concen: 16.409 ng/ul

RT: 16.929 min Scan# 2356

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34




Tgt Ion:266 Resp: 51193


Ion Ratio Lower Upper

266 100

264 58.3 46.6 70.0

268 59.5 47.6 71.4









#72
Phenanthrene
Concen: 17.815 ng/ul
RT: 17.323 min Scan# 2423
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:178 Resp: 383186
Ion Ratio Lower Upper
178 100
179 14.9 11.9 17.9
176 19.8 15.8 23.8



#74
Anthracene
Concen: 17.703 ng/ul
RT: 17.417 min Scan# 2439
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:178 Resp: 385317
Ion Ratio Lower Upper
178 100
179 16.8 13.4 20.2
176 20.2 16.2 24.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.493 ng/uL
RT: 13.334 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:216 Resp: 98097
Ion Ratio Lower Upper
216 100
214 77.8 62.2 93.4
218 51.1 40.9 61.3



#76
Pentachlorobenzene
Concen: 18.226 ng/uL
RT: 14.855 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:250 Resp: 109162
Ion Ratio Lower Upper
250 100
248 64.5 51.6 77.4
252 68.0 54.4 81.6









#77
 Carbazole
 Concen: 18.081 ng/ul
 RT: 17.681 min Scan# 2484
 Delta R.T. 0.000 min
 Lab File: BG058285.D
 Acq: 17 Jul 2023 16:34

Instrument :
 BNA_G
 ClientSampleId :
 SST020444

Tgt Ion:167 Resp: 355923




Ion	Ratio	Lower	Upper
167	100		
166	22.3	17.8	26.8
139	11.7	9.4	14.0




#78
 Di-n-butylphthalate
 Concen: 17.293 ng/ul
 RT: 18.240 min Scan# 2579
 Delta R.T. 0.000 min
 Lab File: BG058285.D
 Acq: 17 Jul 2023 16:34

Tgt Ion:149 Resp: 433637

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.2	10.8
104	4.9	3.9	5.9









#80
 Fluoranthene
 Concen: 17.563 ng/ul
 RT: 19.332 min Scan# 2765
 Delta R.T. 0.000 min
 Lab File: BG058285.D
 Acq: 17 Jul 2023 16:34




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020444


Tgt Ion:	202	Resp:	470855
Ion	Ratio	Lower	Upper
202	100		
101	10.7	8.6	12.8
100	9.0	7.2	10.8



#82
 Pyrene
 Concen: 17.629 ng/ul
 RT: 19.697 min Scan# 2827
 Delta R.T. 0.000 min
 Lab File: BG058285.D
 Acq: 17 Jul 2023 16:34

Tgt Ion:	202	Resp:	475756
Ion	Ratio	Lower	Upper
202	100		
101	12.7	10.2	15.2
100	11.1	8.9	13.3









#83
Butylbenzylphthalate
Concen: 16.963 ng/ul
RT: 20.578 min Scan# 2977
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:	149	Resp:	185636
Ion Ratio	Lower	Upper	
149	100		
91	74.9	59.9	89.9
206	21.1	16.9	25.3



#84
3,3'-Dichlorobenzidine
Concen: 17.888 ng/ul
RT: 21.430 min Scan# 3122
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:	252	Resp:	159087
Ion Ratio	Lower	Upper	
252	100		
254	63.1	50.5	75.7
126	8.1	6.5	9.7





#85

Benzo(a)anthracene

Concen: 17.514 ng/ul

RT: 21.512 min Scan# 3136

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:228 Resp: 461981

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 27.3 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.822 ng/ul

RT: 21.412 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34




Tgt Ion:149 Resp: 263812


Ion Ratio Lower Upper

149 100

167 27.0 21.6 32.4

279 3.7 3.0 4.4





#87

Chrysene

Concen: 17.800 ng/ul

RT: 21.577 min Scan# 3147

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:228 Resp: 440134

Ion Ratio Lower Upper

228 100

226 30.0 24.0 36.0

229 19.8 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 16.588 ng/ul




RT: 22.587 min Scan# 3319


Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Tgt Ion:149 Resp: 456967









#90
Benzo(b)fluoranthene
Concen: 16.696 ng/ul
RT: 23.669 min Scan# 3503
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:252 Resp: 463518
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 8.4 6.7 10.1



#91
Benzo(k)fluoranthene
Concen: 17.174 ng/ul
RT: 23.733 min Scan# 3514
Delta R.T. 0.000 min
Lab File: BG058285.D
Acq: 17 Jul 2023 16:34

Tgt Ion:252 Resp: 467313
Ion Ratio Lower Upper
252 100
253 22.9 18.3 27.5
125 9.2 7.4 11.0





#93

Benzo(a)pyrene

Concen: 17.163 ng/ul

RT: 24.520 min Scan# 3

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:252 Resp: 421719

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 9.6 7.7 11.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.804 ng/ul

RT: 28.228 min Scan# 4279

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34




Tgt Ion:276 Resp: 548309


Ion Ratio Lower Upper

276 100

138 17.2 13.8 20.6

277 24.0 19.2 28.8





#95

Dibenzo(a,h)anthracene

Concen: 16.967 ng/ul

RT: 28.293 min Scan# 4290

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:278 Resp: 456527

Ion Ratio Lower Upper

278 100

139 13.0 10.4 15.6

279 23.0 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 16.647 ng/ul

RT: 29.350 min Scan# 4470

Delta R.T. 0.000 min

Lab File: BG058285.D

Acq: 17 Jul 2023 16:34

Tgt Ion:276 Resp: 443670

Ion Ratio Lower Upper

276 100

138 16.9 13.5 20.3

277 24.3 19.4 29.2

