

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071723\
 Data File : BG058288.D
 Acq On : 17 Jul 2023 18:40
 Operator : CG/JU
 Sample : SSTD16097
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160447

Quant Time: Jul 18 01:32:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 18 01:24:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.900	152	53882	20.000	ng/ul	0.00
20) Naphthalene-d8	10.709	136	253893	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.539	164	183622	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.289	188	424779	20.000	ng/ul	0.00
79) Chrysene-d12	21.549	240	341025	20.000	ng/ul	0.02
88) Perylene-d12	24.680	264	436081	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.341	96	85568	55.609	ng/uL	0.00
4) Pyridine-d5	3.752	84	653427	139.875	ng/ul	0.00
7) Phenol-d5	7.078	99	774606	140.768	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.236	67	462335	139.421	ng/ul	0.01
11) 2-Chlorophenol-d4	7.442	132	530271	139.721	ng/ul	0.01
15) 4-Methylphenol-d8	8.629	113	583775	140.598	ng/ul	0.02
21) Nitrobenzene-d5	9.075	128	286473	138.437	ng/ul	0.01
24) 2-Nitrophenol-d4	9.792	143	322494	148.283	ng/ul	0.01
28) 2,4-Dichlorophenol-d3	10.333	165	579097	138.802	ng/ul	0.00
31) 4-Chloroaniline-d4	10.850	131	780450	123.381	ng/ul	0.01
46) Dimethylphthalate-d6	13.958	166	1660217	114.592	ng/ul	0.02
49) Acenaphthylene-d8	14.240	160	1847911	116.884	ng/ul	0.00
54) 4-Nitrophenol-d4	14.774	143	314323	122.798	ng/ul	0.04
60) Fluorene-d10	15.538	176	1414424	114.955	ng/ul	0.01
65) 4,6-Dinitro-2-methylph...	15.679	200	334260	144.917	ng/ul	0.03
73) Anthracene-d10	17.395	188	2182326	110.717	ng/ul	0.01
81) Pyrene-d10	19.680	212	2464747	114.757	ng/ul	0.01
92) Benzo(a)pyrene-d12	24.487	264	2769162	124.640	ng/ul	0.04
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.376	88	95609	50.295	ng/uL#	97
5) Pyridine	3.770	79	665023	138.939	ng/ul	95
6) Benzaldehyde	7.042	77	288559	157.024	ng/ul	96
8) Phenol	7.107	94	776138	138.993	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.330	93	573252	132.957	ng/ul	96
12) 2-Chlorophenol	7.471	128	543160	137.532	ng/ul	96
13) 2-Methylphenol	8.352	108	613128	139.809	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.435	45	939570	141.568	ng/ul	98
16) Acetophenone	8.734	105	892053	129.493	ng/ul	95
17) N-Nitroso-di-n-propyla...	8.746	70	474598	128.874	ng/ul	93
18) 4-Methylphenol	8.699	108	653962	137.909	ng/ul	98
19) Hexachloroethane	8.981	117	210556	139.071	ng/ul	93
22) Nitrobenzene	9.116	77	719678	133.969	ng/ul	99
23) Isophorone	9.657	82	1494668	130.394	ng/ul	99
25) 2-Nitrophenol	9.827	139	330106	141.452	ng/ul	96
26) 2,4-Dimethylphenol	9.886	107	692366	131.936	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.121	93	826512	127.956	ng/ul	97
29) 2,4-Dichlorophenol	10.362	162	534805	131.840	ng/ul	97
30) Naphthalene	10.761	128	1713119	120.065	ng/ul	97
32) 4-Chloroaniline	10.879	127	775129	119.842	ng/ul	99
33) Hexachlorobutadiene	11.038	225	380245	132.405	ng/ul	98
34) Caprolactam	11.696	113	121016	74.203	ng/ul	94
35) 4-Chloro-3-methylphenol	12.013	107	678716	136.208	ng/ul	96
36) 2-Methylnaphthalene	12.365	142	1170835	121.731	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071723\
 Data File : BG058288.D
 Acq On : 17 Jul 2023 18:40
 Operator : CG/JU
 Sample : SSTD16097
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160447


Quant Time: Jul 18 01:32:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 18 01:24:07 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.589	142	1165459	120.834	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.736	216	707943	126.514	ng/ul	99
40) Hexachlorocyclopentadiene	12.712	237	493413	139.989	ng/ul	99
41) 2,4,6-Trichlorophenol	12.982	196	518709	136.768	ng/ul	100
42) 2,4,5-Trichlorophenol	13.059	196	556383	135.649	ng/ul	98
43) 1,1'-Biphenyl	13.382	154	1478598	111.739	ng/ul	96
44) 2-Chloronaphthalene	13.423	162	1244523	117.185	ng/ul	94
45) 2-Nitroaniline	13.640	65	514564	139.357	ng/ul	99
47) Dimethylphthalate	14.005	163	1631321	112.074	ng/ul#	96
48) 2,6-Dinitrotoluene	14.134	165	392395	131.335	ng/ul	97
50) Acenaphthylene	14.275	152	1859586	108.922	ng/ul	96
51) 3-Nitroaniline	14.463	138	339100	134.302	ng/ul	93
52) Acenaphthene	14.610	153	1339445	114.334	ng/ul	98
53) 2,4-Dinitrophenol	14.675	184	258975	166.262	ng/ul	97
55) 4-Nitrophenol	14.792	109	245284	129.354	ng/ul	99
56) Dibenzofuran	14.951	168	1804892	108.380	ng/ul	96
57) 2,4-Dinitrotoluene	14.927	165	524879	123.993	ng/ul#	83
58) 2,3,4,6-Tetrachlorophenol	15.174	232	495875	132.804	ng/ul#	99
59) Diethylphthalate	15.368	149	1645191	108.293	ng/ul	97
61) Fluorene	15.597	166	1438351	105.265	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.585	204	816869	112.982	ng/ul	98
63) 4-Nitroaniline	15.644	138	317542	140.390	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.691	198	335212	142.112	ng/ul	97
67) N-Nitrosodiphenylamine	15.803	169	1304546	115.398	ng/ul	98
68) 4-Bromophenyl-phenylether	16.478	248	600220	128.727	ng/ul	96
69) Hexachlorobenzene	16.596	284	672181	128.760	ng/ul	96
70) Atrazine	16.760	200	540883	117.003	ng/ul	95
71) Pentachlorophenol	16.942	266	441707	140.937	ng/ul	93
72) Phenanthrene	17.336	178	2322723	107.498	ng/ul#	93
74) Anthracene	17.430	178	2250251	102.913	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.347	216	753179	133.695	ng/ul	98
76) Pentachlorobenzene	14.868	250	794580	132.064	ng/ul	98
77) Carbazole	17.700	167	2159198	109.187	ng/ul	94
78) Di-n-butylphthalate	18.247	149	2555317	101.440	ng/ul#	95
80) Fluoranthene	19.346	202	2707187	108.777	ng/ul	97
82) Pyrene	19.716	202	2627892	104.898	ng/ul	97
83) Butylbenzylphthalate	20.591	149	1192347	117.373	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.443	252	960863	116.385	ng/ul	95
85) Benzo(a)anthracene	21.525	228	2867271	117.098	ng/ul#	92
86) Bis(2-ethylhexyl)phtha...	21.420	149	1616289	111.027	ng/ul	96
87) Chrysene	21.596	228	2705976	117.887	ng/ul	92
89) Di-n-octyl phthalate	22.601	149	2903109	108.350	ng/ul	100
90) Benzo(b)fluoranthene	23.699	252	3260147	120.742	ng/ul	97
91) Benzo(k)fluoranthene	23.770	252	3039492	114.847	ng/ul	97
93) Benzo(a)pyrene	24.563	252	2843823	118.995	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	28.306	276	4097804	129.124	ng/ul	98
95) Dibenzo(a,h)anthracene	28.370	278	3265279	124.775	ng/ul	96
96) Benzo(g,h,i)perylene	29.445	276	3374301	130.175	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_G
ClientSampleId :
SSTD160447

Quant Time: Jul 18 01:32:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 18 01:24:07 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 50.295 ng/uL
RT: 3.376 min Scan# 49
Delta R.T. 0.001 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion: 88 Resp: 95609
Ion Ratio Lower Upper
88 100
43 30.1 19.6 29.4#
58 86.6 70.2 105.4



#5
Pyridine
Concen: 138.939 ng/uL
RT: 3.770 min Scan# 116
Delta R.T. 0.001 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion: 79 Resp: 665023
Ion Ratio Lower Upper
79 100
52 97.2 72.3 108.5
51 41.3 33.4 50.0









#6
Benzaldehyde
Concen: 157.024 ng/ul
RT: 7.042 min Scan# 673
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion: 77 Resp: 288559
Ion Ratio Lower Upper
77 100
105 100.4 78.2 117.4
106 95.2 71.6 107.4



#8
Phenol
Concen: 138.993 ng/ul
RT: 7.107 min Scan# 684
Delta R.T. 0.019 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion: 94 Resp: 776138
Ion Ratio Lower Upper
94 100
65 31.4 23.8 35.8
66 38.8 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 132.957 ng/ul
RT: 7.330 min Scan# 71
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion: 93 Resp: 573252
Ion Ratio Lower Upper
93 100
63 90.0 67.7 101.5
95 33.4 27.0 40.6



#12
2-Chlorophenol
Concen: 137.532 ng/ul
RT: 7.471 min Scan# 746
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:128 Resp: 543160
Ion Ratio Lower Upper
128 100
64 60.3 45.6 68.4
130 32.7 25.2 37.8





#13

2-Methylphenol

Concen: 139.809 ng/ul

RT: 8.352 min Scan# 894

Delta R.T. 0.013 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :





SSTD160447

Tgt Ion:108 Resp: 613128

Ion Ratio Lower Upper

108 100

107 89.0 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 141.568 ng/ul

RT: 8.435 min Scan# 910

Delta R.T. 0.007 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40




Tgt Ion: 45 Resp: 939570


Ion Ratio Lower Upper

45 100

77 12.9 9.4 14.2

79 10.0 7.4 11.2









#16
Acetophenone
Concen: 129.493 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:105 Resp: 892053
Ion Ratio Lower Upper
105 100
77 80.1 60.7 91.1
51 32.6 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 128.874 ng/ul
RT: 8.746 min Scan# 963
Delta R.T. 0.042 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion: 70 Resp: 474598
Ion Ratio Lower Upper
70 100
42 64.9 46.6 69.8
101 9.1 7.3 10.9
130 18.9 14.2 21.2









#18
4-Methylphenol
Concen: 137.909 ng/ul
RT: 8.699 min Scan# 911
Delta R.T. 0.031 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:108 Resp: 653962
Ion Ratio Lower Upper
108 100
107 111.1 90.6 136.0



#19
Hexachloroethane
Concen: 139.071 ng/ul
RT: 8.981 min Scan# 1003
Delta R.T. 0.001 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:117 Resp: 210556
Ion Ratio Lower Upper
117 100
201 116.1 103.6 155.4
199 74.8 60.1 90.1





#22

Nitrobenzene

Concen: 133.969 ng/ul

RT: 9.116 min Scan# 1108

Delta R.T. 0.013 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :

SSTD160447





Tgt Ion: 77 Resp: 719678

Ion Ratio Lower Upper

77 100

123 42.3 34.0 51.0

65 14.5 11.0 16.6



#23

Isophorone

Concen: 130.394 ng/ul

RT: 9.657 min Scan# 1118

Delta R.T. 0.037 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40




Tgt Ion: 82 Resp: 1494668


Ion Ratio Lower Upper

82 100

95 6.7 5.1 7.7

138 15.5 11.9 17.9









#25
2-Nitrophenol
Concen: 141.452 ng/ul
RT: 9.827 min Scan# 1147
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:139 Resp: 330106
Ion Ratio Lower Upper
139 100
65 59.1 49.2 73.8
109 29.6 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 131.936 ng/ul
RT: 9.886 min Scan# 1157
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:107 Resp: 692366
Ion Ratio Lower Upper
107 100
121 47.7 39.1 58.7
122 79.5 68.9 103.3









#27
Bis(2-Chloroethoxy)methane
Concen: 127.956 ng/ul
RT: 10.121 min Scan# 1197
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion: 93 Resp: 826512
Ion Ratio Lower Upper
93 100
95 33.9 25.8 38.8
123 10.7 9.3 13.9



#29
2,4-Dichlorophenol
Concen: 131.840 ng/ul
RT: 10.362 min Scan# 1238
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:162 Resp: 534805
Ion Ratio Lower Upper
162 100
164 70.4 58.4 87.6
98 39.3 29.6 44.4









#30
Naphthalene
Concen: 120.065 ng/ul
RT: 10.761 min Scan# 11
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:128 Resp: 1713119
Ion Ratio Lower Upper
128 100
129 12.1 8.8 13.2
127 14.5 10.8 16.2



#32
4-Chloroaniline
Concen: 119.842 ng/ul
RT: 10.879 min Scan# 1326
Delta R.T. 0.019 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:127 Resp: 775129
Ion Ratio Lower Upper
127 100
129 33.4 26.1 39.1









#33
Hexachlorobutadiene
Concen: 132.405 ng/ul
RT: 11.038 min Scan# 11
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:225 Resp: 380245
Ion Ratio Lower Upper
225 100
223 63.6 49.7 74.5
227 65.3 53.2 79.8



#34
Caprolactam
Concen: 74.203 ng/ul
RT: 11.696 min Scan# 1465
Delta R.T. 0.060 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:113 Resp: 121016
Ion Ratio Lower Upper
113 100
55 235.7 179.8 269.6
56 173.4 132.2 198.4









#35
4-Chloro-3-methylphenol
Concen: 136.208 ng/ul
RT: 12.013 min Scan# 1519
Delta R.T. 0.025 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:107 Resp: 678716
Ion Ratio Lower Upper
107 100
144 24.5 19.0 28.6
142 73.5 55.8 83.6



#36
2-Methylnaphthalene
Concen: 121.731 ng/ul
RT: 12.365 min Scan# 1579
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:142 Resp: 1170835
Ion Ratio Lower Upper
142 100
141 89.6 72.6 108.8









#37
1-Methylnaphthalene
Concen: 120.834 ng/ul
RT: 12.589 min Scan# 1
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:142 Resp: 1165459
Ion Ratio Lower Upper
142 100
141 93.7 75.8 113.6
116 4.0 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 126.514 ng/ul
RT: 12.736 min Scan# 1642
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:216 Resp: 707943
Ion Ratio Lower Upper
216 100
214 79.3 62.2 93.4
179 19.8 16.2 24.2
108 15.7 12.2 18.4









#40
Hexachlorocyclopentadiene
Concen: 139.989 ng/ul
RT: 12.712 min Scan# 1638
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:237 Resp: 493413
Ion Ratio Lower Upper
237 100
235 64.9 50.9 76.3
272 13.3 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 136.768 ng/ul
RT: 12.982 min Scan# 1684
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:196 Resp: 518709
Ion Ratio Lower Upper
196 100
198 97.2 77.7 116.5
200 32.1 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 135.649 ng/ul
RT: 13.059 min Scan# 1
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:196 Resp: 556383
Ion Ratio Lower Upper
196 100
198 96.3 74.8 112.2
200 31.0 24.4 36.6



#43
1,1'-Biphenyl
Concen: 111.739 ng/ul
RT: 13.382 min Scan# 1752
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:154 Resp: 1478598
Ion Ratio Lower Upper
154 100
153 42.6 31.8 47.6
76 15.5 11.7 17.5









#44
2-Chloronaphthalene
Concen: 117.185 ng/ul
RT: 13.423 min Scan# 1758
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:162 Resp: 1244523
Ion Ratio Lower Upper
162 100
164 37.1 26.2 39.2
127 40.3 29.8 44.6



#45
2-Nitroaniline
Concen: 139.357 ng/ul
RT: 13.640 min Scan# 1796
Delta R.T. 0.025 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion: 65 Resp: 514564
Ion Ratio Lower Upper
65 100
92 63.6 51.4 77.2
138 89.2 71.6 107.4









#47
Dimethylphthalate
Concen: 112.074 ng/ul
RT: 14.005 min Scan# 1858
Delta R.T. 0.019 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:163 Resp: 1631321
Ion Ratio Lower Upper
163 100
194 4.1 2.6 3.8#
164 11.6 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 131.335 ng/ul
RT: 14.134 min Scan# 1880
Delta R.T. 0.025 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:165 Resp: 392395
Ion Ratio Lower Upper
165 100
89 58.3 48.9 73.3
121 20.0 16.2 24.4









#50
Acenaphthylene
Concen: 108.922 ng/ul
RT: 14.275 min Scan# 1904
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:152 Resp: 1859586
Ion Ratio Lower Upper
152 100
151 22.6 16.4 24.6
153 15.0 10.6 16.0



#51
3-Nitroaniline
Concen: 134.302 ng/ul
RT: 14.463 min Scan# 1936
Delta R.T. 0.019 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:138 Resp: 339100
Ion Ratio Lower Upper
138 100
108 11.6 8.6 13.0
92 126.1 108.4 162.6





#52

Acenaphthene

Concen: 114.334 ng/ul

RT: 14.610 min Scan# 1961

Delta R.T. 0.007 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :

SSTD160447





Tgt Ion:153 Resp: 1339445

Ion Ratio Lower Upper

153 100

152 49.8 39.1 58.7

154 87.6 68.7 103.1



#53

2,4-Dinitrophenol

Concen: 166.262 ng/ul

RT: 14.675 min Scan# 1972

Delta R.T. 0.025 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40




Tgt Ion:184 Resp: 258975


Ion Ratio Lower Upper

184 100

63 68.6 52.9 79.3

154 57.4 47.8 71.8









#55
4-Nitrophenol
Concen: 129.354 ng/ul
RT: 14.792 min Scan# 1986
Delta R.T. 0.037 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:109 Resp: 245284
Ion Ratio Lower Upper
109 100
139 127.5 100.3 150.5
65 151.1 120.6 180.8



#56
Dibenzofuran
Concen: 108.380 ng/ul
RT: 14.951 min Scan# 2019
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:168 Resp: 1804892
Ion Ratio Lower Upper
168 100
139 36.9 27.8 41.8









#57
2,4-Dinitrotoluene
Concen: 123.993 ng/ul
RT: 14.927 min Scan# 2011
Delta R.T. 0.025 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:165 Resp: 524879
Ion Ratio Lower Upper
165 100
63 53.1 33.8 50.6#
182 2.3 1.2 1.8#



#58
2,3,4,6-Tetrachlorophenol
Concen: 132.804 ng/ul
RT: 15.174 min Scan# 2057
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:232 Resp: 495875
Ion Ratio Lower Upper
232 100
131 37.7 29.0 43.6
130 2.2 1.1 1.7#
166 28.7 23.0 34.6









#59
Diethylphthalate
Concen: 108.293 ng/ul
RT: 15.368 min Scan# 2087
Delta R.T. 0.019 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:149 Resp: 1645191
Ion Ratio Lower Upper
149 100
177 24.2 18.4 27.6
150 13.6 10.0 15.0



#61
Fluorene
Concen: 105.265 ng/ul
RT: 15.597 min Scan# 2129
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:166 Resp: 1438351
Ion Ratio Lower Upper
166 100
165 99.9 80.2 120.2
167 14.7 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 112.982 ng/ul
RT: 15.585 min Scan# 2127
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:204 Resp: 816869
Ion Ratio Lower Upper
204 100
206 34.4 26.0 39.0
141 51.5 42.2 63.2



#63
4-Nitroaniline
Concen: 140.390 ng/ul
RT: 15.644 min Scan# 2137
Delta R.T. 0.037 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:138 Resp: 317542
Ion Ratio Lower Upper
138 100
92 55.6 45.0 67.4
108 83.0 60.8 91.2









#66
4,6-Dinitro-2-methylphenol
Concen: 142.112 ng/ul
RT: 15.691 min Scan# 2145
Delta R.T. 0.025 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:198 Resp: 335212
Ion Ratio Lower Upper
198 100
182 3.6 3.4 5.2
77 24.6 18.4 27.6



#67
N-Nitrosodiphenylamine
Concen: 115.398 ng/ul
RT: 15.803 min Scan# 2164
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:169 Resp: 1304546
Ion Ratio Lower Upper
169 100
168 71.2 55.7 83.5
167 39.7 31.2 46.8









#68
4-Bromophenyl-phenylether
Concen: 128.727 ng/ul
RT: 16.478 min Scan# 21
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:248 Resp: 600220
Ion Ratio Lower Upper
248 100
250 98.9 73.8 110.6
141 60.1 48.2 72.2



#69
Hexachlorobenzene
Concen: 128.760 ng/ul
RT: 16.596 min Scan# 2299
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:284 Resp: 672181
Ion Ratio Lower Upper
284 100
142 27.7 20.5 30.7
249 33.2 25.2 37.8





#70

Atrazine

Concen: 117.003 ng/ul

RT: 16.760 min Scan# 21

Delta R.T. 0.025 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :

SSTD160447





Tgt Ion:200 Resp: 540883

Ion Ratio Lower Upper

200 100

173 25.7 19.8 29.8

215 49.8 36.2 54.4



#71

Pentachlorophenol

Concen: 140.937 ng/ul

RT: 16.942 min Scan# 2358

Delta R.T. 0.013 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40




Tgt Ion:266 Resp: 441707


Ion Ratio Lower Upper

266 100

264 63.8 46.6 70.0

268 63.7 47.6 71.4









#72
Phenanthrene
Concen: 107.498 ng/ul
RT: 17.336 min Scan# 2423
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:178 Resp: 2322723
Ion Ratio Lower Upper
178 100
179 18.2 11.9 17.9#
176 22.4 15.8 23.8



#74
Anthracene
Concen: 102.913 ng/ul
RT: 17.430 min Scan# 2441
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:178 Resp: 2250251
Ion Ratio Lower Upper
178 100
179 19.2 13.4 20.2
176 22.3 16.2 24.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 133.695 ng/uL
RT: 13.347 min Scan# 1
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:216 Resp: 753179
Ion Ratio Lower Upper
216 100
214 78.8 62.2 93.4
218 49.5 40.9 61.3



#76
Pentachlorobenzene
Concen: 132.064 ng/uL
RT: 14.868 min Scan# 2005
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:250 Resp: 794580
Ion Ratio Lower Upper
250 100
248 63.6 51.6 77.4
252 65.3 54.4 81.6





#77

Carbazole

Concen: 109.187 ng/ul

RT: 17.700 min Scan# 2484

Delta R.T. 0.019 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :

SSTD160447





Tgt Ion:167 Resp: 2159198

Ion Ratio Lower Upper

167 100

166 25.0 17.8 26.8

139 13.8 9.4 14.0



#78

Di-n-butylphthalate

Concen: 101.440 ng/ul

RT: 18.247 min Scan# 2580

Delta R.T. 0.007 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40




Tgt Ion:149 Resp: 2555317


Ion Ratio Lower Upper

149 100

150 11.1 7.2 10.8#

104 6.2 3.9 5.9#









#80
Fluoranthene
Concen: 108.777 ng/ul
RT: 19.346 min Scan# 2765
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:202 Resp: 2707187
Ion Ratio Lower Upper
202 100
101 12.2 8.6 12.8
100 9.7 7.2 10.8



#82
Pyrene
Concen: 104.898 ng/ul
RT: 19.716 min Scan# 2830
Delta R.T. 0.019 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:202 Resp: 2627892
Ion Ratio Lower Upper
202 100
101 14.1 10.2 15.2
100 11.9 8.9 13.3





#83

Butylbenzylphthalate

Concen: 117.373 ng/ul

RT: 20.591 min Scan# 2979

Delta R.T. 0.013 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :

SSTD160447





Tgt Ion:149 Resp: 1192347

Ion Ratio Lower Upper

149 100

91 72.5 59.9 89.9

206 24.9 16.9 25.3



#84

3,3'-Dichlorobenzidine

Concen: 116.385 ng/ul

RT: 21.443 min Scan# 3124

Delta R.T. 0.013 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40




Tgt Ion:252 Resp: 960863


Ion Ratio Lower Upper

252 100

254 66.8 50.5 75.7

126 9.1 6.5 9.7









#85
Benzo(a)anthracene
Concen: 117.098 ng/ul
RT: 21.525 min Scan# 3136
Delta R.T. 0.013 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:228 Resp: 2867271
Ion Ratio Lower Upper
228 100
229 23.8 15.7 23.5#
226 31.3 21.8 32.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 111.027 ng/ul
RT: 21.420 min Scan# 3120
Delta R.T. 0.007 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:149 Resp: 1616289
Ion Ratio Lower Upper
149 100
167 29.1 21.6 32.4
279 4.4 3.0 4.4





#87

Chrysene

Concen: 117.887 ng/ul

RT: 21.596 min Scan# 31

Delta R.T. 0.019 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :

SSTD160447





Tgt Ion:228 Resp: 2705976

Ion Ratio Lower Upper

228 100

226 34.7 24.0 36.0

229 23.5 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 108.350 ng/ul




RT: 22.601 min Scan# 3321


Delta R.T. 0.013 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Tgt Ion:149 Resp: 2903109









#90
Benzo(b)fluoranthene
Concen: 120.742 ng/ul
RT: 23.699 min Scan# 3503
Delta R.T. 0.031 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40




Instrument :
BNA_G
ClientSampleId :
SSTD160447


Tgt Ion:252 Resp: 3260147
Ion Ratio Lower Upper
252 100
253 24.3 17.9 26.9
125 9.2 6.7 10.1



#91
Benzo(k)fluoranthene
Concen: 114.847 ng/ul
RT: 23.770 min Scan# 3520
Delta R.T. 0.037 min
Lab File: BG058288.D
Acq: 17 Jul 2023 18:40

Tgt Ion:252 Resp: 3039492
Ion Ratio Lower Upper
252 100
253 25.0 18.3 27.5
125 8.9 7.4 11.0





#93

Benzo(a)pyrene

Concen: 118.995 ng/ul

RT: 24.563 min Scan# 3

Delta R.T. 0.042 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40

Instrument :

BNA_G

ClientSampleId :

SSTD160447





Tgt Ion:252 Resp: 2843823

Ion Ratio Lower Upper

252 100

253 24.1 17.4 26.0

125 9.9 7.7 11.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 129.124 ng/ul

RT: 28.306 min Scan# 4292

Delta R.T. 0.078 min

Lab File: BG058288.D

Acq: 17 Jul 2023 18:40




Tgt Ion:276 Resp: 4097804


Ion Ratio Lower Upper

276 100

138 17.0 13.8 20.6

277 25.2 19.2 28.8









#95
 Dibenzo(a,h)anthracene
 Concen: 124.775 ng/ul
 RT: 28.370 min Scan# 41
 Delta R.T. 0.078 min
 Lab File: BG058288.D
 Acq: 17 Jul 2023 18:40

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160447

Tgt Ion:278 Resp: 3265279
 Ion Ratio Lower Upper
 278 100
 139 12.9 10.4 15.6
 279 25.6 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 130.175 ng/ul
 RT: 29.445 min Scan# 4486
 Delta R.T. 0.095 min
 Lab File: BG058288.D
 Acq: 17 Jul 2023 18:40

Tgt Ion:276 Resp: 3374301
 Ion Ratio Lower Upper
 276 100
 138 17.7 13.5 20.3
 277 24.8 19.4 29.2

