


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG071818\
 Data File : BG035736.D
 Acq On : 19 Jul 2018 3:21
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 31 12:30:06 2018



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.3	23061	PASS
68	69	0.00	2	1.7	462	PASS
69	198	0.00	100	31.9	27945	PASS
70	69	0.00	2	0.4	105	PASS
127	198	10	80	38.5	33773	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	87720	PASS
199	198	5	9	6.8	5993	PASS
275	198	10	60	30.1	26400	PASS
365	198	1	100	3.4	2977	PASS
441	198	0.01	100	10.4	9153	PASS
442	198	50	100	63.7	55877	PASS
443	442	15	24	18.0	10058	PASS