

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG072318\
 Data File : BG035838.D
 Acq On : 23 Jul 2018 21:50
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 24 02:28:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 20 15:28:54 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.613	0.538	12.2	93	0.00
3	Pyridine	1.601	1.449	9.5	86	0.00
4	n-Nitrosodimethylamine	0.687	0.593	13.7	86	0.00
5 S	2-Fluorophenol	1.205	1.169	3.0	95	0.00
6	Aniline	2.330	2.181	6.4	92	0.00
7 S	Phenol-d6	1.797	1.790	0.4	97	0.00
8	2-Chlorophenol	1.225	1.232	-0.6	99	0.00
9	Benzaldehyde	1.103	1.002	9.2	88	0.00
10 C	Phenol	1.816	1.778	2.1	95	0.00
11	bis(2-Chloroethyl)ether	1.422	1.369	3.7	95	0.00
12	1,3-Dichlorobenzene	1.480	1.499	-1.3	101	0.00
13 C	1,4-Dichlorobenzene	1.503	1.524	-1.4	101	0.00
14	1,2-Dichlorobenzene	1.444	1.452	-0.6	100	0.00
15	Benzyl Alcohol	1.632	1.531	6.2	92	0.00
16	2,2'-oxybis(1-Chloropropane	1.192	1.272	-6.7	106	0.00
17	2-Methylphenol	1.235	1.199	2.9	94	0.00
18	Hexachloroethane	0.575	0.558	3.0	98	0.00
19 P	n-Nitroso-di-n-propylamine	1.513	1.355	10.4	90	0.00
20	3+4-Methylphenols	1.713	1.715	-0.1	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
22	Acetophenone	0.622	0.572	8.0	92	0.00
23 S	Nitrobenzene-d5	0.479	0.452	5.6	92	0.00
24	Nitrobenzene	0.481	0.441	8.3	91	0.00
25	Isophorone	0.889	0.795	10.6	88	0.00
26 C	2-Nitrophenol	0.171	0.183	-7.0	102	0.00
27	2,4-Dimethylphenol	0.289	0.282	2.4	94	0.00
28	bis(2-Chloroethoxy)methane	0.490	0.462	5.7	93	0.00
29 C	2,4-Dichlorophenol	0.330	0.343	-3.9	98	0.00
30	1,2,4-Trichlorobenzene	0.405	0.411	-1.5	101	0.00
31	Naphthalene	1.042	1.011	3.0	99	0.00
32	Benzoic acid	0.195	0.154	21.0	77	0.00
33	4-Chloroaniline	0.454	0.423	6.8	93	0.00
34 C	Hexachlorobutadiene	0.316	0.323	-2.2	102	0.00
35	Caprolactam	0.142	0.128	9.9	92	0.00
36 C	4-Chloro-3-methylphenol	0.443	0.444	-0.2	97	0.00
37	2-Methylnaphthalene	0.776	0.767	1.2	97	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
39	1,2,4,5-Tetrachlorobenzene	0.755	0.762	-0.9	102	0.00
40 P	Hexachlorocyclopentadiene	0.396	0.367	7.3	90	0.00
41 S	2,4,6-Tribromophenol	0.264	0.285	-8.0	106	0.00
42 C	2,4,6-Trichlorophenol	0.441	0.474	-7.5	101	0.00
43	2,4,5-Trichlorophenol	0.494	0.517	-4.7	106	0.00
44 S	2-Fluorobiphenyl	1.493	1.457	2.4	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG072318\
 Data File : BG035838.D
 Acq On : 23 Jul 2018 21:50
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 24 02:28:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 20 15:28:54 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.551	1.520	2.0	97	0.00
46	2-Chloronaphthalene	1.206	1.178	2.3	99	0.00
47	2-Nitroaniline	0.426	0.428	-0.5	98	0.00
48	Acenaphthylene	2.020	2.009	0.5	99	0.00
49	Dimethylphthalate	1.752	1.714	2.2	99	0.00
50	2,6-Dinitrotoluene	0.357	0.381	-6.7	104	0.00
51 C	Acenaphthene	1.259	1.235	1.9	99	0.00
52	3-Nitroaniline	0.365	0.369	-1.1	97	0.00
53 P	2,4-Dinitrophenol	0.158	0.108	31.6#	67	0.00
54	Dibenzofuran	2.002	1.992	0.5	99	0.00
55 P	4-Nitrophenol	0.260	0.251	3.5	93	0.00
56	2,4-Dinitrotoluene	0.512	0.573	-11.9	105	0.00
57	Fluorene	1.678	1.686	-0.5	102	0.00
58	2,3,4,6-Tetrachlorophenol	0.473	0.506	-7.0	105	0.00
59	Diethylphthalate	1.802	1.781	1.2	99	0.00
60	4-Chlorophenyl-phenylether	0.986	0.994	-0.8	102	0.00
61	4-Nitroaniline	0.382	0.389	-1.8	97	0.00
62	Azobenzene	1.777	1.662	6.5	94	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
64	4,6-Dinitro-2-methylphenol	0.111	0.110	0.9	101	0.00
65 c	n-Nitrosodiphenylamine	0.569	0.575	-1.1	102	0.00
66	4-Bromophenyl-phenylether	0.232	0.236	-1.7	102	0.00
67	Hexachlorobenzene	0.239	0.249	-4.2	107	0.00
68	Atrazine	0.234	0.227	3.0	95	0.00
69 C	Pentachlorophenol	0.146	0.115	21.2#	78	0.00
70	Phenanthrene	0.998	1.014	-1.6	105	0.00
71	Anthracene	0.994	1.000	-0.6	103	0.00
72	Carbazole	0.944	0.939	0.5	101	0.00
73	Di-n-butylphthalate	1.115	1.119	-0.4	101	0.00
74 C	Fluoranthene	1.344	1.342	0.1	102	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
76	Benzidine	0.526	0.416	20.9	87	0.00
77	Pyrene	1.265	1.222	3.4	103	0.00
78 S	Terphenyl-d14	0.907	0.883	2.6	103	0.00
79	Butylbenzylphthalate	0.452	0.457	-1.1	104	0.00
80	Benzo(a)anthracene	1.246	1.214	2.6	104	0.00
81	3,3'-Dichlorobenzidine	0.435	0.436	-0.2	104	0.00
82	Chrysene	1.155	1.123	2.8	104	0.00
83	Bis(2-ethylhexyl)phthalate	0.615	0.636	-3.4	106	0.00
84 c	Di-n-octyl phthalate	1.029	1.083	-5.2	107	0.00
85	Indeno(1,2,3-cd)pyrene	1.279	1.275	0.3	104	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Benzo(b)fluoranthene	1.216	1.227	-0.9	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG072318\
Data File : BG035838.D
Acq On : 23 Jul 2018 21:50
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 24 02:28:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 20 15:28:54 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.188	1.204	-1.3	105	0.00
89 C	Benzo(a)pyrene	1.131	1.150	-1.7	105	0.00
90	Dibenzo(a,h)anthracene	1.110	1.122	-1.1	104	0.00
91	Benzo(a,h,i)perylene	1.086	1.071	1.4	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1