



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072423\
Data File : BG058428.D
Acq On : 24 Jul 2023 10:05
Operator : CG/JU
Sample : DFTPP522
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP522

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
Title : SVOA CALIBRATION
Last Update : Mon Jul 24 22:10:54 2023



AutoFind: Scans 2486, 2487, 2488; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	40192	PASS
68	69	0.00	2	1.8	624	PASS
69	198	0.00	100	33.9	34181	PASS
70	69	0.00	2	0.8	263	PASS
127	198	10	80	39.2	39483	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	100739	PASS
199	198	5	9	6.8	6829	PASS
275	198	10	60	26.3	26501	PASS
365	198	1	100	3.7	3681	PASS
441	198	0.01	100	12.7	12843	PASS
442	198	50	100	75.8	76395	PASS
443	442	15	24	19.3	14770	PASS

m/z	Abundance
41.00	196.0
42.90	394.0
43.90	1763.0
45.00	256.0
50.00	151.0
55.00	364.0
57.00	235.0
73.10	368.0
81.00	218.0
105.00	152.0
207.00	450.0
208.90	167.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522

m/z	Abundance
37.90	293.0
39.00	2362.0
39.90	338.0
41.00	238.0
43.00	513.0
43.90	1566.0
44.90	174.0
49.00	317.0
50.00	8457.0
51.00	37128.0
52.00	1912.0
52.90	152.0
55.00	683.0
56.00	1122.0
57.00	3000.0
58.00	181.0
60.90	647.0
62.00	455.0
63.00	1601.0
64.10	204.0
65.00	668.0
66.90	160.0
68.00	598.0
68.90	29984.0
69.90	303.0
73.00	513.0
74.00	3200.0
75.00	4845.0
76.00	1673.0
77.00	36880.0
78.00	2743.0
79.00	2301.0
79.90	1935.0
81.00	2231.0
82.00	736.0
83.00	804.0
83.90	162.0
84.90	589.0
85.90	673.0
86.90	308.0
91.00	482.0
91.90	667.0
93.00	3534.0
94.00	249.0
97.00	184.0
97.90	3088.0
99.00	2513.0
99.90	300.0
100.90	1330.0
103.00	593.0
103.90	575.0
105.00	842.0
106.00	289.0
107.00	10619.0
108.00	1680.0
109.90	20696.0
111.00	2983.0
111.90	398.0
115.10	156.0
116.00	642.0
117.00	6348.0
117.90	552.0
121.90	641.0
123.00	889.0
123.90	447.0
124.90	374.0
125.80	211.0
127.00	33792.0
128.00	2449.0
129.00	12707.0
130.00	966.0
131.00	177.0
134.00	332.0
135.00	1258.0
135.90	447.0
137.00	482.0
140.00	182.0
141.00	1833.0
141.90	587.0
143.00	645.0
144.90	204.0
145.90	243.0
146.90	928.0
148.00	2042.0
148.90	625.0
150.10	163.0
151.00	194.0
151.60	250.0
153.00	618.0
153.90	345.0
155.00	1185.0
156.00	1506.0

Instrument :

BNA_G

ClientSampleId :

DFTPP522

157.00	320.0
157.90	254.0
159.00	229.0
160.00	527.0
161.00	770.0
162.00	346.0
165.00	779.0
166.00	622.0
167.00	3292.0
167.90	1499.0
169.10	392.0
170.90	159.0
172.00	366.0
172.90	455.0
173.90	784.0
175.00	1416.0
176.00	479.0
177.00	560.0
178.00	269.0
179.00	2893.0
180.00	1795.0
181.00	974.0
184.00	197.0
184.90	1083.0
186.00	10284.0
187.00	3375.0
187.90	277.0
188.80	672.0
190.90	451.0
192.00	883.0
193.00	983.0
194.00	195.0
195.10	170.0
196.00	2210.0
197.90	80600.0
198.90	5554.0
199.90	439.0
201.60	352.0
203.00	581.0
204.00	2884.0
205.00	5170.0
206.00	21312.0
207.00	3335.0
207.90	775.0
209.00	304.0
211.00	707.0
212.10	230.0
214.80	292.0
215.90	472.0
216.90	5255.0
217.90	656.0
221.00	3030.0
221.90	723.0
222.90	1290.0
224.00	11711.0
225.00	2574.0
225.90	389.0
227.00	4823.0
228.00	533.0
228.90	1145.0
230.00	165.0
230.90	243.0
233.90	362.0
235.00	342.0
235.90	219.0
237.00	275.0
239.90	164.0
241.00	197.0
241.90	590.0
242.90	551.0
244.00	8685.0
245.00	1226.0
246.00	1624.0
247.00	348.0
248.90	319.0
252.30	188.0
253.00	232.0
255.00	45544.0
256.00	6686.0
257.00	626.0
258.00	2424.0
259.00	499.0
265.00	1480.0
265.90	265.0
270.90	162.0
273.00	1285.0
274.00	3325.0
275.00	19176.0
276.00	2699.0
276.90	1479.0
278.00	205.0
280.80	156.0
282.90	329.0
284.10	186.0
285.00	397.0
290.10	178.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522

292.90	472.0
294.00	163.0
296.00	5327.0
297.00	919.0
302.00	156.0
303.10	491.0
304.00	161.0
313.90	244.0
314.90	601.0
315.90	420.0
320.90	208.0
323.00	1756.0
324.00	375.0
326.90	264.0
328.00	276.0
332.90	162.0
334.10	986.0
334.90	347.0
341.00	309.0
346.00	412.0
351.90	804.0
353.00	368.0
354.00	563.0
355.10	184.0
364.90	2364.0
366.00	401.0
372.00	824.0
372.90	198.0
383.00	376.0
390.00	151.0
402.00	362.0
403.00	505.0
403.90	261.0
421.10	460.0
422.00	351.0
422.90	2626.0
423.90	646.0
441.00	7744.0
442.00	47440.0
443.00	8902.0
444.00	718.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522

m/z	Abundance
38.00	410.0
39.00	3246.0
39.90	208.0
41.00	333.0
43.10	299.0
44.00	1649.0
44.90	213.0
49.00	363.0
50.00	10940.0
51.00	45944.0
52.00	2358.0
55.00	523.0
56.00	1244.0
56.90	3190.0
57.90	162.0
59.90	155.0
60.90	392.0
61.90	618.0
63.00	1819.0
63.90	221.0
65.00	842.0
68.00	738.0
68.90	39296.0
69.90	218.0
73.00	442.0
74.00	3485.0
75.00	6107.0
76.00	2189.0
77.00	46592.0
78.00	3351.0
79.00	2612.0
79.90	2472.0
80.90	3277.0
81.90	676.0
83.00	925.0
83.90	250.0
84.90	656.0
86.00	1075.0
87.00	446.0
88.90	170.0
91.00	829.0
92.00	836.0
93.00	4693.0
94.00	274.0
94.80	191.0
95.90	289.0
96.90	206.0
98.00	3553.0
99.00	2833.0
99.90	207.0
101.00	1918.0
103.00	501.0
104.00	992.0
105.00	995.0
106.00	502.0
107.00	13851.0
108.00	2012.0
110.00	25992.0
111.00	3134.0
111.90	313.0
116.00	711.0
117.00	10287.0
118.00	845.0
122.00	592.0
123.00	1057.0
123.90	525.0
124.90	598.0
127.00	45160.0
128.00	3517.0
129.00	16776.0
130.00	1528.0
131.00	362.0
132.10	200.0
133.90	451.0
135.00	1422.0
136.00	625.0
137.00	746.0
138.00	218.0
139.00	151.0
140.00	156.0
140.90	2319.0
142.10	794.0
142.90	482.0
144.00	169.0
144.80	212.0
145.90	468.0
147.00	1244.0
148.00	2061.0
148.90	413.0
150.00	194.0
151.00	394.0
152.90	566.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522

153.90	470.0
155.00	1142.0
156.00	2144.0
156.90	467.0
157.90	485.0
159.00	335.0
160.00	847.0
160.90	1120.0
161.90	307.0
163.90	155.0
164.90	927.0
166.00	754.0
167.00	4603.0
168.00	2394.0
169.00	358.0
169.90	152.0
172.00	458.0
172.90	749.0
174.00	1037.0
175.00	1845.0
176.00	685.0
177.00	913.0
178.00	247.0
178.90	3650.0
180.00	2833.0
181.00	1208.0
182.00	323.0
183.90	171.0
185.00	1778.0
186.00	15499.0
187.00	4751.0
188.10	495.0
188.90	957.0
189.90	175.0
191.00	602.0
192.00	1550.0
193.00	1447.0
194.00	421.0
194.90	290.0
196.00	3346.0
197.90	116456.0
198.90	8124.0
199.90	738.0
201.40	378.0
203.00	733.0
204.00	4217.0
205.00	6956.0
206.00	29800.0
207.00	4320.0
208.00	979.0
208.90	402.0
210.00	385.0
210.90	934.0
211.80	226.0
213.00	212.0
214.90	326.0
215.90	690.0
216.90	7532.0
218.00	890.0
221.00	4126.0
221.90	961.0
223.00	1606.0
224.00	17968.0
225.00	4580.0
225.90	435.0
226.90	6800.0
228.00	903.0
228.90	1317.0
230.00	275.0
231.00	736.0
234.00	552.0
234.90	335.0
236.00	239.0
236.90	481.0
238.90	320.0
240.10	201.0
241.00	267.0
242.00	773.0
243.00	822.0
244.00	13359.0
245.00	1745.0
246.00	2664.0
246.90	455.0
249.00	434.0
249.90	163.0
251.00	171.0
251.90	224.0
253.00	325.0
255.00	70432.0
256.00	10200.0
257.10	805.0
258.00	3464.0
258.90	773.0
259.90	160.0
261.00	162.0
265.00	1299.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522

266.00	249.0
271.00	208.0
272.00	204.0
273.00	1799.0
274.00	5973.0
275.00	31040.0
276.00	3980.0
276.90	2364.0
278.00	486.0
278.90	155.0
281.00	190.0
283.10	352.0
284.00	215.0
284.90	472.0
293.00	578.0
295.00	215.0
296.00	8783.0
296.90	1170.0
302.00	221.0
303.00	1060.0
303.90	292.0
314.00	457.0
315.00	1035.0
316.00	701.0
320.90	329.0
322.10	163.0
323.00	2995.0
324.00	511.0
326.90	551.0
327.80	236.0
331.90	170.0
332.90	230.0
334.00	1667.0
334.90	410.0
340.90	316.0
346.00	425.0
352.00	749.0
352.90	590.0
354.10	982.0
355.00	218.0
364.90	4413.0
365.90	475.0
370.90	270.0
372.00	1809.0
373.00	451.0
382.80	466.0
390.00	331.0
402.00	652.0
403.00	970.0
403.90	216.0
421.00	836.0
421.90	703.0
423.00	5151.0
424.00	964.0
425.10	154.0
441.00	14924.0
442.00	86168.0
443.00	17904.0
444.00	1692.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522

m/z	Abundance
37.90	512.0
39.00	2824.0
39.90	259.0
41.00	296.0
43.00	414.0
44.00	1445.0
44.90	226.0
48.90	301.0
50.00	8610.0
51.00	37504.0
52.00	1896.0
54.90	367.0
55.90	994.0
56.90	2679.0
58.00	243.0
61.00	806.0
62.00	571.0
63.00	1495.0
64.00	195.0
65.00	749.0
67.10	225.0
68.00	537.0
68.90	33264.0
69.90	268.0
72.90	308.0
74.00	2940.0
75.00	5485.0
76.00	1884.0
77.00	40216.0
78.00	2465.0
79.00	2309.0
79.90	1751.0
80.90	2818.0
81.90	813.0
82.90	788.0
84.00	186.0
85.00	612.0
86.00	708.0
87.00	332.0
90.90	597.0
91.90	545.0
93.00	4021.0
94.00	270.0
96.00	258.0
97.00	241.0
97.90	2835.0
99.00	2581.0
99.90	212.0
100.90	1754.0
102.00	171.0
102.90	407.0
104.00	717.0
105.00	868.0
106.00	360.0
107.00	11356.0
108.00	1725.0
109.90	22672.0
111.00	3227.0
111.90	454.0
116.00	594.0
116.90	7697.0
118.00	580.0
120.10	213.0
122.00	842.0
122.90	1209.0
123.90	543.0
125.00	443.0
127.00	39496.0
128.00	3429.0
129.00	14626.0
129.90	1500.0
130.90	208.0
131.80	233.0
133.90	560.0
135.00	1143.0
136.00	511.0
137.00	679.0
141.00	1670.0
141.90	724.0
142.90	454.0
144.80	161.0
146.00	321.0
147.00	914.0
147.90	2200.0
148.90	473.0
151.00	324.0
151.70	251.0
151.90	259.0
153.00	760.0
154.00	501.0
155.00	1221.0
156.00	1832.0

Instrument :

BNA_G

ClientSampleId :

DFTPP522

157.00	344.0
157.70	309.0
158.90	293.0
160.00	760.0
161.00	1172.0
162.00	272.0
163.00	151.0
164.90	761.0
166.00	773.0
167.00	4867.0
168.00	2111.0
169.00	287.0
170.90	193.0
172.00	571.0
173.00	516.0
173.90	980.0
175.00	1844.0
175.90	433.0
177.00	759.0
177.90	272.0
178.90	3908.0
180.00	2339.0
180.90	1081.0
181.90	226.0
183.10	210.0
184.00	268.0
185.00	1820.0
186.00	12857.0
187.00	3566.0
188.00	467.0
188.90	733.0
189.90	169.0
190.90	406.0
192.00	934.0
193.00	1428.0
194.00	354.0
194.90	212.0
196.00	3202.0
197.90	105160.0
198.90	6809.0
199.90	623.0
201.30	363.0
202.00	247.0
203.00	675.0
204.00	3780.0
205.00	6527.0
206.00	28824.0
207.00	4323.0
208.00	920.0
209.00	292.0
211.00	952.0
214.90	312.0
215.90	653.0
216.90	7401.0
217.90	859.0
221.00	4015.0
221.80	938.0
223.00	1884.0
224.00	16264.0
225.00	4232.0
226.10	508.0
227.00	6668.0
227.90	850.0
228.90	1358.0
230.00	212.0
231.00	579.0
232.90	183.0
233.90	445.0
234.90	524.0
236.00	273.0
236.90	597.0
239.00	374.0
240.00	188.0
240.90	362.0
242.00	876.0
243.00	872.0
244.00	13442.0
245.00	2070.0
246.00	2550.0
246.90	339.0
248.90	587.0
251.00	162.0
253.00	332.0
255.00	66920.0
256.00	9550.0
257.00	751.0
258.00	3154.0
259.00	572.0
264.10	152.0
264.90	1579.0
265.90	248.0
266.90	162.0
271.90	277.0
273.00	2084.0
274.00	5742.0
275.00	29288.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522

276.00	4345.0
277.00	2523.0
277.90	365.0
282.90	326.0
283.90	306.0
285.00	488.0
288.80	159.0
289.90	156.0
293.00	735.0
294.10	213.0
294.80	152.0
296.00	8623.0
296.90	1307.0
302.00	170.0
303.00	859.0
304.00	340.0
307.80	244.0
310.00	202.0
314.10	403.0
315.00	901.0
316.00	558.0
320.90	331.0
323.00	2720.0
324.00	498.0
326.80	487.0
327.90	283.0
331.90	250.0
333.00	370.0
334.00	1931.0
335.00	463.0
341.00	492.0
346.00	478.0
352.00	1151.0
352.90	539.0
354.00	957.0
355.20	168.0
365.00	4265.0
366.00	623.0
370.80	308.0
372.00	1670.0
373.00	311.0
383.00	459.0
383.80	166.0
389.90	284.0
391.00	241.0
401.90	533.0
402.90	888.0
404.00	327.0
421.00	1030.0
422.00	807.0
423.00	5498.0
424.00	1087.0
441.00	15860.0
442.00	95576.0
443.00	17504.0
444.00	1735.0

Instrument :
BNA_G
ClientSampleId :
DFTPP522