

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG072718\
 Data File : BG035944.D
 Acq On : 27 Jul 2018 19:21
 Operator : JU/SJ
 Sample : J4192-01MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Quant Time: Jul 27 23:15:29 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 25 11:04:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.03	152	33786	20.00	ng	0.00
21) Naphthalene-d8	10.83	136	138441	20.00	ng	0.00
38) Acenaphthene-d10	14.65	164	109840	20.00	ng	0.00
63) Phenanthrene-d10	17.40	188	296748	20.00	ng	0.00
75) Chrysene-d12	21.66	240	301698	20.00	ng	0.00
86) Perylene-d12	24.86	264	299362	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.61	112	293362	144.16	ng	0.00
7) Phenol-d6	7.20	99	452549	149.05	ng	0.00
23) Nitrobenzene-d5	9.19	82	294087	88.74	ng	0.00
41) 2,4,6-Tribromophenol	16.14	330	221490	152.99	ng	0.00
44) 2-Fluorobiphenyl	13.27	172	703244	85.78	ng	0.00
78) Terphenyl-d14	20.00	244	1261487	92.17	ng	0.00

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.51	88	29931	28.898	ng	# 75
3) Pyridine	3.91	79	83807	30.994	ng	# 71
4) n-Nitrosodimethylamine	3.83	42	41287	35.561	ng	# 76
6) Aniline	7.35	93	126149	32.055	ng	98
8) 2-Chlorophenol	7.59	128	98635	47.656	ng	98
9) Benzaldehyde	7.16	77	70447	37.815	ng	95
10) Phenol	7.22	94	162439	52.955	ng	93
11) bis(2-Chloroethyl)ether	7.44	93	97709	40.673	ng	90
12) 1,3-Dichlorobenzene	7.91	146	101452	40.591	ng	94
13) 1,4-Dichlorobenzene	8.06	146	107756	42.432	ng	95
14) 1,2-Dichlorobenzene	8.38	146	104610	42.880	ng	96
15) Benzyl Alcohol	8.27	79	117995	42.796	ng	90
16) 2,2'-oxybis(1-Chloropropan	8.55	45	115175	57.187	ng	78
17) 2-Methylphenol	8.47	107	102164	48.986	ng	# 93
18) Hexachloroethane	9.10	117	39622	40.806	ng	90
19) n-Nitroso-di-n-propylamine	8.82	70	97961	38.315	ng	# 83
20) 3+4-Methylphenols	8.80	107	140844	48.680	ng	94
22) Acetophenone	8.85	105	177912	41.326	ng	# 92
24) Nitrobenzene	9.23	77	140866	42.266	ng	91
25) Isophorone	9.75	82	288528	46.901	ng	98
26) 2-Nitrophenol	9.94	139	60715	51.294	ng	# 84
27) 2,4-Dimethylphenol	10.00	122	109144	54.473	ng	90
28) bis(2-Chloroethoxy)methane	10.23	93	151040	44.557	ng	97
29) 2,4-Dichlorophenol	10.48	162	121534	53.138	ng	94
30) 1,2,4-Trichlorobenzene	10.69	180	128537	45.831	ng	99
31) Naphthalene	10.88	128	319131	44.253	ng	98
32) Benzoic acid	10.15	122	13438	9.963	ng	95
33) 4-Chloroaniline	10.99	127	54772	17.426	ng	94
34) Hexachlorobutadiene	11.16	225	93259	42.643	ng	94
35) Caprolactam	11.77	113	24626	25.090	ng	# 62
36) 4-Chloro-3-methylphenol	12.11	107	154018	50.239	ng	95
37) 2-Methylnaphthalene	12.48	142	263178	48.984	ng	96
39) 1,2,4,5-Tetrachlorobenzene	12.85	216	173933	41.955	ng	99
40) Hexachlorocyclopentadiene	12.83	237	211094	97.090	ng	92

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG072718\
 Data File : BG035944.D
 Acq On : 27 Jul 2018 19:21
 Operator : JU/SJ
 Sample : J4192-01MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1


Quant Time: Jul 27 23:15:29 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 25 11:04:14 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.09	196	117470	48.452	ng	95
43) 2,4,5-Trichlorophenol	13.17	196	129409	47.713	ng	98
45) 1,1'-Biphenyl	13.48	154	367418	43.145	ng	99
46) 2-Chloronaphthalene	13.53	162	293210	44.265	ng	98
47) 2-Nitroaniline	13.73	65	102567	43.798	ng	97
48) Acenaphthylene	14.37	152	489892	44.151	ng	98
49) Dimethylphthalate	14.11	163	455563	47.338	ng	99
50) 2,6-Dinitrotoluene	14.22	165	97418	49.710	ng	91
51) Acenaphthene	14.71	154	290392	42.012	ng	99
52) 3-Nitroaniline	14.56	138	55162	27.540	ng	# 97
53) 2,4-Dinitrophenol	14.76	184	66256	66.918	ng	# 64
54) Dibenzofuran	15.05	168	484228	44.050	ng	93
55) 4-Nitrophenol	14.88	139	127756	89.562	ng	# 85
56) 2,4-Dinitrotoluene	15.01	165	136734	48.634	ng	# 86
57) Fluorene	15.70	166	409467	44.442	ng	94
58) 2,3,4,6-Tetrachlorophenol	15.28	232	125572	48.288	ng	# 93
59) Diethylphthalate	15.46	149	414941	41.932	ng	97
60) 4-Chlorophenyl-phenylether	15.69	204	233269	43.077	ng	96
61) 4-Nitroaniline	15.72	138	92052	43.935	ng	# 80
62) Azobenzene	15.98	77	419237	42.951	ng	95
64) 4,6-Dinitro-2-methylphenol	15.78	198	75636	45.788	ng	86
65) n-Nitrosodiphenylamine	15.90	169	366558	43.397	ng	97
66) 4-Bromophenyl-phenylether	16.58	248	157841	45.847	ng	92
67) Hexachlorobenzene	16.70	284	163856	46.216	ng	93
68) Atrazine	16.85	200	159203	45.778	ng	95
69) Pentachlorophenol	17.05	266	145851	67.540	ng	99
70) Phenanthrene	17.44	178	657088	44.376	ng	98
71) Anthracene	17.52	178	661507	44.866	ng	97
72) Carbazole	17.80	167	599229	42.796	ng	98
73) Di-n-butylphthalate	18.35	149	683238	41.292	ng	# 98
74) Fluoranthene	19.44	202	838254	42.028	ng	97
76) Benzidine	19.62	184	326549	41.170	ng	98
77) Pyrene	19.80	202	824990	43.246	ng	98
79) Butylbenzylphthalate	20.69	149	309723	45.401	ng	97
80) Benzo(a)anthracene	21.64	228	838141	44.580	ng	99
81) 3,3'-Dichlorobenzidine	21.55	252	186489	28.448	ng	99
82) Chrysene	21.70	228	791677	45.440	ng	98
83) Bis(2-ethylhexyl)phthalate	21.54	149	435848	46.995	ng	# 98
84) Di-n-octyl phthalate	22.74	149	740977	47.716	ng	# 92
85) Indeno(1,2,3-cd)pyrene	28.49	276	916694	47.524	ng	# 94
87) Benzo(b)fluoranthene	23.84	252	820983	45.121	ng	# 96
88) Benzo(k)fluoranthene	23.91	252	773568	43.487	ng	# 97
89) Benzo(a)pyrene	24.70	252	778705	46.003	ng	# 97
90) Dibenzo(a,h)anthracene	28.56	278	762506	45.883	ng	# 95
91) Benzo(g,h,i)perylene	29.64	276	760597	46.772	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG072718\
 Data File : BG035944.D
 Acq On : 27 Jul 2018 19:21
 Operator : JU/SJ
 Sample : J4192-01MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1




Quant Time: Jul 27 23:15:29 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 25 11:04:14 2018
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.03 min Scan# 806
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
152	33786		
152	100		
150	132.5	126.6	189.8
115	62.6	53.0	79.4



#2
 1,4-Dioxane
 Concen: 28.898 ng
 RT: 3.51 min Scan# 38
 Delta R.T. -0.01 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
88	29931		
88	100		
58	67.2	79.6	119.4#
43	31.2	27.4	41.0









#3
 Pyridine
 Concen: 30.994 ng
 RT: 3.91 min Scan# 105
 Delta R.T. -0.01 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
79	100		
52	57.4	69.9	104.9#
51	28.6	34.0	51.0#



#4
 n-Nitrosodimethylamine
 Concen: 35.561 ng
 RT: 3.83 min Scan# 91
 Delta R.T. -0.01 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
42	100		
74	152.5	102.5	153.7
44	5.6	18.9	28.3#









#5
 2-Fluorophenol
 Concen: 144.157 ng
 RT: 5.61 min Scan# 394
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
112	100		
64	62.8	56.3	84.5
63	37.2	30.1	45.1



#6
 Aniline
 Concen: 32.055 ng
 RT: 7.35 min Scan# 691
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
93	100		
66	40.9	33.7	50.5
65	21.0	16.1	24.1









#7
 Phenol-d6
 Concen: 149.050 ng
 RT: 7.20 min Scan# 665
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
99	100		
42	16.5	14.1	21.1
71	42.2	26.1	39.1#



#8
 2-Chlorophenol
 Concen: 47.656 ng
 RT: 7.59 min Scan# 732
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
128	100		
130	31.9	14.0	54.0
64	57.3	37.7	77.7








#9
Benzaldehyde
Concen: 37.815 ng
RT: 7.16 min Scan# 659
Delta R.T. -0.00 min
Lab File: BG035944.D
Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
77	100		
105	85.3	71.3	111.3
106	81.7	64.2	104.2



#10
Phenol
Concen: 52.955 ng
RT: 7.22 min Scan# 669
Delta R.T. 0.00 min
Lab File: BG035944.D
Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
94	100		
65	33.3	11.9	51.9
66	48.5	22.2	62.2








#11
 bis(2-Chloroethyl)ether
 Concen: 40.673 ng
 RT: 7.44 min Scan# 707
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Ion Ratio	Lower	Upper
93	97709	100		
63		74.7	67.1	107.1
95		31.7	11.5	51.5



#12
 1,3-Dichlorobenzene
 Concen: 40.591 ng
 RT: 7.91 min Scan# 787
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Ion Ratio	Lower	Upper
146	101452	100		
148		67.2	51.4	77.2
75		38.8	26.6	39.8









#13
 1,4-Dichlorobenzene
 Concen: 42.432 ng
 RT: 8.06 min Scan# 812
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
146	107756		
148	61.7	53.7	80.5
111	42.9	35.2	52.8



#14
 1,2-Dichlorobenzene
 Concen: 42.880 ng
 RT: 8.38 min Scan# 866
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
146	104610		
148	62.8	49.3	73.9
111	47.2	34.3	51.5








#15
 Benzyl Alcohol
 Concen: 42.796 ng
 RT: 8.27 min Scan# 847
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
79	117995		
108	63.2	57.2	85.8
77	64.2	46.1	69.1



#16
 2,2'-oxybis(1-Chloropropane)
 Concen: 57.187 ng
 RT: 8.55 min Scan# 895
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
45	115175		
77	20.8	0.0	30.2
79	12.5	0.0	27.9








#17
 2-Methylphenol
 Concen: 48.986 ng
 RT: 8.47 min Scan# 881
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
107	102164		
107	100		
108	105.0	83.9	125.9
77	55.7	37.2	55.8
79	55.6	36.2	54.2#



#18
 Hexachloroethane
 Concen: 40.806 ng
 RT: 9.10 min Scan# 989
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
117	39622		
117	100		
119	79.6	76.7	115.1
201	115.1	95.7	143.5








#19
 n-Nitroso-di-n-propylamine
 Concen: 38.315 ng
 RT: 8.82 min Scan# 942
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21

Tgt Ion	Resp	Lower	Upper
70	100		
42	43.2	49.1	73.7#
101	10.7	8.5	12.7
130	18.7	13.4	20.0




Abundance Ion 70.00 (69.70 to 70.70): BGC
 Ion 42.00 (41.70 to 42.70): BGC
 Ion 101.00 (100.70 to 101.70): E
 Ion 130.00 (129.70 to 130.70): E






#20
 3+4-Methylphenols
 Concen: 48.680 ng
 RT: 8.80 min Scan# 937
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21

Tgt Ion	Resp	Lower	Upper
107	100		
108	83.2	61.6	101.6
77	39.5	13.9	53.9
79	35.5	8.8	48.8




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 108.00 (107.70 to 108.70): E
 Ion 77.00 (76.70 to 77.70): BGC
 Ion 79.00 (78.70 to 79.70): BGC








#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.83 min Scan# 1283
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21

Tgt Ion	Resp	Lower	Upper
136	138441		
137	12.0	9.2	13.8
54	7.0	10.3	15.5#
68	5.0	4.5	6.7




Abundance Ion 136.00 (135.70 to 136.70): E
 Ion 137.00 (136.70 to 137.70): E
 Ion 54.00 (53.70 to 54.70): BG
 Ion 68.00 (67.70 to 68.70): BG






#22
 Acetophenone
 Concen: 41.326 ng
 RT: 8.85 min Scan# 946
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21

Tgt Ion	Resp	Lower	Upper
105	177912		
71	1.3	3.0	4.4#
51	25.8	26.1	39.1#
120	21.1	17.4	26.2






Abundance Ion 105.00 (104.70 to 105.70): E
 Ion 71.00 (70.70 to 71.70): BG
 Ion 51.00 (50.70 to 51.70): BG
 Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 88.741 ng
 RT: 9.19 min Scan# 1004
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
82	100		
128	35.3	29.7	44.5
54	44.0	43.1	64.7



#24
 Nitrobenzene
 Concen: 42.266 ng
 RT: 9.23 min Scan# 1011
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
77	100		
123	34.9	33.0	49.6
65	13.2	13.0	19.6









#25
 Isophorone
 Concen: 46.901 ng
 RT: 9.75 min Scan# 1099
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
82	288528		
95	7.7	6.2	9.2
138	13.7	12.1	18.1



#26
 2-Nitrophenol
 Concen: 51.294 ng
 RT: 9.94 min Scan# 1132
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
139	60715		
109	46.2	24.2	36.2#
65	65.8	47.4	71.0









#27
 2,4-Dimethylphenol
 Concen: 54.473 ng
 RT: 10.00 min Scan# 1142
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
122	109144		
107	141.0	100.2	150.2
121	55.3	44.4	66.6



#28
 bis(2-Chloroethoxy)methane
 Concen: 44.557 ng
 RT: 10.23 min Scan# 1181
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
93	151040		
95	32.3	24.3	36.5
123	10.6	8.4	12.6









#29
 2,4-Dichlorophenol
 Concen: 53.138 ng
 RT: 10.48 min Scan# 1224
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
162	121534		
164	61.2	48.0	88.0
98	39.9	21.1	61.1



#30
 1,2,4-Trichlorobenzene
 Concen: 45.831 ng
 RT: 10.69 min Scan# 1259
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
180	128537		
182	96.0	77.5	116.3
145	27.6	23.4	35.2









#31
 Naphthalene
 Concen: 44.253 ng
 RT: 10.88 min Scan# 1292
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
128	319131		
129	10.2	8.6	12.8
127	13.6	11.6	17.4



#32
 Benzoic acid
 Concen: 9.963 ng
 RT: 10.15 min Scan# 1167
 Delta R.T. -0.02 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
122	13438		
105	138.9	123.5	163.5
77	113.3	85.7	125.7








#33
 4-Chloroaniline
 Concen: 17.426 ng
 RT: 10.99 min Scan# 1311
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
127	100		
129	31.9	24.0	36.0
65	38.7	27.0	40.4
92	22.4	16.6	25.0



#34
 Hexachlorobutadiene
 Concen: 42.643 ng
 RT: 11.16 min Scan# 1339
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
225	100		
223	65.2	49.7	74.5
227	65.8	48.1	72.1








#35
 Caprolactam
 Concen: 25.090 ng
 RT: 11.77 min Scan# 1443
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
113	100		
55	133.3	186.9	226.9#
56	118.7	130.9	170.9#



#36
 4-Chloro-3-methylphenol
 Concen: 50.239 ng
 RT: 12.11 min Scan# 1502
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
107	100		
144	21.5	18.2	27.4
142	69.1	59.0	88.4









#37
 2-Methylnaphthalene
 Concen: 48.984 ng
 RT: 12.48 min Scan# 1564
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
142	263178		
141	82.9	67.1	100.7
115	45.3	30.7	46.1



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.65 min Scan# 1933
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
164	109840		
162	92.0	78.8	118.2
160	41.6	35.4	53.0








#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 41.955 ng
 RT: 12.85 min Scan# 1627
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
216	173933		
214	78.8	63.0	94.4
179	20.4	17.0	25.6
108	16.7	12.8	19.2



#40
 Hexachlorocyclopentadiene
 Concen: 97.090 ng
 RT: 12.83 min Scan# 1623
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
237	211094		
235	62.6	50.4	90.4
272	11.8	0.0	31.8








#41
 2,4,6-Tribromophenol
 Concen: 152.988 ng
 RT: 16.14 min Scan# 2187
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
330	100		
332	94.7	77.0	115.6
141	39.1	25.2	37.8#



#42
 2,4,6-Trichlorophenol
 Concen: 48.452 ng
 RT: 13.09 min Scan# 1668
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
196	100		
198	91.6	78.2	117.2
200	31.6	24.6	36.8








#43
 2,4,5-Trichlorophenol
 Concen: 47.713 ng
 RT: 13.17 min Scan# 1681
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
196	129409		
198	95.8	76.2	114.4
97	46.4	38.7	58.1
132	23.5	20.2	30.2



#44
 2-Fluorobiphenyl
 Concen: 85.776 ng
 RT: 13.27 min Scan# 1699
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
172	703244		
171	35.6	30.0	45.0
170	24.2	19.5	29.3








#45
 1,1'-Biphenyl
 Concen: 43.145 ng
 RT: 13.48 min Scan# 1735
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
154	367418		
153	40.2	19.8	59.8
76	13.2	0.0	31.9



#46
 2-Chloronaphthalene
 Concen: 44.265 ng
 RT: 13.53 min Scan# 1743
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
162	293210		
127	36.7	30.7	46.1
164	32.8	26.0	39.0






#47
 2-Nitroaniline
 Concen: 43.798 ng
 RT: 13.73 min Scan# 1777
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
65	102567		
92	73.1	54.6	82.0
138	91.7	72.9	109.3



#48
 Acenaphthylene
 Concen: 44.151 ng
 RT: 14.37 min Scan# 1886
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
152	489892		
151	20.3	17.2	25.8
153	13.1	10.2	15.4









#49
 Dimethylphthalate
 Concen: 47.338 ng
 RT: 14.11 min Scan# 1841
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
163	100		
194	4.9	3.4	5.2
164	10.6	8.2	12.4



#50
 2,6-Dinitrotoluene
 Concen: 49.710 ng
 RT: 14.22 min Scan# 1861
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
165	100		
63	54.3	52.7	79.1
89	58.8	49.2	73.8








#51
 Acenaphthene
 Concen: 42.012 ng
 RT: 14.71 min Scan# 1944
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
154	100		
153	114.0	90.4	135.6
152	53.5	41.6	62.4



#52
 3-Nitroaniline
 Concen: 27.540 ng
 RT: 14.56 min Scan# 1918
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
138	100		
108	13.8	17.5	26.3#
92	133.5	105.8	158.8








#53
 2,4-Dinitrophenol
 Concen: 66.918 ng
 RT: 14.76 min Scan# 1953
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
184	100		
63	59.3	59.8	89.8#
154	74.1	101.8	152.8#



#54
 Dibenzofuran
 Concen: 44.050 ng
 RT: 15.05 min Scan# 2001
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
168	100		
139	41.3	28.6	43.0
169	14.5	11.2	16.8








#55
 4-Nitrophenol
 Concen: 89.562 ng
 RT: 14.88 min Scan# 1972
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
139	127756		
109	115.7	62.2	102.2#
65	117.4	97.2	137.2



#56
 2,4-Dinitrotoluene
 Concen: 48.634 ng
 RT: 15.01 min Scan# 1995
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
165	136734		
63	41.5	42.0	63.0#
89	71.3	66.9	100.3
182	1.9	2.6	3.8#









#57
 Fluorene
 Concen: 44.442 ng
 RT: 15.70 min Scan# 2112
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
166	409467		
165	94.0	80.6	121.0
167	13.5	10.4	15.6



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 48.288 ng
 RT: 15.28 min Scan# 2040
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
232	125572		
131	36.9	33.5	50.3
130	2.1	2.2	3.2#
166	27.2	24.8	37.2









#59
 Diethylphthalate
 Concen: 41.932 ng
 RT: 15.46 min Scan# 2072
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
149	414941		
177	24.9	18.5	27.7
150	13.4	10.2	15.2



#60
 4-Chlorophenyl-phenylether
 Concen: 43.077 ng
 RT: 15.69 min Scan# 2110
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
204	233269		
206	33.6	26.2	39.2
141	53.4	45.8	68.6









#61
 4-Nitroaniline
 Concen: 43.935 ng
 RT: 15.72 min Scan# 2116
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Ion Ratio	Lower	Upper
138	100			
92	63.9	40.1	80.1	
108	113.7	65.4	105.4	



#62
 Azobenzene
 Concen: 42.951 ng
 RT: 15.98 min Scan# 2160
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Ion Ratio	Lower	Upper
77	100			
182	28.6	7.4	47.4	
105	19.8	0.3	40.3	
51	26.1	11.6	51.6	









#63
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.40 min Scan# 2401
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
188	100		
94	5.7	6.9	10.3#
80	6.9	7.7	11.5#



#64
 4,6-Dinitro-2-methylphenol
 Concen: 45.788 ng
 RT: 15.78 min Scan# 2126
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
198	100		
51	35.9	30.6	70.6
105	40.6	23.8	63.8









#65
 n-Nitrosodiphenylamine
 Concen: 43.397 ng
 RT: 15.90 min Scan# 2147
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
169	100		
168	66.2	55.7	83.5
167	36.7	30.1	45.1



#66
 4-Bromophenyl-phenylether
 Concen: 45.847 ng
 RT: 16.58 min Scan# 2262
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
248	100		
250	102.4	77.1	115.7
141	72.8	50.8	76.2








#67
 Hexachlorobenzene
 Concen: 46.216 ng
 RT: 16.70 min Scan# 2283
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
284	163856		
142	27.9	26.8	40.2
249	35.6	26.3	39.5



#68
 Atrazine
 Concen: 45.778 ng
 RT: 16.85 min Scan# 2308
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
200	159203		
173	22.8	5.2	45.2
215	46.8	30.4	70.4








#69
 Pentachlorophenol
 Concen: 67.540 ng
 RT: 17.05 min Scan# 2342
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
266	145851		
268	65.8	51.1	76.7
264	62.1	49.6	74.4



#70
 Phenanthrene
 Concen: 44.376 ng
 RT: 17.44 min Scan# 2408
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
178	657088		
176	20.4	15.7	23.5
179	17.0	12.5	18.7









#71
 Anthracene
 Concen: 44.866 ng
 RT: 17.52 min Scan# 2423
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
178	100		
176	19.3	16.0	24.0
179	18.0	12.5	18.7



#72
 Carbazole
 Concen: 42.796 ng
 RT: 17.80 min Scan# 2469
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
167	100		
166	22.5	18.2	27.4
139	13.3	9.4	14.2








#73
 Di-n-butylphthalate
 Concen: 41.292 ng
 RT: 18.35 min Scan# 2563
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Ion Ratio	Lower	Upper
149	683238	100		
150		9.7	7.8	11.6
104		6.4	3.9	5.9



#74
 Fluoranthene
 Concen: 42.028 ng
 RT: 19.44 min Scan# 2749
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Ion Ratio	Lower	Upper
202	838254	100		
101		7.1	0.0	28.9
203		18.4	0.0	37.5








#75
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.66 min Scan# 3127
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
240	301698		
120	5.0	6.8	10.2#
236	26.5	20.9	31.3



#76
 Benzidine
 Concen: 41.170 ng
 RT: 19.62 min Scan# 2780
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
184	326549		
185	14.6	11.1	16.7
183	12.1	10.6	16.0








#77
 Pyrene
 Concen: 43.246 ng
 RT: 19.80 min Scan# 2811
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
202	824990		
200	21.7	16.9	25.3
203	18.3	13.9	20.9



#78
 Terphenyl-d14
 Concen: 92.168 ng
 RT: 20.00 min Scan# 2844
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
244	1261487		
212	9.3	5.8	8.8#
122	7.4	7.0	10.4








#79
 Butylbenzylphthalate
 Concen: 45.401 ng
 RT: 20.69 min Scan# 2961
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Ion Ratio	Lower	Upper
149	309723	100		
91		76.2	62.2	93.2
206		27.7	18.6	27.8



#80
 Benzo(a)anthracene
 Concen: 44.580 ng
 RT: 21.64 min Scan# 3123
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Ion Ratio	Lower	Upper
228	838141	100		
226		27.5	22.7	34.1
229		20.5	16.2	24.2








#81
 3,3'-Dichlorobenzidine
 Concen: 28.448 ng
 RT: 21.55 min Scan# 3108
 Delta R.T. -0.01 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
252	186489		
254	63.4	50.8	76.2
126	8.0	7.4	11.2



#82
 Chrysene
 Concen: 45.440 ng
 RT: 21.70 min Scan# 3134
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
228	791677		
226	31.8	24.9	37.3
229	20.5	15.0	22.4








#83
 Bis(2-ethylhexyl)phthalate
 Concen: 46.995 ng
 RT: 21.54 min Scan# 3106
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
149	100		
167	31.2	24.3	36.5
279	6.5	4.0	6.0#



#84
 Di-n-octyl phthalate
 Concen: 47.716 ng
 RT: 22.74 min Scan# 3310
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
149	100		
167	1.8	1.4	2.0
43	7.5	8.9	13.3#








#85
 Indeno(1,2,3-cd)pyrene
 Concen: 47.524 ng
 RT: 28.49 min Scan# 4290
 Delta R.T. -0.01 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
276	916694		
138	14.5	16.0	24.0#
277	25.9	20.0	30.0



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 24.86 min Scan# 3671
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21




Tgt Ion	Resp	Lower	Upper
264	299362		
260	22.5	17.4	26.0
265	20.8	17.7	26.5









#87
 Benzo(b)fluoranthene
 Concen: 45.121 ng
 RT: 23.84 min Scan# 3498
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
252	100		
253	22.1	18.2	27.4
125	5.5	7.2	10.8#



#88
 Benzo(k)fluoranthene
 Concen: 43.487 ng
 RT: 23.91 min Scan# 3510
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21





Tgt Ion	Resp	Lower	Upper
252	100		
253	21.6	17.4	26.2
125	5.2	6.6	9.8#








#89
 Benzo(a)pyrene
 Concen: 46.003 ng
 RT: 24.70 min Scan# 3645
 Delta R.T. -0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21


Tgt Ion	Resp	Lower	Upper
252	100		
253	22.0	18.3	27.5
125	6.6	7.3	10.9#



#90
 Dibenzo(a,h)anthracene
 Concen: 45.883 ng
 RT: 28.56 min Scan# 4301
 Delta R.T. -0.01 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21

Tgt Ion	Resp	Lower	Upper
278	100		
139	8.2	9.8	14.6#
279	24.2	18.4	27.6





#91
 Benzo(g,h,i)perylene
 Concen: 46.772 ng
 RT: 29.64 min Scan# 4485
 Delta R.T. 0.00 min
 Lab File: BG035944.D
 Acq: 27 Jul 2018 19:21

Tgt Ion	Resp	Lower	Upper
276	100		
277	23.5	18.3	27.5
138	11.3	13.9	20.9#

