

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG072718\  
 Data File : BG035941.D  
 Acq On : 27 Jul 2018 17:23  
 Operator : JU/SJ  
 Sample : J4192-05  
 Misc :  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 FK-BPM-03-001-C

Quant Time: Jul 28 00:10:23 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 25 11:04:14 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)    |
|-----------------------------|-------|------|----------|--------|-------|-------------|
| 1) 1,4-Dichlorobenzene-d4   | 8.02  | 152  | 30250    | 20.00  | ng    | 0.00        |
| 21) Naphthalene-d8          | 10.83 | 136  | 117142   | 20.00  | ng    | 0.00        |
| 38) Acenaphthene-d10        | 14.65 | 164  | 91174    | 20.00  | ng    | 0.00        |
| 63) Phenanthrene-d10        | 17.39 | 188  | 268460   | 20.00  | ng    | 0.00        |
| 75) Chrysene-d12            | 21.65 | 240  | 278325   | 20.00  | ng    | 0.00        |
| 86) Perylene-d12            | 24.85 | 264  | 262773   | 20.00  | ng    | 0.00        |
| System Monitoring Compounds |       |      |          |        |       |             |
| 5) 2-Fluorophenol           | 5.60  | 112  | 170277   | 93.45  | ng    | 0.00        |
| 7) Phenol-d6                | 7.19  | 99   | 252390   | 92.84  | ng    | 0.00        |
| 23) Nitrobenzene-d5         | 9.19  | 82   | 161915   | 57.74  | ng    | 0.00        |
| 41) 2,4,6-Tribromophenol    | 16.14 | 330  | 120378   | 100.17 | ng    | 0.00        |
| 44) 2-Fluorobiphenyl        | 13.27 | 172  | 417931   | 61.41  | ng    | 0.00        |
| 78) Terphenyl-d14           | 20.00 | 244  | 821267   | 65.04  | ng    | 0.00        |
| Target Compounds            |       |      |          |        |       |             |
| 10) Phenol                  | 7.22  | 94   | 7844     | 2.856  | ng    | Qvalue # 73 |
| 49) Dimethylphthalate       | 14.10 | 163  | 36853    | 4.613  | ng    | 100         |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG072718\  
Data File : BG035941.D  
Acq On : 27 Jul 2018 17:23  
Operator : JU/SJ  
Sample : J4192-05  
Misc :  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
FK-BPM-03-001-C

Quant Time: Jul 28 00:10:23 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jul 25 11:04:14 2018  
Response via : Initial Calibration

