


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG072718\
 Data File : BG035948.D
 Acq On : 27 Jul 2018 22:36
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 31 12:30:06 2018



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.3	13108	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	29.2	15150	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	36.7	18993	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	51797	PASS
199	198	5	9	6.7	3464	PASS
275	198	10	60	28.4	14709	PASS
365	198	1	100	3.5	1830	PASS
441	198	0.01	100	10.0	5167	PASS
442	198	50	100	63.9	33114	PASS
443	442	15	24	18.0	5969	PASS