



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072723\
Data File : BG058495.D
Acq On : 27 Jul 2023 9:21
Operator : MA/JU
Sample : DFTPP525
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP525

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG071723.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 26 17:43:09 2023



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.4	40541	PASS
68	69	0.00	2	1.7	577	PASS
69	198	0.00	100	30.9	34395	PASS
70	69	0.00	2	0.7	230	PASS
127	198	10	80	38.3	42643	PASS
197	198	0.00	2	0.5	583	PASS
198	198	100	100	100.0	111307	PASS
199	198	5	9	6.8	7548	PASS
275	198	10	60	28.8	32003	PASS
365	198	1	100	3.7	4147	PASS
441	198	0.01	100	14.8	16522	PASS
442	198	50	100	87.5	97421	PASS
443	442	15	24	19.6	19064	PASS

Scan 2470 (17.600 min): BG058495.D\data.ms
DFTPP525

m/z	Abundance
43.90	546.0
73.00	164.0
206.90	174.0

Instrument :

BNA_G

ClientSampleId :

DFTPP525

m/z	Abundance
38.00	690.0
39.00	2703.0
39.90	205.0
41.00	216.0
44.00	721.0
48.90	303.0
50.00	10297.0
51.00	46456.0
52.00	2378.0
55.00	299.0
56.00	1254.0
56.90	2981.0
60.90	442.0
62.00	612.0
63.00	1772.0
64.10	256.0
65.00	802.0
68.00	903.0
68.90	37728.0
70.00	235.0
73.00	400.0
74.00	3643.0
75.00	6277.0
76.00	2384.0
77.00	45872.0
78.00	3063.0
79.00	2720.0
79.90	2107.0
80.90	3099.0
82.00	799.0
82.90	725.0
84.90	587.0
86.00	1063.0
86.90	540.0
88.00	162.0
90.90	602.0
91.90	930.0
92.90	4274.0
94.00	302.0
96.00	248.0
97.90	3312.0
99.00	3182.0
100.00	263.0
101.00	1720.0
102.90	588.0
103.90	1015.0
105.00	916.0
105.90	479.0
107.00	13563.0
108.00	2105.0
110.00	25768.0
111.00	3786.0
111.90	414.0
114.90	165.0
116.00	739.0
117.00	8984.0
117.90	711.0
120.00	278.0
121.90	645.0
123.00	1238.0
124.00	582.0
125.00	371.0
127.00	44696.0
128.00	3226.0
129.00	17072.0
129.90	1378.0
130.90	249.0
134.00	391.0
135.00	1200.0
136.00	449.0
137.00	635.0
138.00	269.0
139.00	204.0
140.00	338.0
140.90	2152.0
141.90	754.0
143.00	421.0
146.00	342.0
147.00	1237.0
148.00	2168.0
148.90	688.0
149.90	317.0
151.00	278.0
153.00	645.0
154.00	469.0
155.00	1188.0
156.00	2319.0
157.00	345.0
158.00	481.0
158.90	439.0
160.00	602.0
160.90	960.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

161.90	344.0
163.90	217.0
165.00	894.0
165.90	762.0
167.00	5252.0
168.00	2299.0
168.90	602.0
170.00	237.0
171.00	172.0
171.80	342.0
172.90	753.0
174.00	1149.0
175.00	1697.0
176.00	607.0
177.00	822.0
178.00	407.0
179.00	3475.0
180.00	2616.0
181.00	1000.0
181.90	263.0
182.90	151.0
184.10	219.0
185.00	1815.0
186.00	13573.0
187.00	4233.0
188.10	627.0
189.00	912.0
190.00	194.0
190.90	303.0
192.00	1163.0
193.00	1583.0
194.10	315.0
196.00	2914.0
196.80	767.0
197.90	108080.0
198.90	7034.0
199.90	671.0
201.40	494.0
202.90	699.0
204.00	3954.0
205.00	6345.0
206.00	28592.0
207.00	3825.0
207.90	850.0
208.90	280.0
210.30	533.0
211.00	1130.0
212.00	290.0
214.90	342.0
216.10	441.0
216.90	7016.0
218.00	1076.0
219.90	172.0
221.00	4333.0
223.00	1834.0
224.00	15802.0
225.00	3973.0
225.90	475.0
227.00	6685.0
228.00	1151.0
228.90	1510.0
230.00	162.0
231.00	613.0
233.90	439.0
235.10	414.0
235.80	241.0
236.00	241.0
237.00	786.0
237.90	180.0
238.80	192.0
240.80	419.0
242.00	802.0
243.00	1033.0
244.00	12717.0
245.00	1428.0
246.00	2264.0
246.90	481.0
249.00	371.0
252.90	368.0
253.90	329.0
255.00	65240.0
256.00	9698.0
256.90	728.0
258.00	4064.0
258.90	518.0
259.90	182.0
263.90	210.0
265.00	1233.0
266.00	184.0
272.00	208.0
272.90	2101.0
274.00	4809.0
275.00	28104.0
276.00	3944.0
277.00	2421.0
277.90	397.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

282.80	227.0
284.00	170.0
285.00	543.0
292.90	456.0
293.90	223.0
296.00	7492.0
296.90	1084.0
302.00	155.0
302.90	955.0
303.90	194.0
314.00	436.0
315.00	1030.0
315.90	579.0
321.00	282.0
321.90	169.0
323.00	2680.0
324.00	370.0
327.00	430.0
327.90	281.0
331.90	177.0
332.90	366.0
334.00	1646.0
335.10	380.0
341.10	294.0
346.00	530.0
352.00	683.0
353.00	709.0
354.00	745.0
355.10	171.0
359.10	209.0
365.00	3244.0
365.90	556.0
370.90	175.0
372.00	1302.0
373.10	394.0
383.00	382.0
383.90	191.0
390.00	197.0
400.80	158.0
402.00	542.0
402.90	1162.0
404.00	298.0
421.10	888.0
421.90	549.0
423.00	4230.0
424.00	1090.0
441.00	12535.0
442.00	74512.0
443.00	15072.0
444.00	1415.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

m/z	Abundance
37.00	209.0
38.00	559.0
39.00	2942.0
39.90	187.0
40.90	156.0
43.90	810.0
49.00	402.0
50.00	10846.0
51.00	45920.0
52.00	2323.0
54.80	170.0
55.90	1539.0
56.90	3048.0
61.00	673.0
62.00	683.0
63.00	2036.0
63.90	313.0
65.00	899.0
68.00	829.0
68.90	40136.0
69.90	248.0
72.90	438.0
74.00	4038.0
75.00	6421.0
76.00	2195.0
77.00	49664.0
78.00	3299.0
79.00	2426.0
79.90	2144.0
81.00	3292.0
81.90	790.0
82.90	949.0
85.00	863.0
85.90	1101.0
87.00	343.0
88.00	155.0
91.00	698.0
92.00	846.0
93.00	4648.0
94.00	341.0
96.00	250.0
98.00	3670.0
99.00	2971.0
100.00	216.0
101.00	2215.0
102.00	182.0
102.90	478.0
103.90	1192.0
104.90	1292.0
106.00	304.0
107.00	14717.0
108.00	2230.0
109.90	28008.0
111.00	3572.0
112.00	505.0
112.80	176.0
115.00	169.0
115.90	838.0
117.00	10276.0
117.90	932.0
121.90	805.0
123.00	1225.0
124.00	472.0
124.90	656.0
127.00	49344.0
128.00	3635.0
129.00	18176.0
130.00	1375.0
131.00	462.0
131.90	169.0
133.90	549.0
134.90	1077.0
136.00	584.0
137.00	818.0
137.90	171.0
139.00	220.0
140.00	393.0
140.90	2169.0
141.90	890.0
143.00	488.0
145.90	474.0
147.00	1191.0
148.00	2662.0
148.90	591.0
150.00	307.0
150.90	363.0
151.90	297.0
153.00	627.0
153.90	545.0
155.00	1587.0
156.00	2068.0
157.00	640.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

158.00	629.0
158.90	408.0
159.90	810.0
161.00	1459.0
161.90	341.0
164.10	196.0
164.90	1194.0
166.00	832.0
167.00	6436.0
168.00	3112.0
169.00	421.0
169.90	256.0
171.10	169.0
172.00	384.0
173.00	699.0
174.00	1127.0
175.00	2340.0
176.00	466.0
176.90	1386.0
178.00	413.0
178.90	4069.0
180.00	2813.0
181.00	1515.0
182.00	285.0
184.10	297.0
185.00	2281.0
186.00	16688.0
187.00	4677.0
188.00	503.0
188.90	966.0
189.90	235.0
191.00	471.0
192.00	1297.0
193.00	1736.0
193.90	472.0
195.00	305.0
196.00	3823.0
196.70	983.0
197.90	130216.0
198.90	8760.0
200.00	717.0
201.50	794.0
202.90	924.0
204.00	4795.0
205.00	7735.0
206.00	34936.0
207.00	4786.0
208.00	1235.0
209.00	537.0
209.90	416.0
211.00	1252.0
211.80	337.0
214.90	400.0
216.00	589.0
216.90	8813.0
218.00	1318.0
221.00	5440.0
221.80	1119.0
222.90	2484.0
224.00	18712.0
225.00	5134.0
226.00	441.0
227.00	8263.0
228.00	1082.0
228.90	1711.0
229.90	279.0
231.10	850.0
233.90	595.0
234.90	503.0
236.00	352.0
236.90	610.0
239.00	332.0
239.80	219.0
240.90	578.0
242.00	1225.0
243.00	1230.0
244.00	15863.0
245.00	2300.0
246.00	3153.0
246.90	648.0
247.80	226.0
248.90	535.0
250.10	292.0
250.80	156.0
252.10	175.0
253.00	542.0
255.00	81984.0
256.00	12577.0
256.90	869.0
258.00	4041.0
259.00	585.0
265.00	2301.0
266.00	277.0
271.00	177.0
273.00	2435.0
274.00	6706.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

275.00	38088.0
276.00	4909.0
277.00	2726.0
278.00	542.0
282.90	428.0
284.00	278.0
284.90	720.0
289.10	198.0
289.80	180.0
292.90	766.0
293.90	327.0
294.90	171.0
296.00	10850.0
297.00	1889.0
300.70	183.0
302.10	270.0
303.00	1312.0
304.00	266.0
307.90	217.0
312.80	186.0
313.90	569.0
314.90	1193.0
316.00	680.0
320.90	480.0
322.00	223.0
323.00	3156.0
324.10	494.0
327.00	570.0
328.00	285.0
332.10	316.0
332.90	228.0
334.00	2529.0
335.10	621.0
340.90	514.0
345.90	670.0
346.90	232.0
352.00	997.0
353.00	1015.0
354.00	1388.0
355.00	301.0
365.00	5053.0
365.90	680.0
371.00	308.0
372.00	2180.0
373.10	505.0
382.90	569.0
383.90	246.0
390.10	303.0
391.10	221.0
392.00	162.0
402.00	833.0
403.00	1247.0
404.10	450.0
421.00	1068.0
422.00	1067.0
423.00	6569.0
424.00	1364.0
424.90	180.0
441.00	18800.0
442.00	114072.0
443.00	21656.0
444.00	2075.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

m/z	Abundance
38.00	201.0
39.00	2163.0
39.90	181.0
43.90	611.0
49.00	349.0
50.00	6960.0
51.00	29248.0
52.00	1604.0
55.00	158.0
56.00	764.0
57.00	2296.0
61.00	458.0
62.00	385.0
63.00	1205.0
64.00	260.0
65.00	580.0
68.90	25320.0
69.90	208.0
73.00	301.0
74.00	2514.0
75.00	4006.0
76.00	1568.0
77.00	32208.0
78.00	2117.0
78.90	1987.0
79.90	1260.0
80.90	2257.0
81.90	625.0
83.00	482.0
84.90	703.0
86.00	727.0
86.90	402.0
90.90	476.0
91.90	559.0
93.00	2955.0
94.10	253.0
98.00	2619.0
98.90	2260.0
99.80	159.0
101.00	1357.0
102.90	335.0
103.90	703.0
105.00	644.0
106.00	321.0
107.00	9837.0
108.00	1711.0
110.00	18568.0
111.00	3029.0
111.90	225.0
116.00	401.0
117.00	6560.0
117.90	589.0
122.00	480.0
123.00	995.0
124.00	502.0
124.90	413.0
127.00	33888.0
128.00	2892.0
129.00	13158.0
129.90	1364.0
130.90	253.0
133.90	314.0
134.90	1006.0
135.90	304.0
137.00	430.0
137.90	205.0
138.80	176.0
140.00	182.0
141.00	1377.0
141.90	507.0
142.80	281.0
144.80	180.0
145.90	242.0
147.00	800.0
148.00	1826.0
148.90	331.0
151.00	351.0
153.00	560.0
153.90	501.0
155.00	973.0
156.00	1632.0
157.00	232.0
158.00	307.0
159.00	221.0
159.90	526.0
161.00	822.0
161.90	324.0
165.00	623.0
166.00	723.0
167.00	3945.0
168.00	2058.0
169.00	350.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

169.90	180.0
170.80	180.0
171.90	387.0
173.00	401.0
174.00	818.0
175.00	1679.0
176.00	352.0
176.90	942.0
178.00	266.0
178.90	2942.0
180.00	2013.0
181.00	1150.0
182.00	212.0
183.90	230.0
185.00	1388.0
186.00	11773.0
187.00	3342.0
188.00	320.0
189.00	681.0
190.00	227.0
190.90	306.0
191.90	877.0
192.90	1059.0
194.00	317.0
195.00	156.0
196.00	2672.0
197.90	95624.0
198.90	6849.0
200.00	414.0
201.40	410.0
203.00	563.0
204.00	3588.0
205.00	5314.0
206.00	25176.0
207.00	3629.0
208.00	674.0
208.90	259.0
210.90	951.0
212.10	154.0
215.00	175.0
215.90	512.0
216.90	6086.0
218.00	974.0
221.00	4109.0
221.90	911.0
222.90	1367.0
224.00	14339.0
225.00	3119.0
226.00	442.0
227.00	5752.0
228.00	846.0
229.00	1240.0
229.80	203.0
231.00	523.0
234.00	324.0
234.90	432.0
236.00	245.0
237.00	510.0
238.90	240.0
240.90	444.0
242.00	802.0
243.00	967.0
244.00	12155.0
245.00	1801.0
246.00	2411.0
247.00	422.0
247.90	223.0
249.00	371.0
250.90	238.0
252.80	318.0
253.90	380.0
255.00	62728.0
256.00	8737.0
257.00	622.0
257.90	3538.0
259.00	688.0
265.00	1422.0
265.80	198.0
271.00	195.0
272.00	219.0
273.00	1978.0
274.00	5345.0
275.00	29816.0
276.00	3990.0
277.00	2254.0
277.90	193.0
282.10	155.0
283.00	294.0
283.90	190.0
284.90	387.0
291.90	208.0
293.00	781.0
294.00	172.0
296.00	8307.0
297.00	1233.0
303.00	1028.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525

304.00	303.0
308.90	152.0
314.00	341.0
314.90	1126.0
315.90	605.0
317.00	156.0
321.00	356.0
323.00	3092.0
324.00	466.0
326.90	485.0
328.00	264.0
332.00	213.0
333.00	305.0
334.00	1705.0
335.00	481.0
341.00	374.0
346.00	512.0
347.00	209.0
352.00	1049.0
352.90	585.0
354.00	1048.0
355.00	186.0
365.00	4145.0
365.90	532.0
370.80	210.0
371.10	212.0
372.00	1612.0
373.00	389.0
382.90	463.0
390.90	171.0
401.90	795.0
403.00	1226.0
404.00	496.0
405.00	170.0
420.90	1049.0
422.00	823.0
423.00	5944.0
424.00	1222.0
424.90	197.0
441.00	18232.0
442.00	103680.0
443.00	20464.0
444.00	1830.0
445.10	151.0

Instrument :
BNA_G
ClientSampleId :
DFTPP525