


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072820\
Data File : BG046093.D
Acq On : 28 Jul 2020 8:01
Operator : CG/JU
Sample : DFTPP28
Misc : MDL-ME-05
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP28

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG072320MA.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 28 10:34:48 2020



AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.9	57445	PASS
68	69	0.00	2	1.3	720	PASS
69	198	0.00	100	32.2	54698	PASS
70	69	0.00	2	0.6	343	PASS
127	198	10	80	41.3	70021	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	169621	PASS
199	198	5	9	7.0	11853	PASS
275	198	10	60	29.4	49861	PASS
365	198	1	100	4.3	7330	PASS
441	198	0.01	100	13.9	23571	PASS
442	198	50	100	84.6	143552	PASS
443	442	15	24	19.2	27560	PASS

Scan 2494 (17.743 min): BG046093.D
DFTPP28

m/z	Abundance
40.00	453.0
44.00	776.0
96.20	188.0
96.40	183.0
191.00	155.0
192.90	157.0
207.00	1332.0
208.00	423.0
209.00	194.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

m/z	Abundance
37.10	198.0
38.10	644.0
39.10	4097.0
40.00	710.0
44.00	944.0
49.00	419.0
50.10	13669.0
51.10	57808.0
52.10	3060.0
55.10	252.0
56.00	1725.0
57.00	3881.0
61.10	688.0
62.00	983.0
63.10	2086.0
64.10	316.0
65.10	1131.0
66.00	233.0
68.00	1000.0
69.00	52200.0
70.00	328.0
73.00	652.0
74.10	4857.0
75.00	9177.0
76.10	3099.0
77.10	60064.0
78.10	3856.0
79.10	3647.0
80.00	2697.0
81.00	4004.0
82.00	822.0
83.00	1018.0
84.10	225.0
85.10	711.0
86.00	1011.0
87.10	520.0
88.20	298.0
91.00	1134.0
92.00	1074.0
93.00	7172.0
94.00	500.0
96.00	329.0
97.20	219.0
98.00	4835.0
99.00	4793.0
100.00	488.0
101.00	2428.0
101.90	157.0
103.00	765.0
104.00	1302.0
105.10	1329.0
106.10	417.0
107.00	18352.0
108.00	2869.0
110.00	33184.0
111.00	6089.0
111.90	899.0
113.10	174.0
115.00	177.0
116.00	1296.0
117.00	14827.0
118.00	1074.0
118.90	191.0
120.00	269.0
122.00	1080.0
123.10	2112.0
124.10	838.0
125.00	948.0
127.00	63864.0
128.00	4748.0
129.00	25424.0
130.00	2177.0
131.00	335.0
132.10	168.0
133.00	223.0
134.10	798.0
135.00	2314.0
136.00	701.0
137.10	1443.0
138.00	223.0
140.00	566.0
141.00	3153.0
142.00	1238.0
143.00	581.0
143.90	245.0
146.00	680.0
147.00	1978.0
148.00	4247.0
149.10	1037.0
149.90	284.0
151.00	354.0
152.20	158.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

153.00	1203.0
154.00	900.0
155.00	2103.0
156.00	3375.0
157.10	762.0
157.90	470.0
159.10	525.0
160.00	1111.0
161.00	1496.0
162.00	423.0
162.90	172.0
163.90	260.0
165.00	1475.0
166.00	889.0
167.10	7074.0
168.00	4500.0
169.00	875.0
170.00	253.0
171.00	319.0
172.00	659.0
173.00	807.0
173.90	1514.0
175.00	2679.0
176.00	1001.0
177.00	1558.0
178.00	639.0
179.00	5213.0
180.00	3465.0
181.00	1944.0
181.90	313.0
183.00	206.0
184.00	474.0
185.10	3001.0
186.00	20088.0
187.00	5008.0
188.10	738.0
189.00	1282.0
189.90	236.0
191.00	849.0
192.00	1677.0
193.00	2055.0
194.00	664.0
195.00	368.0
196.00	4357.0
198.00	142976.0
199.00	10341.0
200.00	949.0
201.40	748.0
203.00	1151.0
204.00	5637.0
205.00	9598.0
206.00	41752.0
207.00	6278.0
208.00	1737.0
209.00	807.0
211.00	1627.0
211.90	268.0
214.90	533.0
215.90	979.0
217.00	10881.0
218.00	1363.0
218.90	156.0
219.90	213.0
221.00	6129.0
222.10	1504.0
223.00	2643.0
224.00	21488.0
225.00	5181.0
226.10	930.0
227.00	8610.0
228.00	1674.0
229.00	2041.0
229.90	312.0
231.00	805.0
232.10	161.0
234.00	583.0
235.00	787.0
236.00	474.0
236.90	903.0
238.10	226.0
239.00	318.0
240.10	351.0
241.00	610.0
242.10	1156.0
243.00	1573.0
244.10	16416.0
245.10	2671.0
246.00	3715.0
247.00	709.0
248.00	201.0
248.90	556.0
250.00	255.0
251.00	238.0
251.90	169.0
253.10	546.0
255.00	88752.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

256.00	13373.0
257.00	1201.0
258.00	5314.0
259.10	759.0
265.00	2223.0
266.00	309.0
271.00	283.0
271.90	175.0
273.00	2573.0
274.00	6975.0
275.00	39272.0
276.00	4964.0
277.00	3215.0
278.00	548.0
281.10	199.0
283.10	469.0
284.10	343.0
285.10	475.0
289.00	286.0
289.90	200.0
293.10	858.0
293.70	162.0
296.00	11329.0
297.00	1731.0
297.90	163.0
300.90	188.0
302.10	288.0
303.10	1425.0
304.00	409.0
307.90	179.0
308.90	187.0
313.00	164.0
314.10	556.0
315.00	1283.0
316.00	607.0
321.00	394.0
323.10	3670.0
324.00	802.0
327.00	726.0
327.90	354.0
332.20	306.0
333.00	352.0
334.00	2526.0
335.00	673.0
341.00	500.0
346.00	710.0
352.00	1290.0
353.10	702.0
354.00	1097.0
355.10	178.0
365.00	5110.0
366.00	845.0
370.90	313.0
372.00	2133.0
373.00	547.0
383.00	431.0
383.90	188.0
402.00	593.0
403.10	1166.0
404.00	548.0
421.00	969.0
421.90	719.0
423.00	5596.0
424.10	1635.0
424.90	163.0
441.00	14193.0
442.00	92288.0
443.00	18208.0
444.10	1798.0
445.10	151.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

m/z	Abundance
37.00	287.0
38.00	860.0
39.10	4749.0
40.00	620.0
41.00	423.0
44.10	789.0
50.10	17584.0
51.10	68288.0
52.10	3447.0
53.20	289.0
55.10	309.0
56.00	2149.0
57.10	4189.0
57.90	205.0
61.00	804.0
62.00	1162.0
63.10	2717.0
64.00	309.0
65.00	1125.0
69.00	64848.0
70.00	451.0
73.20	769.0
74.00	6062.0
75.00	10576.0
76.10	3616.0
77.10	75656.0
78.10	4803.0
79.00	4483.0
80.00	4185.0
81.00	5159.0
82.10	1180.0
83.00	1212.0
84.00	268.0
85.10	705.0
86.10	1940.0
87.00	842.0
87.70	199.0
89.00	158.0
91.00	1146.0
92.00	1557.0
93.00	7675.0
94.10	453.0
95.00	197.0
95.90	391.0
97.00	171.0
98.00	6481.0
99.00	5913.0
100.00	573.0
101.00	2717.0
103.00	1262.0
104.00	2050.0
105.00	1795.0
106.10	774.0
107.00	22400.0
108.00	3774.0
110.00	41904.0
111.00	6659.0
112.00	743.0
113.10	267.0
116.00	1383.0
117.00	18304.0
118.00	1404.0
119.00	167.0
120.10	479.0
122.00	1738.0
123.00	2465.0
124.00	1273.0
125.10	1191.0
127.00	82832.0
128.10	6247.0
129.00	31616.0
130.00	2844.0
131.10	474.0
132.00	310.0
133.00	218.0
134.00	983.0
135.00	2681.0
136.00	741.0
137.00	1668.0
137.70	226.0
139.90	469.0
141.00	4436.0
142.00	1533.0
143.00	944.0
144.00	262.0
144.90	251.0
146.10	946.0
147.00	2550.0
148.00	5045.0
149.10	1162.0
150.10	263.0
151.00	516.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

151.70	330.0
153.00	1423.0
154.00	1349.0
155.10	2479.0
156.00	3570.0
157.10	741.0
157.90	919.0
159.00	545.0
160.00	1467.0
161.00	2275.0
162.10	621.0
162.90	264.0
163.80	315.0
165.00	1790.0
166.00	1490.0
167.00	9113.0
168.00	5270.0
169.10	1026.0
170.10	297.0
171.00	433.0
172.00	918.0
173.10	1128.0
174.00	1645.0
175.00	4048.0
176.10	989.0
176.90	1579.0
178.10	787.0
179.00	8316.0
180.00	5490.0
181.10	2237.0
182.10	448.0
183.00	169.0
184.10	743.0
185.00	3728.0
186.10	28568.0
187.00	8179.0
188.00	676.0
189.00	1775.0
190.00	339.0
191.00	930.0
192.00	2086.0
193.00	3261.0
194.00	572.0
195.10	384.0
196.10	5597.0
198.00	202048.0
199.00	13921.0
200.00	1117.0
201.40	890.0
203.00	1289.0
204.00	7950.0
205.10	13234.0
206.10	55064.0
207.00	8767.0
208.00	2185.0
209.00	668.0
210.10	957.0
211.00	1809.0
212.10	331.0
215.00	559.0
216.00	1072.0
217.00	14891.0
218.10	1597.0
219.10	185.0
221.00	8608.0
222.00	1758.0
223.00	3762.0
224.00	30096.0
225.00	8689.0
226.10	794.0
227.00	13569.0
227.90	1894.0
229.00	2687.0
230.00	439.0
231.00	1349.0
232.00	216.0
232.90	205.0
234.00	934.0
235.00	891.0
236.00	682.0
237.10	1218.0
238.00	274.0
239.00	545.0
240.00	551.0
241.00	737.0
242.00	1450.0
243.00	2150.0
244.00	25000.0
245.10	3484.0
246.00	4838.0
247.00	1000.0
249.00	783.0
250.10	167.0
251.00	285.0
252.00	299.0
252.90	666.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

255.00	135424.0
256.00	18952.0
257.00	1467.0
258.10	7644.0
259.10	1831.0
261.00	163.0
262.80	175.0
264.10	195.0
265.00	3079.0
266.00	404.0
270.00	154.0
271.00	319.0
272.10	417.0
273.00	3745.0
274.00	9758.0
275.00	58328.0
276.00	8453.0
277.00	5230.0
277.90	795.0
279.00	176.0
280.90	213.0
283.10	738.0
283.90	400.0
285.00	1031.0
285.90	182.0
289.00	216.0
289.90	214.0
291.90	319.0
293.00	1315.0
294.30	271.0
294.80	266.0
296.00	18232.0
297.00	2747.0
301.10	225.0
302.10	379.0
303.10	2100.0
304.10	578.0
304.90	162.0
308.00	384.0
310.00	389.0
313.00	169.0
314.10	784.0
315.00	1907.0
316.00	1173.0
317.00	206.0
321.10	561.0
322.00	252.0
323.10	5704.0
324.00	985.0
327.00	912.0
328.00	604.0
332.00	390.0
333.00	433.0
334.00	3027.0
335.10	792.0
336.00	182.0
341.10	510.0
342.00	277.0
346.00	1248.0
347.00	348.0
352.00	1737.0
353.00	1298.0
354.00	1746.0
355.00	554.0
365.00	8896.0
366.10	1198.0
370.00	155.0
370.90	386.0
372.00	3150.0
373.00	1048.0
383.10	927.0
384.00	282.0
390.10	603.0
391.10	432.0
391.80	162.0
401.00	190.0
402.00	1177.0
403.00	1823.0
404.00	725.0
421.10	1513.0
421.90	1588.0
423.10	10479.0
424.00	2420.0
425.00	361.0
441.00	27344.0
442.00	168640.0
443.00	32872.0
444.00	3274.0
445.20	164.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

m/z	Abundance
38.00	202.0
39.10	1457.0
40.00	472.0
44.00	726.0
50.10	4570.0
51.10	19176.0
52.00	1073.0
56.00	494.0
57.00	1251.0
61.00	367.0
62.10	366.0
63.00	763.0
64.10	160.0
65.10	362.0
67.90	175.0
69.00	18496.0
73.00	280.0
74.00	1713.0
75.10	3354.0
76.10	1062.0
77.10	23240.0
78.10	1616.0
79.10	1649.0
80.10	1164.0
81.00	2036.0
82.10	460.0
83.00	440.0
84.90	171.0
86.00	604.0
87.20	182.0
88.10	194.0
91.10	326.0
92.00	563.0
93.00	2623.0
94.00	169.0
96.10	301.0
98.00	1961.0
99.00	1375.0
101.00	1035.0
102.00	191.0
103.00	255.0
104.00	539.0
104.90	529.0
107.00	7107.0
108.00	1244.0
110.00	13296.0
111.00	2064.0
115.90	395.0
117.00	6568.0
118.10	482.0
122.10	391.0
123.00	842.0
124.00	564.0
125.00	285.0
127.00	28072.0
128.00	1999.0
129.00	10916.0
130.00	1002.0
130.90	275.0
133.90	217.0
135.00	1012.0
136.00	446.0
137.00	393.0
139.90	172.0
141.00	1863.0
142.10	623.0
143.10	315.0
145.90	348.0
147.00	884.0
148.00	2046.0
149.10	526.0
150.20	176.0
151.00	266.0
151.40	206.0
153.00	558.0
154.00	463.0
155.10	881.0
156.00	1612.0
157.00	332.0
158.00	395.0
159.10	206.0
160.00	785.0
160.90	820.0
162.10	363.0
165.00	628.0
166.00	484.0
167.00	3623.0
168.00	1983.0
169.10	412.0
171.10	211.0
171.90	313.0
173.10	471.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

174.00	640.0
175.10	1234.0
176.10	447.0
177.00	492.0
178.10	241.0
179.00	2964.0
180.00	1934.0
181.00	976.0
184.00	255.0
185.00	1763.0
186.00	9613.0
187.00	2571.0
188.00	235.0
189.10	874.0
191.00	290.0
191.90	1049.0
193.10	895.0
194.10	295.0
196.00	2163.0
198.00	81432.0
199.00	4960.0
199.90	462.0
201.50	702.0
203.10	638.0
204.00	2978.0
205.10	4945.0
206.00	21808.0
207.10	4174.0
208.00	1247.0
209.10	579.0
210.00	376.0
210.90	706.0
212.10	202.0
215.20	355.0
216.00	416.0
217.00	5747.0
218.10	681.0
219.90	201.0
221.10	3187.0
222.10	607.0
223.00	1258.0
224.00	12768.0
225.10	3163.0
226.20	365.0
227.00	5945.0
228.00	930.0
229.00	1052.0
230.10	203.0
231.00	655.0
233.00	276.0
234.00	408.0
235.10	446.0
236.10	287.0
237.00	562.0
238.00	180.0
239.10	277.0
239.90	251.0
241.00	235.0
242.00	960.0
243.10	693.0
244.10	10278.0
245.10	1252.0
246.00	1895.0
246.90	343.0
248.90	490.0
251.00	199.0
252.90	232.0
255.00	58848.0
256.00	8464.0
257.00	847.0
258.00	3524.0
259.00	514.0
263.80	180.0
265.00	1529.0
271.00	178.0
272.00	237.0
273.00	1707.0
274.00	5243.0
275.00	27704.0
276.00	3314.0
277.00	2323.0
277.90	402.0
281.10	331.0
282.10	193.0
283.00	362.0
284.20	160.0
285.00	396.0
293.00	510.0
294.10	171.0
296.00	8412.0
296.90	968.0
303.10	1012.0
303.90	227.0
314.10	430.0
315.00	944.0
316.00	704.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

321.00	357.0
323.10	3229.0
324.00	517.0
327.00	543.0
327.90	439.0
331.90	239.0
334.00	1940.0
335.10	382.0
341.10	324.0
346.00	700.0
352.00	929.0
353.00	763.0
354.00	915.0
354.90	229.0
355.30	188.0
359.00	160.0
365.00	4679.0
365.90	675.0
371.20	233.0
372.00	1467.0
373.00	342.0
383.00	592.0
390.10	269.0
391.80	155.0
402.10	665.0
403.10	1369.0
404.00	458.0
421.00	1063.0
422.10	754.0
423.00	5789.0
424.00	1241.0
441.00	15821.0
442.00	104368.0
443.00	20240.0
444.00	1869.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28