

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072823\
 Data File : BG058518.D
 Acq On : 28 Jul 2023 14:51
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG072823

Quant Time: Jul 28 22:35:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	0.00
2	1,4-Dioxane	40.000	39.604	1.0	90	0.00
3	Pyridine	40.000	40.839	-2.1	91	0.00
4	n-Nitrosodimethylamine	40.000	39.826	0.4	89	0.00
5 S	2-Fluorophenol	80.000	80.813	-1.0	90	0.00
6	Aniline	40.000	41.942	-4.9	91	0.00
7 S	Phenol-d6	80.000	82.980	-3.7	92	0.00
8	2-Chlorophenol	40.000	40.923	-2.3	92	0.00
9	Benzaldehyde	40.000	39.560	1.1	93	0.00
10 C	Phenol	40.000	41.237	-3.1	91	0.00
11	bis(2-Chloroethyl)ether	40.000	40.594	-1.5	92	0.00
12	1,3-Dichlorobenzene	40.000	40.004	-0.0	90	0.00
13 C	1,4-Dichlorobenzene	40.000	40.652	-1.6	92	0.00
14	1,2-Dichlorobenzene	40.000	40.724	-1.8	92	0.00
15	Benzyl Alcohol	40.000	43.600	-9.0	92	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.597	-1.5	92	0.00
17	2-Methylphenol	40.000	41.392	-3.5	92	0.00
18	Hexachloroethane	40.000	40.711	-1.8	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.521	-6.3	95	0.00
20	3+4-Methylphenols	40.000	42.621	-6.6	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	41.116	-2.8	94	0.00
23 S	Nitrobenzene-d5	80.000	82.321	-2.9	94	0.00
24	Nitrobenzene	40.000	40.699	-1.7	93	0.00
25	Isophorone	40.000	40.845	-2.1	95	0.00
26 C	2-Nitrophenol	40.000	41.335	-3.3	92	0.00
27	2,4-Dimethylphenol	40.000	41.759	-4.4	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.199	-3.0	93	0.00
29 C	2,4-Dichlorophenol	40.000	41.072	-2.7	93	0.00
30	1,2,4-Trichlorobenzene	40.000	40.023	-0.1	93	0.00
31	Naphthalene	40.000	40.188	-0.5	94	0.00
32	Benzoic acid	40.000	41.207	-3.0	92	0.00
33	4-Chloroaniline	40.000	42.665	-6.7	94	0.00
34 C	Hexachlorobutadiene	40.000	39.950	0.1	93	0.00
35	Caprolactam	40.000	43.070	-7.7	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.127	-5.3	96	0.00
37	2-Methylnaphthalene	40.000	40.513	-1.3	94	0.00
38	1-Methylnaphthalene	40.000	40.616	-1.5	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.557	1.1	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.726	3.2	96	0.00
42 S	2,4,6-Tribromophenol	80.000	85.119	-6.4	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.884	0.3	93	0.00
44	2,4,5-Trichlorophenol	40.000	40.946	-2.4	98	0.00
45 S	2-Fluorobiphenyl	80.000	78.572	1.8	96	0.00
46	1,1'-Biphenyl	40.000	39.637	0.9	95	0.00
47	2-Chloronaphthalene	40.000	39.427	1.4	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072823\
 Data File : BG058518.D
 Acq On : 28 Jul 2023 14:51
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG072823

Quant Time: Jul 28 22:35:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.985	-5.0	96	0.00
49	Acenaphthylene	40.000	39.693	0.8	95	0.00
50	Dimethylphthalate	40.000	40.429	-1.1	97	0.00
51	2,6-Dinitrotoluene	40.000	41.433	-3.6	98	0.00
52 C	Acenaphthene	40.000	40.041	-0.1	96	0.00
53	3-Nitroaniline	40.000	43.448	-8.6	99	0.00
54 P	2,4-Dinitrophenol	40.000	41.862	-4.7	103	0.00
55	Dibenzofuran	40.000	39.837	0.4	96	0.00
56 P	4-Nitrophenol	40.000	42.888	-7.2	98	0.00
57	2,4-Dinitrotoluene	40.000	42.691	-6.7	99	0.00
58	Fluorene	40.000	40.273	-0.7	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.977	-4.9	99	0.00
60	Diethylphthalate	40.000	40.529	-1.3	97	0.00
61	4-Chlorophenyl-phenylether	40.000	40.552	-1.4	98	0.00
62	4-Nitroaniline	40.000	45.669	-14.2	101	0.00
63	Azobenzene	40.000	40.491	-1.2	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.707	-9.3	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.580	1.1	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.508	1.2	97	0.00
68	Hexachlorobenzene	40.000	39.931	0.2	100	0.00
69	Atrazine	40.000	40.435	-1.1	99	0.00
70 C	Pentachlorophenol	40.000	43.644	-9.1	101	0.00
71	Phenanthrene	40.000	39.862	0.3	99	0.00
72	Anthracene	40.000	39.903	0.2	98	0.00
73	Carbazole	40.000	41.339	-3.3	99	0.00
74	Di-n-butylphthalate	40.000	41.202	-3.0	99	0.00
75 C	Fluoranthene	40.000	40.912	-2.3	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzidine	40.000	49.035	-22.6	101	0.00
78	Pyrene	40.000	39.478	1.3	99	0.00
79 S	Terphenyl-d14	80.000	79.056	1.2	101	0.00
80	Butylbenzylphthalate	40.000	40.857	-2.1	101	0.00
81	Benzo(a)anthracene	40.000	39.921	0.2	100	0.00
82	3,3'-Dichlorobenzidine	40.000	40.578	-1.4	99	0.00
83	Chrysene	40.000	39.934	0.2	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.523	-1.3	100	0.00
85 c	Di-n-octyl phthalate	40.000	40.353	-0.9	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.586	-1.5	99	0.00
88	Benzo(b)fluoranthene	40.000	40.846	-2.1	100	0.00
89	Benzo(k)fluoranthene	40.000	40.002	-0.0	99	0.00
90 C	Benzo(a)pyrene	40.000	40.494	-1.2	100	0.00
91	Dibenzo(a,h)anthracene	40.000	41.008	-2.5	100	0.00
92	Benzo(g,h,i)perylene	40.000	41.204	-3.0	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072823\
Data File : BG058518.D
Acq On : 28 Jul 2023 14:51
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG072823

Quant Time: Jul 28 22:35:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0