



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG073021\
 Data File : BG049576.D
 Acq On : 30 Jul 2021 13:24
 Operator : CG/JU
 Sample : DFTPP408
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP408

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG072421.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jul 26 11:36:47 2021



AutoFind: Scans 2526, 2527, 2528; Background Corrected with Scan 2521

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.0	36288	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	45.5	40280	PASS
70	69	0.00	2	0.6	243	PASS
127	198	10	80	52.8	46731	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	88496	PASS
199	198	5	9	7.0	6225	PASS
275	198	10	60	32.4	28691	PASS
365	198	1	100	5.4	4803	PASS
441	198	0.01	100	14.6	12897	PASS
442	198	50	100	89.3	79056	PASS
443	442	15	24	20.0	15808	PASS

m/z	Abundance
44.05	225.0
49.95	160.0
62.95	196.0
73.15	209.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

m/z	Abundance
37.05	172.0
38.25	482.0
39.15	2506.0
50.15	9371.0
51.15	33688.0
52.25	1835.0
52.95	161.0
55.05	240.0
56.05	1280.0
57.15	2985.0
58.15	256.0
61.15	525.0
62.15	371.0
63.15	1610.0
64.15	500.0
65.15	1221.0
69.15	38536.0
70.15	180.0
73.05	384.0
74.15	4288.0
75.15	6931.0
76.15	2162.0
77.15	37800.0
78.15	2311.0
79.15	2726.0
80.05	2032.0
81.15	3211.0
82.05	913.0
83.15	362.0
85.15	576.0
86.15	1170.0
87.05	419.0
91.05	1046.0
92.05	1093.0
93.15	6221.0
94.15	426.0
96.15	232.0
98.15	4048.0
99.15	2990.0
100.25	262.0
101.15	1762.0
103.15	645.0
104.15	1308.0
105.05	947.0
107.15	13425.0
108.15	2563.0
110.15	19784.0
111.05	3937.0
112.05	295.0
113.05	156.0
116.15	568.0
117.15	11373.0
118.05	822.0
120.15	186.0
122.05	474.0
123.15	1567.0
124.05	664.0
125.15	560.0
127.15	40888.0
128.15	2347.0
129.15	16896.0
130.15	1889.0
131.15	319.0
132.15	190.0
132.85	168.0
134.05	621.0
135.05	1634.0
136.15	523.0
137.15	775.0
140.25	355.0
141.05	2612.0
142.05	794.0
143.15	772.0
144.15	192.0
146.15	708.0
147.15	1470.0
148.05	3124.0
149.15	666.0
150.05	176.0
151.15	265.0
153.15	693.0
154.15	477.0
155.15	1710.0
156.15	1996.0
157.15	351.0
158.05	551.0
159.15	501.0
160.05	1100.0
161.05	1260.0
161.85	170.0
162.25	269.0
165.05	1235.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

166.05	655.0
167.15	4051.0
168.15	2009.0
169.15	242.0
171.15	156.0
172.15	419.0
173.15	327.0
174.05	995.0
175.15	1747.0
176.25	412.0
177.05	925.0
178.25	384.0
179.05	3401.0
180.15	2388.0
181.15	1276.0
182.05	186.0
183.95	287.0
185.15	1624.0
186.15	12051.0
187.15	3398.0
187.95	436.0
189.05	1076.0
190.05	186.0
191.05	322.0
192.15	1208.0
193.15	1304.0
193.95	288.0
196.15	1843.0
198.05	72152.0
199.05	4944.0
200.15	618.0
201.65	151.0
203.15	686.0
204.15	3901.0
205.15	5980.0
206.15	23184.0
207.15	2808.0
208.15	1022.0
209.05	509.0
210.05	477.0
211.15	866.0
212.15	419.0
213.05	154.0
214.95	425.0
215.25	442.0
216.05	776.0
217.05	6910.0
218.15	844.0
221.15	2359.0
222.05	741.0
223.05	1317.0
224.15	13538.0
225.05	3584.0
226.05	365.0
227.15	6066.0
228.15	762.0
229.15	1034.0
231.05	766.0
234.15	197.0
234.95	363.0
236.05	209.0
237.15	438.0
239.15	282.0
240.05	249.0
241.05	470.0
242.05	795.0
243.25	535.0
244.15	9573.0
245.05	1637.0
246.15	2345.0
247.05	415.0
249.05	355.0
252.35	235.0
253.45	520.0
255.15	51032.0
256.15	7349.0
257.05	492.0
258.05	3574.0
259.05	859.0
259.95	200.0
264.05	328.0
265.15	1415.0
266.05	260.0
271.05	166.0
271.95	303.0
273.15	1245.0
274.15	3687.0
275.15	21296.0
276.15	2265.0
277.15	2419.0
278.15	313.0
283.15	458.0
284.05	203.0
285.25	531.0
292.95	423.0
294.15	291.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

296.15	7673.0
297.15	1273.0
302.15	394.0
303.05	1070.0
303.95	323.0
309.95	226.0
314.05	286.0
315.05	786.0
316.25	482.0
320.95	178.0
323.15	1886.0
324.15	386.0
327.15	392.0
328.15	307.0
331.95	296.0
333.25	172.0
334.05	1403.0
335.05	303.0
341.25	354.0
345.95	516.0
347.15	185.0
351.95	467.0
353.05	645.0
354.05	501.0
365.05	3057.0
366.15	347.0
369.95	170.0
372.05	714.0
372.85	159.0
373.25	225.0
383.95	214.0
391.25	205.0
402.05	617.0
403.15	552.0
421.05	428.0
421.95	380.0
423.05	3398.0
424.25	438.0
439.75	168.0
441.05	8381.0
442.15	52152.0
443.05	11144.0
444.05	900.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

m/z	Abundance
37.05	161.0
38.15	662.0
39.25	2362.0
40.25	224.0
44.15	244.0
45.15	161.0
49.15	566.0
50.15	12058.0
51.15	41088.0
52.15	2341.0
55.05	295.0
56.05	1210.0
57.15	2706.0
61.15	913.0
62.25	671.0
63.15	2337.0
64.15	565.0
65.25	1001.0
69.15	47080.0
70.05	219.0
71.05	206.0
73.25	342.0
74.15	5257.0
75.15	8164.0
76.25	2550.0
77.15	51368.0
78.15	3217.0
79.15	4358.0
80.15	2912.0
81.15	4834.0
82.15	1116.0
83.15	493.0
85.15	880.0
86.15	912.0
87.15	677.0
87.95	253.0
89.25	227.0
91.15	1203.0
92.05	954.0
93.15	7339.0
94.05	413.0
96.05	356.0
98.15	5479.0
99.15	3693.0
100.15	405.0
101.15	2028.0
102.45	156.0
103.05	541.0
104.15	1656.0
105.15	1985.0
106.25	562.0
107.15	18288.0
108.15	2688.0
109.15	770.0
110.15	27952.0
111.15	4685.0
112.05	744.0
113.05	216.0
115.15	208.0
116.25	1497.0
117.15	15691.0
118.15	990.0
118.95	252.0
122.15	1134.0
123.05	1863.0
124.15	860.0
125.05	641.0
127.15	53416.0
128.15	3471.0
129.15	23680.0
130.15	2127.0
131.15	196.0
132.25	324.0
132.95	157.0
134.05	734.0
135.15	2468.0
136.05	793.0
137.05	1120.0
138.05	198.0
139.95	441.0
141.15	3161.0
142.05	1105.0
143.15	700.0
145.05	152.0
146.05	345.0
147.15	1647.0
148.15	3233.0
149.05	793.0
150.25	431.0
151.25	424.0
151.95	418.0
153.15	908.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

154.15	664.0
155.15	2367.0
156.15	2802.0
157.15	367.0
158.15	726.0
159.05	407.0
159.95	994.0
161.15	1520.0
162.05	281.0
162.85	217.0
163.85	150.0
165.05	880.0
166.15	905.0
167.15	5072.0
168.15	2336.0
169.05	447.0
170.15	482.0
171.35	336.0
172.15	634.0
173.15	897.0
174.15	1339.0
175.15	2368.0
176.05	612.0
177.05	1322.0
178.05	646.0
179.05	5304.0
180.05	4166.0
181.05	1293.0
182.05	396.0
184.15	472.0
185.15	2185.0
186.15	14460.0
187.15	5870.0
188.15	543.0
189.15	1023.0
190.15	307.0
191.15	702.0
192.15	1211.0
193.15	2060.0
194.15	224.0
195.05	408.0
196.15	3046.0
198.05	102696.0
199.05	7230.0
200.15	874.0
201.55	346.0
203.15	924.0
204.15	5412.0
205.15	7419.0
206.15	33288.0
207.15	4846.0
208.05	1553.0
209.05	344.0
210.15	591.0
211.25	1363.0
213.15	335.0
213.95	156.0
215.05	545.0
216.15	1071.0
217.15	11135.0
218.15	1410.0
221.15	2798.0
222.05	1440.0
223.15	1785.0
224.15	19032.0
225.15	4908.0
225.95	518.0
227.15	10350.0
228.15	1627.0
229.15	1809.0
230.25	208.0
231.15	965.0
232.15	227.0
232.95	225.0
234.15	685.0
235.15	571.0
235.95	509.0
237.25	571.0
239.05	372.0
240.25	396.0
241.05	515.0
242.15	1233.0
243.05	1298.0
244.15	13746.0
245.15	1975.0
246.15	2898.0
246.95	532.0
248.15	308.0
249.25	345.0
250.25	218.0
251.35	224.0
252.35	302.0
253.25	575.0
255.15	77232.0
256.15	10487.0
257.15	1356.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

258.15	5315.0
259.15	1001.0
265.15	2324.0
266.05	333.0
271.05	184.0
272.15	299.0
273.05	2572.0
274.15	5152.0
275.15	33848.0
276.15	4489.0
277.15	3491.0
278.15	480.0
282.95	366.0
283.95	232.0
285.15	759.0
285.95	257.0
290.95	304.0
293.05	606.0
294.75	309.0
296.15	12468.0
297.15	1717.0
301.05	198.0
302.25	208.0
303.15	878.0
304.15	290.0
314.05	528.0
315.05	1483.0
315.95	605.0
317.05	368.0
321.05	307.0
322.15	216.0
323.15	3889.0
324.15	718.0
327.05	599.0
328.25	396.0
332.05	345.0
332.95	474.0
334.15	2507.0
335.15	304.0
341.15	784.0
342.15	213.0
346.15	737.0
347.15	259.0
352.05	965.0
353.15	727.0
354.05	1019.0
355.15	356.0
358.35	158.0
365.05	6206.0
366.15	905.0
370.85	363.0
371.35	208.0
372.05	1677.0
373.25	475.0
383.05	561.0
384.15	374.0
389.95	185.0
390.25	229.0
390.95	339.0
391.85	169.0
392.35	162.0
401.05	160.0
402.15	922.0
403.15	1375.0
403.95	386.0
421.05	956.0
422.15	809.0
423.05	4794.0
423.95	1389.0
439.35	194.0
441.05	14052.0
442.15	90368.0
443.15	16496.0
444.05	1676.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

m/z	Abundance
37.25	228.0
38.15	332.0
39.15	1946.0
41.15	206.0
50.15	8807.0
51.15	34088.0
52.15	1867.0
55.05	163.0
56.15	885.0
57.15	2596.0
61.15	462.0
61.85	379.0
63.15	1684.0
65.15	688.0
69.15	35224.0
70.05	331.0
73.15	197.0
74.15	4550.0
75.15	6039.0
76.25	2189.0
77.15	38488.0
78.15	2514.0
79.15	2544.0
80.05	1643.0
81.15	3228.0
82.05	832.0
83.15	791.0
85.05	643.0
86.15	858.0
87.05	554.0
91.05	686.0
92.05	656.0
93.15	5765.0
94.05	373.0
98.15	4313.0
99.05	2974.0
100.15	243.0
101.05	1589.0
103.15	558.0
104.15	1488.0
105.15	1735.0
107.15	14767.0
108.15	2159.0
110.15	20320.0
111.15	3055.0
112.15	232.0
116.15	938.0
117.15	12812.0
118.05	936.0
122.15	815.0
123.05	1338.0
124.15	498.0
125.15	537.0
127.15	45888.0
128.15	3602.0
129.15	19752.0
130.15	1603.0
131.25	332.0
131.95	186.0
134.05	557.0
135.05	1624.0
136.05	623.0
137.15	874.0
137.85	211.0
138.95	155.0
140.15	418.0
141.15	2884.0
142.05	601.0
143.05	665.0
144.05	170.0
145.15	183.0
146.05	289.0
147.15	1616.0
148.15	3195.0
149.05	542.0
150.15	176.0
151.15	459.0
152.15	153.0
153.05	798.0
154.15	735.0
155.15	1475.0
156.15	1963.0
157.05	543.0
158.15	579.0
158.95	259.0
160.15	1310.0
161.15	987.0
162.15	474.0
165.15	1352.0
166.15	912.0
167.15	5058.0
168.15	2047.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

169.05	430.0
171.05	257.0
172.25	858.0
173.05	728.0
174.05	1170.0
175.15	2119.0
176.15	893.0
177.15	724.0
178.05	345.0
179.05	4004.0
180.15	2681.0
181.05	961.0
183.95	313.0
185.15	2180.0
186.15	15340.0
187.15	4535.0
188.15	481.0
189.05	1362.0
190.15	299.0
191.15	420.0
192.05	1396.0
193.15	1167.0
195.15	159.0
196.15	2305.0
198.05	90640.0
199.05	6502.0
200.05	436.0
200.95	294.0
201.65	298.0
203.05	972.0
204.15	4064.0
205.15	7335.0
206.15	26976.0
207.15	3695.0
208.15	948.0
208.95	308.0
210.05	292.0
211.15	1053.0
211.75	181.0
213.25	167.0
215.15	475.0
216.15	677.0
217.05	8607.0
218.05	968.0
220.05	156.0
221.15	2857.0
222.15	1265.0
223.15	1817.0
224.15	16472.0
225.15	4264.0
226.15	606.0
227.15	7994.0
228.05	821.0
229.15	1776.0
229.85	153.0
231.15	624.0
233.05	243.0
234.15	580.0
235.15	616.0
236.15	228.0
237.15	846.0
238.25	199.0
239.15	495.0
240.15	272.0
240.95	385.0
242.15	978.0
243.15	1252.0
244.15	13189.0
245.15	1897.0
246.15	3149.0
247.25	580.0
248.25	317.0
249.15	537.0
251.05	277.0
252.25	185.0
253.05	476.0
255.15	72648.0
256.15	10966.0
257.15	840.0
258.05	3991.0
259.15	581.0
263.85	192.0
265.15	2260.0
266.05	454.0
267.15	160.0
270.15	228.0
272.05	218.0
273.15	3026.0
274.15	5472.0
275.15	30928.0
276.15	3346.0
277.15	2855.0
277.95	549.0
282.25	230.0
283.05	533.0
285.05	904.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408

289.15	227.0
293.15	704.0
295.05	655.0
296.15	11189.0
297.15	1764.0
301.05	218.0
303.15	906.0
304.15	332.0
314.15	766.0
315.15	1292.0
316.15	826.0
316.95	302.0
321.05	160.0
323.15	4166.0
324.25	383.0
325.05	170.0
327.05	576.0
327.85	231.0
328.25	152.0
332.35	214.0
332.85	251.0
333.35	352.0
334.15	2854.0
335.15	490.0
341.05	224.0
341.95	249.0
346.05	972.0
347.15	179.0
352.05	946.0
353.25	663.0
354.05	1051.0
365.05	5146.0
365.95	878.0
370.85	226.0
372.05	1958.0
373.05	344.0
383.05	760.0
385.05	212.0
390.05	171.0
390.85	213.0
391.05	213.0
402.05	771.0
403.15	771.0
404.25	318.0
421.05	845.0
422.15	674.0
423.05	5594.0
424.15	1398.0
434.25	213.0
438.95	152.0
441.15	16258.0
442.15	94648.0
443.05	19784.0
444.05	1965.0

Instrument :
BNA_G
ClientSampleId :
DFTPP408