

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG073118\
 Data File : BG036002.D
 Acq On : 31 Jul 2018 15:32
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 31 17:22:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 12:30:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.613	0.544	11.3	89	0.00
3	Pyridine	1.601	1.548	3.3	87	0.00
4	n-Nitrosodimethylamine	0.687	0.589	14.3	81	0.00
5 S	2-Fluorophenol	1.205	1.202	0.2	92	0.00
6	Aniline	2.330	2.222	4.6	89	0.00
7 S	Phenol-d6	1.797	1.806	-0.5	93	0.00
8	2-Chlorophenol	1.225	1.259	-2.8	95	0.00
9	Benzaldehyde	1.103	0.981	11.1	81	0.00
10 C	Phenol	1.816	1.853	-2.0	94	0.00
11	bis(2-Chloroethyl)ether	1.422	1.385	2.6	91	0.00
12	1,3-Dichlorobenzene	1.480	1.516	-2.4	97	0.00
13 C	1,4-Dichlorobenzene	1.503	1.536	-2.2	96	0.00
14	1,2-Dichlorobenzene	1.444	1.484	-2.8	97	0.00
15	Benzyl Alcohol	1.632	1.582	3.1	90	0.00
16	2,2'-oxybis(1-Chloropropane	1.192	1.096	8.1	86	0.00
17	2-Methylphenol	1.235	1.250	-1.2	92	0.00
18	Hexachloroethane	0.575	0.572	0.5	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.513	1.453	4.0	91	0.00
20	3+4-Methylphenols	1.713	1.760	-2.7	91	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.622	0.615	1.1	93	0.00
23 S	Nitrobenzene-d5	0.479	0.471	1.7	91	0.00
24	Nitrobenzene	0.481	0.459	4.6	90	0.00
25	Isophorone	0.889	0.846	4.8	89	0.00
26 C	2-Nitrophenol	0.171	0.194	-13.5	102	0.00
27	2,4-Dimethylphenol	0.289	0.289	0.0	91	0.00
28	bis(2-Chloroethoxy)methane	0.490	0.481	1.8	91	0.00
29 C	2,4-Dichlorophenol	0.330	0.352	-6.7	95	0.00
30	1,2,4-Trichlorobenzene	0.405	0.423	-4.4	98	0.00
31	Naphthalene	1.042	1.018	2.3	94	0.00
32	Benzoic acid	0.195	0.164	15.9	77	0.00
33	4-Chloroaniline	0.454	0.441	2.9	92	0.00
34 C	Hexachlorobutadiene	0.316	0.337	-6.6	100	0.00
35	Caprolactam	0.142	0.122	14.1	83	0.00
36 C	4-Chloro-3-methylphenol	0.443	0.441	0.5	91	0.00
37	2-Methylnaphthalene	0.776	0.783	-0.9	94	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
39	1,2,4,5-Tetrachlorobenzene	0.755	0.773	-2.4	98	0.00
40 P	Hexachlorocyclopentadiene	0.396	0.400	-1.0	93	0.00
41 S	2,4,6-Tribromophenol	0.264	0.274	-3.8	97	0.00
42 C	2,4,6-Trichlorophenol	0.441	0.479	-8.6	97	0.00
43	2,4,5-Trichlorophenol	0.494	0.529	-7.1	102	0.00
44 S	2-Fluorobiphenyl	1.493	1.537	-2.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG073118\
 Data File : BG036002.D
 Acq On : 31 Jul 2018 15:32
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 31 17:22:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 12:30:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.551	1.591	-2.6	96	0.00
46	2-Chloronaphthalene	1.206	1.233	-2.2	97	0.00
47	2-Nitroaniline	0.426	0.418	1.9	90	0.00
48	Acenaphthylene	2.020	2.019	0.0	94	0.00
49	Dimethylphthalate	1.752	1.704	2.7	93	0.00
50	2,6-Dinitrotoluene	0.357	0.385	-7.8	99	0.00
51 C	Acenaphthene	1.259	1.241	1.4	94	0.00
52	3-Nitroaniline	0.365	0.354	3.0	88	0.00
53 P	2,4-Dinitrophenol	0.158	0.238	-50.6#	140	0.00
54	Dibenzofuran	2.002	2.011	-0.4	95	0.00
55 P	4-Nitrophenol	0.260	0.233	10.4	82	0.00
56	2,4-Dinitrotoluene	0.512	0.543	-6.1	94	0.00
57	Fluorene	1.678	1.658	1.2	95	0.00
58	2,3,4,6-Tetrachlorophenol	0.473	0.459	3.0	90	0.00
59	Diethylphthalate	1.802	1.720	4.6	90	0.00
60	4-Chlorophenyl-phenylether	0.986	0.995	-0.9	97	0.00
61	4-Nitroaniline	0.382	0.366	4.2	87	0.00
62	Azobenzene	1.777	1.617	9.0	86	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
64	4,6-Dinitro-2-methylphenol	0.111	0.127	-14.4	100	0.00
65 c	n-Nitrosodiphenylamine	0.569	0.599	-5.3	91	0.00
66	4-Bromophenyl-phenylether	0.232	0.257	-10.8	95	0.00
67	Hexachlorobenzene	0.239	0.257	-7.5	94	0.00
68	Atrazine	0.234	0.210	10.3	75	0.00
69 C	Pentachlorophenol	0.146	0.153	-4.8	88	0.00
70	Phenanthrene	0.998	1.015	-1.7	90	0.00
71	Anthracene	0.994	1.009	-1.5	89	0.00
72	Carbazole	0.944	0.933	1.2	86	0.00
73	Di-n-butylphthalate	1.115	1.095	1.8	85	0.00
74 C	Fluoranthene	1.344	1.298	3.4	84	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
76	Benzidine	0.526	0.518	1.5	87	0.00
77	Pyrene	1.265	1.270	-0.4	87	0.00
78 S	Terphenyl-d14	0.907	0.929	-2.4	88	0.00
79	Butylbenzylphthalate	0.452	0.487	-7.7	89	0.00
80	Benzo(a)anthracene	1.246	1.255	-0.7	87	0.00
81	3,3'-Dichlorobenzidine	0.435	0.467	-7.4	89	0.00
82	Chrysene	1.155	1.153	0.2	86	0.00
83	Bis(2-ethylhexyl)phthalate	0.615	0.671	-9.1	90	0.00
84 c	Di-n-octyl phthalate	1.029	1.136	-10.4	91	0.00
85	Indeno(1,2,3-cd)pyrene	1.279	1.330	-4.0	88	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	84	-0.01
87	Benzo(b)fluoranthene	1.216	1.238	-1.8	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG073118\
Data File : BG036002.D
Acq On : 31 Jul 2018 15:32
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 31 17:22:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 12:30:06 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.188	1.184	0.3	85	0.00
89 C	Benzo(a)pyrene	1.131	1.142	-1.0	86	0.00
90	Dibenzo(a,h)anthracene	1.110	1.144	-3.1	88	-0.01
91	Benzo(a,h,i)perylene	1.086	1.119	-3.0	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0