


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG073118\
 Data File : BG036001.D
 Acq On : 31 Jul 2018 14:53
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 31 12:30:06 2018



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.7	10546	PASS
68	69	0.00	2	1.0	124	PASS
69	198	0.00	100	30.2	12902	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	38.3	16343	PASS
197	198	0.00	2	0.4	181	PASS
198	198	100	100	100.0	42717	PASS
199	198	5	9	6.9	2945	PASS
275	198	10	60	28.9	12324	PASS
365	198	1	100	3.1	1305	PASS
441	198	0.01	100	10.5	4472	PASS
442	198	50	100	64.4	27493	PASS
443	442	15	24	18.9	5192	PASS