

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080119\
 Data File : BG041874.D
 Acq On : 1 Aug 2019 21:58
 Operator : HP/JU
 Sample : K4044-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-1

Quant Time: Aug 02 03:55:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 08:57:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.04	152	78930	20.00	ng	0.00
21) Naphthalene-d8	10.87	136	296400	20.00	ng	0.00
39) Acenaphthene-d10	14.70	164	196091	20.00	ng	0.00
64) Phenanthrene-d10	17.45	188	459357	20.00	ng	0.00
76) Chrysene-d12	21.73	240	622072	20.00	ng	0.00
87) Perylene-d12	25.00	264	741892	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.59	112	514694	126.87	ng	0.00
7) Phenol-d6	7.19	99	622985	113.91	ng	0.00
23) Nitrobenzene-d5	9.20	82	469780	109.07	ng	0.00
42) 2,4,6-Tribromophenol	16.18	330	330952	148.59	ng	0.00
45) 2-Fluorobiphenyl	13.32	172	1130090	101.64	ng	0.00
79) Terphenyl-d14	20.06	244	2226183	81.89	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080119\
Data File : BG041874.D
Acq On : 1 Aug 2019 21:58
Operator : HP/JU
Sample : K4044-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-1

Quant Time: Aug 02 03:55:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 08:57:26 2019
Response via : Initial Calibration

