

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG080218\
 Data File : BG036052.D
 Acq On : 2 Aug 2018 9:25
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 09:19:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 12:30:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.613	0.529	13.7	88	0.00
3	Pyridine	1.601	1.415	11.6	81	0.00
4	n-Nitrosodimethylamine	0.687	0.579	15.7	81	0.00
5 S	2-Fluorophenol	1.205	1.148	4.7	90	0.00
6	Aniline	2.330	2.152	7.6	88	0.00
7 S	Phenol-d6	1.797	1.695	5.7	89	0.00
8	2-Chlorophenol	1.225	1.215	0.8	94	0.00
9	Benzaldehyde	1.103	0.990	10.2	84	0.00
10 C	Phenol	1.816	1.749	3.7	90	0.00
11	bis(2-Chloroethyl)ether	1.422	1.340	5.8	89	0.00
12	1,3-Dichlorobenzene	1.480	1.446	2.3	94	0.00
13 C	1,4-Dichlorobenzene	1.503	1.505	-0.1	96	0.00
14	1,2-Dichlorobenzene	1.444	1.412	2.2	94	0.00
15	Benzyl Alcohol	1.632	1.478	9.4	86	0.00
16	2,2'-oxybis(1-Chloropropane	1.192	1.002	15.9	80	0.00
17	2-Methylphenol	1.235	1.186	4.0	89	0.00
18	Hexachloroethane	0.575	0.566	1.6	96	0.00
19 P	n-Nitroso-di-n-propylamine	1.513	1.344	11.2	86	0.00
20	3+4-Methylphenols	1.713	1.692	1.2	89	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	-0.01
22	Acetophenone	0.622	0.582	6.4	89	0.00
23 S	Nitrobenzene-d5	0.479	0.454	5.2	88	0.00
24	Nitrobenzene	0.481	0.447	7.1	88	0.00
25	Isophorone	0.889	0.810	8.9	85	0.00
26 C	2-Nitrophenol	0.171	0.181	-5.8	96	-0.01
27	2,4-Dimethylphenol	0.289	0.286	1.0	91	0.00
28	bis(2-Chloroethoxy)methane	0.490	0.448	8.6	85	0.00
29 C	2,4-Dichlorophenol	0.330	0.339	-2.7	92	0.00
30	1,2,4-Trichlorobenzene	0.405	0.414	-2.2	96	0.00
31	Naphthalene	1.042	1.000	4.0	93	0.00
32	Benzoic acid	0.195	0.158	19.0	75	0.02
33	4-Chloroaniline	0.454	0.431	5.1	90	0.00
34 C	Hexachlorobutadiene	0.316	0.324	-2.5	97	0.00
35	Caprolactam	0.142	0.133	6.3	91	0.00
36 C	4-Chloro-3-methylphenol	0.443	0.441	0.5	91	0.00
37	2-Methylnaphthalene	0.776	0.772	0.5	93	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
39	1,2,4,5-Tetrachlorobenzene	0.755	0.733	2.9	97	0.00
40 P	Hexachlorocyclopentadiene	0.396	0.354	10.6	85	0.00
41 S	2,4,6-Tribromophenol	0.264	0.287	-8.7	106	0.00
42 C	2,4,6-Trichlorophenol	0.441	0.453	-2.7	95	0.00
43	2,4,5-Trichlorophenol	0.494	0.508	-2.8	103	0.00
44 S	2-Fluorobiphenyl	1.493	1.431	4.2	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG080218\
 Data File : BG036052.D
 Acq On : 2 Aug 2018 9:25
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 09:19:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 31 12:30:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.551	1.494	3.7	94	0.00
46	2-Chloronaphthalene	1.206	1.177	2.4	97	0.00
47	2-Nitroaniline	0.426	0.417	2.1	94	0.00
48	Acenaphthylene	2.020	1.979	2.0	96	0.00
49	Dimethylphthalate	1.752	1.714	2.2	98	0.00
50	2,6-Dinitrotoluene	0.357	0.373	-4.5	100	0.00
51 C	Acenaphthene	1.259	1.217	3.3	96	0.00
52	3-Nitroaniline	0.365	0.377	-3.3	98	0.00
53 P	2,4-Dinitrophenol	0.158	0.208	-31.6#	128	0.00
54	Dibenzofuran	2.002	1.998	0.2	98	0.00
55 P	4-Nitrophenol	0.260	0.245	5.8	90	0.00
56	2,4-Dinitrotoluene	0.512	0.570	-11.3	103	0.00
57	Fluorene	1.678	1.675	0.2	100	0.00
58	2,3,4,6-Tetrachlorophenol	0.473	0.491	-3.8	100	0.00
59	Diethylphthalate	1.802	1.795	0.4	98	0.00
60	4-Chlorophenyl-phenylether	0.986	1.006	-2.0	102	0.00
61	4-Nitroaniline	0.382	0.389	-1.8	96	0.00
62	Azobenzene	1.777	1.653	7.0	92	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
64	4,6-Dinitro-2-methylphenol	0.111	0.121	-9.0	113	0.00
65 c	n-Nitrosodiphenylamine	0.569	0.548	3.7	100	0.00
66	4-Bromophenyl-phenylether	0.232	0.230	0.9	102	0.00
67	Hexachlorobenzene	0.239	0.241	-0.8	106	0.00
68	Atrazine	0.234	0.216	7.7	93	0.00
69 C	Pentachlorophenol	0.146	0.133	8.9	92	0.00
70	Phenanthrene	0.998	0.959	3.9	101	0.00
71	Anthracene	0.994	0.958	3.6	101	0.00
72	Carbazole	0.944	0.927	1.8	102	0.00
73	Di-n-butylphthalate	1.115	1.080	3.1	100	0.00
74 C	Fluoranthene	1.344	1.292	3.9	100	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
76	Benzidine	0.526	0.509	3.2	112	0.00
77	Pyrene	1.265	1.171	7.4	104	0.00
78 S	Terphenyl-d14	0.907	0.861	5.1	106	0.00
79	Butylbenzylphthalate	0.452	0.453	-0.2	108	0.00
80	Benzo(a)anthracene	1.246	1.196	4.0	107	0.00
81	3,3'-Dichlorobenzidine	0.435	0.451	-3.7	112	0.00
82	Chrysene	1.155	1.119	3.1	108	0.00
83	Bis(2-ethylhexyl)phthalate	0.615	0.635	-3.3	111	0.00
84 c	Di-n-octyl phthalate	1.029	1.081	-5.1	112	0.00
85	Indeno(1,2,3-cd)pyrene	1.279	1.298	-1.5	111	0.00
86 I	Perylene-d12	1.000	1.000	0.0	109	0.00
87	Benzo(b)fluoranthene	1.216	1.202	1.2	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG080218\
Data File : BG036052.D
Acq On : 2 Aug 2018 9:25
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 03 09:19:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 31 12:30:06 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.188	1.159	2.4	109	0.00
89 C	Benzo(a)pyrene	1.131	1.118	1.1	110	0.00
90	Dibenzo(a,h)anthracene	1.110	1.117	-0.6	112	0.00
91	Benzo(a,h,i)perylene	1.086	1.085	0.1	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0