

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/4/2023	BNA_G	<u>BG058601.D</u>
Compound Name	Response	Retention Time
DDT	520480	20.675
DDD	13865	20.24
DDE	55	19.1
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
13920	534400	2.60