

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080423\
 Data File : BG058602.D
 Acq On : 4 Aug 2023 8:38
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 09:11:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.634	0.568	10.4	79	0.00
3	Pyridine	1.729	1.675	3.1	84	0.00
4	n-Nitrosodimethylamine	0.732	0.681	7.0	81	0.00
5 S	2-Fluorophenol	1.309	1.245	4.9	83	0.00
6	Aniline	2.242	2.229	0.6	84	0.00
7 S	Phenol-d6	1.821	1.797	1.3	85	0.00
8	2-Chlorophenol	1.284	1.261	1.8	86	0.00
9	Benzaldehyde	0.989	0.916	7.4	85	0.00
10 C	Phenol	1.779	1.743	2.0	84	0.00
11	bis(2-Chloroethyl)ether	1.402	1.387	1.1	88	0.00
12	1,3-Dichlorobenzene	1.371	1.299	5.3	83	0.00
13 C	1,4-Dichlorobenzene	1.376	1.317	4.3	84	0.00
14	1,2-Dichlorobenzene	1.316	1.261	4.2	84	0.00
15	Benzyl Alcohol	1.155	1.287	-11.4	91	0.00
16	2,2'-oxybis(1-Chloropropane	2.278	2.308	-1.3	89	0.00
17	2-Methylphenol	1.300	1.302	-0.2	87	0.00
18	Hexachloroethane	0.500	0.485	3.0	85	0.00
19 P	n-Nitroso-di-n-propylamine	1.176	1.214	-3.2	90	0.00
20	3+4-Methylphenols	1.759	1.825	-3.8	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
22	Acetophenone	0.536	0.519	3.2	89	0.00
23 S	Nitrobenzene-d5	0.407	0.392	3.7	88	0.00
24	Nitrobenzene	0.414	0.395	4.6	87	0.00
25	Isophorone	0.841	0.815	3.1	90	0.00
26 C	2-Nitrophenol	0.180	0.176	2.2	87	0.00
27	2,4-Dimethylphenol	0.299	0.293	2.0	89	0.00
28	bis(2-Chloroethoxy)methane	0.458	0.452	1.3	89	0.00
29 C	2,4-Dichlorophenol	0.310	0.300	3.2	87	0.00
30	1,2,4-Trichlorobenzene	0.358	0.334	6.7	86	0.00
31	Naphthalene	1.054	0.998	5.3	88	0.00
32	Benzoic acid	0.207	0.215	-3.9	85	0.00
33	4-Chloroaniline	0.445	0.451	-1.3	89	0.00
34 C	Hexachlorobutadiene	0.230	0.215	6.5	86	0.00
35	Caprolactam	0.115	0.114	0.9	89	0.00
36 C	4-Chloro-3-methylphenol	0.384	0.368	4.2	87	0.00
37	2-Methylnaphthalene	0.754	0.720	4.5	88	0.00
38	1-Methylnaphthalene	0.717	0.673	6.1	88	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	0.614	0.588	4.2	88	0.00
41 P	Hexachlorocyclopentadiene	0.236	0.236	0.0	87	0.00
42 S	2,4,6-Tribromophenol	0.328	0.309	5.8	85	0.00
43 C	2,4,6-Trichlorophenol	0.422	0.409	3.1	87	0.00
44	2,4,5-Trichlorophenol	0.497	0.470	5.4	87	0.00
45 S	2-Fluorobiphenyl	1.335	1.278	4.3	90	0.00
46	1,1'-Biphenyl	1.374	1.317	4.1	89	0.00
47	2-Chloronaphthalene	1.071	1.035	3.4	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080423\
 Data File : BG058602.D
 Acq On : 4 Aug 2023 8:38
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 09:11:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.409	0.404	1.2	87	0.00
49	Acenaphthylene	1.793	1.702	5.1	87	0.00
50	Dimethylphthalate	1.502	1.454	3.2	89	0.00
51	2,6-Dinitrotoluene	0.330	0.321	2.7	88	0.00
52 C	Acenaphthene	1.036	0.992	4.2	89	0.00
53	3-Nitroaniline	0.330	0.330	0.0	88	0.00
54 P	2,4-Dinitrophenol	0.150	0.164	-9.3	92	0.00
55	Dibenzofuran	1.780	1.691	5.0	88	0.00
56 P	4-Nitrophenol	0.222	0.227	-2.3	85	0.00
57	2,4-Dinitrotoluene	0.457	0.450	1.5	88	0.00
58	Fluorene	1.414	1.336	5.5	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.424	0.407	4.0	87	0.00
60	Diethylphthalate	1.530	1.489	2.7	90	0.00
61	4-Chlorophenyl-phenylether	0.788	0.748	5.1	88	0.00
62	4-Nitroaniline	0.312	0.335	-7.4	91	0.00
63	Azobenzene	1.507	1.442	4.3	88	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	0.110	0.114	-3.6	90	0.00
66 c	n-Nitrosodiphenylamine	0.521	0.499	4.2	89	0.00
67	4-Bromophenyl-phenylether	0.227	0.213	6.2	87	0.00
68	Hexachlorobenzene	0.240	0.224	6.7	88	0.00
69	Atrazine	0.176	0.174	1.1	92	0.00
70 C	Pentachlorophenol	0.145	0.141	2.8	84	0.00
71	Phenanthrene	0.948	0.892	5.9	88	0.00
72	Anthracene	0.973	0.919	5.5	87	0.00
73	Carbazole	0.858	0.852	0.7	89	0.00
74	Di-n-butylphthalate	1.075	1.089	-1.3	92	0.00
75 C	Fluoranthene	1.139	1.117	1.9	90	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
77	Benzydine	0.303	0.393	-29.7#	102	0.00
78	Pyrene	1.361	1.298	4.6	91	0.00
79 S	Terphenyl-d14	1.064	1.022	3.9	93	0.00
80	Butylbenzylphthalate	0.559	0.551	1.4	92	0.00
81	Benzo(a)anthracene	1.376	1.325	3.7	91	0.00
82	3,3'-Dichlorobenzidine	0.465	0.467	-0.4	93	0.00
83	Chrysene	1.320	1.270	3.8	92	0.00
84	Bis(2-ethylhexyl)phthalate	0.817	0.812	0.6	93	0.00
85 c	Di-n-octyl phthalate	1.436	1.438	-0.1	94	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	1.331	1.295	2.7	91	0.01
88	Benzo(b)fluoranthene	1.085	1.070	1.4	92	0.00
89	Benzo(k)fluoranthene	1.090	1.050	3.7	91	0.00
90 C	Benzo(a)pyrene	1.065	1.039	2.4	92	0.00
91	Dibenzo(a,h)anthracene	1.100	1.068	2.9	91	0.01
92	Benzo(g,h,i)perylene	1.103	1.074	2.6	91	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080423\
Data File : BG058602.D
Acq On : 4 Aug 2023 8:38
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 04 09:11:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0