

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080524\
 Data File : BG062282.D
 Acq On : 6 Aug 2024 23:45
 Operator : MA/JU
 Sample : PB162282BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK282

Quant Time: Aug 07 01:54:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG080524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 07 01:46:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.964	152	55838	20.000	ng/u1	0.00
20) Naphthalene-d8	10.754	136	270427	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.579	164	208755	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	442703	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	433501	20.000	ng/u1	0.00
88) Perylene-d12	24.653	264	486574	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.423	96	7220	5.575	ng/uL	0.00
4) Pyridine-d5	3.828	84	117314	28.796	ng/u1	0.00
7) Phenol-d5	7.112	99	168184	31.557	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.288	67	102768	31.325	ng/u1	0.00
11) 2-Chlorophenol-d4	7.494	132	122080	31.971	ng/u1	0.00
15) 4-Methylphenol-d8	8.651	113	141588	31.895	ng/u1	0.00
21) Nitrobenzene-d5	9.115	128	64035	38.402	ng/u1	0.00
24) 2-Nitrophenol-d4	9.838	143	69288	44.171	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.367	165	145738	31.988	ng/u1	0.00
31) 4-Chloroaniline-d4	10.884	131	207744	31.438	ng/u1	0.00
46) Dimethylphthalate-d6	13.991	166	495760	30.781	ng/u1	0.00
49) Acenaphthylene-d8	14.273	160	564928	30.935	ng/u1	0.00
54) 4-Nitrophenol-d4	14.755	143	68800	30.512	ng/u1	0.00
60) Fluorene-d10	15.571	176	451691	31.244	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.677	200	46276	35.296	ng/u1	0.00
73) Anthracene-d10	17.416	188	687707	32.207	ng/u1	0.00
81) Pyrene-d10	19.701	212	836236	32.455	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.448	264	845975	32.524	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080524\
Data File : BG062282.D
Acq On : 6 Aug 2024 23:45
Operator : MA/JU
Sample : PB162282BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK282

Quant Time: Aug 07 01:54:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG080524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 07 01:46:39 2024
Response via : Initial Calibration

