


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG080620\
Data File : BG046174.D
Acq On : 6 Aug 2020 9:12
Operator : CG/JU
Sample : DFTPP29
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP29

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG072320MA.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 06 10:37:54 2020



AutoFind: Scans 2498, 2499, 2500; Background Corrected with Scan 2492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.0	59341	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	27.9	61301	PASS
70	69	0.00	2	0.5	326	PASS
127	198	10	80	37.2	81866	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	220074	PASS
199	198	5	9	7.3	15987	PASS
275	198	10	60	31.2	68618	PASS
365	198	1	100	4.1	9116	PASS
441	198	0.01	100	16.2	35762	PASS
442	198	50	100	98.1	215808	PASS
443	442	15	24	19.7	42578	PASS

Scan 2492 (17.732 min): BG046174.D
DFTPP29

m/z	Abundance
40.00	527.0
44.00	1297.0
48.90	186.0
73.00	170.0
80.80	162.0
207.00	859.0
208.00	226.0
281.10	210.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

m/z	Abundance
37.00	343.0
38.10	728.0
39.10	4651.0
40.00	547.0
41.00	348.0
44.00	1569.0
45.10	375.0
49.10	553.0
50.10	15840.0
51.10	64912.0
52.10	3426.0
55.00	331.0
56.00	2089.0
57.00	4863.0
58.00	362.0
61.00	802.0
62.10	799.0
63.00	2470.0
64.10	306.0
65.00	1430.0
67.20	292.0
69.00	67984.0
70.10	400.0
73.00	552.0
74.10	6224.0
75.00	10754.0
76.10	3555.0
77.10	78528.0
78.10	4780.0
79.00	5044.0
80.00	3640.0
81.00	5075.0
82.00	1288.0
83.10	1486.0
83.90	288.0
85.00	1053.0
86.10	1639.0
87.10	726.0
87.90	296.0
91.10	1192.0
91.90	1310.0
93.10	7870.0
93.90	536.0
96.10	374.0
98.00	5968.0
99.00	4952.0
99.90	538.0
101.10	3413.0
103.10	869.0
104.10	1912.0
105.00	1788.0
107.00	23192.0
108.00	3366.0
110.00	44368.0
111.00	7249.0
112.00	820.0
113.00	210.0
116.10	1208.0
117.00	17752.0
118.00	1208.0
119.20	183.0
120.10	271.0
122.00	1912.0
123.00	2443.0
124.00	1165.0
125.10	1166.0
127.00	83320.0
128.00	6211.0
129.00	31808.0
130.10	2848.0
130.90	414.0
132.00	379.0
132.90	254.0
134.00	985.0
135.00	2932.0
136.00	960.0
137.00	1589.0
137.90	288.0
139.00	231.0
139.90	602.0
141.00	4053.0
142.00	1252.0
143.10	873.0
144.00	237.0
145.00	227.0
146.00	935.0
147.00	2760.0
148.00	4823.0
149.10	973.0
149.90	220.0
151.10	691.0
152.00	474.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

153.00	1417.0
154.10	1293.0
155.10	2567.0
156.00	4819.0
157.10	889.0
158.00	942.0
158.90	714.0
160.00	1280.0
161.00	2098.0
162.10	608.0
164.00	379.0
165.00	2050.0
166.00	1573.0
167.00	8821.0
168.00	5363.0
169.10	981.0
170.10	396.0
171.00	498.0
172.00	974.0
173.00	1085.0
174.00	1811.0
175.10	3346.0
176.10	1184.0
177.00	2249.0
178.00	738.0
179.00	7055.0
180.00	5130.0
181.10	2058.0
182.10	569.0
182.80	295.0
183.10	272.0
184.00	720.0
185.00	3779.0
186.00	27192.0
187.00	7390.0
188.10	678.0
189.10	1788.0
190.00	187.0
191.20	865.0
192.00	2093.0
193.10	3070.0
194.00	437.0
195.10	628.0
196.00	5440.0
198.00	212672.0
199.00	15567.0
200.00	943.0
201.40	911.0
203.10	1471.0
204.10	7578.0
205.10	13791.0
206.10	57840.0
207.00	7939.0
208.00	2542.0
209.10	740.0
210.00	612.0
211.00	1827.0
212.00	380.0
215.00	682.0
216.10	1411.0
217.00	15039.0
218.00	1886.0
219.00	300.0
221.00	8340.0
221.90	2294.0
223.00	3479.0
224.00	29280.0
225.00	8039.0
226.10	1172.0
227.00	12646.0
228.00	1778.0
229.00	2848.0
230.00	393.0
230.90	1063.0
232.00	256.0
233.10	371.0
234.00	847.0
235.00	762.0
235.90	494.0
237.00	1435.0
238.00	326.0
239.00	645.0
240.00	499.0
241.00	803.0
242.00	1847.0
243.00	2083.0
244.00	25488.0
245.00	3685.0
246.00	4620.0
246.80	853.0
248.10	327.0
249.00	1170.0
251.00	266.0
252.00	586.0
253.00	815.0
255.00	134720.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

256.00	18624.0
257.00	1520.0
258.00	7529.0
259.00	1330.0
260.10	222.0
264.10	235.0
265.00	2381.0
266.00	357.0
270.00	265.0
271.00	380.0
272.00	314.0
273.00	4235.0
274.00	10251.0
275.00	62072.0
276.00	7630.0
277.00	4878.0
277.90	792.0
279.10	225.0
281.10	255.0
282.00	246.0
283.00	666.0
283.80	417.0
285.00	827.0
286.00	164.0
289.00	210.0
289.80	284.0
291.00	194.0
292.00	213.0
293.00	1347.0
294.00	285.0
294.90	324.0
296.00	16101.0
297.00	2014.0
297.90	152.0
301.00	575.0
302.00	327.0
303.00	1669.0
304.00	539.0
308.10	263.0
309.90	235.0
314.00	879.0
314.90	1740.0
316.00	934.0
321.00	415.0
322.00	250.0
323.10	5648.0
324.00	891.0
326.00	207.0
327.00	1214.0
328.00	665.0
332.00	550.0
333.00	428.0
334.00	3915.0
335.00	863.0
341.00	580.0
342.10	155.0
346.00	1051.0
347.10	252.0
352.00	1914.0
353.00	1183.0
354.00	1902.0
355.00	494.0
359.30	177.0
361.00	173.0
365.00	7894.0
366.00	1048.0
371.00	468.0
372.00	2743.0
373.10	769.0
383.10	757.0
383.80	240.0
390.00	329.0
390.90	244.0
392.00	239.0
402.00	1191.0
403.00	2024.0
404.00	660.0
421.10	1547.0
422.10	1403.0
423.10	10008.0
424.00	2785.0
425.00	247.0
441.00	25584.0
442.00	166784.0
443.00	34056.0
444.10	2781.0
445.20	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

m/z	Abundance
37.10	217.0
38.10	676.0
39.10	5313.0
40.00	763.0
41.10	238.0
44.10	1000.0
49.10	682.0
50.10	16424.0
51.10	70048.0
52.10	3293.0
53.00	234.0
55.10	433.0
56.00	2298.0
57.00	5166.0
58.10	208.0
61.10	909.0
62.00	835.0
63.00	2528.0
64.00	324.0
65.10	1509.0
69.00	70368.0
70.10	341.0
71.00	204.0
73.00	656.0
74.10	7374.0
75.00	11648.0
76.10	4246.0
77.10	87368.0
78.10	5813.0
79.00	5712.0
80.00	4215.0
81.00	6024.0
82.00	1087.0
83.00	1286.0
84.10	316.0
85.00	1061.0
86.00	2036.0
87.00	833.0
88.00	459.0
89.00	324.0
91.00	1163.0
92.00	1532.0
93.00	9013.0
93.90	462.0
94.90	186.0
95.90	453.0
97.00	219.0
98.00	6823.0
99.00	5636.0
100.00	593.0
101.00	4005.0
103.00	998.0
104.00	2432.0
105.00	2003.0
106.10	944.0
107.00	26136.0
108.00	4339.0
109.00	932.0
110.00	51200.0
111.00	7682.0
112.10	886.0
113.10	291.0
115.10	229.0
116.10	1585.0
117.00	20096.0
117.90	1519.0
119.00	235.0
120.00	518.0
122.00	1992.0
123.10	3056.0
124.00	1208.0
125.00	1374.0
127.00	96952.0
128.00	8315.0
129.00	38168.0
130.00	3354.0
131.10	567.0
132.00	288.0
133.00	290.0
134.00	959.0
135.00	2841.0
136.00	1502.0
137.00	1844.0
138.10	510.0
139.20	191.0
140.00	636.0
141.00	4956.0
142.00	1716.0
143.00	1295.0
144.00	261.0
145.10	276.0
146.00	1171.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

147.00	3083.0
148.00	5849.0
149.00	1419.0
150.10	339.0
151.00	535.0
152.00	529.0
153.00	1657.0
154.10	1227.0
155.10	3305.0
156.00	4509.0
157.00	824.0
158.00	1058.0
159.00	888.0
160.00	2010.0
161.00	2866.0
161.90	626.0
163.00	341.0
164.00	313.0
165.00	1726.0
166.00	1772.0
167.00	10162.0
168.00	5279.0
169.00	992.0
170.00	431.0
170.90	757.0
172.00	982.0
173.10	1537.0
174.00	2199.0
175.10	4457.0
176.00	1235.0
177.00	2450.0
178.00	875.0
179.00	9007.0
180.00	6515.0
181.00	2635.0
182.00	658.0
183.00	255.0
184.10	698.0
185.00	4158.0
186.00	33976.0
187.00	10358.0
188.00	896.0
189.00	2320.0
189.80	329.0
191.10	1100.0
192.00	2990.0
193.00	4004.0
194.10	735.0
196.00	7423.0
198.00	253568.0
199.00	19544.0
200.00	1266.0
201.00	530.0
201.50	1383.0
202.90	1799.0
204.10	9395.0
205.00	17224.0
206.00	71304.0
207.00	10167.0
208.00	2333.0
209.00	908.0
210.10	1211.0
211.00	3129.0
213.10	220.0
214.00	169.0
214.90	902.0
216.00	1541.0
217.00	18024.0
218.00	2652.0
219.00	261.0
221.00	12420.0
223.00	4374.0
224.00	39296.0
225.00	10385.0
226.10	1292.0
227.00	17136.0
228.00	1936.0
229.00	3720.0
230.10	577.0
231.00	1542.0
231.90	430.0
232.90	401.0
234.00	1079.0
235.00	1021.0
235.90	802.0
237.00	1515.0
239.00	736.0
239.90	437.0
241.00	870.0
242.00	2049.0
243.00	2579.0
244.00	32672.0
245.10	4227.0
246.00	6362.0
247.00	1256.0
248.10	430.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

249.00	1101.0
249.90	237.0
251.10	275.0
252.00	505.0
253.10	933.0
255.00	174912.0
256.00	25320.0
257.00	1831.0
258.00	9805.0
259.00	1720.0
259.70	162.0
261.00	413.0
264.10	280.0
265.00	4317.0
266.00	748.0
267.10	179.0
270.90	499.0
272.00	490.0
273.00	5250.0
274.00	13442.0
275.00	80912.0
276.00	10855.0
277.00	6095.0
278.00	1171.0
279.00	308.0
281.00	209.0
281.80	154.0
283.00	805.0
284.00	570.0
285.00	1042.0
286.20	330.0
288.90	430.0
290.90	225.0
292.10	378.0
293.00	1903.0
294.10	398.0
296.00	22744.0
297.00	3319.0
298.10	368.0
301.00	382.0
302.10	527.0
303.00	2669.0
304.10	516.0
308.00	295.0
308.90	379.0
310.10	408.0
312.10	195.0
312.90	270.0
314.00	1181.0
315.00	2578.0
316.00	1347.0
317.00	296.0
321.00	1198.0
321.90	331.0
323.00	7276.0
324.00	1218.0
325.00	297.0
327.00	1447.0
327.90	848.0
332.00	626.0
333.10	761.0
334.10	4778.0
335.00	1239.0
341.10	783.0
342.10	217.0
346.00	1670.0
346.90	511.0
349.90	163.0
352.10	2106.0
353.00	1885.0
354.00	2606.0
355.00	444.0
359.10	219.0
363.60	156.0
365.00	10334.0
366.00	1523.0
370.00	158.0
371.00	816.0
372.00	4671.0
373.10	866.0
383.10	1098.0
384.00	432.0
390.10	466.0
391.00	567.0
392.00	270.0
401.00	349.0
402.00	1926.0
403.00	2510.0
403.90	889.0
421.00	1978.0
422.00	1835.0
423.00	15174.0
424.00	2811.0
425.20	490.0
438.80	153.0
441.10	41728.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

442.00	250560.0
443.00	49712.0
444.00	4277.0
444.90	283.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

m/z	Abundance
37.10	252.0
38.00	483.0
39.00	2855.0
40.00	573.0
41.00	250.0
43.00	209.0
44.00	1547.0
49.00	235.0
50.10	11348.0
51.10	43064.0
52.10	2388.0
55.20	296.0
56.00	1371.0
57.00	3182.0
57.90	157.0
61.00	842.0
62.10	522.0
63.00	1649.0
64.10	284.0
65.10	792.0
69.00	45552.0
70.20	237.0
73.00	421.0
74.10	4827.0
75.10	7105.0
76.10	3015.0
77.10	56832.0
78.10	3635.0
79.00	3213.0
80.00	2572.0
81.00	4076.0
82.00	1083.0
83.00	903.0
85.00	709.0
86.10	949.0
87.00	669.0
87.90	235.0
89.90	163.0
90.90	646.0
92.00	811.0
93.00	6307.0
94.00	392.0
95.00	183.0
96.20	288.0
97.00	161.0
98.00	4715.0
99.00	3766.0
100.00	383.0
101.00	2419.0
103.00	847.0
104.00	1461.0
105.00	1384.0
107.00	16824.0
108.10	2650.0
110.00	33560.0
111.00	5297.0
112.00	708.0
112.90	280.0
116.00	1003.0
117.00	14641.0
118.00	947.0
118.90	155.0
119.90	168.0
122.00	1228.0
123.10	2049.0
124.00	903.0
125.10	881.0
127.00	65328.0
128.00	5354.0
129.00	26352.0
130.00	2421.0
131.00	537.0
132.10	281.0
133.00	210.0
133.90	724.0
135.00	1981.0
136.00	856.0
137.00	1164.0
138.00	328.0
139.20	160.0
140.00	343.0
141.10	3974.0
142.00	1091.0
142.90	920.0
144.00	187.0
145.10	341.0
146.00	587.0
147.00	2144.0
148.00	4517.0
148.90	818.0
149.80	197.0
151.10	724.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

152.00	312.0
153.00	1526.0
154.00	1033.0
155.00	1963.0
156.00	3328.0
157.00	766.0
158.00	735.0
158.90	568.0
160.00	1145.0
161.00	1752.0
162.00	629.0
163.00	266.0
164.10	230.0
165.00	1536.0
166.10	1398.0
167.00	8216.0
168.00	3700.0
169.00	838.0
169.90	357.0
171.00	523.0
172.00	1012.0
173.00	981.0
174.10	1748.0
175.10	3467.0
176.00	1103.0
177.00	1734.0
178.00	568.0
179.00	6039.0
180.10	4198.0
181.00	2293.0
182.00	345.0
183.10	241.0
184.00	745.0
185.10	3340.0
186.00	25672.0
187.00	7785.0
188.00	575.0
189.00	1389.0
190.00	247.0
191.00	1050.0
192.00	2371.0
193.10	2119.0
194.00	593.0
194.90	553.0
196.00	5430.0
198.00	193984.0
199.00	12851.0
200.00	1257.0
200.80	150.0
201.60	810.0
203.00	1543.0
204.00	7214.0
205.00	12621.0
206.00	53400.0
207.10	7638.0
208.00	1356.0
209.00	645.0
210.00	592.0
211.00	1810.0
213.00	167.0
215.00	789.0
216.10	1241.0
217.00	13639.0
218.00	1823.0
221.00	8107.0
221.80	1898.0
223.00	3292.0
224.00	30096.0
225.00	7518.0
226.10	1094.0
227.00	12205.0
228.00	1989.0
229.00	2903.0
230.00	327.0
231.00	1134.0
232.10	343.0
233.10	283.0
234.00	983.0
235.00	756.0
236.00	637.0
237.00	837.0
237.90	153.0
238.90	583.0
240.00	248.0
241.00	517.0
242.00	1436.0
243.00	1838.0
244.00	24632.0
245.10	3323.0
246.00	4669.0
247.10	1308.0
248.00	261.0
249.00	772.0
250.10	279.0
251.00	193.0
252.10	349.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29

253.10	815.0
255.00	134208.0
256.00	19136.0
257.00	1593.0
258.00	8066.0
259.00	1184.0
259.90	286.0
260.90	229.0
263.00	164.0
264.00	207.0
265.00	2548.0
266.10	518.0
269.80	171.0
271.00	365.0
272.10	438.0
273.00	4108.0
274.00	11747.0
275.00	62872.0
276.00	8372.0
277.00	5018.0
278.00	807.0
282.10	327.0
283.10	632.0
284.00	385.0
285.10	897.0
286.10	299.0
288.90	344.0
290.10	262.0
291.10	205.0
292.00	335.0
293.00	1362.0
294.00	307.0
296.00	18272.0
297.10	2314.0
300.90	380.0
301.90	309.0
303.00	2156.0
304.00	652.0
307.80	262.0
308.10	270.0
309.00	194.0
310.00	240.0
313.10	302.0
314.00	1108.0
315.00	1989.0
316.00	1112.0
317.10	162.0
321.10	777.0
322.00	308.0
323.00	6512.0
324.00	979.0
324.90	208.0
325.90	296.0
327.10	1119.0
328.00	573.0
332.10	511.0
333.00	547.0
334.00	4565.0
335.10	1006.0
341.00	817.0
342.10	175.0
346.00	1286.0
347.10	259.0
352.10	2114.0
353.00	1309.0
354.10	2208.0
354.90	586.0
365.00	9121.0
366.00	1203.0
369.90	303.0
371.00	742.0
372.10	3287.0
373.10	707.0
383.10	1316.0
383.90	275.0
390.00	682.0
390.90	294.0
392.00	158.0
400.90	275.0
402.10	1657.0
403.00	2212.0
404.00	819.0
405.00	192.0
420.90	2261.0
422.00	1984.0
423.00	14143.0
424.00	2958.0
424.90	261.0
439.10	164.0
441.00	39976.0
442.00	230080.0
443.00	43968.0
444.10	4107.0
445.00	261.0

Instrument :
BNA_G
ClientSampleId :
DFTPP29