

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080723\
 Data File : BG058632.D
 Acq On : 7 Aug 2023 18:22
 Operator : MA/JU
 Sample : 03903-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-07

Quant Time: Aug 08 01:31:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.815	152	34142	20.000	ng	0.00
21) Naphthalene-d8	10.612	136	154734	20.000	ng	0.00
39) Acenaphthene-d10	14.460	164	121339	20.000	ng	0.00
64) Phenanthrene-d10	17.204	188	330769	20.000	ng	0.00
76) Chrysene-d12	21.446	240	304733	20.000	ng	0.00
86) Perylene-d12	24.519	264	373496	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.382	112	230441	103.163	ng	0.00
7) Phenol-d6	6.981	99	327426	105.341	ng	0.00
23) Nitrobenzene-d5	8.972	82	184339	58.524	ng	0.00
42) 2,4,6-Tribromophenol	15.952	330	185654	93.348	ng	0.00
45) 2-Fluorobiphenyl	13.079	172	464275	57.340	ng	0.00
79) Terphenyl-d14	19.824	244	981702	60.581	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080723\
Data File : BG058632.D
Acq On : 7 Aug 2023 18:22
Operator : MA/JU
Sample : 03903-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MW-07

Quant Time: Aug 08 01:31:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration

