

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/8/2023	BNA_G	<u>BG058633.D</u>
Compound Name	Response	Retention Time
DDT	397979	20.673
DDD	12210	20.238
DDE	2837	19.733
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
15047	413026	3.64