

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080824\
 Data File : BG062336.D
 Acq On : 8 Aug 2024 19:34
 Operator : MA/JU
 Sample : P3401-12
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E1GF8

Quant Time: Aug 08 23:35:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG080524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 07 01:46:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	77232	20.000	ng/u1	0.00
20) Naphthalene-d8	10.748	136	374295	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.572	164	293655	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.315	188	627433	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	626780	20.000	ng/u1	-0.01
88) Perylene-d12	24.635	264	734113	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.422	96	6182	3.451	ng/uL	0.00
4) Pyridine-d5	3.828	84	25797	4.578	ng/u1	0.00
7) Phenol-d5	7.112	99	49868	6.765	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.282	67	112370	24.763	ng/u1	0.00
11) 2-Chlorophenol-d4	7.487	132	122644	23.222	ng/u1	0.00
15) 4-Methylphenol-d8	8.651	113	97382	15.860	ng/u1	0.00
21) Nitrobenzene-d5	9.109	128	71464	30.964	ng/u1	0.00
24) 2-Nitrophenol-d4	9.825	143	80379	37.022	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.366	165	164603	26.103	ng/u1	0.00
31) 4-Chloroaniline-d4	10.871	131	18809	2.057	ng/u1	-0.01
46) Dimethylphthalate-d6	13.985	166	563136	24.855	ng/u1	0.00
49) Acenaphthylene-d8	14.267	160	595186	23.169	ng/u1	0.00
54) 4-Nitrophenol-d4	14.748	143	21002	6.621	ng/u1	0.00
60) Fluorene-d10	15.565	176	518800	25.511	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.671	200	69278	37.283	ng/u1	0.00
73) Anthracene-d10	17.409	188	778480	25.724	ng/u1	-0.01
81) Pyrene-d10	19.694	212	971270	26.072	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.429	264	1057348	26.943	ng/u1	-0.01
Target Compounds						
					Qvalue	
16) Acetophenone	8.768	105	16680	1.706	ng/u1	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080824\
Data File : BG062336.D
Acq On : 8 Aug 2024 19:34
Operator : MA/JU
Sample : P3401-12
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E1GF8


Quant Time: Aug 08 23:35:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG080524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 07 01:46:39 2024
Response via : Initial Calibration





#16
 Acetophenone
 Concen: 1.706 ng/ul
 RT: 8.768 min Scan# 91
 Delta R.T. -0.012 min
 Lab File: BG062336.D
 Acq: 8 Aug 2024 19:34

Instrument :
 BNA_G
 ClientSampleId :
 E1GF8



Tgt Ion:	105	Resp:	16680
Ion	Ratio	Lower	Upper
105	100		
77	78.8	61.7	92.5
51	32.5	26.4	39.6

