

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG081019\
 Data File : BG042131.D
 Acq On : 10 Aug 2019 19:12
 Operator : HP/JU
 Sample : K4184-14
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0AA5

Quant Time: Aug 11 00:09:39 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 09 16:22:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.03	152	71926	20.00	ng/ul	0.00
18) Naphthalene-d8	10.85	136	302157	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.68	164	215739	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.44	188	451978	20.00	ng/ul	0.00
77) Chrysene-d12	21.72	240	454252	20.00	ng/ul	0.00
85) Perylene-d12	24.99	264	522937	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.41	96	7081	4.31	ng/uL	0.00
5) Phenol-d5	7.18	99	33212	5.88	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.34	67	56053	17.76	ng/ul	0.00
9) 2-Chlorophenol-d4	7.55	132	91447	18.95	ng/ul	0.00
13) 4-Methylphenol-d8	8.74	113	61619	13.00	ng/ul	0.00
19) Nitrobenzene-d5	9.19	128	50761	22.32	ng/ul	0.00
22) 2-Nitrophenol-d4	9.92	143	64102	24.27	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.47	165	125758	23.59	ng/ul	0.00
29) 4-Chloroaniline-d4	10.99	131	9424	1.59	ng/ul	0.01
43) Dimethylphthalate-d6	14.08	166	414460	24.77	ng/ul	0.00
46) Acenaphthylene-d8	14.38	160	478213	25.02	ng/ul	0.00
51) 4-Nitrophenol-d4	14.89	143	14110	5.02	ng/ul	0.02
57) Fluorene-d10	15.68	176	372979	25.06	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.80	200	66962	23.09	ng/ul	0.00
70) Anthracene-d10	17.54	188	574355	26.49	ng/ul	0.00
78) Pyrene-d10	19.82	212	656282	25.89	ng/ul	0.00
89) Benzo(a)pyrene-d12	24.77	264	724651	26.53	ng/ul	0.00

Target Compounds				Ovalue		
6) Phenol	7.21	94	33129	5.686	ng/ul#	89
44) Dimethylphthalate	14.13	163	46508	2.799	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG081019\
Data File : BG042131.D
Acq On : 10 Aug 2019 19:12
Operator : HP/JU
Sample : K4184-14
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0AA5

Quant Time: Aug 11 00:09:39 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 09 16:22:05 2019
Response via : Initial Calibration

