


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081020\
 Data File : BG046204.D
 Acq On : 10 Aug 2020 12:31
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG073020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 30 18:43:49 2020



AutoFind: Scans 2497, 2498, 2499; Background Corrected with Scan 2491

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.7	26754	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	27.6	25736	PASS
70	69	0.00	2	0.6	152	PASS
127	198	10	80	38.5	35949	PASS
197	198	0.00	2	0.4	336	PASS
198	198	100	100	100.0	93301	PASS
199	198	5	9	6.9	6479	PASS
275	198	10	60	30.2	28192	PASS
365	198	1	100	3.9	3636	PASS
441	198	0.01	100	15.1	14075	PASS
442	198	50	100	92.2	86021	PASS
443	442	15	24	19.9	17114	PASS