

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081023\
 Data File : BG058676.D
 Acq On : 10 Aug 2023 12:46
 Operator : MA/JU
 Sample : 03932-03DL 250X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 NEVINS-SB-01-Z30-32DL

Quant Time: Aug 11 01:53:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.812	152	43917	20.000	ng	0.00
21) Naphthalene-d8	10.615	136	190804	20.000	ng	0.00
39) Acenaphthene-d10	14.457	164	129925	20.000	ng	0.00
64) Phenanthrene-d10	17.207	188	302492	20.000	ng	0.00
76) Chrysene-d12	21.449	240	268460	20.000	ng	0.00
86) Perylene-d12	24.522	264	294050	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0	0.000	ng	
45) 2-Fluorobiphenyl	13.082	172	1610	0.186	ng	0.00
79) Terphenyl-d14	19.827	244	3067	0.215	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	10.667	128	431562	42.914	ng	98
37) 2-Methylnaphthalene	12.277	142	107269	14.912	ng	99
38) 1-Methylnaphthalene	12.501	142	70710	10.341	ng	98
52) Acenaphthene	14.522	154	51687	7.680	ng	96
58) Fluorene	15.509	166	29426	3.203	ng	97
71) Phenanthrene	17.248	178	148774	10.371	ng	97
72) Anthracene	17.342	178	37308	2.535	ng	96
75) Fluoranthene	19.263	202	58724	3.408	ng	97
78) Pyrene	19.628	202	93344	5.109	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081023\
Data File : BG058676.D
Acq On : 10 Aug 2023 12:46
Operator : MA/JU
Sample : 03932-03DL 250X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
NEVINS-SB-01-Z30-32DL

Quant Time: Aug 11 01:53:55 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration

