




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081023\
Data File : BG058677.D
Acq On : 10 Aug 2023 13:32
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 11 03:06:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration



TIC: BG058677.D\data.ms

(70) Pentachlorophenol (C)




16.866min (+ 0.007) 26627.37 ng

response 98790

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.40	61.61
264.00	62.80	63.71
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081023\
Data File : BG058677.D
Acq On : 10 Aug 2023 13:32
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 11 03:06:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration



TIC: BG058677.D\data.ms

(77) Benzidine

19.451min (+ 0.001) 76109.67 ng

response 732873

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.97
183.00	10.90	12.18
0.00	0.00	0.00