

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081023\
 Data File : BG058689.D
 Acq On : 10 Aug 2023 21:53
 Operator : MA/JU
 Sample : 03955-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-23

Quant Time: Aug 11 01:58:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.812	152	36679	20.000	ng	0.00
21) Naphthalene-d8	10.614	136	147686	20.000	ng	0.00
39) Acenaphthene-d10	14.457	164	111151	20.000	ng	0.00
64) Phenanthrene-d10	17.207	188	319693	20.000	ng	0.00
76) Chrysene-d12	21.449	240	289881	20.000	ng	0.00
86) Perylene-d12	24.522	264	315416	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.385	112	283255	118.036	ng	0.00
7) Phenol-d6	6.983	99	362507	108.561	ng	0.00
23) Nitrobenzene-d5	8.975	82	265334	88.259	ng	0.00
42) 2,4,6-Tribromophenol	15.949	330	217785	119.541	ng	0.00
45) 2-Fluorobiphenyl	13.076	172	622266	83.897	ng	0.00
79) Terphenyl-d14	19.827	244	1414955	91.792	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081023\
Data File : BG058689.D
Acq On : 10 Aug 2023 21:53
Operator : MA/JU
Sample : 03955-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-23

Quant Time: Aug 11 01:58:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.812 min Scan# 804
Delta R.T. 0.001 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53




Instrument :
BNA_G
ClientSampleId :
WC-23


Tgt Ion:152 Resp: 36679
Ion Ratio Lower Upper
152 100
150 151.9 120.5 180.7
115 57.6 45.9 68.9



#5
2-Fluorophenol
Concen: 118.036 ng
RT: 5.385 min Scan# 391
Delta R.T. 0.000 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53

Tgt Ion:112 Resp: 283255
Ion Ratio Lower Upper
112 100
64 70.1 55.3 82.9
63 35.4 28.3 42.5









#7
Phenol-d6
Concen: 108.561 ng
RT: 6.983 min Scan# 60
Delta R.T. 0.000 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53




Instrument :
BNA_G
ClientSampleId :
WC-23


Tgt Ion: 99 Resp: 362507
Ion Ratio Lower Upper
99 100
42 17.7 14.6 21.8
71 30.6 25.0 37.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.614 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53

Tgt Ion: 136 Resp: 147686
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 12.6 8.9 13.3
68 6.7 5.6 8.4









#23
Nitrobenzene-d5
Concen: 88.259 ng
RT: 8.975 min Scan# 10
Delta R.T. -0.005 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53




Instrument :
BNA_G
ClientSampleId :
WC-23


Tgt Ion: 82 Resp: 265334
Ion Ratio Lower Upper
82 100
128 38.0 32.2 48.2
54 56.2 46.0 69.0



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.457 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53

Tgt Ion:164 Resp: 111151
Ion Ratio Lower Upper
164 100
162 99.1 78.3 117.5
160 44.3 34.1 51.1





#42

2,4,6-Tribromophenol

Concen: 119.541 ng

RT: 15.949 min Scan# 21

Delta R.T. -0.005 min

Lab File: BG058689.D

Acq: 10 Aug 2023 21:53

Instrument :

BNA_G

ClientSampleId :

WC-23





Tgt Ion:330 Resp: 217785

Ion Ratio Lower Upper

330 100

332 94.3 77.5 116.3

141 26.2 20.7 31.1



#45

2-Fluorobiphenyl

Concen: 83.897 ng

RT: 13.076 min Scan# 1700

Delta R.T. -0.005 min

Lab File: BG058689.D

Acq: 10 Aug 2023 21:53




Tgt Ion:172 Resp: 622266


Ion Ratio Lower Upper

172 100

171 35.1 28.0 42.0

170 22.6 18.4 27.6









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.207 min Scan# 2403
Delta R.T. 0.000 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53




Instrument :
BNA_G
ClientSampleId :
WC-23


Tgt Ion:188 Resp: 319693
Ion Ratio Lower Upper
188 100
94 9.3 7.0 10.6
80 10.2 8.1 12.1



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.449 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BG058689.D
Acq: 10 Aug 2023 21:53

Tgt Ion:240 Resp: 289881
Ion Ratio Lower Upper
240 100
120 9.6 7.4 11.0
236 26.7 22.2 33.4









#79
 Terphenyl-d14
 Concen: 91.792 ng
 RT: 19.827 min Scan# 214
 Delta R.T. 0.000 min
 Lab File: BG058689.D
 Acq: 10 Aug 2023 21:53

Instrument :
 BNA_G
 ClientSampleId :
 WC-23

Tgt Ion:244 Resp: 1414955
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.9 8.9
 122 8.6 6.8 10.2



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 24.522 min Scan# 3648
 Delta R.T. 0.006 min
 Lab File: BG058689.D
 Acq: 10 Aug 2023 21:53

Tgt Ion:264 Resp: 315416
 Ion Ratio Lower Upper
 264 100
 260 24.6 18.4 27.6
 265 21.2 17.9 26.9

