

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081123\
 Data File : BG058694.D
 Acq On : 11 Aug 2023 8:22
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 11 11:40:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.634	0.571	9.9	76	-0.01
3	Pyridine	1.729	1.618	6.4	78	0.00
4	n-Nitrosodimethylamine	0.732	0.713	2.6	82	-0.01
5 S	2-Fluorophenol	1.309	1.218	7.0	78	0.00
6	Aniline	2.242	2.257	-0.7	82	0.00
7 S	Phenol-d6	1.821	1.806	0.8	82	0.00
8	2-Chlorophenol	1.284	1.257	2.1	82	0.00
9	Benzaldehyde	0.989	0.894	9.6	79	0.00
10 C	Phenol	1.779	1.764	0.8	82	0.00
11	bis(2-Chloroethyl)ether	1.402	1.354	3.4	82	0.00
12	1,3-Dichlorobenzene	1.371	1.264	7.8	78	0.00
13 C	1,4-Dichlorobenzene	1.376	1.280	7.0	78	0.00
14	1,2-Dichlorobenzene	1.316	1.230	6.5	79	0.00
15	Benzyl Alcohol	1.155	1.289	-11.6	88	0.00
16	2,2'-oxybis(1-Chloropropane	2.278	2.249	1.3	83	-0.01
17	2-Methylphenol	1.300	1.298	0.2	83	0.00
18	Hexachloroethane	0.500	0.470	6.0	79	0.00
19 P	n-Nitroso-di-n-propylamine	1.176	1.183	-0.6	84	0.00
20	3+4-Methylphenols	1.759	1.783	-1.4	82	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
22	Acetophenone	0.536	0.518	3.4	84	0.00
23 S	Nitrobenzene-d5	0.407	0.394	3.2	84	0.00
24	Nitrobenzene	0.414	0.398	3.9	84	0.00
25	Isophorone	0.841	0.814	3.2	85	0.00
26 C	2-Nitrophenol	0.180	0.171	5.0	80	0.00
27	2,4-Dimethylphenol	0.299	0.292	2.3	84	0.00
28	bis(2-Chloroethoxy)methane	0.458	0.447	2.4	84	0.00
29 C	2,4-Dichlorophenol	0.310	0.303	2.3	83	0.00
30	1,2,4-Trichlorobenzene	0.358	0.333	7.0	81	0.00
31	Naphthalene	1.054	0.999	5.2	83	0.00
32	Benzoic acid	0.207	0.154	25.6#	58	0.00
33	4-Chloroaniline	0.445	0.461	-3.6	86	0.00
34 C	Hexachlorobutadiene	0.230	0.210	8.7	80	0.00
35	Caprolactam	0.115	0.114	0.9	85	0.00
36 C	4-Chloro-3-methylphenol	0.384	0.374	2.6	84	0.00
37	2-Methylnaphthalene	0.754	0.725	3.8	84	0.00
38	1-Methylnaphthalene	0.717	0.681	5.0	85	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	0.614	0.556	9.4	84	0.00
41 P	Hexachlorocyclopentadiene	0.236	0.255	-8.1	94	0.00
42 S	2,4,6-Tribromophenol	0.328	0.297	9.5	82	0.00
43 C	2,4,6-Trichlorophenol	0.422	0.381	9.7	81	0.00
44	2,4,5-Trichlorophenol	0.497	0.444	10.7	82	0.00
45 S	2-Fluorobiphenyl	1.335	1.219	8.7	86	0.00
46	1,1'-Biphenyl	1.374	1.263	8.1	85	0.00
47	2-Chloronaphthalene	1.071	0.989	7.7	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081123\
 Data File : BG058694.D
 Acq On : 11 Aug 2023 8:22
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 11 11:40:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.409	0.402	1.7	87	0.00
49	Acenaphthylene	1.793	1.665	7.1	85	0.00
50	Dimethylphthalate	1.502	1.411	6.1	87	0.00
51	2,6-Dinitrotoluene	0.330	0.313	5.2	86	0.00
52 C	Acenaphthene	1.036	0.959	7.4	86	0.00
53	3-Nitroaniline	0.330	0.320	3.0	85	0.00
54 P	2,4-Dinitrophenol	0.150	0.136	9.3	77	0.00
55	Dibenzofuran	1.780	1.648	7.4	86	0.00
56 P	4-Nitrophenol	0.222	0.212	4.5	80	0.00
57	2,4-Dinitrotoluene	0.457	0.445	2.6	87	0.00
58	Fluorene	1.414	1.313	7.1	87	0.00
59	2,3,4,6-Tetrachlorophenol	0.424	0.382	9.9	82	0.00
60	Diethylphthalate	1.530	1.450	5.2	88	0.00
61	4-Chlorophenyl-phenylether	0.788	0.731	7.2	86	0.00
62	4-Nitroaniline	0.312	0.322	-3.2	88	0.00
63	Azobenzene	1.507	1.414	6.2	86	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	0.110	0.109	0.9	84	0.00
66 c	n-Nitrosodiphenylamine	0.521	0.505	3.1	88	0.00
67	4-Bromophenyl-phenylether	0.227	0.213	6.2	85	0.00
68	Hexachlorobenzene	0.240	0.222	7.5	85	0.00
69	Atrazine	0.176	0.167	5.1	86	0.00
70 C	Pentachlorophenol	0.145	0.136	6.2	79	0.00
71	Phenanthrene	0.948	0.905	4.5	87	0.00
72	Anthracene	0.973	0.934	4.0	87	0.00
73	Carbazole	0.858	0.846	1.4	87	0.00
74	Di-n-butylphthalate	1.075	1.053	2.0	87	0.00
75 C	Fluoranthene	1.139	1.114	2.2	87	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
77	Benzydine	0.303	0.377	-24.4	92	0.00
78	Pyrene	1.361	1.312	3.6	87	0.00
79 S	Terphenyl-d14	1.064	1.026	3.6	89	0.00
80	Butylbenzylphthalate	0.559	0.546	2.3	87	0.00
81	Benzo(a)anthracene	1.376	1.334	3.1	87	0.00
82	3,3'-Dichlorobenzidine	0.465	0.466	-0.2	88	0.00
83	Chrysene	1.320	1.282	2.9	88	0.00
84	Bis(2-ethylhexyl)phthalate	0.817	0.796	2.6	87	0.00
85 c	Di-n-octyl phthalate	1.436	1.406	2.1	87	0.00
86 I	Perylene-d12	1.000	1.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	1.331	1.292	2.9	86	0.01
88	Benzo(b)fluoranthene	1.085	1.076	0.8	88	0.00
89	Benzo(k)fluoranthene	1.090	1.051	3.6	87	0.00
90 C	Benzo(a)pyrene	1.065	1.039	2.4	87	0.00
91	Dibenzo(a,h)anthracene	1.100	1.074	2.4	87	0.01
92	Benzo(g,h,i)perylene	1.103	1.065	3.4	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081123\
Data File : BG058694.D
Acq On : 11 Aug 2023 8:22
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 11 11:40:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0