

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081123\
 Data File : BG058702.D
 Acq On : 11 Aug 2023 14:31
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 11 15:05:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.634	0.570	10.1	80	0.00
3	Pyridine	1.729	1.606	7.1	81	0.00
4	n-Nitrosodimethylamine	0.732	0.731	0.1	87	0.00
5 S	2-Fluorophenol	1.309	1.224	6.5	82	0.00
6	Aniline	2.242	2.226	0.7	85	0.00
7 S	Phenol-d6	1.821	1.783	2.1	85	0.00
8	2-Chlorophenol	1.284	1.254	2.3	86	0.00
9	Benzaldehyde	0.989	0.918	7.2	85	0.00
10 C	Phenol	1.779	1.757	1.2	85	0.00
11	bis(2-Chloroethyl)ether	1.402	1.353	3.5	86	0.00
12	1,3-Dichlorobenzene	1.371	1.270	7.4	82	0.00
13 C	1,4-Dichlorobenzene	1.376	1.277	7.2	82	0.00
14	1,2-Dichlorobenzene	1.316	1.241	5.7	83	0.00
15	Benzyl Alcohol	1.155	1.320	-14.3	94	0.00
16	2,2'-oxybis(1-Chloropropane	2.278	2.308	-1.3	90	-0.01
17	2-Methylphenol	1.300	1.306	-0.5	88	0.00
18	Hexachloroethane	0.500	0.474	5.2	84	0.00
19 P	n-Nitroso-di-n-propylamine	1.176	1.186	-0.9	89	0.00
20	3+4-Methylphenols	1.759	1.822	-3.6	88	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
22	Acetophenone	0.536	0.520	3.0	88	0.00
23 S	Nitrobenzene-d5	0.407	0.396	2.7	88	0.00
24	Nitrobenzene	0.414	0.401	3.1	88	0.00
25	Isophorone	0.841	0.819	2.6	89	0.00
26 C	2-Nitrophenol	0.180	0.179	0.6	88	0.00
27	2,4-Dimethylphenol	0.299	0.295	1.3	89	0.00
28	bis(2-Chloroethoxy)methane	0.458	0.447	2.4	87	0.00
29 C	2,4-Dichlorophenol	0.310	0.300	3.2	86	0.00
30	1,2,4-Trichlorobenzene	0.358	0.340	5.0	87	0.00
31	Naphthalene	1.054	0.991	6.0	87	0.00
32	Benzoic acid	0.207	0.166	19.8	66	0.00
33	4-Chloroaniline	0.445	0.456	-2.5	89	0.00
34 C	Hexachlorobutadiene	0.230	0.217	5.7	87	0.00
35	Caprolactam	0.115	0.110	4.3	86	0.00
36 C	4-Chloro-3-methylphenol	0.384	0.377	1.8	88	0.00
37	2-Methylnaphthalene	0.754	0.722	4.2	88	0.00
38	1-Methylnaphthalene	0.717	0.678	5.4	88	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	0.614	0.577	6.0	88	0.00
41 P	Hexachlorocyclopentadiene	0.236	0.290	-22.9	109	0.00
42 S	2,4,6-Tribromophenol	0.328	0.313	4.6	88	0.00
43 C	2,4,6-Trichlorophenol	0.422	0.394	6.6	85	0.00
44	2,4,5-Trichlorophenol	0.497	0.461	7.2	86	0.00
45 S	2-Fluorobiphenyl	1.335	1.255	6.0	89	0.00
46	1,1'-Biphenyl	1.374	1.303	5.2	89	0.00
47	2-Chloronaphthalene	1.071	1.016	5.1	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081123\
 Data File : BG058702.D
 Acq On : 11 Aug 2023 14:31
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 11 15:05:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.409	0.411	-0.5	90	0.00
49	Acenaphthylene	1.793	1.699	5.2	88	0.00
50	Dimethylphthalate	1.502	1.441	4.1	90	0.00
51	2,6-Dinitrotoluene	0.330	0.317	3.9	89	0.00
52 C	Acenaphthene	1.036	0.970	6.4	88	0.00
53	3-Nitroaniline	0.330	0.331	-0.3	89	0.00
54 P	2,4-Dinitrophenol	0.150	0.146	2.7	84	0.00
55	Dibenzofuran	1.780	1.681	5.6	89	0.00
56 P	4-Nitrophenol	0.222	0.216	2.7	82	0.00
57	2,4-Dinitrotoluene	0.457	0.457	0.0	90	0.00
58	Fluorene	1.414	1.329	6.0	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.424	0.399	5.9	87	0.00
60	Diethylphthalate	1.530	1.466	4.2	90	0.00
61	4-Chlorophenyl-phenylether	0.788	0.751	4.7	90	0.00
62	4-Nitroaniline	0.312	0.333	-6.7	92	0.00
63	Azobenzene	1.507	1.438	4.6	89	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	0.110	0.111	-0.9	89	0.00
66 c	n-Nitrosodiphenylamine	0.521	0.490	6.0	89	0.00
67	4-Bromophenyl-phenylether	0.227	0.213	6.2	89	0.00
68	Hexachlorobenzene	0.240	0.229	4.6	91	0.00
69	Atrazine	0.176	0.166	5.7	89	0.00
70 C	Pentachlorophenol	0.145	0.134	7.6	82	0.00
71	Phenanthrene	0.948	0.889	6.2	89	0.00
72	Anthracene	0.973	0.912	6.3	89	0.00
73	Carbazole	0.858	0.836	2.6	89	0.00
74	Di-n-butylphthalate	1.075	1.041	3.2	90	0.00
75 C	Fluoranthene	1.139	1.098	3.6	90	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzidine	0.303	0.370	-22.1	97	0.00
78	Pyrene	1.361	1.263	7.2	90	0.00
79 S	Terphenyl-d14	1.064	0.997	6.3	93	0.00
80	Butylbenzylphthalate	0.559	0.533	4.7	91	0.00
81	Benzo(a)anthracene	1.376	1.318	4.2	93	0.00
82	3,3'-Dichlorobenzidine	0.465	0.467	-0.4	95	0.00
83	Chrysene	1.320	1.249	5.4	92	0.00
84	Bis(2-ethylhexyl)phthalate	0.817	0.771	5.6	90	0.00
85 c	Di-n-octyl phthalate	1.436	1.365	4.9	91	0.00
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	1.331	1.286	3.4	91	0.02
88	Benzo(b)fluoranthene	1.085	1.049	3.3	91	0.00
89	Benzo(k)fluoranthene	1.090	1.062	2.6	93	0.00
90 C	Benzo(a)pyrene	1.065	1.034	2.9	92	0.00
91	Dibenzo(a,h)anthracene	1.100	1.067	3.0	91	0.02
92	Benzo(g,h,i)perylene	1.103	1.060	3.9	91	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081123\
Data File : BG058702.D
Acq On : 11 Aug 2023 14:31
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 11 15:05:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0