

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081222\
 Data File : BG054629.D
 Acq On : 12 Aug 2022 19:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 13 03:20:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 13 00:32:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	46.397	-16.0	110	0.00
3	Pyridine	40.000	49.296	-23.2	107	0.00
4	n-Nitrosodimethylamine	40.000	51.924	-29.8#	113	0.00
5 S	2-Fluorophenol	80.000	91.797	-14.7	104	0.00
6	Aniline	40.000	44.995	-12.5	101	0.00
7 S	Phenol-d6	80.000	91.920	-14.9	103	0.00
8	2-Chlorophenol	40.000	45.128	-12.8	102	0.00
9	Benzaldehyde	40.000	56.623	-41.6#	122	0.00
10 C	Phenol	40.000	45.921	-14.8	104	0.00
11	bis(2-Chloroethyl)ether	40.000	46.223	-15.6	107	0.00
12	1,3-Dichlorobenzene	40.000	43.734	-9.3	99	0.00
13 C	1,4-Dichlorobenzene	40.000	44.143	-10.4	100	0.00
14	1,2-Dichlorobenzene	40.000	44.370	-10.9	101	0.00
15	Benzyl Alcohol	40.000	44.447	-11.1	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	51.442	-28.6#	115	0.00
17	2-Methylphenol	40.000	45.171	-12.9	101	0.00
18	Hexachloroethane	40.000	46.455	-16.1	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	46.419	-16.0	102	0.00
20	3+4-Methylphenols	40.000	44.554	-11.4	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	41.908	-4.8	99	0.00
23 S	Nitrobenzene-d5	80.000	87.076	-8.8	102	0.00
24	Nitrobenzene	40.000	43.519	-8.8	105	0.00
25	Isophorone	40.000	42.869	-7.2	101	0.00
26 C	2-Nitrophenol	40.000	43.055	-7.6	100	0.00
27	2,4-Dimethylphenol	40.000	41.336	-3.3	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.466	-8.7	102	0.00
29 C	2,4-Dichlorophenol	40.000	40.466	-1.2	96	0.00
30	1,2,4-Trichlorobenzene	40.000	40.381	-1.0	96	0.00
31	Naphthalene	40.000	41.601	-4.0	99	0.00
32	Benzoic acid	40.000	21.069	47.3#	42	0.00
33	4-Chloroaniline	40.000	42.025	-5.1	101	0.00
34 C	Hexachlorobutadiene	40.000	39.287	1.8	93	0.00
35	Caprolactam	40.000	43.287	-8.2	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.914	-4.8	102	0.00
37	2-Methylnaphthalene	40.000	40.869	-2.2	99	0.00
38	1-Methylnaphthalene	40.000	40.416	-1.0	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.599	1.0	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	62.857	-57.1#	180	0.00
42 S	2,4,6-Tribromophenol	80.000	70.548	11.8	85	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.272	4.3	89	0.00
44	2,4,5-Trichlorophenol	40.000	35.389	11.5	86	0.00
45 S	2-Fluorobiphenyl	80.000	81.086	-1.4	96	0.00
46	1,1'-Biphenyl	40.000	40.629	-1.6	97	0.00
47	2-Chloronaphthalene	40.000	41.084	-2.7	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081222\
 Data File : BG054629.D
 Acq On : 12 Aug 2022 19:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 13 03:20:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 13 00:32:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.428	-11.1	104	0.00
49	Acenaphthylene	40.000	40.355	-0.9	96	0.00
50	Dimethylphthalate	40.000	40.045	-0.1	97	0.00
51	2,6-Dinitrotoluene	40.000	41.019	-2.5	99	0.00
52 C	Acenaphthene	40.000	40.887	-2.2	98	0.00
53	3-Nitroaniline	40.000	42.011	-5.0	102	0.00
54 P	2,4-Dinitrophenol	40.000	39.438	1.4	102	0.00
55	Dibenzofuran	40.000	40.005	-0.0	97	0.00
56 P	4-Nitrophenol	40.000	44.445	-11.1	105	0.00
57	2,4-Dinitrotoluene	40.000	41.018	-2.5	97	0.00
58	Fluorene	40.000	40.647	-1.6	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	31.381	21.5	74	0.00
60	Diethylphthalate	40.000	41.078	-2.7	99	0.00
61	4-Chlorophenyl-phenylether	40.000	39.668	0.8	95	0.00
62	4-Nitroaniline	40.000	41.713	-4.3	100	0.00
63	Azobenzene	40.000	42.742	-6.9	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	35.210	12.0	80	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.410	-3.5	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.595	1.0	92	0.00
68	Hexachlorobenzene	40.000	39.952	0.1	93	0.00
69	Atrazine	40.000	40.690	-1.7	97	0.00
70 C	Pentachlorophenol	40.000	42.342	-5.9	102	0.00
71	Phenanthrene	40.000	41.210	-3.0	97	0.00
72	Anthracene	40.000	40.935	-2.3	97	0.00
73	Carbazole	40.000	41.847	-4.6	100	0.00
74	Di-n-butylphthalate	40.000	43.027	-7.6	102	0.00
75 C	Fluoranthene	40.000	40.300	-0.7	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzydine	40.000	37.961	5.1	92	0.00
78	Pyrene	40.000	41.627	-4.1	97	0.00
79 S	Terphenyl-d14	80.000	81.823	-2.3	95	0.00
80	Butylbenzylphthalate	40.000	43.851	-9.6	104	0.00
81	Benzo(a)anthracene	40.000	40.395	-1.0	96	0.00
82	3,3'-Dichlorobenzidine	40.000	40.559	-1.4	94	0.00
83	Chrysene	40.000	40.315	-0.8	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.125	-7.8	102	0.00
85 c	Di-n-octyl phthalate	40.000	43.698	-9.2	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.237	-0.6	91	-0.02
88	Benzo(b)fluoranthene	40.000	41.565	-3.9	93	0.00
89	Benzo(k)fluoranthene	40.000	41.806	-4.5	96	0.00
90 C	Benzo(a)pyrene	40.000	41.595	-4.0	94	0.00
91	Dibenzo(a,h)anthracene	40.000	40.510	-1.3	92	0.00
92	Benzo(g,h,i)perylene	40.000	40.458	-1.1	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081222\
Data File : BG054629.D
Acq On : 12 Aug 2022 19:38
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 13 03:20:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 13 00:32:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0