


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081519\
 Data File : BG042219.D
 Acq On : 15 Aug 2019 8:34
 Operator : HP/JU
 Sample : DFTPP10
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP10

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Aug 19 13:55:01 2019



AutoFind: Scans 2518, 2519, 2520; Background Corrected with Scan 2512

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.1	56237	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	23.8	63442	PASS
70	69	0.00	2	0.5	342	PASS
127	198	10	80	37.1	98962	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	266709	PASS
199	198	5	9	7.1	18978	PASS
275	198	10	60	31.6	84400	PASS
365	198	1	100	3.8	10012	PASS
441	198	0.01	100	14.2	37789	PASS
442	198	50	100	86.4	230421	PASS
443	442	15	24	20.6	47448	PASS

m/z	Abundance
37.10	254.0
38.20	708.0
39.10	4287.0
40.00	254.0
41.10	228.0
44.00	616.0
50.10	16688.0
51.10	65432.0
52.10	3277.0
53.10	196.0
54.90	318.0
56.00	2182.0
57.00	4991.0
61.10	970.0
62.10	1153.0
63.00	3100.0
64.00	538.0
65.10	1459.0
69.00	71392.0
70.00	431.0
73.10	609.0
74.00	7378.0
75.00	12053.0
76.10	4330.0
77.10	86216.0
78.10	5960.0
79.00	6136.0
80.00	4663.0
81.00	6573.0
82.00	1784.0
83.00	1343.0
84.10	284.0
85.00	921.0
86.00	1672.0
87.00	939.0
87.90	364.0
89.00	250.0
91.10	1661.0
92.00	1626.0
93.00	9187.0
94.00	545.0
95.10	187.0
96.20	373.0
98.00	7245.0
99.00	6001.0
100.00	433.0
101.00	3518.0
102.00	216.0
103.00	1312.0
104.00	2543.0
105.00	2088.0
107.00	28416.0
108.00	4609.0
110.00	55248.0
111.00	8121.0
112.00	899.0
113.00	347.0
115.00	197.0
117.00	20360.0
118.00	1499.0
118.90	338.0
119.90	410.0
121.00	214.0
122.00	2184.0
123.00	3269.0
124.00	1445.0
125.00	1332.0
127.00	107888.0
128.00	8435.0
129.00	41960.0
130.00	3853.0
131.00	674.0
132.00	368.0
133.00	213.0
134.00	1399.0
135.00	3850.0
136.00	1414.0
137.00	1547.0
138.00	347.0
138.90	467.0
139.90	714.0
141.00	6174.0
142.00	2175.0
143.00	1427.0
144.10	264.0
144.90	246.0
146.00	930.0
147.00	3137.0
148.00	5513.0
149.00	1317.0
150.10	341.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

151.00	878.0
152.10	463.0
153.00	2264.0
154.00	1297.0
155.00	3483.0
156.00	5198.0
157.10	982.0
157.90	1301.0
158.90	911.0
160.00	1944.0
161.00	2691.0
161.80	755.0
163.00	282.0
164.00	510.0
165.00	2409.0
166.00	1786.0
167.00	11563.0
168.00	5036.0
169.00	995.0
170.00	548.0
170.90	539.0
172.00	1113.0
173.00	1647.0
174.00	2457.0
175.00	5118.0
176.00	1136.0
177.00	2140.0
178.00	1062.0
179.00	9123.0
180.00	6133.0
181.00	3055.0
182.10	503.0
182.90	365.0
184.00	987.0
185.00	4908.0
186.00	36824.0
187.00	9532.0
188.00	1190.0
189.00	2281.0
190.00	410.0
191.00	1020.0
192.00	2932.0
193.00	3549.0
194.10	489.0
195.00	521.0
196.00	8476.0
197.90	277440.0
198.90	20184.0
200.00	1315.0
201.50	1364.0
203.00	1997.0
204.00	9846.0
205.00	17416.0
206.00	74696.0
207.00	10557.0
208.00	2522.0
209.00	631.0
209.90	1085.0
211.00	2946.0
212.00	600.0
214.90	624.0
217.00	19128.0
218.00	2468.0
218.70	174.0
221.00	12899.0
223.00	4718.0
224.00	40264.0
225.00	10110.0
227.00	15097.0
228.00	2413.0
229.00	3401.0
230.00	408.0
231.00	1701.0
232.00	336.0
233.00	242.0
233.90	1121.0
234.90	1376.0
236.00	746.0
237.00	1208.0
238.00	229.0
239.00	542.0
240.00	479.0
241.00	928.0
242.00	2020.0
243.00	2804.0
244.00	33920.0
245.00	5045.0
246.00	5896.0
247.00	1269.0
248.10	369.0
249.00	1097.0
249.90	248.0
250.90	377.0
251.90	345.0
253.10	859.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

255.00	172800.0
256.00	26632.0
257.00	1963.0
257.90	9063.0
258.90	1573.0
260.00	288.0
260.80	337.0
263.90	393.0
265.00	3663.0
265.90	665.0
269.90	254.0
271.00	428.0
272.00	539.0
273.00	5516.0
274.00	14214.0
275.00	81184.0
276.00	11251.0
277.00	5969.0
277.90	1020.0
279.00	206.0
281.00	267.0
281.90	252.0
283.00	743.0
283.90	465.0
285.00	1402.0
285.90	270.0
288.90	401.0
289.90	329.0
291.00	155.0
292.00	315.0
293.00	1582.0
293.90	511.0
295.90	20496.0
297.00	2863.0
297.80	169.0
300.90	372.0
301.90	512.0
302.90	2618.0
304.00	753.0
307.90	435.0
309.00	296.0
310.00	305.0
312.80	192.0
313.90	1032.0
314.90	2439.0
316.00	1264.0
317.00	304.0
321.10	783.0
322.00	476.0
323.00	6345.0
324.00	1446.0
326.90	1356.0
327.90	758.0
331.90	379.0
333.00	604.0
334.00	3888.0
335.00	1220.0
338.90	176.0
341.00	794.0
345.90	1328.0
346.90	285.0
351.90	1968.0
352.90	1378.0
354.00	2259.0
354.90	459.0
359.00	154.0
364.90	9451.0
365.90	1237.0
366.90	170.0
370.90	538.0
371.90	3522.0
373.00	948.0
374.00	156.0
383.00	930.0
383.90	252.0
390.00	394.0
390.90	306.0
392.00	327.0
400.80	244.0
401.90	1260.0
402.90	1971.0
403.90	609.0
405.00	155.0
420.90	1484.0
421.90	1547.0
422.90	11151.0
423.90	2352.0
424.80	296.0
438.50	172.0
439.60	249.0
440.90	31616.0
441.90	193280.0
442.90	40672.0
443.90	3705.0
444.90	221.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

m/z	Abundance
37.10	151.0
38.10	687.0
39.00	2507.0
40.00	247.0
44.00	530.0
50.10	10110.0
51.10	39448.0
52.10	1909.0
53.10	180.0
54.90	162.0
55.20	157.0
56.10	1074.0
57.00	2977.0
61.10	546.0
62.00	631.0
63.10	1532.0
64.10	323.0
65.10	1100.0
69.00	45368.0
70.00	259.0
73.00	502.0
74.10	4286.0
75.00	7135.0
76.10	3071.0
77.10	51552.0
78.10	3960.0
79.00	3483.0
80.00	2651.0
81.00	3748.0
82.00	901.0
83.00	905.0
85.00	732.0
86.00	1095.0
87.00	446.0
88.00	221.0
91.00	877.0
92.00	973.0
93.00	6219.0
93.90	419.0
96.00	350.0
96.90	151.0
98.00	4779.0
99.00	3853.0
99.90	446.0
101.00	2409.0
101.80	171.0
103.00	745.0
104.00	1938.0
105.00	1490.0
107.00	18504.0
108.10	3042.0
110.00	35576.0
111.00	5213.0
112.00	683.0
113.00	205.0
116.00	1138.0
117.00	14413.0
117.90	1032.0
118.90	168.0
120.00	360.0
122.00	1497.0
123.00	2268.0
123.90	1028.0
125.00	897.0
127.00	75488.0
128.00	5764.0
129.00	29128.0
130.00	2381.0
131.00	497.0
132.10	287.0
133.10	231.0
133.90	746.0
135.00	2593.0
136.00	1045.0
137.00	1092.0
138.00	315.0
139.90	498.0
141.00	4052.0
142.00	1391.0
143.00	997.0
143.90	251.0
146.00	735.0
147.00	2432.0
148.00	4228.0
149.00	726.0
150.00	225.0
151.00	668.0
151.60	340.0
153.00	1145.0
154.00	1148.0
155.00	2422.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

156.00	3569.0
157.10	828.0
158.00	840.0
158.90	645.0
160.00	1467.0
161.00	2405.0
162.00	613.0
163.70	160.0
164.90	1771.0
166.00	1347.0
167.00	8419.0
168.00	3594.0
169.00	664.0
169.90	381.0
170.90	443.0
172.00	818.0
173.00	1102.0
174.00	1852.0
175.00	4086.0
176.10	1144.0
177.00	1880.0
178.00	693.0
179.00	7134.0
180.00	5053.0
181.00	2407.0
182.00	309.0
182.70	231.0
184.10	604.0
185.00	3463.0
186.00	27632.0
187.00	8059.0
188.00	616.0
189.00	1957.0
189.90	380.0
191.00	1012.0
192.00	2556.0
193.00	2849.0
194.00	638.0
195.00	432.0
196.00	6821.0
198.00	216576.0
198.90	15048.0
199.90	1025.0
201.40	928.0
202.90	1362.0
204.00	8227.0
205.00	14115.0
206.00	57552.0
207.00	7633.0
208.00	1681.0
209.00	457.0
210.00	940.0
211.00	2214.0
213.00	192.0
214.90	774.0
216.00	1595.0
217.00	14622.0
217.90	2048.0
218.90	290.0
219.90	398.0
221.00	10806.0
223.00	3711.0
224.00	33496.0
225.00	8097.0
226.00	948.0
227.00	13235.0
228.00	1875.0
228.90	2861.0
229.90	416.0
231.00	1366.0
232.00	264.0
232.90	277.0
233.90	934.0
235.00	951.0
236.00	647.0
237.00	1384.0
237.80	189.0
239.00	451.0
239.90	426.0
240.90	766.0
242.00	1616.0
243.00	2341.0
244.00	28112.0
245.00	4265.0
246.00	4940.0
247.00	1137.0
247.90	193.0
249.00	1132.0
249.90	216.0
250.90	285.0
251.90	343.0
252.90	869.0
255.00	144960.0
256.00	21328.0
257.00	1649.0

Instrument :

BNA_G

ClientSampleId :

DFTPP10

257.90	7824.0
258.90	1163.0
260.00	194.0
260.90	299.0
264.10	259.0
265.00	3268.0
265.90	518.0
269.90	231.0
271.00	368.0
272.00	632.0
273.00	5035.0
274.00	12196.0
275.00	73752.0
276.00	9335.0
277.00	5059.0
277.90	1014.0
281.00	156.0
282.10	213.0
283.00	537.0
284.00	427.0
285.00	924.0
286.00	231.0
288.90	304.0
289.90	275.0
292.00	252.0
293.00	1126.0
294.00	436.0
295.90	18544.0
296.90	2423.0
300.60	247.0
301.90	365.0
302.90	2281.0
304.00	622.0
308.00	353.0
308.90	195.0
310.00	303.0
312.90	170.0
314.00	1078.0
314.90	1850.0
315.90	1438.0
316.90	224.0
320.10	160.0
321.00	720.0
322.00	394.0
323.00	6431.0
323.90	1073.0
326.90	1156.0
328.00	738.0
328.90	156.0
331.90	535.0
333.00	748.0
334.00	3822.0
334.90	798.0
335.90	197.0
338.90	163.0
340.10	151.0
340.90	760.0
341.90	182.0
345.90	1408.0
346.90	384.0
351.00	214.0
352.00	2294.0
353.00	1470.0
354.00	2065.0
354.90	425.0
359.00	175.0
364.90	9206.0
365.90	1435.0
370.00	254.0
370.90	530.0
372.00	3821.0
373.00	892.0
383.00	901.0
383.90	303.0
389.90	530.0
390.90	298.0
400.90	235.0
401.90	1299.0
402.90	2400.0
404.00	528.0
420.90	1857.0
421.90	1874.0
422.90	13361.0
423.90	2393.0
424.90	274.0
438.00	155.0
440.90	38424.0
441.90	230784.0
442.90	46512.0
443.90	4480.0
444.90	267.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

m/z	Abundance
37.10	150.0
38.10	614.0
39.10	4487.0
40.00	283.0
41.10	256.0
44.00	596.0
50.10	16608.0
51.10	63832.0
52.10	3345.0
55.00	285.0
56.00	2119.0
57.00	4883.0
58.00	198.0
61.00	945.0
62.00	990.0
63.10	3153.0
64.10	401.0
65.10	1348.0
69.00	73568.0
70.10	338.0
72.90	598.0
74.00	7781.0
75.00	12138.0
76.10	4068.0
77.10	85840.0
78.10	5625.0
79.00	5747.0
80.00	4392.0
81.00	6443.0
82.00	1503.0
83.10	1511.0
83.90	178.0
85.00	1124.0
86.00	1908.0
87.10	905.0
88.00	271.0
89.10	173.0
91.00	1444.0
92.00	1543.0
93.00	10147.0
94.00	577.0
96.10	489.0
98.00	7582.0
99.00	5972.0
100.10	582.0
101.00	3718.0
101.90	334.0
103.00	1147.0
104.00	2454.0
105.00	2493.0
106.00	945.0
107.00	30080.0
108.00	4401.0
110.00	56080.0
111.00	8111.0
112.10	1100.0
112.90	263.0
115.00	178.0
116.00	2076.0
117.00	20632.0
118.00	1468.0
119.00	263.0
120.00	411.0
121.20	190.0
122.00	2229.0
123.00	3254.0
124.00	1440.0
125.00	1296.0
127.00	113512.0
128.00	8442.0
129.00	45296.0
130.00	3955.0
131.00	674.0
131.90	462.0
132.60	204.0
134.00	1342.0
135.00	3982.0
136.00	1322.0
137.10	1760.0
138.00	439.0
139.00	486.0
140.10	669.0
141.00	5899.0
142.10	1894.0
143.00	1198.0
144.00	321.0
145.00	309.0
146.00	1113.0
147.00	3214.0
148.00	5856.0
149.00	1540.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

149.90	627.0
151.20	953.0
152.00	372.0
153.00	2034.0
154.00	1453.0
155.00	3268.0
156.00	5395.0
157.00	1109.0
158.00	1424.0
159.00	932.0
160.00	2238.0
161.00	2668.0
162.00	1015.0
164.20	498.0
165.00	2326.0
166.00	1826.0
167.00	11728.0
168.00	4725.0
169.00	1209.0
170.00	442.0
171.00	385.0
172.00	1377.0
173.00	1645.0
174.00	2726.0
175.00	5299.0
176.00	1529.0
177.00	2536.0
178.00	895.0
179.00	9978.0
180.00	7540.0
181.00	3132.0
182.10	503.0
182.90	233.0
184.00	878.0
185.00	5336.0
186.00	41552.0
187.00	12008.0
188.00	1300.0
189.00	2255.0
189.90	447.0
191.00	1185.0
192.00	3478.0
193.00	3670.0
194.00	689.0
195.00	720.0
196.00	9680.0
197.90	306112.0
198.90	21704.0
199.90	1869.0
201.40	1376.0
203.00	2125.0
204.00	10659.0
205.00	18904.0
206.00	84184.0
207.00	10899.0
208.00	2654.0
209.00	784.0
210.10	1491.0
211.00	3082.0
213.00	171.0
213.80	164.0
215.00	1215.0
216.10	1722.0
216.90	20136.0
218.00	2444.0
219.10	308.0
221.00	15045.0
223.00	4927.0
224.00	47488.0
225.00	12257.0
226.10	1357.0
227.00	18848.0
227.90	2742.0
229.00	4129.0
230.00	580.0
231.00	1645.0
231.90	338.0
233.00	303.0
233.90	1217.0
235.00	1374.0
236.00	1047.0
237.00	1883.0
238.00	293.0
239.00	982.0
240.00	736.0
240.90	1003.0
242.00	2320.0
243.00	2802.0
244.00	40392.0
245.00	5347.0
246.00	7274.0
247.00	1325.0
248.00	276.0
248.90	1296.0
249.90	528.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

251.00	410.0
252.00	321.0
253.00	779.0
255.00	196864.0
256.00	31248.0
256.90	2227.0
258.00	11428.0
259.00	1804.0
259.90	370.0
261.00	453.0
262.80	182.0
264.10	449.0
264.90	4134.0
266.00	593.0
269.90	252.0
270.90	481.0
272.00	885.0
273.00	6728.0
274.00	17104.0
275.00	98264.0
276.00	13390.0
276.90	7147.0
277.90	1458.0
278.90	331.0
281.00	211.0
282.00	193.0
282.90	1212.0
283.90	696.0
285.00	1435.0
286.00	287.0
288.90	252.0
289.80	319.0
291.00	267.0
291.90	382.0
292.90	1545.0
293.90	593.0
296.00	24312.0
297.00	3520.0
298.00	268.0
301.00	380.0
302.00	633.0
303.00	3458.0
304.00	970.0
307.90	476.0
308.90	235.0
309.90	558.0
313.00	318.0
314.00	1074.0
315.00	2821.0
316.00	1728.0
317.00	275.0
321.00	710.0
321.90	506.0
323.00	8415.0
324.00	1777.0
326.90	1485.0
327.90	830.0
331.90	574.0
333.00	1031.0
334.00	5806.0
334.90	1278.0
339.00	164.0
340.00	214.0
340.90	895.0
342.00	349.0
345.90	1573.0
346.80	298.0
351.90	2760.0
353.00	1842.0
354.00	2721.0
355.00	551.0
358.90	188.0
364.90	11379.0
365.90	1785.0
369.90	202.0
370.90	641.0
372.00	4317.0
373.00	1205.0
382.90	1272.0
384.00	395.0
390.00	610.0
391.00	599.0
392.00	326.0
394.70	156.0
401.00	297.0
401.90	1982.0
402.90	2968.0
403.90	964.0
405.00	162.0
420.90	1972.0
422.00	1803.0
422.90	15083.0
424.00	2922.0
424.90	330.0
438.20	187.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

438.50	176.0
439.20	281.0
440.90	43328.0
441.90	267200.0
442.90	55160.0
443.90	5384.0
444.90	435.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

Scan 2512 (17.846 min): BG042219.D
DFTPP10

m/z	Abundance
40.00	172.0
44.00	689.0
206.90	271.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10