

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081523\
 Data File : BG058728.D
 Acq On : 15 Aug 2023 12:14
 Operator : MA/JU
 Sample : 03973-12
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SB-06-COMP

Quant Time: Aug 15 13:06:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 22:29:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.807	152	53843	20.000	ng	0.00
21) Naphthalene-d8	10.610	136	246491	20.000	ng	0.00
39) Acenaphthene-d10	14.459	164	171548	20.000	ng	0.00
64) Phenanthrene-d10	17.202	188	456306	20.000	ng	0.00
76) Chrysene-d12	21.444	240	416703	20.000	ng	0.00
86) Perylene-d12	24.517	264	474140	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	232512	66.004	ng	0.00
7) Phenol-d6	6.979	99	322835	65.861	ng	0.00
23) Nitrobenzene-d5	8.971	82	199241	39.708	ng	0.00
42) 2,4,6-Tribromophenol	15.951	330	183742	65.347	ng	0.00
45) 2-Fluorobiphenyl	13.078	172	548117	47.882	ng	0.00
79) Terphenyl-d14	19.823	244	956419	43.162	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081523\
Data File : BG058728.D
Acq On : 15 Aug 2023 12:14
Operator : MA/JU
Sample : 03973-12
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SB-06-COMP

Quant Time: Aug 15 13:06:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration

