


Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG081619\  
Data File : BG042253.D  
Acq On : 16 Aug 2019 10:48  
Operator : HP/JU  
Sample : DFTPP11  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\SOM-EPA-BG081619MA.M  
Title : SVOA CALIBRATION  
Last Update : Fri Aug 16 15:15:09 2019



AutoFind: Scans 2515, 2516, 2517; Background Corrected with Scan 2508

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 21.6      | 55784   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 24.1      | 62037   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 342     | PASS             |
| 127         | 198          | 10           | 80           | 36.9      | 95024   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 257664  | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 17917   | PASS             |
| 275         | 198          | 10           | 60           | 30.7      | 79090   | PASS             |
| 365         | 198          | 1            | 100          | 3.7       | 9548    | PASS             |
| 441         | 198          | 0.01         | 100          | 14.0      | 35986   | PASS             |
| 442         | 198          | 50           | 100          | 86.1      | 221973  | PASS             |
| 443         | 442          | 15           | 24           | 20.2      | 44776   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 185.0     |
| 38.10  | 580.0     |
| 39.10  | 4418.0    |
| 40.00  | 365.0     |
| 43.20  | 200.0     |
| 44.00  | 749.0     |
| 49.00  | 708.0     |
| 50.10  | 16054.0   |
| 51.10  | 65072.0   |
| 52.10  | 3320.0    |
| 53.00  | 237.0     |
| 55.00  | 307.0     |
| 56.00  | 2254.0    |
| 57.00  | 4839.0    |
| 58.00  | 269.0     |
| 61.00  | 930.0     |
| 62.00  | 1101.0    |
| 63.00  | 2719.0    |
| 64.10  | 550.0     |
| 65.10  | 1474.0    |
| 67.10  | 166.0     |
| 69.00  | 70760.0   |
| 70.10  | 500.0     |
| 73.10  | 552.0     |
| 74.00  | 6682.0    |
| 75.00  | 11202.0   |
| 76.10  | 4521.0    |
| 77.10  | 81136.0   |
| 78.10  | 5604.0    |
| 79.00  | 5588.0    |
| 80.00  | 4535.0    |
| 81.00  | 5914.0    |
| 82.00  | 1676.0    |
| 83.00  | 1351.0    |
| 83.90  | 301.0     |
| 85.00  | 1059.0    |
| 86.00  | 2014.0    |
| 87.00  | 843.0     |
| 88.00  | 440.0     |
| 89.00  | 169.0     |
| 91.00  | 1513.0    |
| 92.00  | 1650.0    |
| 93.00  | 9327.0    |
| 94.10  | 584.0     |
| 96.10  | 399.0     |
| 98.00  | 7018.0    |
| 99.00  | 5145.0    |
| 100.00 | 559.0     |
| 101.00 | 3127.0    |
| 101.90 | 167.0     |
| 103.00 | 1039.0    |
| 104.00 | 2749.0    |
| 105.00 | 2494.0    |
| 107.00 | 27440.0   |
| 108.00 | 4598.0    |
| 110.00 | 50832.0   |
| 111.00 | 7898.0    |
| 112.00 | 1050.0    |
| 113.00 | 371.0     |
| 115.00 | 203.0     |
| 116.00 | 1687.0    |
| 117.00 | 19280.0   |
| 118.00 | 1530.0    |
| 119.00 | 271.0     |
| 120.00 | 388.0     |
| 121.00 | 156.0     |
| 122.00 | 1751.0    |
| 123.00 | 3209.0    |
| 123.90 | 1295.0    |
| 125.00 | 1367.0    |
| 127.00 | 103872.0  |
| 128.00 | 8670.0    |
| 129.00 | 40504.0   |
| 130.00 | 3231.0    |
| 131.00 | 639.0     |
| 131.90 | 360.0     |
| 133.00 | 268.0     |
| 133.90 | 1117.0    |
| 135.00 | 3363.0    |
| 136.00 | 1284.0    |
| 137.00 | 1698.0    |
| 137.90 | 450.0     |
| 139.00 | 289.0     |
| 140.00 | 498.0     |
| 141.00 | 5728.0    |
| 142.00 | 1644.0    |
| 143.00 | 1097.0    |
| 144.10 | 376.0     |
| 145.00 | 417.0     |
| 145.90 | 960.0     |
| 147.00 | 3106.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

|        |          |
|--------|----------|
| 148.00 | 5668.0   |
| 149.00 | 1110.0   |
| 150.10 | 367.0    |
| 151.00 | 927.0    |
| 151.90 | 506.0    |
| 153.00 | 1751.0   |
| 154.00 | 1547.0   |
| 155.00 | 3433.0   |
| 156.00 | 5292.0   |
| 157.00 | 935.0    |
| 157.90 | 1231.0   |
| 159.00 | 862.0    |
| 160.00 | 1907.0   |
| 161.00 | 2730.0   |
| 162.00 | 921.0    |
| 163.10 | 232.0    |
| 164.00 | 316.0    |
| 165.00 | 2311.0   |
| 166.00 | 1950.0   |
| 167.00 | 11025.0  |
| 168.00 | 4338.0   |
| 169.00 | 946.0    |
| 170.00 | 512.0    |
| 170.90 | 517.0    |
| 171.90 | 1139.0   |
| 173.00 | 1344.0   |
| 174.00 | 2517.0   |
| 175.00 | 5035.0   |
| 176.00 | 1190.0   |
| 177.00 | 1787.0   |
| 178.00 | 718.0    |
| 179.00 | 8968.0   |
| 180.00 | 6511.0   |
| 181.00 | 3090.0   |
| 182.10 | 447.0    |
| 183.10 | 295.0    |
| 184.00 | 775.0    |
| 185.00 | 4830.0   |
| 186.00 | 34272.0  |
| 187.00 | 9845.0   |
| 188.00 | 1172.0   |
| 189.00 | 2233.0   |
| 189.90 | 495.0    |
| 191.00 | 1075.0   |
| 192.00 | 2862.0   |
| 193.00 | 3407.0   |
| 194.00 | 865.0    |
| 196.00 | 7550.0   |
| 197.90 | 264576.0 |
| 198.90 | 18984.0  |
| 200.00 | 1296.0   |
| 201.50 | 1313.0   |
| 203.00 | 1926.0   |
| 204.00 | 9618.0   |
| 205.00 | 16576.0  |
| 206.00 | 70848.0  |
| 207.00 | 10814.0  |
| 208.00 | 2412.0   |
| 209.00 | 653.0    |
| 210.20 | 1122.0   |
| 211.00 | 2659.0   |
| 213.00 | 277.0    |
| 215.00 | 600.0    |
| 216.00 | 1697.0   |
| 216.90 | 17560.0  |
| 218.00 | 2509.0   |
| 219.00 | 284.0    |
| 221.00 | 12098.0  |
| 221.70 | 2953.0   |
| 223.00 | 4315.0   |
| 224.00 | 40568.0  |
| 225.00 | 10780.0  |
| 226.10 | 1387.0   |
| 227.00 | 15003.0  |
| 228.00 | 2503.0   |
| 229.00 | 3573.0   |
| 230.00 | 462.0    |
| 231.00 | 1567.0   |
| 231.90 | 245.0    |
| 233.00 | 290.0    |
| 234.00 | 1132.0   |
| 235.00 | 1332.0   |
| 235.90 | 808.0    |
| 237.00 | 1323.0   |
| 238.00 | 155.0    |
| 238.90 | 503.0    |
| 240.00 | 738.0    |
| 241.00 | 817.0    |
| 242.00 | 1983.0   |
| 243.00 | 2604.0   |
| 244.00 | 32200.0  |
| 245.00 | 4297.0   |
| 246.00 | 5674.0   |
| 246.90 | 983.0    |
| 248.00 | 396.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

|        |          |
|--------|----------|
| 248.90 | 964.0    |
| 249.90 | 293.0    |
| 251.00 | 219.0    |
| 252.00 | 328.0    |
| 253.00 | 921.0    |
| 255.00 | 158208.0 |
| 256.00 | 23704.0  |
| 257.00 | 2002.0   |
| 258.00 | 8409.0   |
| 259.00 | 1534.0   |
| 259.90 | 282.0    |
| 260.80 | 263.0    |
| 264.00 | 315.0    |
| 264.90 | 3603.0   |
| 265.90 | 421.0    |
| 270.00 | 194.0    |
| 271.00 | 327.0    |
| 272.90 | 5194.0   |
| 274.00 | 13133.0  |
| 275.00 | 77384.0  |
| 276.00 | 10567.0  |
| 277.00 | 5900.0   |
| 277.90 | 928.0    |
| 278.90 | 276.0    |
| 280.80 | 221.0    |
| 282.00 | 248.0    |
| 283.00 | 799.0    |
| 284.00 | 554.0    |
| 285.00 | 1106.0   |
| 286.00 | 219.0    |
| 288.90 | 223.0    |
| 290.00 | 198.0    |
| 291.00 | 263.0    |
| 291.90 | 331.0    |
| 293.00 | 1387.0   |
| 294.00 | 505.0    |
| 295.90 | 19144.0  |
| 296.90 | 2952.0   |
| 297.90 | 279.0    |
| 301.00 | 305.0    |
| 302.00 | 433.0    |
| 303.00 | 2304.0   |
| 303.90 | 759.0    |
| 304.90 | 210.0    |
| 308.10 | 259.0    |
| 309.00 | 214.0    |
| 309.90 | 294.0    |
| 313.10 | 219.0    |
| 314.00 | 845.0    |
| 314.90 | 2043.0   |
| 316.00 | 1029.0   |
| 317.00 | 151.0    |
| 320.00 | 155.0    |
| 321.00 | 539.0    |
| 321.90 | 503.0    |
| 323.00 | 6490.0   |
| 324.00 | 1213.0   |
| 325.00 | 233.0    |
| 325.90 | 177.0    |
| 326.90 | 1284.0   |
| 327.90 | 622.0    |
| 331.90 | 708.0    |
| 332.90 | 474.0    |
| 334.00 | 4111.0   |
| 334.90 | 1051.0   |
| 338.80 | 208.0    |
| 341.00 | 786.0    |
| 341.90 | 257.0    |
| 345.90 | 1348.0   |
| 346.80 | 352.0    |
| 352.00 | 2040.0   |
| 352.90 | 1444.0   |
| 354.00 | 2370.0   |
| 354.90 | 373.0    |
| 358.80 | 219.0    |
| 364.90 | 8533.0   |
| 365.90 | 1457.0   |
| 370.00 | 268.0    |
| 370.90 | 382.0    |
| 372.00 | 3085.0   |
| 372.90 | 801.0    |
| 382.90 | 724.0    |
| 383.90 | 289.0    |
| 389.90 | 522.0    |
| 390.90 | 271.0    |
| 391.90 | 151.0    |
| 400.90 | 271.0    |
| 402.00 | 1402.0   |
| 402.90 | 1858.0   |
| 403.90 | 775.0    |
| 404.80 | 184.0    |
| 421.00 | 1397.0   |
| 421.90 | 1437.0   |
| 422.90 | 10411.0  |
| 423.90 | 2456.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

|        |          |
|--------|----------|
| 424.80 | 220.0    |
| 440.90 | 28800.0  |
| 441.90 | 186944.0 |
| 442.90 | 39184.0  |
| 443.90 | 3458.0   |
| 444.90 | 276.0    |

|                  |
|------------------|
| Instrument :     |
| BNA_G            |
| ClientSampleId : |
| DFTPP11          |

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 193.0     |
| 38.10  | 657.0     |
| 39.10  | 3929.0    |
| 40.10  | 348.0     |
| 41.10  | 238.0     |
| 42.90  | 151.0     |
| 44.00  | 861.0     |
| 49.10  | 735.0     |
| 50.10  | 17032.0   |
| 51.10  | 64600.0   |
| 52.10  | 3597.0    |
| 53.00  | 203.0     |
| 55.00  | 409.0     |
| 56.00  | 2163.0    |
| 57.00  | 4688.0    |
| 58.00  | 338.0     |
| 61.00  | 840.0     |
| 62.10  | 1103.0    |
| 63.00  | 3054.0    |
| 64.00  | 403.0     |
| 65.10  | 1497.0    |
| 66.00  | 261.0     |
| 67.00  | 158.0     |
| 69.00  | 71784.0   |
| 70.00  | 293.0     |
| 73.10  | 848.0     |
| 74.10  | 6999.0    |
| 75.00  | 11527.0   |
| 76.10  | 4586.0    |
| 77.10  | 84288.0   |
| 78.10  | 5791.0    |
| 79.00  | 5519.0    |
| 80.00  | 4927.0    |
| 81.00  | 6262.0    |
| 82.00  | 1638.0    |
| 83.00  | 1631.0    |
| 84.00  | 254.0     |
| 85.00  | 1196.0    |
| 86.00  | 2006.0    |
| 87.00  | 871.0     |
| 88.00  | 339.0     |
| 91.00  | 1549.0    |
| 92.00  | 1499.0    |
| 93.00  | 9219.0    |
| 94.00  | 708.0     |
| 95.00  | 298.0     |
| 96.00  | 514.0     |
| 98.00  | 7021.0    |
| 99.00  | 5788.0    |
| 100.10 | 561.0     |
| 101.00 | 3722.0    |
| 102.10 | 392.0     |
| 103.00 | 1286.0    |
| 104.00 | 2808.0    |
| 105.00 | 2104.0    |
| 107.00 | 28984.0   |
| 108.00 | 4311.0    |
| 110.00 | 55352.0   |
| 111.00 | 9021.0    |
| 112.00 | 757.0     |
| 112.90 | 294.0     |
| 116.10 | 1891.0    |
| 117.00 | 21360.0   |
| 118.00 | 1392.0    |
| 118.90 | 298.0     |
| 120.00 | 571.0     |
| 121.10 | 249.0     |
| 122.00 | 1672.0    |
| 123.00 | 3222.0    |
| 124.00 | 1467.0    |
| 125.00 | 1398.0    |
| 127.00 | 109392.0  |
| 128.00 | 8787.0    |
| 129.00 | 43568.0   |
| 130.00 | 3478.0    |
| 131.00 | 851.0     |
| 131.90 | 368.0     |
| 132.50 | 184.0     |
| 133.00 | 172.0     |
| 134.00 | 1376.0    |
| 135.00 | 3419.0    |
| 136.00 | 1510.0    |
| 137.00 | 1794.0    |
| 138.00 | 633.0     |
| 138.90 | 202.0     |
| 140.00 | 722.0     |
| 141.00 | 6223.0    |
| 142.00 | 1753.0    |
| 143.00 | 1275.0    |
| 143.90 | 319.0     |
| 144.90 | 333.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

|        |          |
|--------|----------|
| 146.00 | 1165.0   |
| 147.00 | 3492.0   |
| 148.00 | 6370.0   |
| 149.00 | 1503.0   |
| 150.00 | 288.0    |
| 151.10 | 902.0    |
| 152.10 | 438.0    |
| 153.00 | 2015.0   |
| 154.00 | 1446.0   |
| 155.00 | 3679.0   |
| 156.00 | 4858.0   |
| 156.90 | 788.0    |
| 158.00 | 1134.0   |
| 158.90 | 831.0    |
| 160.00 | 2392.0   |
| 161.00 | 3000.0   |
| 162.00 | 898.0    |
| 163.10 | 202.0    |
| 163.90 | 483.0    |
| 165.00 | 2536.0   |
| 165.90 | 1703.0   |
| 167.00 | 11706.0  |
| 168.00 | 4852.0   |
| 169.00 | 847.0    |
| 170.00 | 391.0    |
| 171.00 | 542.0    |
| 172.00 | 1390.0   |
| 173.00 | 1564.0   |
| 174.00 | 2711.0   |
| 175.00 | 5414.0   |
| 176.00 | 1511.0   |
| 177.00 | 2409.0   |
| 178.00 | 860.0    |
| 179.00 | 9954.0   |
| 180.00 | 7434.0   |
| 181.00 | 3287.0   |
| 182.00 | 660.0    |
| 183.00 | 415.0    |
| 184.00 | 962.0    |
| 185.00 | 4891.0   |
| 186.00 | 38456.0  |
| 187.00 | 11243.0  |
| 188.00 | 1280.0   |
| 189.00 | 2450.0   |
| 189.90 | 309.0    |
| 191.00 | 1066.0   |
| 192.00 | 3422.0   |
| 193.00 | 4150.0   |
| 194.10 | 914.0    |
| 195.00 | 639.0    |
| 196.00 | 9512.0   |
| 197.90 | 297216.0 |
| 198.90 | 20552.0  |
| 200.00 | 1526.0   |
| 201.50 | 1610.0   |
| 202.90 | 2006.0   |
| 204.00 | 10387.0  |
| 205.00 | 18792.0  |
| 206.00 | 82616.0  |
| 207.00 | 11233.0  |
| 208.00 | 2860.0   |
| 209.00 | 719.0    |
| 211.00 | 3031.0   |
| 213.00 | 247.0    |
| 213.90 | 159.0    |
| 215.00 | 875.0    |
| 216.00 | 1853.0   |
| 216.90 | 21072.0  |
| 217.90 | 2839.0   |
| 218.90 | 268.0    |
| 221.00 | 15136.0  |
| 221.70 | 3351.0   |
| 223.00 | 5186.0   |
| 224.00 | 47776.0  |
| 225.00 | 11870.0  |
| 226.00 | 1835.0   |
| 227.00 | 18296.0  |
| 228.00 | 2689.0   |
| 229.00 | 4029.0   |
| 230.00 | 814.0    |
| 231.00 | 2196.0   |
| 231.90 | 278.0    |
| 232.90 | 393.0    |
| 234.00 | 1289.0   |
| 235.00 | 1352.0   |
| 235.90 | 508.0    |
| 237.00 | 1537.0   |
| 238.00 | 221.0    |
| 239.00 | 860.0    |
| 239.90 | 657.0    |
| 240.90 | 1198.0   |
| 242.00 | 2872.0   |
| 243.00 | 2582.0   |
| 244.00 | 38376.0  |
| 245.00 | 5554.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

|        |          |
|--------|----------|
| 245.90 | 7203.0   |
| 247.00 | 1260.0   |
| 248.00 | 329.0    |
| 248.90 | 1534.0   |
| 249.90 | 349.0    |
| 250.90 | 425.0    |
| 251.90 | 410.0    |
| 253.00 | 1199.0   |
| 255.00 | 195520.0 |
| 256.00 | 29688.0  |
| 257.00 | 2083.0   |
| 258.00 | 10728.0  |
| 259.00 | 1726.0   |
| 260.00 | 262.0    |
| 261.00 | 481.0    |
| 264.00 | 488.0    |
| 265.00 | 4176.0   |
| 266.00 | 676.0    |
| 269.90 | 302.0    |
| 270.90 | 535.0    |
| 273.00 | 6424.0   |
| 274.00 | 16496.0  |
| 275.00 | 92752.0  |
| 276.00 | 13164.0  |
| 276.90 | 6681.0   |
| 277.90 | 1214.0   |
| 279.00 | 301.0    |
| 281.10 | 201.0    |
| 281.90 | 217.0    |
| 283.00 | 862.0    |
| 283.90 | 485.0    |
| 285.00 | 1381.0   |
| 285.90 | 290.0    |
| 288.90 | 251.0    |
| 289.90 | 320.0    |
| 290.90 | 226.0    |
| 291.90 | 352.0    |
| 292.90 | 1476.0   |
| 293.90 | 586.0    |
| 294.90 | 628.0    |
| 295.90 | 24880.0  |
| 296.90 | 3221.0   |
| 297.90 | 298.0    |
| 301.00 | 379.0    |
| 302.00 | 437.0    |
| 303.00 | 2806.0   |
| 303.90 | 983.0    |
| 308.00 | 392.0    |
| 308.90 | 282.0    |
| 309.90 | 428.0    |
| 312.80 | 247.0    |
| 314.00 | 1158.0   |
| 314.90 | 2843.0   |
| 315.90 | 1759.0   |
| 316.90 | 326.0    |
| 320.90 | 743.0    |
| 322.00 | 587.0    |
| 323.00 | 8163.0   |
| 324.00 | 1733.0   |
| 325.80 | 287.0    |
| 326.90 | 1444.0   |
| 327.90 | 784.0    |
| 329.00 | 163.0    |
| 331.90 | 627.0    |
| 332.90 | 840.0    |
| 334.00 | 5432.0   |
| 334.90 | 1488.0   |
| 336.00 | 172.0    |
| 338.80 | 162.0    |
| 340.90 | 1143.0   |
| 342.00 | 356.0    |
| 345.90 | 1712.0   |
| 347.00 | 311.0    |
| 350.90 | 204.0    |
| 351.90 | 2409.0   |
| 353.00 | 1642.0   |
| 354.00 | 3098.0   |
| 354.90 | 565.0    |
| 359.00 | 275.0    |
| 364.90 | 11142.0  |
| 365.90 | 1560.0   |
| 367.00 | 190.0    |
| 370.90 | 550.0    |
| 372.00 | 4329.0   |
| 373.00 | 1274.0   |
| 382.90 | 1189.0   |
| 383.90 | 318.0    |
| 389.90 | 409.0    |
| 390.90 | 446.0    |
| 392.00 | 322.0    |
| 400.80 | 313.0    |
| 401.90 | 1658.0   |
| 402.90 | 2513.0   |
| 403.90 | 707.0    |
| 404.80 | 154.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11



|        |          |
|--------|----------|
| 421.00 | 1837.0   |
| 421.90 | 2123.0   |
| 422.90 | 14683.0  |
| 423.90 | 3086.0   |
| 425.00 | 289.0    |
| 439.10 | 216.0    |
| 440.90 | 43432.0  |
| 441.90 | 259072.0 |
| 442.90 | 51872.0  |
| 443.90 | 5281.0   |
| 444.90 | 348.0    |

**Instrument :**  
BNA\_G  
**ClientSampleId :**  
DFTPP11

| m/z    | Abundance |
|--------|-----------|
| 38.10  | 539.0     |
| 39.10  | 2545.0    |
| 40.00  | 247.0     |
| 41.00  | 203.0     |
| 44.00  | 791.0     |
| 49.00  | 516.0     |
| 50.10  | 10029.0   |
| 51.10  | 37680.0   |
| 52.00  | 2069.0    |
| 55.00  | 165.0     |
| 56.10  | 1130.0    |
| 57.00  | 2853.0    |
| 61.00  | 481.0     |
| 62.10  | 603.0     |
| 63.10  | 1828.0    |
| 64.10  | 265.0     |
| 65.00  | 1125.0    |
| 69.00  | 43568.0   |
| 69.90  | 235.0     |
| 73.00  | 511.0     |
| 74.10  | 4878.0    |
| 75.00  | 6563.0    |
| 77.10  | 50496.0   |
| 78.10  | 3772.0    |
| 79.00  | 3399.0    |
| 80.00  | 2343.0    |
| 81.00  | 3731.0    |
| 82.00  | 1111.0    |
| 83.00  | 768.0     |
| 85.00  | 625.0     |
| 86.00  | 1382.0    |
| 87.00  | 442.0     |
| 87.90  | 211.0     |
| 89.00  | 174.0     |
| 91.00  | 865.0     |
| 92.00  | 1153.0    |
| 93.00  | 6089.0    |
| 94.00  | 376.0     |
| 96.00  | 285.0     |
| 98.00  | 4660.0    |
| 99.00  | 3488.0    |
| 100.00 | 417.0     |
| 101.00 | 2634.0    |
| 103.00 | 774.0     |
| 104.00 | 1742.0    |
| 105.00 | 1437.0    |
| 106.10 | 486.0     |
| 107.00 | 18424.0   |
| 108.00 | 2846.0    |
| 110.00 | 35040.0   |
| 111.00 | 5157.0    |
| 112.00 | 671.0     |
| 113.10 | 171.0     |
| 116.00 | 1206.0    |
| 117.00 | 13740.0   |
| 118.00 | 760.0     |
| 119.10 | 164.0     |
| 119.90 | 331.0     |
| 122.00 | 1398.0    |
| 123.00 | 2007.0    |
| 124.00 | 931.0     |
| 125.00 | 988.0     |
| 127.00 | 71808.0   |
| 128.00 | 5555.0    |
| 129.00 | 27848.0   |
| 130.00 | 2457.0    |
| 131.00 | 502.0     |
| 131.90 | 475.0     |
| 133.00 | 259.0     |
| 134.00 | 706.0     |
| 135.00 | 2484.0    |
| 136.00 | 979.0     |
| 137.00 | 917.0     |
| 137.90 | 301.0     |
| 140.00 | 383.0     |
| 141.00 | 3835.0    |
| 141.90 | 1416.0    |
| 142.90 | 932.0     |
| 144.00 | 282.0     |
| 144.90 | 353.0     |
| 145.90 | 649.0     |
| 147.00 | 2175.0    |
| 148.00 | 4283.0    |
| 149.10 | 955.0     |
| 149.90 | 292.0     |
| 151.00 | 544.0     |
| 151.60 | 458.0     |
| 153.00 | 1436.0    |
| 154.00 | 917.0     |
| 155.00 | 2354.0    |
| 156.00 | 3498.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

|        |          |
|--------|----------|
| 157.00 | 588.0    |
| 158.00 | 762.0    |
| 159.00 | 668.0    |
| 160.00 | 1567.0   |
| 161.00 | 1965.0   |
| 162.00 | 587.0    |
| 163.00 | 203.0    |
| 164.00 | 318.0    |
| 165.00 | 1838.0   |
| 166.00 | 1169.0   |
| 167.00 | 7729.0   |
| 168.00 | 3307.0   |
| 169.00 | 605.0    |
| 170.00 | 280.0    |
| 171.00 | 315.0    |
| 171.90 | 824.0    |
| 173.00 | 1087.0   |
| 174.00 | 1968.0   |
| 175.00 | 3836.0   |
| 175.90 | 1177.0   |
| 177.00 | 1737.0   |
| 178.10 | 514.0    |
| 179.00 | 7137.0   |
| 180.00 | 5032.0   |
| 181.00 | 1884.0   |
| 182.00 | 424.0    |
| 182.80 | 360.0    |
| 184.00 | 562.0    |
| 185.00 | 3433.0   |
| 186.00 | 27472.0  |
| 187.00 | 7663.0   |
| 188.00 | 786.0    |
| 188.90 | 1834.0   |
| 190.00 | 386.0    |
| 191.00 | 817.0    |
| 192.00 | 2008.0   |
| 193.00 | 2704.0   |
| 194.00 | 503.0    |
| 196.00 | 6683.0   |
| 197.90 | 211200.0 |
| 198.90 | 14217.0  |
| 200.00 | 1200.0   |
| 201.30 | 1090.0   |
| 202.90 | 1483.0   |
| 204.00 | 8059.0   |
| 205.00 | 13000.0  |
| 206.00 | 56984.0  |
| 207.00 | 7325.0   |
| 208.00 | 1563.0   |
| 208.90 | 507.0    |
| 210.20 | 1004.0   |
| 210.90 | 1928.0   |
| 212.80 | 185.0    |
| 214.90 | 635.0    |
| 216.00 | 1373.0   |
| 216.90 | 13930.0  |
| 217.90 | 1913.0   |
| 221.00 | 10531.0  |
| 221.80 | 2130.0   |
| 223.00 | 3468.0   |
| 224.00 | 32272.0  |
| 225.00 | 8935.0   |
| 226.00 | 957.0    |
| 227.00 | 12822.0  |
| 227.90 | 1834.0   |
| 229.00 | 2849.0   |
| 230.10 | 376.0    |
| 231.00 | 1121.0   |
| 232.90 | 380.0    |
| 233.90 | 1074.0   |
| 235.00 | 1097.0   |
| 235.90 | 610.0    |
| 237.00 | 1049.0   |
| 239.00 | 801.0    |
| 239.90 | 468.0    |
| 241.00 | 796.0    |
| 242.00 | 1581.0   |
| 243.00 | 2218.0   |
| 244.00 | 27024.0  |
| 245.00 | 3949.0   |
| 246.00 | 5080.0   |
| 246.90 | 957.0    |
| 247.90 | 335.0    |
| 248.90 | 970.0    |
| 250.00 | 253.0    |
| 250.90 | 253.0    |
| 252.00 | 429.0    |
| 253.00 | 706.0    |
| 255.00 | 139328.0 |
| 256.00 | 20728.0  |
| 257.00 | 1715.0   |
| 257.90 | 7482.0   |
| 259.00 | 1210.0   |
| 259.90 | 269.0    |
| 260.80 | 325.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

|        |          |
|--------|----------|
| 264.00 | 284.0    |
| 264.90 | 3171.0   |
| 266.00 | 444.0    |
| 269.90 | 151.0    |
| 270.90 | 350.0    |
| 271.90 | 529.0    |
| 272.90 | 4514.0   |
| 274.00 | 10885.0  |
| 275.00 | 67136.0  |
| 276.00 | 9386.0   |
| 276.90 | 5159.0   |
| 277.90 | 845.0    |
| 278.80 | 178.0    |
| 282.00 | 227.0    |
| 282.90 | 717.0    |
| 284.00 | 592.0    |
| 285.00 | 845.0    |
| 286.00 | 182.0    |
| 288.90 | 235.0    |
| 289.90 | 238.0    |
| 290.90 | 222.0    |
| 291.90 | 268.0    |
| 292.90 | 1011.0   |
| 293.90 | 447.0    |
| 295.90 | 17880.0  |
| 296.90 | 1965.0   |
| 298.00 | 157.0    |
| 300.90 | 206.0    |
| 302.00 | 416.0    |
| 303.00 | 2122.0   |
| 303.90 | 709.0    |
| 307.90 | 365.0    |
| 308.90 | 176.0    |
| 310.00 | 235.0    |
| 313.00 | 187.0    |
| 313.90 | 700.0    |
| 314.90 | 2129.0   |
| 315.90 | 1107.0   |
| 316.90 | 205.0    |
| 320.90 | 764.0    |
| 321.90 | 406.0    |
| 323.00 | 6973.0   |
| 323.90 | 1354.0   |
| 324.90 | 174.0    |
| 326.90 | 1040.0   |
| 327.90 | 523.0    |
| 331.90 | 392.0    |
| 332.90 | 731.0    |
| 334.00 | 4253.0   |
| 334.90 | 1155.0   |
| 340.90 | 901.0    |
| 341.90 | 193.0    |
| 345.90 | 1337.0   |
| 346.90 | 241.0    |
| 351.90 | 2163.0   |
| 353.00 | 1246.0   |
| 354.00 | 2059.0   |
| 354.90 | 299.0    |
| 364.90 | 8969.0   |
| 365.90 | 1182.0   |
| 369.70 | 151.0    |
| 371.00 | 690.0    |
| 371.90 | 3200.0   |
| 372.90 | 909.0    |
| 382.90 | 1156.0   |
| 383.90 | 293.0    |
| 389.90 | 411.0    |
| 391.00 | 301.0    |
| 391.90 | 311.0    |
| 400.90 | 467.0    |
| 401.90 | 1507.0   |
| 402.90 | 1862.0   |
| 403.90 | 751.0    |
| 404.90 | 164.0    |
| 420.90 | 1637.0   |
| 421.90 | 1556.0   |
| 422.90 | 11660.0  |
| 423.90 | 2766.0   |
| 424.90 | 335.0    |
| 436.50 | 200.0    |
| 439.50 | 195.0    |
| 440.90 | 35728.0  |
| 441.90 | 219904.0 |
| 442.90 | 43272.0  |
| 443.90 | 4155.0   |
| 445.00 | 240.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11

Scan 2508 (17.823 min): BG042253.D  
DFTPP11

| m/z    | Abundance |
|--------|-----------|
| 44.10  | 770.0     |
| 49.00  | 296.0     |
| 83.90  | 289.0     |
| 207.00 | 584.0     |
| 208.00 | 153.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP11