


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081619\
 Data File : BG042253.D
 Acq On : 16 Aug 2019 10:48
 Operator : HP/JU
 Sample : DFTPP11
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG081619MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 16 15:15:09 2019



AutoFind: Scans 2515, 2516, 2517; Background Corrected with Scan 2508

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.6	55784	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	24.1	62037	PASS
70	69	0.00	2	0.6	342	PASS
127	198	10	80	36.9	95024	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	257664	PASS
199	198	5	9	7.0	17917	PASS
275	198	10	60	30.7	79090	PASS
365	198	1	100	3.7	9548	PASS
441	198	0.01	100	14.0	35986	PASS
442	198	50	100	86.1	221973	PASS
443	442	15	24	20.2	44776	PASS

m/z	Abundance
37.10	185.0
38.10	580.0
39.10	4418.0
40.00	365.0
43.20	200.0
44.00	749.0
49.00	708.0
50.10	16054.0
51.10	65072.0
52.10	3320.0
53.00	237.0
55.00	307.0
56.00	2254.0
57.00	4839.0
58.00	269.0
61.00	930.0
62.00	1101.0
63.00	2719.0
64.10	550.0
65.10	1474.0
67.10	166.0
69.00	70760.0
70.10	500.0
73.10	552.0
74.00	6682.0
75.00	11202.0
76.10	4521.0
77.10	81136.0
78.10	5604.0
79.00	5588.0
80.00	4535.0
81.00	5914.0
82.00	1676.0
83.00	1351.0
83.90	301.0
85.00	1059.0
86.00	2014.0
87.00	843.0
88.00	440.0
89.00	169.0
91.00	1513.0
92.00	1650.0
93.00	9327.0
94.10	584.0
96.10	399.0
98.00	7018.0
99.00	5145.0
100.00	559.0
101.00	3127.0
101.90	167.0
103.00	1039.0
104.00	2749.0
105.00	2494.0
107.00	27440.0
108.00	4598.0
110.00	50832.0
111.00	7898.0
112.00	1050.0
113.00	371.0
115.00	203.0
116.00	1687.0
117.00	19280.0
118.00	1530.0
119.00	271.0
120.00	388.0
121.00	156.0
122.00	1751.0
123.00	3209.0
123.90	1295.0
125.00	1367.0
127.00	103872.0
128.00	8670.0
129.00	40504.0
130.00	3231.0
131.00	639.0
131.90	360.0
133.00	268.0
133.90	1117.0
135.00	3363.0
136.00	1284.0
137.00	1698.0
137.90	450.0
139.00	289.0
140.00	498.0
141.00	5728.0
142.00	1644.0
143.00	1097.0
144.10	376.0
145.00	417.0
145.90	960.0
147.00	3106.0

Instrument :
BNA_G
ClientSampleId :

148.00	5668.0
149.00	1110.0
150.10	367.0
151.00	927.0
151.90	506.0
153.00	1751.0
154.00	1547.0
155.00	3433.0
156.00	5292.0
157.00	935.0
157.90	1231.0
159.00	862.0
160.00	1907.0
161.00	2730.0
162.00	921.0
163.10	232.0
164.00	316.0
165.00	2311.0
166.00	1950.0
167.00	11025.0
168.00	4338.0
169.00	946.0
170.00	512.0
170.90	517.0
171.90	1139.0
173.00	1344.0
174.00	2517.0
175.00	5035.0
176.00	1190.0
177.00	1787.0
178.00	718.0
179.00	8968.0
180.00	6511.0
181.00	3090.0
182.10	447.0
183.10	295.0
184.00	775.0
185.00	4830.0
186.00	34272.0
187.00	9845.0
188.00	1172.0
189.00	2233.0
189.90	495.0
191.00	1075.0
192.00	2862.0
193.00	3407.0
194.00	865.0
196.00	7550.0
197.90	264576.0
198.90	18984.0
200.00	1296.0
201.50	1313.0
203.00	1926.0
204.00	9618.0
205.00	16576.0
206.00	70848.0
207.00	10814.0
208.00	2412.0
209.00	653.0
210.20	1122.0
211.00	2659.0
213.00	277.0
215.00	600.0
216.00	1697.0
216.90	17560.0
218.00	2509.0
219.00	284.0
221.00	12098.0
221.70	2953.0
223.00	4315.0
224.00	40568.0
225.00	10780.0
226.10	1387.0
227.00	15003.0
228.00	2503.0
229.00	3573.0
230.00	462.0
231.00	1567.0
231.90	245.0
233.00	290.0
234.00	1132.0
235.00	1332.0
235.90	808.0
237.00	1323.0
238.00	155.0
238.90	503.0
240.00	738.0
241.00	817.0
242.00	1983.0
243.00	2604.0
244.00	32200.0
245.00	4297.0
246.00	5674.0
246.90	983.0
248.00	396.0

Instrument :
BNA_G
ClientSampleId :

248.90	964.0
249.90	293.0
251.00	219.0
252.00	328.0
253.00	921.0
255.00	158208.0
256.00	23704.0
257.00	2002.0
258.00	8409.0
259.00	1534.0
259.90	282.0
260.80	263.0
264.00	315.0
264.90	3603.0
265.90	421.0
270.00	194.0
271.00	327.0
272.90	5194.0
274.00	13133.0
275.00	77384.0
276.00	10567.0
277.00	5900.0
277.90	928.0
278.90	276.0
280.80	221.0
282.00	248.0
283.00	799.0
284.00	554.0
285.00	1106.0
286.00	219.0
288.90	223.0
290.00	198.0
291.00	263.0
291.90	331.0
293.00	1387.0
294.00	505.0
295.90	19144.0
296.90	2952.0
297.90	279.0
301.00	305.0
302.00	433.0
303.00	2304.0
303.90	759.0
304.90	210.0
308.10	259.0
309.00	214.0
309.90	294.0
313.10	219.0
314.00	845.0
314.90	2043.0
316.00	1029.0
317.00	151.0
320.00	155.0
321.00	539.0
321.90	503.0
323.00	6490.0
324.00	1213.0
325.00	233.0
325.90	177.0
326.90	1284.0
327.90	622.0
331.90	708.0
332.90	474.0
334.00	4111.0
334.90	1051.0
338.80	208.0
341.00	786.0
341.90	257.0
345.90	1348.0
346.80	352.0
352.00	2040.0
352.90	1444.0
354.00	2370.0
354.90	373.0
358.80	219.0
364.90	8533.0
365.90	1457.0
370.00	268.0
370.90	382.0
372.00	3085.0
372.90	801.0
382.90	724.0
383.90	289.0
389.90	522.0
390.90	271.0
391.90	151.0
400.90	271.0
402.00	1402.0
402.90	1858.0
403.90	775.0
404.80	184.0
421.00	1397.0
421.90	1437.0
422.90	10411.0
423.90	2456.0

Instrument :
BNA_G
ClientSampleId :

424.80	220.0
440.90	28800.0
441.90	186944.0
442.90	39184.0
443.90	3458.0
444.90	276.0

Instrument :
BNA_G
ClientSampleId :

m/z	Abundance
37.00	193.0
38.10	657.0
39.10	3929.0
40.10	348.0
41.10	238.0
42.90	151.0
44.00	861.0
49.10	735.0
50.10	17032.0
51.10	64600.0
52.10	3597.0
53.00	203.0
55.00	409.0
56.00	2163.0
57.00	4688.0
58.00	338.0
61.00	840.0
62.10	1103.0
63.00	3054.0
64.00	403.0
65.10	1497.0
66.00	261.0
67.00	158.0
69.00	71784.0
70.00	293.0
73.10	848.0
74.10	6999.0
75.00	11527.0
76.10	4586.0
77.10	84288.0
78.10	5791.0
79.00	5519.0
80.00	4927.0
81.00	6262.0
82.00	1638.0
83.00	1631.0
84.00	254.0
85.00	1196.0
86.00	2006.0
87.00	871.0
88.00	339.0
91.00	1549.0
92.00	1499.0
93.00	9219.0
94.00	708.0
95.00	298.0
96.00	514.0
98.00	7021.0
99.00	5788.0
100.10	561.0
101.00	3722.0
102.10	392.0
103.00	1286.0
104.00	2808.0
105.00	2104.0
107.00	28984.0
108.00	4311.0
110.00	55352.0
111.00	9021.0
112.00	757.0
112.90	294.0
116.10	1891.0
117.00	21360.0
118.00	1392.0
118.90	298.0
120.00	571.0
121.10	249.0
122.00	1672.0
123.00	3222.0
124.00	1467.0
125.00	1398.0
127.00	109392.0
128.00	8787.0
129.00	43568.0
130.00	3478.0
131.00	851.0
131.90	368.0
132.50	184.0
133.00	172.0
134.00	1376.0
135.00	3419.0
136.00	1510.0
137.00	1794.0
138.00	633.0
138.90	202.0
140.00	722.0
141.00	6223.0
142.00	1753.0
143.00	1275.0
143.90	319.0
144.90	333.0

Instrument :
BNA_G
ClientSampleId :

146.00	1165.0
147.00	3492.0
148.00	6370.0
149.00	1503.0
150.00	288.0
151.10	902.0
152.10	438.0
153.00	2015.0
154.00	1446.0
155.00	3679.0
156.00	4858.0
156.90	788.0
158.00	1134.0
158.90	831.0
160.00	2392.0
161.00	3000.0
162.00	898.0
163.10	202.0
163.90	483.0
165.00	2536.0
165.90	1703.0
167.00	11706.0
168.00	4852.0
169.00	847.0
170.00	391.0
171.00	542.0
172.00	1390.0
173.00	1564.0
174.00	2711.0
175.00	5414.0
176.00	1511.0
177.00	2409.0
178.00	860.0
179.00	9954.0
180.00	7434.0
181.00	3287.0
182.00	660.0
183.00	415.0
184.00	962.0
185.00	4891.0
186.00	38456.0
187.00	11243.0
188.00	1280.0
189.00	2450.0
189.90	309.0
191.00	1066.0
192.00	3422.0
193.00	4150.0
194.10	914.0
195.00	639.0
196.00	9512.0
197.90	297216.0
198.90	20552.0
200.00	1526.0
201.50	1610.0
202.90	2006.0
204.00	10387.0
205.00	18792.0
206.00	82616.0
207.00	11233.0
208.00	2860.0
209.00	719.0
211.00	3031.0
213.00	247.0
213.90	159.0
215.00	875.0
216.00	1853.0
216.90	21072.0
217.90	2839.0
218.90	268.0
221.00	15136.0
221.70	3351.0
223.00	5186.0
224.00	47776.0
225.00	11870.0
226.00	1835.0
227.00	18296.0
228.00	2689.0
229.00	4029.0
230.00	814.0
231.00	2196.0
231.90	278.0
232.90	393.0
234.00	1289.0
235.00	1352.0
235.90	508.0
237.00	1537.0
238.00	221.0
239.00	860.0
239.90	657.0
240.90	1198.0
242.00	2872.0
243.00	2582.0
244.00	38376.0
245.00	5554.0

Instrument :
BNA_G
ClientSampleId :

245.90	7203.0
247.00	1260.0
248.00	329.0
248.90	1534.0
249.90	349.0
250.90	425.0
251.90	410.0
253.00	1199.0
255.00	195520.0
256.00	29688.0
257.00	2083.0
258.00	10728.0
259.00	1726.0
260.00	262.0
261.00	481.0
264.00	488.0
265.00	4176.0
266.00	676.0
269.90	302.0
270.90	535.0
273.00	6424.0
274.00	16496.0
275.00	92752.0
276.00	13164.0
276.90	6681.0
277.90	1214.0
279.00	301.0
281.10	201.0
281.90	217.0
283.00	862.0
283.90	485.0
285.00	1381.0
285.90	290.0
288.90	251.0
289.90	320.0
290.90	226.0
291.90	352.0
292.90	1476.0
293.90	586.0
294.90	628.0
295.90	24880.0
296.90	3221.0
297.90	298.0
301.00	379.0
302.00	437.0
303.00	2806.0
303.90	983.0
308.00	392.0
308.90	282.0
309.90	428.0
312.80	247.0
314.00	1158.0
314.90	2843.0
315.90	1759.0
316.90	326.0
320.90	743.0
322.00	587.0
323.00	8163.0
324.00	1733.0
325.80	287.0
326.90	1444.0
327.90	784.0
329.00	163.0
331.90	627.0
332.90	840.0
334.00	5432.0
334.90	1488.0
336.00	172.0
338.80	162.0
340.90	1143.0
342.00	356.0
345.90	1712.0
347.00	311.0
350.90	204.0
351.90	2409.0
353.00	1642.0
354.00	3098.0
354.90	565.0
359.00	275.0
364.90	11142.0
365.90	1560.0
367.00	190.0
370.90	550.0
372.00	4329.0
373.00	1274.0
382.90	1189.0
383.90	318.0
389.90	409.0
390.90	446.0
392.00	322.0
400.80	313.0
401.90	1658.0
402.90	2513.0
403.90	707.0
404.80	154.0

Instrument :
BNA_G
ClientSampleId :

421.00	1837.0
421.90	2123.0
422.90	14683.0
423.90	3086.0
425.00	289.0
439.10	216.0
440.90	43432.0
441.90	259072.0
442.90	51872.0
443.90	5281.0
444.90	348.0

Instrument :
BNA_G
ClientSampleId :

m/z	Abundance
38.10	539.0
39.10	2545.0
40.00	247.0
41.00	203.0
44.00	791.0
49.00	516.0
50.10	10029.0
51.10	37680.0
52.00	2069.0
55.00	165.0
56.10	1130.0
57.00	2853.0
61.00	481.0
62.10	603.0
63.10	1828.0
64.10	265.0
65.00	1125.0
69.00	43568.0
69.90	235.0
73.00	511.0
74.10	4878.0
75.00	6563.0
77.10	50496.0
78.10	3772.0
79.00	3399.0
80.00	2343.0
81.00	3731.0
82.00	1111.0
83.00	768.0
85.00	625.0
86.00	1382.0
87.00	442.0
87.90	211.0
89.00	174.0
91.00	865.0
92.00	1153.0
93.00	6089.0
94.00	376.0
96.00	285.0
98.00	4660.0
99.00	3488.0
100.00	417.0
101.00	2634.0
103.00	774.0
104.00	1742.0
105.00	1437.0
106.10	486.0
107.00	18424.0
108.00	2846.0
110.00	35040.0
111.00	5157.0
112.00	671.0
113.10	171.0
116.00	1206.0
117.00	13740.0
118.00	760.0
119.10	164.0
119.90	331.0
122.00	1398.0
123.00	2007.0
124.00	931.0
125.00	988.0
127.00	71808.0
128.00	5555.0
129.00	27848.0
130.00	2457.0
131.00	502.0
131.90	475.0
133.00	259.0
134.00	706.0
135.00	2484.0
136.00	979.0
137.00	917.0
137.90	301.0
140.00	383.0
141.00	3835.0
141.90	1416.0
142.90	932.0
144.00	282.0
144.90	353.0
145.90	649.0
147.00	2175.0
148.00	4283.0
149.10	955.0
149.90	292.0
151.00	544.0
151.60	458.0
153.00	1436.0
154.00	917.0
155.00	2354.0
156.00	3498.0

Instrument :
BNA_G
ClientSampleId :

157.00	588.0
158.00	762.0
159.00	668.0
160.00	1567.0
161.00	1965.0
162.00	587.0
163.00	203.0
164.00	318.0
165.00	1838.0
166.00	1169.0
167.00	7729.0
168.00	3307.0
169.00	605.0
170.00	280.0
171.00	315.0
171.90	824.0
173.00	1087.0
174.00	1968.0
175.00	3836.0
175.90	1177.0
177.00	1737.0
178.10	514.0
179.00	7137.0
180.00	5032.0
181.00	1884.0
182.00	424.0
182.80	360.0
184.00	562.0
185.00	3433.0
186.00	27472.0
187.00	7663.0
188.00	786.0
188.90	1834.0
190.00	386.0
191.00	817.0
192.00	2008.0
193.00	2704.0
194.00	503.0
196.00	6683.0
197.90	211200.0
198.90	14217.0
200.00	1200.0
201.30	1090.0
202.90	1483.0
204.00	8059.0
205.00	13000.0
206.00	56984.0
207.00	7325.0
208.00	1563.0
208.90	507.0
210.20	1004.0
210.90	1928.0
212.80	185.0
214.90	635.0
216.00	1373.0
216.90	13930.0
217.90	1913.0
221.00	10531.0
221.80	2130.0
223.00	3468.0
224.00	32272.0
225.00	8935.0
226.00	957.0
227.00	12822.0
227.90	1834.0
229.00	2849.0
230.10	376.0
231.00	1121.0
232.90	380.0
233.90	1074.0
235.00	1097.0
235.90	610.0
237.00	1049.0
239.00	801.0
239.90	468.0
241.00	796.0
242.00	1581.0
243.00	2218.0
244.00	27024.0
245.00	3949.0
246.00	5080.0
246.90	957.0
247.90	335.0
248.90	970.0
250.00	253.0
250.90	253.0
252.00	429.0
253.00	706.0
255.00	139328.0
256.00	20728.0
257.00	1715.0
257.90	7482.0
259.00	1210.0
259.90	269.0
260.80	325.0

Instrument :
BNA_G
ClientSampleId :

264.00	284.0
264.90	3171.0
266.00	444.0
269.90	151.0
270.90	350.0
271.90	529.0
272.90	4514.0
274.00	10885.0
275.00	67136.0
276.00	9386.0
276.90	5159.0
277.90	845.0
278.80	178.0
282.00	227.0
282.90	717.0
284.00	592.0
285.00	845.0
286.00	182.0
288.90	235.0
289.90	238.0
290.90	222.0
291.90	268.0
292.90	1011.0
293.90	447.0
295.90	17880.0
296.90	1965.0
298.00	157.0
300.90	206.0
302.00	416.0
303.00	2122.0
303.90	709.0
307.90	365.0
308.90	176.0
310.00	235.0
313.00	187.0
313.90	700.0
314.90	2129.0
315.90	1107.0
316.90	205.0
320.90	764.0
321.90	406.0
323.00	6973.0
323.90	1354.0
324.90	174.0
326.90	1040.0
327.90	523.0
331.90	392.0
332.90	731.0
334.00	4253.0
334.90	1155.0
340.90	901.0
341.90	193.0
345.90	1337.0
346.90	241.0
351.90	2163.0
353.00	1246.0
354.00	2059.0
354.90	299.0
364.90	8969.0
365.90	1182.0
369.70	151.0
371.00	690.0
371.90	3200.0
372.90	909.0
382.90	1156.0
383.90	293.0
389.90	411.0
391.00	301.0
391.90	311.0
400.90	467.0
401.90	1507.0
402.90	1862.0
403.90	751.0
404.90	164.0
420.90	1637.0
421.90	1556.0
422.90	11660.0
423.90	2766.0
424.90	335.0
436.50	200.0
439.50	195.0
440.90	35728.0
441.90	219904.0
442.90	43272.0
443.90	4155.0
445.00	240.0

Instrument :
BNA_G
ClientSampleId :

Scan 2508 (17.823 min): BG042253.D
DFTPP11

m/z	Abundance
44.10	770.0
49.00	296.0
83.90	289.0
207.00	584.0
208.00	153.0

Instrument :
BNA_G
ClientSampleId :