




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081723\
Data File : BG058753.D
Acq On : 17 Aug 2023 7:41
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 09:22:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration



TIC: BG058753.D\data.ms

(70) Pentachlorophenol (C)




16.861min (+ 0.002) 8720.45 ng

response 91859

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.40	62.80
264.00	62.80	62.24
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081723\
Data File : BG058753.D
Acq On : 17 Aug 2023 7:41
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 09:22:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 22:29:07 2023
Response via : Initial Calibration



TIC: BG058753.D\data.ms

(77) Benzidine

19.452min (+ 0.001) 27495.39 ng

response 583301

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.42
183.00	10.90	11.98
0.00	0.00	0.00