

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081823\
 Data File : BG058779.D
 Acq On : 18 Aug 2023 16:49
 Operator : MA/JU
 Sample : PB154469BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB154469BL

Quant Time: Aug 21 02:12:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG081823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 01:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.812	152	24453	20.000	ng	0.00
21) Naphthalene-d8	10.614	136	100199	20.000	ng	0.00
39) Acenaphthene-d10	14.463	164	70588	20.000	ng	0.00
64) Phenanthrene-d10	17.206	188	183259	20.000	ng	0.00
76) Chrysene-d12	21.449	240	185891	20.000	ng	0.00
86) Perylene-d12	24.533	264	220147	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.385	112	217073	142.856	ng	0.00
7) Phenol-d6	6.983	99	279715	131.204	ng	0.00
23) Nitrobenzene-d5	8.975	82	176315	84.600	ng	0.00
42) 2,4,6-Tribromophenol	15.949	330	138963	125.553	ng	0.00
45) 2-Fluorobiphenyl	13.076	172	408683	84.209	ng	0.00
79) Terphenyl-d14	19.827	244	887226	85.700	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081823\
Data File : BG058779.D
Acq On : 18 Aug 2023 16:49
Operator : MA/JU
Sample : PB154469BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB154469BL

Quant Time: Aug 21 02:12:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG081823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 01:40:57 2023
Response via : Initial Calibration

