


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG082019\
 Data File : BG042361.D
 Acq On : 20 Aug 2019 16:28
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG081919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 19 13:40:37 2019



AutoFind: Scans 2512, 2513, 2514; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.0	22222	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	22.8	24053	PASS
70	69	0.00	2	0.7	166	PASS
127	198	10	80	35.9	37960	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	105634	PASS
199	198	5	9	6.9	7311	PASS
275	198	10	60	29.9	31608	PASS
365	198	1	100	3.5	3653	PASS
441	198	0.01	100	12.2	12889	PASS
442	198	50	100	81.7	86272	PASS
443	442	15	24	19.4	16701	PASS