

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082123\
 Data File : BG058788.D
 Acq On : 21 Aug 2023 11:01
 Operator : MA/JU
 Sample : 04086-16
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 B-PP18B

Quant Time: Aug 22 01:24:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG081823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 01:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.809	152	30147	20.000	ng	0.00
21) Naphthalene-d8	10.611	136	129135	20.000	ng	0.00
39) Acenaphthene-d10	14.460	164	92409	20.000	ng	0.00
64) Phenanthrene-d10	17.203	188	230421	20.000	ng	0.00
76) Chrysene-d12	21.446	240	209886	20.000	ng	0.00
86) Perylene-d12	24.518	264	251641	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.382	112	138311	73.831	ng	0.00
7) Phenol-d6	6.980	99	186184	70.837	ng	0.00
23) Nitrobenzene-d5	8.972	82	114889	42.774	ng	0.00
42) 2,4,6-Tribromophenol	15.952	330	95988	66.246	ng	0.00
45) 2-Fluorobiphenyl	13.073	172	276756	43.560	ng	0.00
79) Terphenyl-d14	19.818	244	522365	44.689	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082123\
Data File : BG058788.D
Acq On : 21 Aug 2023 11:01
Operator : MA/JU
Sample : 04086-16
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
B-PP18B

Quant Time: Aug 22 01:24:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG081823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 01:40:57 2023
Response via : Initial Calibration

