

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082218\
 Data File : BG036333.D
 Acq On : 22 Aug 2018 8:48
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02020

Quant Time: Aug 22 10:22:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG082118MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 21 20:07:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.566	0.555	1.9	100	0.00
3 S	1,4-Dioxane-d8	0.537	0.493	8.2	86	0.00
4	Benzaldehyde	1.247	1.399	-12.2	96	0.00
5 S	Phenol-d5	1.861	1.906	-2.4	95	0.00
6	Phenol	1.860	1.895	-1.9	96	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.275	1.304	-2.3	95	0.00
8	Bis(2-Chloroethyl)ether	1.381	1.488	-7.7	100	0.00
9 S	2-Chlorophenol-d4	1.344	1.400	-4.2	98	0.00
10	2-Chlorophenol	1.344	1.378	-2.5	99	0.00
11	2-Methylphenol	1.417	1.463	-3.2	98	0.00
12	2,2'-oxybis(1-Chloropropane	3.040	3.098	-1.9	95	0.00
13 S	4-Methylphenol-d8	1.420	1.464	-3.1	98	0.00
14	Acetophenone	2.177	2.330	-7.0	99	0.00
15 P	N-Nitroso-di-n-propylamine	1.312	1.318	-0.5	95	0.00
16	4-Methylphenol	1.482	1.570	-5.9	99	0.00
17	Hexachloroethane	0.535	0.555	-3.7	98	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
19 S	Nitrobenzene-d5	0.122	0.133	-9.0	113	0.00
20	Nitrobenzene	0.346	0.377	-9.0	110	0.00
21	Isophorone	0.840	0.825	1.8	95	0.00
22 S	2-Nitrophenol-d4	0.123	0.140	-13.8	120	0.00
23 C	2-Nitrophenol	0.130	0.142	-9.2	110	0.00
24	2,4-Dimethylphenol	0.397	0.392	1.3	96	0.00
25	Bis(2-Chloroethoxy)methane	0.454	0.461	-1.5	97	0.00
26 S	2,4-Dichlorophenol-d3	0.329	0.327	0.6	96	0.00
27 C	2,4-Dichlorophenol	0.311	0.315	-1.3	96	0.00
28	Naphthalene	1.025	1.020	0.5	95	0.00
29 S	4-Chloroaniline-d4	0.363	0.356	1.9	90	0.00
30	4-Chloroaniline	0.354	0.351	0.8	91	0.00
31 C	Hexachlorobutadiene	0.222	0.225	-1.4	97	0.00
32	Caprolactam	0.129	0.132	-2.3	96	0.00
33 C	4-Chloro-3-methylphenol	0.366	0.367	-0.3	95	0.00
34	2-Methylnaphthalene	0.765	0.770	-0.7	97	0.00
35	1-Methylnaphthalene	0.746	0.754	-1.1	96	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
37	1,2,4,5-Tetrachlorobenzene	0.681	0.675	0.9	95	0.00
38	Hexachlorocyclopentadiene	0.408	0.408	0.0	101	0.00
39 C	2,4,6-Trichlorophenol	0.421	0.421	0.0	98	0.00
40	2,4,5-Trichlorophenol	0.436	0.450	-3.2	102	0.00
41	1,1'-Biphenyl	1.565	1.573	-0.5	98	0.00
42	2-Chloronaphthalene	1.201	1.221	-1.7	98	0.00
43	2-Nitroaniline	0.314	0.356	-13.4	120	0.00
44 S	Dimethylphthalate-d6	1.672	1.714	-2.5	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082218\
 Data File : BG036333.D
 Acq On : 22 Aug 2018 8:48
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02020

Quant Time: Aug 22 10:22:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG082118MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 21 20:07:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.605	1.664	-3.7	100	0.00
46	2,6-Dinitrotoluene	0.224	0.267	-19.2	123	0.00
47 S	Acenaphthylene-d8	2.016	2.053	-1.8	99	0.00
48	Acenaphthylene	1.901	1.929	-1.5	97	0.00
49	3-Nitroaniline	0.234	0.276	-17.9	117	0.00
50 C	Acenaphthene	1.361	1.368	-0.5	97	0.00
51	2,4-Dinitrophenol	0.114	0.116	-1.8	121	0.00
52 S	4-Nitrophenol-d4	0.247	0.285	-15.4	118	0.00
53	4-Nitrophenol	0.230	0.254	-10.4	114	0.00
54	Dibenzofuran	1.884	1.903	-1.0	98	0.00
55	2,4-Dinitrotoluene	0.310	0.382	-23.2#	127	0.00
56	2,3,4,6-Tetrachlorophenol	0.408	0.420	-2.9	102	0.00
57	Diethylphthalate	1.629	1.693	-3.9	100	0.00
58 S	Fluorene-d10	1.454	1.482	-1.9	98	0.00
59	Fluorene	1.519	1.559	-2.6	98	0.00
60	4-Chlorophenyl-phenylether	0.827	0.839	-1.5	97	0.00
61	4-Nitroaniline	0.296	0.353	-19.3	119	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.072	0.071	1.4	125	0.00
64	4,6-Dinitro-2-methylphenol	0.075	0.078	-4.0	125	0.00
65	N-Nitrosodiphenylamine	0.560	0.549	2.0	98	0.00
66	4-Bromophenyl-phenylether	0.221	0.214	3.2	97	0.00
67	Hexachlorobenzene	0.221	0.216	2.3	99	0.00
68	Atrazine	0.231	0.244	-5.6	102	0.00
69 C	Pentachlorophenol	0.134	0.136	-1.5	104	0.00
70	Phenanthrene	1.039	1.076	-3.6	102	0.00
71 S	Anthracene-d10	0.933	0.957	-2.6	102	0.00
72	Anthracene	1.046	1.077	-3.0	103	0.00
73	1,2,3,4-Tetrachlorobenzene	0.306	0.289	5.6	96	0.00
74	Pentachlorobenzene	0.266	0.261	1.9	98	0.00
75	Carbazole	0.879	0.986	-12.2	104	0.00
76	Di-n-butylphthalate	1.149	1.226	-6.7	105	0.00
77 C	Fluoranthene	1.205	1.376	-14.2	104	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
79 S	Pyrene-d10	1.025	1.052	-2.6	105	0.00
80	Pyrene	1.212	1.234	-1.8	104	0.00
81	Butylbenzylphthalate	0.493	0.517	-4.9	109	0.00
82	3,3'-Dichlorobenzidine	0.412	0.422	-2.4	99	0.00
83	Benzo(a)anthracene	1.264	1.270	-0.5	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.703	0.721	-2.6	104	0.00
85	Chrysene	1.159	1.170	-0.9	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Di-n-octyl phthalate	1.124	1.194	-6.2	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082218\
Data File : BG036333.D
Acq On : 22 Aug 2018 8:48
Operator : JU/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02020

Quant Time: Aug 22 10:22:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG082118MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 21 20:07:00 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.184	1.198	-1.2	104	0.00
89	Benzo(k)fluoranthene	1.142	1.132	0.9	102	0.00
90 S	Benzo(a)pyrene-d12	1.085	1.084	0.1	102	0.00
91 C	Benzo(a)pyrene	1.121	1.124	-0.3	102	0.00
92	Indeno(1,2,3-cd)pyrene	1.334	1.351	-1.3	104	0.00
93	Dibenzo(a,h)anthracene	1.082	1.081	0.1	103	0.00
94	Benzo(g,h,i)perylene	1.100	1.117	-1.5	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0