


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG082218\
Data File : BG036332.D
Acq On : 22 Aug 2018 8:08
Operator : JU/SJ
Sample : DFTPP05
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP05

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG082118MA.M
Title : SVOA CALIBRATION
Last Update : Wed Aug 22 18:34:10 2018



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2478

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.0	73005	PASS
68	69	0.00	2	0.7	459	PASS
69	198	0.00	100	38.9	66021	PASS
70	69	0.00	2	0.4	281	PASS
127	198	10	80	46.0	78074	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	169642	PASS
199	198	5	9	7.3	12395	PASS
275	198	10	60	27.0	45858	PASS
365	198	1	100	3.8	6438	PASS
441	198	0.01	100	9.4	15993	PASS
442	198	50	100	60.5	102629	PASS
443	442	15	24	19.7	20254	PASS

m/z	Abundance
35.90	155.0
37.00	238.0
38.10	820.0
39.10	5485.0
40.00	902.0
41.10	319.0
43.00	151.0
44.10	1862.0
45.00	191.0
49.10	489.0
50.10	16448.0
51.10	70592.0
52.10	3206.0
55.10	652.0
56.10	2122.0
57.10	4670.0
58.10	207.0
61.10	718.0
62.10	863.0
63.10	2265.0
64.10	330.0
65.10	1663.0
66.20	190.0
67.00	183.0
68.10	1379.0
69.00	60280.0
69.90	342.0
71.10	162.0
73.10	798.0
74.10	6096.0
75.10	10030.0
76.10	3534.0
77.10	68848.0
78.10	4820.0
79.10	4936.0
80.00	3843.0
81.10	4999.0
82.00	1271.0
83.10	993.0
85.10	978.0
86.00	1403.0
87.10	856.0
88.10	382.0
89.00	225.0
91.10	933.0
92.10	1321.0
93.10	7580.0
94.10	507.0
96.10	514.0
98.10	5485.0
99.00	4454.0
100.00	498.0
101.00	2612.0
103.10	1101.0
104.10	2146.0
105.10	1538.0
106.10	842.0
107.10	21744.0
108.10	3092.0
110.00	36544.0
111.10	5810.0
112.10	949.0
113.00	227.0
116.10	1652.0
117.10	17760.0
118.10	1241.0
119.00	164.0
120.10	298.0
122.00	1117.0
123.00	2140.0
124.10	1171.0
125.00	989.0
127.10	68824.0
128.10	5530.0
129.00	28840.0
130.00	2522.0
131.00	626.0
132.00	235.0
134.00	1044.0
135.00	2656.0
136.00	793.0
137.00	988.0
138.10	241.0
138.90	166.0
139.90	359.0
141.00	4046.0
142.00	1253.0
143.00	863.0
144.00	271.0
145.00	187.0
146.00	595.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

147.00	2211.0
148.10	4203.0
149.10	787.0
150.00	217.0
151.10	353.0
151.90	268.0
153.00	1410.0
154.10	861.0
155.10	2191.0
156.10	3469.0
157.10	688.0
158.00	761.0
159.00	536.0
160.00	1070.0
161.00	1984.0
162.00	484.0
163.20	263.0
165.00	1474.0
166.00	1124.0
167.10	6515.0
168.10	3426.0
169.10	668.0
170.00	358.0
170.90	259.0
172.00	728.0
173.00	1037.0
174.10	1939.0
175.10	3097.0
176.00	719.0
177.00	1593.0
178.10	630.0
179.00	5790.0
180.10	3760.0
181.00	1951.0
182.00	449.0
184.10	557.0
185.10	2915.0
186.10	20048.0
187.00	6955.0
188.10	865.0
189.00	1801.0
190.10	228.0
191.00	732.0
192.10	1879.0
193.10	1805.0
194.10	575.0
195.00	239.0
196.00	3748.0
198.00	138496.0
199.00	10237.0
200.10	895.0
200.80	451.0
201.50	699.0
202.20	293.0
203.10	1393.0
204.10	6236.0
205.10	10275.0
206.10	41808.0
207.10	5690.0
208.10	1794.0
209.00	532.0
210.00	660.0
211.10	1707.0
211.90	377.0
215.00	467.0
216.10	1107.0
217.00	12126.0
218.00	1661.0
221.10	5980.0
222.00	1402.0
223.10	2685.0
224.10	24104.0
225.10	5154.0
226.10	710.0
227.00	10569.0
228.00	1612.0
229.00	2179.0
230.10	352.0
231.20	626.0
232.20	220.0
234.10	623.0
235.10	684.0
236.00	431.0
237.10	600.0
239.10	569.0
240.10	473.0
241.10	669.0
242.00	949.0
243.10	1371.0
244.10	16928.0
245.10	2397.0
246.10	4131.0
247.00	980.0
249.10	450.0
251.00	190.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

252.10	225.0
252.90	446.0
255.00	84448.0
256.10	13583.0
257.00	709.0
258.10	5472.0
259.10	757.0
265.00	2193.0
266.00	313.0
267.10	157.0
270.10	190.0
271.00	210.0
272.10	172.0
273.10	2718.0
274.10	7083.0
275.10	35392.0
276.10	4839.0
277.00	3347.0
278.00	704.0
283.10	575.0
284.00	253.0
285.10	755.0
286.00	196.0
290.10	170.0
293.00	600.0
295.00	300.0
296.10	10493.0
297.00	1656.0
302.10	160.0
303.10	1239.0
304.10	293.0
310.10	164.0
314.10	411.0
315.00	1403.0
316.10	708.0
317.20	207.0
321.10	514.0
322.00	159.0
323.10	2939.0
324.10	708.0
326.90	642.0
328.10	313.0
332.10	296.0
333.10	334.0
334.10	1567.0
335.10	361.0
341.10	429.0
346.00	480.0
352.10	1107.0
353.00	570.0
354.10	1008.0
355.10	267.0
365.00	4615.0
366.00	641.0
371.10	210.0
372.10	1421.0
373.20	324.0
383.00	514.0
384.00	173.0
389.90	241.0
402.00	523.0
403.10	957.0
404.00	330.0
421.00	583.0
422.10	663.0
423.10	3571.0
424.10	615.0
441.10	9291.0
442.10	64680.0
443.10	13116.0
444.10	1266.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

m/z	Abundance
36.10	184.0
37.10	215.0
38.10	716.0
39.10	6436.0
40.10	929.0
41.20	349.0
44.10	1513.0
45.00	233.0
49.10	765.0
50.10	21936.0
51.10	86464.0
52.10	4218.0
55.00	550.0
56.10	2965.0
57.10	6511.0
58.00	347.0
61.10	933.0
62.10	1513.0
63.10	3791.0
64.10	577.0
65.10	1589.0
67.30	171.0
69.10	80232.0
70.10	457.0
71.10	158.0
73.10	569.0
74.10	7959.0
75.10	13079.0
76.10	4106.0
77.10	90200.0
78.10	5972.0
79.10	6125.0
80.10	4364.0
81.10	7029.0
82.10	1639.0
83.10	1877.0
84.00	154.0
85.00	1065.0
86.10	1744.0
87.10	1105.0
88.10	339.0
89.10	223.0
91.10	1770.0
92.10	1695.0
93.00	10298.0
94.00	663.0
95.10	284.0
96.10	743.0
98.10	8313.0
99.10	5661.0
100.10	531.0
101.00	3554.0
102.10	194.0
103.00	1231.0
104.10	2608.0
105.10	2449.0
106.10	810.0
107.10	27768.0
108.10	4873.0
110.10	48048.0
111.10	7276.0
112.00	1125.0
113.00	298.0
116.10	1512.0
117.10	22008.0
118.10	1932.0
119.00	229.0
120.00	316.0
121.10	280.0
122.00	1936.0
123.10	3091.0
124.00	1621.0
125.10	1400.0
127.10	94112.0
128.10	7824.0
129.10	37320.0
130.10	3170.0
131.10	681.0
132.10	351.0
133.10	258.0
134.10	903.0
135.10	3545.0
136.00	1173.0
137.00	1403.0
138.10	276.0
139.00	218.0
140.10	604.0
141.00	5146.0
142.10	1554.0
143.10	1479.0
144.00	300.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

145.00	279.0
146.10	683.0
147.00	2476.0
148.10	5515.0
149.00	1269.0
150.00	553.0
151.00	553.0
151.60	299.0
152.10	194.0
153.00	1991.0
154.00	1142.0
155.00	3304.0
156.10	4309.0
157.10	768.0
158.00	1124.0
159.10	750.0
160.10	1972.0
161.00	2527.0
162.10	802.0
163.10	245.0
164.10	309.0
165.00	2390.0
166.00	1543.0
167.10	9599.0
168.10	4365.0
169.10	643.0
170.00	503.0
171.00	476.0
172.00	1232.0
173.10	1682.0
174.10	2431.0
175.10	3978.0
176.10	1067.0
177.00	1984.0
178.00	701.0
179.00	8050.0
180.10	5469.0
181.00	2155.0
182.10	415.0
183.10	242.0
184.10	869.0
185.10	4390.0
186.10	29728.0
187.10	9188.0
188.10	930.0
189.00	2149.0
189.90	486.0
191.00	1336.0
192.00	2444.0
193.10	2986.0
194.10	712.0
195.00	432.0
196.10	5989.0
198.00	200448.0
199.00	15528.0
200.00	1237.0
200.90	323.0
201.50	797.0
203.00	1746.0
204.10	8334.0
205.10	14980.0
206.10	57968.0
207.10	8342.0
208.00	2182.0
209.00	749.0
210.10	891.0
211.10	2028.0
212.10	377.0
213.00	217.0
215.10	784.0
216.00	1348.0
217.00	17224.0
218.00	1836.0
219.10	191.0
221.00	9179.0
221.90	2063.0
223.00	3801.0
224.10	34336.0
225.10	8004.0
226.10	1042.0
227.10	15430.0
228.00	1833.0
229.00	3026.0
230.00	327.0
231.00	1159.0
232.10	224.0
232.90	214.0
234.00	862.0
235.10	1233.0
236.00	650.0
237.10	890.0
238.00	168.0
239.10	783.0
240.00	449.0
241.00	761.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

242.10	1827.0
243.10	1780.0
244.10	24232.0
245.10	3006.0
246.10	5248.0
247.00	1311.0
248.00	278.0
249.10	891.0
249.90	180.0
250.90	253.0
252.00	372.0
253.00	502.0
255.10	126688.0
256.10	18520.0
257.10	1773.0
258.10	8260.0
259.00	1444.0
260.10	157.0
261.10	217.0
264.10	344.0
265.10	3774.0
266.10	650.0
271.00	372.0
272.00	375.0
273.00	4163.0
274.10	9352.0
275.10	53296.0
276.10	6817.0
277.10	4949.0
278.00	744.0
278.90	197.0
282.10	209.0
283.10	601.0
284.00	360.0
285.10	965.0
286.20	225.0
289.00	309.0
289.80	157.0
292.00	360.0
293.00	1081.0
294.20	295.0
295.00	318.0
296.10	16142.0
297.10	2256.0
301.10	195.0
302.10	543.0
303.10	1864.0
304.00	641.0
308.00	282.0
310.00	200.0
313.10	165.0
314.10	937.0
315.10	1852.0
316.10	683.0
317.00	290.0
321.10	376.0
322.00	232.0
323.10	5661.0
324.10	1302.0
327.00	1010.0
328.10	470.0
332.10	387.0
333.10	590.0
334.10	3038.0
335.10	809.0
340.10	169.0
341.00	599.0
342.10	214.0
346.00	1224.0
346.80	194.0
351.00	184.0
352.10	1440.0
353.00	1188.0
354.10	1404.0
355.20	323.0
365.00	7427.0
366.00	1016.0
370.00	255.0
371.10	504.0
372.10	2632.0
373.10	743.0
383.00	874.0
384.10	233.0
390.00	364.0
391.10	165.0
392.00	300.0
402.10	647.0
403.10	1289.0
404.00	433.0
421.10	1378.0
422.10	1128.0
423.10	6654.0
424.10	1720.0
425.00	169.0
441.10	18696.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

442.10	115496.0
443.10	23080.0
444.10	1948.0
445.10	216.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

m/z	Abundance
37.10	164.0
38.10	589.0
39.10	3774.0
40.00	925.0
41.10	337.0
43.10	232.0
44.10	1327.0
49.00	400.0
50.10	14856.0
51.10	61960.0
52.10	3452.0
53.00	157.0
55.10	446.0
56.10	2033.0
57.10	4452.0
58.20	268.0
61.10	770.0
62.00	805.0
63.10	2626.0
64.10	206.0
65.10	1168.0
69.00	57552.0
70.10	386.0
71.10	220.0
73.10	501.0
74.10	5831.0
75.10	9022.0
76.10	3334.0
77.10	67552.0
78.10	4357.0
79.00	4250.0
80.10	3525.0
81.00	4966.0
82.10	1170.0
83.00	1067.0
84.00	228.0
85.10	1000.0
86.00	1237.0
87.00	475.0
88.10	223.0
91.00	1138.0
92.10	1176.0
93.10	8592.0
94.00	492.0
96.10	511.0
98.00	6004.0
99.10	4704.0
100.00	481.0
101.10	2473.0
103.00	1053.0
104.10	1727.0
105.10	1799.0
106.10	682.0
107.10	21624.0
108.10	3649.0
109.00	803.0
110.00	36216.0
111.10	5915.0
112.00	601.0
117.10	17272.0
118.10	1049.0
118.90	189.0
120.00	265.0
122.00	1483.0
123.10	2437.0
124.00	775.0
125.10	974.0
127.10	71288.0
128.10	5300.0
129.10	29472.0
130.00	2716.0
131.10	457.0
132.00	276.0
133.00	260.0
134.00	993.0
135.10	2706.0
136.00	1210.0
137.10	1278.0
138.10	249.0
138.90	171.0
139.80	353.0
141.00	4131.0
142.10	1462.0
143.00	1024.0
144.00	203.0
145.10	319.0
146.10	989.0
147.00	2154.0
148.10	5283.0
149.00	807.0
150.00	225.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

151.10	510.0
151.80	319.0
153.00	1589.0
154.00	1040.0
155.10	2053.0
156.10	2906.0
157.10	477.0
158.00	994.0
159.00	496.0
160.10	1215.0
161.10	1898.0
162.10	693.0
163.10	277.0
164.00	429.0
165.00	1818.0
166.10	1300.0
167.10	7807.0
168.10	3427.0
169.10	663.0
170.00	280.0
171.10	676.0
172.00	818.0
173.00	1538.0
174.00	2034.0
175.10	3727.0
176.10	811.0
177.10	1767.0
178.10	653.0
179.00	7148.0
180.10	4587.0
181.10	2586.0
182.00	450.0
184.10	497.0
185.10	3649.0
186.10	25160.0
187.10	7829.0
188.00	745.0
189.00	1844.0
190.00	264.0
191.00	883.0
192.10	1816.0
193.10	2486.0
194.10	658.0
194.90	156.0
196.00	4477.0
198.00	169984.0
199.00	11420.0
200.10	1017.0
200.80	332.0
201.50	1002.0
203.00	1550.0
204.10	8208.0
205.10	13494.0
206.10	50296.0
207.10	6642.0
208.00	1857.0
208.90	672.0
210.30	839.0
211.10	1980.0
213.10	188.0
215.00	669.0
216.00	1306.0
217.00	15146.0
218.00	1510.0
219.00	278.0
219.80	183.0
221.10	7099.0
222.00	1949.0
223.00	2877.0
224.10	31568.0
225.10	8490.0
226.00	849.0
227.00	13267.0
228.00	1847.0
229.00	2288.0
230.10	370.0
231.00	1226.0
232.00	182.0
233.00	191.0
234.00	683.0
235.00	828.0
236.00	439.0
237.10	971.0
238.10	183.0
239.10	307.0
240.00	564.0
241.00	476.0
242.10	1694.0
243.10	1416.0
244.10	21824.0
245.10	2837.0
246.00	4967.0
247.00	1006.0
248.10	213.0
249.00	791.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

251.10	404.0
252.10	350.0
253.10	623.0
255.10	116504.0
256.10	15960.0
257.00	1444.0
258.00	7757.0
259.00	1120.0
260.10	185.0
261.00	252.0
264.10	167.0
265.00	3292.0
266.00	489.0
270.20	218.0
271.00	276.0
272.10	495.0
273.00	3398.0
274.10	9894.0
275.10	48888.0
276.00	6283.0
277.00	4962.0
278.10	870.0
281.00	179.0
283.00	608.0
284.10	303.0
285.10	831.0
289.00	190.0
291.10	163.0
292.00	202.0
293.10	1026.0
294.00	378.0
295.10	511.0
296.10	16066.0
297.00	2131.0
301.10	188.0
302.00	309.0
303.10	1786.0
304.10	422.0
308.30	249.0
309.10	150.0
310.00	246.0
313.00	165.0
314.10	961.0
315.00	2373.0
316.10	954.0
317.10	359.0
321.00	709.0
322.00	255.0
323.10	5437.0
324.10	989.0
327.00	772.0
328.10	498.0
332.10	368.0
333.00	415.0
334.10	3148.0
335.00	752.0
341.10	521.0
342.10	203.0
346.10	947.0
346.90	284.0
352.10	1496.0
353.10	1164.0
354.10	1383.0
355.10	421.0
365.10	7272.0
366.10	875.0
370.10	185.0
371.10	323.0
372.10	2477.0
373.10	500.0
383.10	797.0
384.00	176.0
390.00	209.0
391.10	226.0
392.10	169.0
402.10	1125.0
403.10	1341.0
404.10	538.0
421.10	1252.0
422.10	1121.0
423.10	7588.0
424.10	1666.0
435.80	160.0
441.10	19992.0
442.10	127712.0
443.10	24568.0
444.10	2078.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

Scan 2478 (17.848 min): BG036332.D
DFTPP05

m/z	Abundance
40.00	500.0
41.10	313.0
43.10	216.0
44.10	1382.0
56.20	205.0
57.10	285.0
71.10	213.0
207.10	397.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05