

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG082318\
 Data File : BG036358.D
 Acq On : 23 Aug 2018 1:07
 Operator : JU/SJ
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Aug 23 02:11:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 23 01:15:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.07	152	51950	20.00	ng	0.00
21) Naphthalene-d8	10.88	136	237734	20.00	ng	0.00
38) Acenaphthene-d10	14.70	164	152397	20.00	ng	0.00
63) Phenanthrene-d10	17.44	188	367220	20.00	ng	0.00
75) Chrysene-d12	21.72	240	356877	20.00	ng	0.00
86) Perylene-d12	24.94	264	379033	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.63	112	394987	125.48	ng	0.00
7) Phenol-d6	7.23	99	567449	121.61	ng	0.00
23) Nitrobenzene-d5	9.24	82	536415	97.00	ng	0.00
41) 2,4,6-Tribromophenol	16.19	330	225593	112.29	ng	0.00
44) 2-Fluorobiphenyl	13.33	172	1189818	104.78	ng	0.00
78) Terphenyl-d14	20.05	244	1712350	106.09	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.54	88	86849	54.596	ng	99
3) Pyridine	3.94	79	267020	64.070	ng	99
4) n-Nitrosodimethylamine	3.85	42	115995	64.201	ng	95
6) Aniline	7.40	93	362800	60.134	ng	98
8) 2-Chlorophenol	7.64	128	212716	65.771	ng	97
9) Benzaldehyde	7.21	77	169420	58.792	ng	97
10) Phenol	7.25	94	297051	62.749	ng	96
11) bis(2-Chloroethyl)ether	7.49	93	233669	62.901	ng	100
12) 1,3-Dichlorobenzene	7.96	146	238808	61.717	ng	99
13) 1,4-Dichlorobenzene	8.11	146	244425	62.239	ng	99
14) 1,2-Dichlorobenzene	8.43	146	232838	61.772	ng	95
15) Benzyl Alcohol	8.31	79	239398	57.301	ng	95
16) 2,2'-oxybis(1-Chloropropan	8.61	45	461248	122.750	ng	99
17) 2-Methylphenol	8.51	107	201657	63.043	ng	99
18) Hexachloroethane	9.16	117	87975	58.833	ng	97
19) n-Nitroso-di-n-propylamine	8.88	70	213934	55.229	ng	99
20) 3+4-Methylphenols	8.84	107	283491	64.008	ng	98
22) Acetophenone	8.90	105	360603	49.629	ng	# 96
24) Nitrobenzene	9.28	77	289689	51.482	ng	98
25) Isophorone	9.81	82	519947	50.295	ng	97
26) 2-Nitrophenol	9.99	139	121717	60.606	ng	97
27) 2,4-Dimethylphenol	10.04	122	203465	59.068	ng	98
28) bis(2-Chloroethoxy)methane	10.29	93	312030	54.084	ng	99
29) 2,4-Dichlorophenol	10.52	162	233536	59.734	ng	96
30) 1,2,4-Trichlorobenzene	10.74	180	261976	54.887	ng	98
31) Naphthalene	10.93	128	687303	55.412	ng	99
32) Benzoic acid	10.19	122	165957	70.821	ng	92
33) 4-Chloroaniline	11.03	127	323737	59.504	ng	99
34) Hexachlorobutadiene	11.22	225	176423	48.086	ng	97
35) Caprolactam	11.82	113	92480	54.971	ng	97
36) 4-Chloro-3-methylphenol	12.14	107	270320	52.399	ng	99
37) 2-Methylnaphthalene	12.53	142	509949	55.481	ng	97
39) 1,2,4,5-Tetrachlorobenzene	12.90	216	313267	54.435	ng	99
40) Hexachlorocyclopentadiene	12.88	237	171509	57.124	ng	94

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG082318\
 Data File : BG036358.D
 Acq On : 23 Aug 2018 1:07
 Operator : JU/SJ
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Aug 23 02:11:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 23 01:15:05 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.13	196	202313	60.212	ng	100
43) 2,4,5-Trichlorophenol	13.20	196	209974	55.557	ng	95
45) 1,1'-Biphenyl	13.54	154	721537	60.234	ng	98
46) 2-Chloronaphthalene	13.58	162	558065	59.781	ng	99
47) 2-Nitroaniline	13.77	65	196367	60.278	ng	99
48) Acenaphthylene	14.42	152	864190	55.897	ng	99
49) Dimethylphthalate	14.16	163	722652	54.077	ng	100
50) 2,6-Dinitrotoluene	14.27	165	154518	56.919	ng	97
51) Acenaphthene	14.77	154	560490	57.832	ng	99
52) 3-Nitroaniline	14.59	138	173718	61.726	ng	95
53) 2,4-Dinitrophenol	14.80	184	63386	54.012	ng	95
54) Dibenzofuran	15.10	168	858313	55.964	ng	100
55) 4-Nitrophenol	14.89	139	134892	65.663	ng	96
56) 2,4-Dinitrotoluene	15.05	165	208640	54.437	ng	95
57) Fluorene	15.75	166	640355	50.370	ng	98
58) 2,3,4,6-Tetrachlorophenol	15.32	232	200547	55.767	ng	99
59) Diethylphthalate	15.52	149	718854	52.496	ng	97
60) 4-Chlorophenyl-phenylether	15.74	204	401615	53.544	ng	100
61) 4-Nitroaniline	15.76	138	172789	58.787	ng	98
62) Azobenzene	16.03	77	727508	53.704	ng	99
64) 4,6-Dinitro-2-methylphenol	15.82	198	104413	52.611	ng	97
65) n-Nitrosodiphenylamine	15.95	169	636955	60.532	ng	100
66) 4-Bromophenyl-phenylether	16.63	248	256523	60.032	ng	98
67) Hexachlorobenzene	16.75	284	260842	59.173	ng	99
68) Atrazine	16.90	200	204754	48.105	ng	99
69) Pentachlorophenol	17.09	266	191655	69.461	ng	96
70) Phenanthrene	17.49	178	1065896	57.730	ng	98
71) Anthracene	17.58	178	1061631	57.753	ng	98
72) Carbazole	17.84	167	1042360	59.213	ng	99
73) Di-n-butylphthalate	18.41	149	1164561	56.399	ng	98
74) Fluoranthene	19.49	202	1270655	51.720	ng	98
76) Benzidine	19.67	184	574134	59.372	ng	98
77) Pyrene	19.85	202	1281346	56.520	ng	97
79) Butylbenzylphthalate	20.75	149	542759	66.008	ng	97
80) Benzo(a)anthracene	21.69	228	1254256	56.353	ng	97
81) 3,3'-Dichlorobenzidine	21.61	252	515971	65.070	ng	98
82) Chrysene	21.76	228	1186030	57.214	ng	96
83) Bis(2-ethylhexyl)phthalate	21.62	149	735021	65.609	ng	100
84) Di-n-octyl phthalate	22.85	149	1264247	67.312	ng	100
85) Indeno(1,2,3-cd)pyrene	28.62	276	1458243	63.214	ng	99
87) Benzo(b)fluoranthene	23.92	252	1272019	55.412	ng	97
88) Benzo(k)fluoranthene	23.99	252	1205106	54.003	ng	99
89) Benzo(a)pyrene	24.79	252	1217113	57.038	ng	99
90) Dibenzo(a,h)anthracene	28.70	278	1161341	55.573	ng	97
91) Benzo(g,h,i)perylene	29.77	276	1184857	57.758	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG082318\
 Data File : BG036358.D
 Acq On : 23 Aug 2018 1:07
 Operator : JU/SJ
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 Client Sampled :

Quant Time: Aug 23 02:11:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 23 01:15:05 2018
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.07 min Scan# 814
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
152	51950		
150	159.7	119.5	179.3
115	60.0	44.5	66.7



#2
 1,4-Dioxane
 Concen: 54.596 ng
 RT: 3.54 min Scan# 43
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
88	86849		
58	90.9	72.6	109.0
43	34.6	29.0	43.4









#3
 Pyridine
 Concen: 64.070 ng
 RT: 3.94 min Scan# 110
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
79	100		
52	90.5	72.3	108.5
51	44.4	33.8	50.6



#4
 n-Nitrosodimethylamine
 Concen: 64.201 ng
 RT: 3.85 min Scan# 95
 Delta R.T. -0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
42	100		
74	127.0	96.6	145.0
44	7.6	7.0	10.4









#5
 2-Fluorophenol
 Concen: 125.478 ng
 RT: 5.63 min Scan# 399
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
112	394987		
64	71.5	57.2	85.8
63	37.9	30.8	46.2



#6
 Aniline
 Concen: 60.134 ng
 RT: 7.40 min Scan# 699
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
93	362800		
66	42.9	35.0	52.4
65	23.9	17.6	26.4









#7
 Phenol-d6
 Concen: 121.613 ng
 RT: 7.23 min Scan# 670
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
99	100		
42	20.7	16.5	24.7
71	36.3	29.4	44.2



#8
 2-Chlorophenol
 Concen: 65.771 ng
 RT: 7.64 min Scan# 740
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
128	100		
130	34.3	10.0	50.0
64	56.3	36.0	76.0









#9
Benzaldehyde
Concen: 58.792 ng
RT: 7.21 min Scan# 667
Delta R.T. -0.00 min
Lab File: BG036358.D
Acq: 23 Aug 2018 1:07




Instrument :
BNA_G
ClientSampled :


Tgt Ion	Resp	Lower	Upper
77	100		
105	94.6	70.5	110.5
106	88.8	69.8	109.8



#10
Phenol
Concen: 62.749 ng
RT: 7.25 min Scan# 674
Delta R.T. -0.00 min
Lab File: BG036358.D
Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
94	100		
65	33.7	11.5	51.5
66	42.5	19.5	59.5









#11
 bis(2-Chloroethyl)ether
 Concen: 62.901 ng
 RT: 7.49 min Scan# 715
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
93	233669		
Ion Ratio			
93	100		
63	89.7	69.9	109.9
95	31.5	11.7	51.7



#12
 1,3-Dichlorobenzene
 Concen: 61.717 ng
 RT: 7.96 min Scan# 795
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
146	238808		
Ion Ratio			
146	100		
148	66.7	53.8	80.6
75	34.2	27.2	40.8









#13
 1,4-Dichlorobenzene
 Concen: 62.239 ng
 RT: 8.11 min Scan# 820
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
146	100		
148	68.3	53.9	80.9
111	41.1	33.8	50.6



#14
 1,2-Dichlorobenzene
 Concen: 61.772 ng
 RT: 8.43 min Scan# 875
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
146	100		
148	63.7	55.0	82.6
111	44.3	34.6	51.8









#15
Benzyl Alcohol
Concen: 57.301 ng
RT: 8.31 min Scan# 854
Delta R.T. 0.01 min
Lab File: BG036358.D
Acq: 23 Aug 2018 1:07

Instrument :
BNA_G
ClientSampled :

Tgt Ion	Resp	Lower	Upper
79	100		
108	74.2	55.9	83.9
77	65.7	50.0	75.0




Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 108.00 (107.70 to 108.70): BGC
Ion 77.00 (76.70 to 77.70): BGC






#16
2,2'-oxybis(1-Chloropropane)
Concen: 122.750 ng
RT: 8.61 min Scan# 905
Delta R.T. 0.01 min
Lab File: BG036358.D
Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
45	100		
77	10.7	0.0	30.5
79	8.2	0.0	27.9



Abundance Ion 45.00 (44.70 to 45.70): BGC
Ion 77.00 (76.70 to 77.70): BGC
Ion 79.00 (78.70 to 79.70): BGC









#17
 2-Methylphenol
 Concen: 63.043 ng
 RT: 8.51 min Scan# 888
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
107	100		
108	109.2	86.7	130.1
77	48.5	39.0	58.4
79	48.3	36.6	54.8



#18
 Hexachloroethane
 Concen: 58.833 ng
 RT: 9.16 min Scan# 999
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07


Tgt Ion	Resp	Lower	Upper
117	100		
119	99.1	76.8	115.2
201	110.6	85.7	128.5








#19
 n-Nitroso-di-n-propylamine
 Concen: 55.229 ng
 RT: 8.88 min Scan# 952
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Instrument :
 BNA_G
 ClientSampled :




Tgt Ion: 70 Resp: 213934

Ion	Ratio	Lower	Upper
70	100		
42	70.8	56.2	84.4
101	10.3	7.8	11.6
130	20.3	16.2	24.2






#20
 3+4-Methylphenols
 Concen: 64.008 ng
 RT: 8.84 min Scan# 944
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07



Tgt Ion: 107 Resp: 283491

Ion	Ratio	Lower	Upper
107	100		
108	89.1	68.0	108.0
77	36.4	15.6	55.6
79	32.5	9.9	49.9






#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.88 min Scan# 1292
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :

Tgt Ion:136 Resp: 237734

Ion	Ratio	Lower	Upper
136	100		
137	10.5	8.8	13.2
54	11.2	9.2	13.8
68	6.6	5.8	8.6




Abundance
 Ion 136.00 (135.70 to 136.70): E
 Ion 137.00 (136.70 to 137.70): E
 Ion 54.00 (53.70 to 54.70): BGC
 Ion 68.00 (67.70 to 68.70): BGC





#22
 Acetophenone
 Concen: 49.629 ng
 RT: 8.90 min Scan# 954
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07


Tgt Ion:105 Resp: 360603

Ion	Ratio	Lower	Upper
105	100		
71	6.3	2.4	3.6#
51	34.0	25.9	38.9
120	22.3	19.4	29.0



Abundance
 Ion 105.00 (104.70 to 105.70): E
 Ion 71.00 (70.70 to 71.70): BGC
 Ion 51.00 (50.70 to 51.70): BGC
 Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 96.997 ng
 RT: 9.24 min Scan# 1012
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
82	536415		
128	41.2	33.3	49.9
54	58.1	45.1	67.7



#24
 Nitrobenzene
 Concen: 51.482 ng
 RT: 9.28 min Scan# 1019
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
77	289689		
123	42.0	32.4	48.6
65	14.8	11.5	17.3









#25
 Isophorone
 Concen: 50.295 ng
 RT: 9.81 min Scan# 1109
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
82	519947		
95	7.4	5.8	8.8
138	17.8	12.9	19.3



#26
 2-Nitrophenol
 Concen: 60.606 ng
 RT: 9.99 min Scan# 1140
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
139	121717		
109	32.1	27.4	41.0
65	57.9	47.8	71.6









#27
 2,4-Dimethylphenol
 Concen: 59.068 ng
 RT: 10.04 min Scan# 1149
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
122	203465		
107	121.3	96.0	144.0
121	58.4	48.2	72.4



#28
 bis(2-Chloroethoxy)methane
 Concen: 54.084 ng
 RT: 10.29 min Scan# 1191
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
93	312030		
95	31.1	25.2	37.8
123	11.8	8.6	12.8









#29
 2,4-Dichlorophenol
 Concen: 59.734 ng
 RT: 10.52 min Scan# 1230
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
162	100		
164	65.2	47.6	87.6
98	36.5	19.2	59.2



#30
 1,2,4-Trichlorobenzene
 Concen: 54.887 ng
 RT: 10.74 min Scan# 1268
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
180	100		
182	93.1	76.0	114.0
145	29.3	22.9	34.3









#31
 Naphthalene
 Concen: 55.412 ng
 RT: 10.93 min Scan# 1301
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
128	687303		
129	12.0	9.4	14.2
127	14.5	11.0	16.4



#32
 Benzoic acid
 Concen: 70.821 ng
 RT: 10.19 min Scan# 1174
 Delta R.T. 0.03 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
122	165957		
105	135.8	108.1	148.1
77	106.8	76.3	116.3









#33
 4-Chloroaniline
 Concen: 59.504 ng
 RT: 11.03 min Scan# 1318
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
127	323737		
129	32.0	25.9	38.9
65	35.8	27.5	41.3
92	20.4	16.2	24.2



#34
 Hexachlorobutadiene
 Concen: 48.086 ng
 RT: 11.22 min Scan# 1349
 Delta R.T. -0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
225	176423		
223	60.8	51.8	77.8
227	62.0	49.8	74.6









#35
 Caprolactam
 Concen: 54.971 ng
 RT: 11.82 min Scan# 1451
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
113	100		
55	225.7	197.7	237.7
56	152.3	131.7	171.7



#36
 4-Chloro-3-methylphenol
 Concen: 52.399 ng
 RT: 12.14 min Scan# 1507
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
107	100		
144	24.7	19.5	29.3
142	76.6	62.4	93.6









#37
 2-Methylnaphthalene
 Concen: 55.481 ng
 RT: 12.53 min Scan# 1573
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
142	509949		
141	88.2	68.5	102.7
115	38.4	29.8	44.8



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.70 min Scan# 1942
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
164	152397		
162	103.7	80.7	121.1
160	45.4	35.3	52.9









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 54.435 ng
 RT: 12.90 min Scan# 1635
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
216	313267		
214	77.9	61.0	91.6
179	19.7	15.7	23.5
108	17.1	13.1	19.7



#40
 Hexachlorocyclopentadiene
 Concen: 57.124 ng
 RT: 12.88 min Scan# 1632
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
237	171509		
235	67.3	42.2	82.2
272	12.5	0.0	32.1









#41
 2,4,6-Tribromophenol
 Concen: 112.291 ng
 RT: 16.19 min Scan# 2195
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
330	100		
332	96.8	75.4	113.2
141	38.8	30.7	46.1



#42
 2,4,6-Trichlorophenol
 Concen: 60.212 ng
 RT: 13.13 min Scan# 1675
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
196	100		
198	96.3	77.2	115.8
200	32.3	26.3	39.5









#43
 2,4,5-Trichlorophenol
 Concen: 55.557 ng
 RT: 13.20 min Scan# 1686
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
196	100		
198	97.5	75.6	113.4
97	51.6	37.2	55.8
132	27.5	19.4	29.2




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BG
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 104.783 ng
 RT: 13.33 min Scan# 1708
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
172	100		
171	39.7	31.3	46.9
170	27.4	21.0	31.6



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 60.234 ng
 RT: 13.54 min Scan# 1744
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
154	100		
153	42.6	24.0	64.0
76	14.1	0.0	32.9



#46
 2-Chloronaphthalene
 Concen: 59.781 ng
 RT: 13.58 min Scan# 1751
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
162	100		
127	38.9	30.7	46.1
164	34.1	26.7	40.1









#47
 2-Nitroaniline
 Concen: 60.278 ng
 RT: 13.77 min Scan# 1784
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
65	196367		
92	61.1	49.9	74.9
138	94.1	76.2	114.4



#48
 Acenaphthylene
 Concen: 55.897 ng
 RT: 14.42 min Scan# 1895
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
152	864190		
151	20.8	16.9	25.3
153	13.7	11.2	16.8








#49
 Dimethylphthalate
 Concen: 54.077 ng
 RT: 14.16 min Scan# 1850
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
163	100		
194	4.9	4.0	6.0
164	10.1	8.1	12.1



#50
 2,6-Dinitrotoluene
 Concen: 56.919 ng
 RT: 14.27 min Scan# 1869
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
165	100		
63	60.3	50.1	75.1
89	57.1	47.8	71.6









#51
 Acenaphthene
 Concen: 57.832 ng
 RT: 14.77 min Scan# 1953
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
154	100		
153	114.3	91.8	137.6
152	55.8	43.5	65.3



#52
 3-Nitroaniline
 Concen: 61.726 ng
 RT: 14.59 min Scan# 1924
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
138	100		
108	11.1	9.4	14.2
92	126.8	96.8	145.2









#53
 2,4-Dinitrophenol
 Concen: 54.012 ng
 RT: 14.80 min Scan# 1959
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
184	100		
63	57.4	47.3	70.9
154	57.8	51.1	76.7



#54
 Dibenzofuran
 Concen: 55.964 ng
 RT: 15.10 min Scan# 2010
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
168	100		
139	39.0	31.0	46.6
169	14.3	11.1	16.7









#55
 4-Nitrophenol
 Concen: 65.663 ng
 RT: 14.89 min Scan# 1974
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
139	100		
109	86.2	63.9	103.9
65	114.9	90.0	130.0



#56
 2,4-Dinitrotoluene
 Concen: 54.437 ng
 RT: 15.05 min Scan# 2002
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
165	100		
63	47.0	40.1	60.1
89	74.3	63.7	95.5
182	3.2	2.5	3.7









#57
 Fluorene
 Concen: 50.370 ng
 RT: 15.75 min Scan# 2120
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
166	100		
165	98.0	77.0	115.4
167	14.2	10.6	16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 55.767 ng
 RT: 15.32 min Scan# 2047
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
232	100		
131	41.7	33.8	50.8
130	2.3	2.1	3.1
166	28.0	22.3	33.5









#59
 Diethylphthalate
 Concen: 52.496 ng
 RT: 15.52 min Scan# 2082
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
149	718854		
177	25.6	19.1	28.7
150	14.2	10.7	16.1



#60
 4-Chlorophenyl-phenylether
 Concen: 53.544 ng
 RT: 15.74 min Scan# 2119
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
204	401615		
206	33.7	27.0	40.4
141	55.2	43.8	65.6








#61
 4-Nitroaniline
 Concen: 58.787 ng
 RT: 15.76 min Scan# 2123
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
138	100		
92	55.5	38.5	78.5
108	91.7	70.3	110.3



#62
 Azobenzene
 Concen: 53.704 ng
 RT: 16.03 min Scan# 2169
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
77	100		
182	28.3	6.7	46.7
105	19.6	0.0	39.8
51	32.1	12.4	52.4








#63
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.44 min Scan# 2409
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
188	367220		
94	7.7	6.7	10.1
80	9.4	8.1	12.1



#64
 4,6-Dinitro-2-methylphenol
 Concen: 52.611 ng
 RT: 15.82 min Scan# 2132
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
198	104413		
51	48.7	26.5	66.5
105	38.9	20.2	60.2









#65
 n-Nitrosodiphenylamine
 Concen: 60.532 ng
 RT: 15.95 min Scan# 2155
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
169	100		
168	71.8	57.6	86.4
167	39.0	31.0	46.4



#66
 4-Bromophenyl-phenylether
 Concen: 60.032 ng
 RT: 16.63 min Scan# 2271
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
248	100		
250	96.1	78.1	117.1
141	68.3	53.7	80.5









#67
 Hexachlorobenzene
 Concen: 59.173 ng
 RT: 16.75 min Scan# 2290
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
284	100		
142	35.7	27.8	41.6
249	34.2	27.4	41.2



#68
 Atrazine
 Concen: 48.105 ng
 RT: 16.90 min Scan# 2316
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
200	100		
173	23.3	3.9	43.9
215	49.8	30.4	70.4









#69
 Pentachlorophenol
 Concen: 69.461 ng
 RT: 17.09 min Scan# 2348
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
266	191655		
268	61.8	51.4	77.2
264	65.3	48.9	73.3



#70
 Phenanthrene
 Concen: 57.730 ng
 RT: 17.49 min Scan# 2416
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
178	1065896		
176	20.7	15.4	23.2
179	17.2	13.4	20.2








#71
 Anthracene
 Concen: 57.753 ng
 RT: 17.58 min Scan# 2432
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion:178 Resp: 1061631





Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.4	23.2
179	17.1	13.0	19.4




#72
 Carbazole
 Concen: 59.213 ng
 RT: 17.84 min Scan# 2477
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion:167 Resp: 1042360

Ion	Ratio	Lower	Upper
167	100		
166	24.7	19.5	29.3
139	14.2	10.7	16.1









#73
 Di-n-butylphthalate
 Concen: 56.399 ng
 RT: 18.41 min Scan# 2573
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion:149 Resp: 1164561




Ion	Ratio	Lower	Upper
149	100		
150	10.5	8.1	12.1
104	6.5	4.6	6.8




#74
 Fluoranthene
 Concen: 51.720 ng
 RT: 19.49 min Scan# 2757
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion:202 Resp: 1270655

Ion	Ratio	Lower	Upper
202	100		
101	10.1	0.0	29.4
203	19.9	0.0	38.9









#75
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.72 min Scan# 3136
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
240	356877		
120	7.8	6.9	10.3
236	26.9	21.2	31.8



#76
 Benzidine
 Concen: 59.372 ng
 RT: 19.67 min Scan# 2787
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
184	574134		
185	16.7	12.6	18.8
183	13.4	10.9	16.3









#77
 Pyrene
 Concen: 56.520 ng
 RT: 19.85 min Scan# 2818
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
202	1281346		
200	24.7	18.3	27.5
203	19.8	15.1	22.7



#78
 Terphenyl-d14
 Concen: 106.091 ng
 RT: 20.05 min Scan# 2853
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
244	1712350		
212	9.5	7.0	10.6
122	10.1	7.1	10.7









#79
 Butylbenzylphthalate
 Concen: 66.008 ng
 RT: 20.75 min Scan# 2971
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
149	100		
91	74.9	62.0	93.0
206	26.8	19.6	29.4



#80
 Benzo(a)anthracene
 Concen: 56.353 ng
 RT: 21.69 min Scan# 3132
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
228	100		
226	30.5	23.4	35.2
229	22.8	17.1	25.7









#81
 3,3'-Dichlorobenzidine
 Concen: 65.070 ng
 RT: 21.61 min Scan# 3118
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
252	515971		
254	65.5	51.0	76.4
126	9.2	7.4	11.0



#82
 Chrysene
 Concen: 57.214 ng
 RT: 21.76 min Scan# 3144
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
228	1186030		
226	34.7	25.5	38.3
229	22.2	16.9	25.3









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 65.609 ng
 RT: 21.62 min Scan# 3120
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
149	735021		
167	29.4	23.4	35.0
279	5.8	4.6	7.0



#84
 Di-n-octyl phthalate
 Concen: 67.312 ng
 RT: 22.85 min Scan# 3329
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
149	1264247		
167	1.7	1.3	1.9
43	12.6	10.0	15.0








#85
 Indeno(1,2,3-cd)pyrene
 Concen: 63.214 ng
 RT: 28.62 min Scan# 4311
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
276	1458243		
138	12.6	10.6	15.8
277	25.6	20.6	30.8



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 24.94 min Scan# 3684
 Delta R.T. -0.00 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion	Resp	Lower	Upper
264	379033		
260	23.5	19.6	29.4
265	21.7	17.4	26.0








#87
 Benzo(b)fluoranthene
 Concen: 55.412 ng
 RT: 23.92 min Scan# 3511
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion: 252 Resp: 1272019





Ion	Ratio	Lower	Upper
252	100		
253	23.2	17.1	25.7
125	7.9	6.6	10.0




#88
 Benzo(k)fluoranthene
 Concen: 54.003 ng
 RT: 23.99 min Scan# 3523
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion: 252 Resp: 1205106

Ion	Ratio	Lower	Upper
252	100		
253	23.2	18.0	27.0
125	9.0	7.0	10.6









#89
 Benzo(a)pyrene
 Concen: 57.038 ng
 RT: 24.79 min Scan# 3660
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion: 252 Resp: 1217113




Ion	Ratio	Lower	Upper
252	100		
253	21.9	18.2	27.2
125	9.2	7.3	10.9




#90
 Dibenzo(a,h)anthracene
 Concen: 55.573 ng
 RT: 28.70 min Scan# 4324
 Delta R.T. 0.01 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Tgt Ion: 278 Resp: 1161341

Ion	Ratio	Lower	Upper
278	100		
139	12.8	11.0	16.4
279	24.9	18.7	28.1





#91
 Benzo(a,h,i)perylene
 Concen: 57.758 ng
 RT: 29.77 min Scan# 4507
 Delta R.T. 0.02 min
 Lab File: BG036358.D
 Acq: 23 Aug 2018 1:07

Instrument :
 BNA_G
 ClientSampled :

Tot Ion: 276 Resp: 1184857

Ion	Ratio	Lower	Upper
276	100		
277	24.2	18.4	27.6
138	16.2	14.6	22.0

