

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082423\
 Data File : BG058847.D
 Acq On : 24 Aug 2023 20:45
 Operator : MA/JU
 Sample : 04091-20
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-F

Quant Time: Aug 25 02:59:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG081823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 01:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.806	152	16384	20.000	ng	0.00
21) Naphthalene-d8	10.609	136	66096	20.000	ng	0.00
39) Acenaphthene-d10	14.457	164	49003	20.000	ng	0.00
64) Phenanthrene-d10	17.201	188	127683	20.000	ng	0.00
76) Chrysene-d12	21.449	240	116384	20.000	ng	0.00
86) Perylene-d12	24.522	264	134193	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.379	112	145605	143.015	ng	0.00
7) Phenol-d6	6.978	99	178219	124.766	ng	0.00
23) Nitrobenzene-d5	8.969	82	125643	91.392	ng	0.00
42) 2,4,6-Tribromophenol	15.949	330	114932	149.581	ng	0.00
45) 2-Fluorobiphenyl	13.070	172	319703	94.891	ng	0.00
79) Terphenyl-d14	19.821	244	670428	103.435	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082423\
Data File : BG058847.D
Acq On : 24 Aug 2023 20:45
Operator : MA/JU
Sample : 04091-20
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-F

Quant Time: Aug 25 02:59:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG081823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 01:40:57 2023
Response via : Initial Calibration

